

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007718.D
 Acq On : 27 Oct 2021 14:17
 Operator : CG/JU
 Sample : M4304-11
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 P073-TW001-01

Quant Time: Oct 27 14:43:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 16:57:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	183377	20.000	ng/u1	0.00
20) Naphthalene-d8	10.734	136	754518	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.569	164	489938	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.322	188	1070883	20.000	ng/u1	-0.01
79) Chrysene-d12	21.416	240	1147287	20.000	ng/u1	0.00
88) Perylene-d12	23.863	264	1201741	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	21263	5.267	ng/uL	0.00
4) Pyridine-d5	3.787	84	113097	9.264	ng/u1	0.00
7) Phenol-d5	7.093	99	124634	8.948	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.258	67	332171	42.954	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.463	132	351789	32.474	ng/u1	-0.01
15) 4-Methylphenol-d8	8.640	113	230459	21.130	ng/u1	-0.01
21) Nitrobenzene-d5	9.087	128	215040	41.822	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.816	143	215270	38.332	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.352	165	400835	35.066	ng/u1	-0.02
31) 4-Chloroaniline-d4	10.869	131	521374	34.495	ng/u1	-0.01
46) Dimethylphthalate-d6	13.981	166	1458000	43.933	ng/u1	-0.01
49) Acenaphthylene-d8	14.263	160	1709780	42.034	ng/u1	0.00
54) 4-Nitrophenol-d4	14.757	143	47005	7.727	ng/u1	-0.02
60) Fluorene-d10	15.563	176	1290598	46.129	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.681	200	198746	34.696	ng/u1	0.00
73) Anthracene-d10	17.422	188	2267343	51.520	ng/u1	-0.01
81) Pyrene-d10	19.663	212	2772231	51.901	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.704	264	2877650	48.745	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
Data File : BP007718.D
Acq On : 27 Oct 2021 14:17
Operator : CG/JU
Sample : M4304-11
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
P073-TW001-01

Quant Time: Oct 27 14:43:37 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 25 16:57:16 2021
Response via : Initial Calibration

