

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110222\
 Data File : BP012472.D
 Acq On : 03 Nov 2022 03:28
 Operator : CG/JU
 Sample : PB148597BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS597

Quant Time: Nov 03 05:20:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 03 00:09:04 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
 Supervised By :mohammad ahmed 11/03/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	329408	20.000	ng/u1	0.00
20) Naphthalene-d8	10.581	136	1548472	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.440	164	1029794	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.198	188	2268064	20.000	ng/u1	0.00
79) Chrysene-d12	21.298	240	2134314	20.000	ng/u1	0.00
88) Perylene-d12	23.680	264	2010804	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	39811	4.866	ng/uL	0.00
4) Pyridine-d5	3.658	84	294739	12.542	ng/u1	0.00
7) Phenol-d5	6.963	99	872885	29.544	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.122	67	486019	27.556	ng/u1	0.00
11) 2-Chlorophenol-d4	7.316	132	650841	30.165	ng/u1	0.00
15) 4-Methylphenol-d8	8.505	113	714383	30.571	ng/u1	0.00
21) Nitrobenzene-d5	8.952	128	343354	34.193	ng/u1	0.00
24) 2-Nitrophenol-d4	9.675	143	390573	37.474	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.210	165	729429	32.294	ng/u1	0.00
31) 4-Chloroaniline-d4	10.734	131	611339	18.178	ng/u1	0.00
46) Dimethylphthalate-d6	13.851	166	2165146	30.713	ng/u1	0.00
49) Acenaphthylene-d8	14.128	160	2408847	29.901	ng/u1	0.00
54) 4-Nitrophenol-d4	14.663	143	434853	32.593	ng/u1	0.00
60) Fluorene-d10	15.434	176	1824692	30.232	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	385857	32.676	ng/u1	0.00
73) Anthracene-d10	17.298	188	2931144	29.994	ng/u1	0.00
81) Pyrene-d10	19.539	212	3495585	32.621	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.521	264	3128850	31.846	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.270	88	80481	9.253	ng/uL	98
5) Pyridine	3.675	79	494387	21.406	ng/u1	99
6) Benzaldehyde	6.934	77	470064	33.734	ng/u1	99
8) Phenol	6.987	94	839021	27.312	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.216	93	646505	26.222	ng/u1	97
12) 2-Chlorophenol	7.346	128	646396	27.521	ng/u1	99
13) 2-Methylphenol	8.240	108	648455	27.448	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.322	45	813047	24.590	ng/u1	99
16) Acetophenone	8.616	105	1002144	27.612	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.605	70	503974	28.844	ng/u1	98
18) 4-Methylphenol	8.569	108	723985	28.038	ng/u1	99
19) Hexachloroethane	8.857	117	242180	26.810	ng/u1	96
22) Nitrobenzene	8.999	77	747432	28.529	ng/u1	98
23) Isophorone	9.522	82	1494409	27.258	ng/u1	99
25) 2-Nitrophenol	9.704	139	402721	32.225	ng/u1	98
26) 2,4-Dimethylphenol	9.769	107	750611	27.354	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.004	93	948537	27.037	ng/u1	99
29) 2,4-Dichlorophenol	10.240	162	688755	29.374	ng/u1	100
30) Naphthalene	10.634	128	2161788	26.342	ng/u1	100
32) 4-Chloroaniline	10.757	127	748009	21.615	ng/u1	99
33) Hexachlorobutadiene	10.910	225	403417	27.654	ng/u1	99
34) Caprolactam	11.569	113	245322m	30.748	ng/u1	
35) 4-Chloro-3-methylphenol	11.893	107	775888	30.115	ng/u1	99
36) 2-Methylnaphthalene	12.251	142	1505305	26.755	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110222\
 Data File : BP012472.D
 Acq On : 03 Nov 2022 03:28
 Operator : CG/JU
 Sample : PB148597BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS597


Quant Time: Nov 03 05:20:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 03 00:09:04 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
 Supervised By :mohammad ahmed 11/03/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.469	142	1504762	26.781	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.616	216	811858	26.840	ng/u1	100
40) Hexachlorocyclopentadiene	12.587	237	316494	22.616	ng/u1	99
41) 2,4,6-Trichlorophenol	12.869	196	568622	29.409	ng/u1	99
42) 2,4,5-Trichlorophenol	12.940	196	627289	29.746	ng/u1	98
43) 1,1'-Biphenyl	13.269	154	2027704	26.778	ng/u1	99
44) 2-Chloronaphthalene	13.310	162	1585355	26.643	ng/u1	100
45) 2-Nitroaniline	13.528	65	485621	35.212	ng/u1	97
47) Dimethylphthalate	13.904	163	2107629	28.377	ng/u1	99
48) 2,6-Dinitrotoluene	14.028	165	446854	36.234	ng/u1	99
50) Acenaphthylene	14.157	152	2490044	27.561	ng/u1	100
51) 3-Nitroaniline	14.363	138	500948	33.175	ng/u1	100
52) Acenaphthene	14.498	153	1705492	26.862	ng/u1	100
53) 2,4-Dinitrophenol	14.575	184	223872	26.881	ng/u1	98
55) 4-Nitrophenol	14.681	109	311462	30.852	ng/u1	91
56) Dibenzofuran	14.840	168	2373540	27.383	ng/u1	98
57) 2,4-Dinitrotoluene	14.822	165	646848	34.823	ng/u1	96
58) 2,3,4,6-Tetrachlorophenol	15.069	232	542094	30.287	ng/u1	98
59) Diethylphthalate	15.269	149	2159320	29.878	ng/u1	99
61) Fluorene	15.492	166	1906099	27.676	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.487	204	950128	28.125	ng/u1	100
63) 4-Nitroaniline	15.534	138	530893m	38.914	ng/u1	
66) 4,6-Dinitro-2-methylph...	15.587	198	371958	29.564	ng/u1	99
67) N-Nitrosodiphenylamine	15.704	169	1720920	26.940	ng/u1	100
68) 4-Bromophenyl-phenylether	16.387	248	660028	28.972	ng/u1	98
69) Hexachlorobenzene	16.492	284	796731	29.159	ng/u1	99
70) Atrazine	16.669	200	308655	13.948	ng/u1	98
71) Pentachlorophenol	16.851	266	469057	27.952	ng/u1	99
72) Phenanthrene	17.239	178	3290149	27.459	ng/u1	99
74) Anthracene	17.334	178	3240732	27.225	ng/u1	100
75) 1,2,3,4-Tetrachloroben...	13.228	216	860390	26.519	ng/uL	99
76) Pentachlorobenzene	14.751	250	853204	24.903	ng/uL	100
77) Carbazole	17.610	167	3114846	29.381	ng/u1	100
78) Di-n-butylphthalate	18.157	149	3893153	31.609	ng/u1	99
80) Fluoranthene	19.216	202	4036643	30.511	ng/u1	97
82) Pyrene	19.569	202	4053684	29.581	ng/u1	97
83) Butylbenzylphthalate	20.445	149	1873217	36.665	ng/u1	95
84) 3,3'-Dichlorobenzidine	21.216	252	1322790	28.531	ng/u1	100
85) Benzo(a)anthracene	21.280	228	3875178	28.460	ng/u1	99
86) Bis(2-ethylhexyl)phtha...	21.198	149	2697821	34.534	ng/u1	100
87) Chrysene	21.333	228	3669405	28.826	ng/u1	99
89) Di-n-octyl phthalate	22.116	149	4540071	37.853	ng/u1	100
90) Benzo(b)fluoranthene	22.951	252	3853327	30.263	ng/u1	99
91) Benzo(k)fluoranthene	22.998	252	3707026	30.549	ng/u1	98
93) Benzo(a)pyrene	23.574	252	3239601	29.135	ng/u1	98
94) Indeno(1,2,3-cd)pyrene	26.151	276	3801251	27.435	ng/u1	97
95) Dibenzo(a,h)anthracene	26.168	278	3287113	26.495	ng/u1	99
96) Benzo(g,h,i)perylene	26.915	276	3149626	25.715	ng/u1	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 9.253 ng/uL
RT: 3.270 min Scan# 48
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion: 88 Resp: 80481
Ion Ratio Lower Upper
88 100
43 26.0 20.6 31.0
58 70.5 58.0 87.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#5
Pyridine
Concen: 21.406 ng/uL
RT: 3.675 min Scan# 117
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion: 79 Resp: 494387
Ion Ratio Lower Upper
79 100
52 72.2 57.4 86.0
51 33.4 25.7 38.5









#6
Benzaldehyde
Concen: 33.734 ng/ul
RT: 6.934 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Instrument :
BNA_P
ClientSampleId :
SLCS597

Tgt Ion: 77 Resp: 470064
Ion Ratio Lower Upper
77 100
105 101.5 81.0 121.6
106 97.8 79.2 118.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#8
Phenol
Concen: 27.312 ng/ul
RT: 6.987 min Scan# 680
Delta R.T. -0.005 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28


Tgt Ion: 94 Resp: 839021
Ion Ratio Lower Upper
94 100
65 26.4 20.8 31.2
66 33.9 26.2 39.2





#10
Bis(2-Chloroethyl)ether
Concen: 26.222 ng/ul
RT: 7.216 min Scan# 719
Delta R.T. -0.005 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion: 93 Resp: 646505
Ion Ratio Lower Upper
93 100
63 71.3 59.9 89.9
95 31.8 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#12
2-Chlorophenol
Concen: 27.521 ng/ul
RT: 7.346 min Scan# 741
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:128 Resp: 646396
Ion Ratio Lower Upper
128 100
64 44.7 35.1 52.7
130 31.6 25.8 38.6





#13

2-Methylphenol

Concen: 27.448 ng/ul

RT: 8.240 min Scan# 893

Delta R.T. 0.000 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS597

Tgt Ion:108 Resp: 64845

Ion Ratio Lower Upper





108 100

107 87.5 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 24.590 ng/ul

RT: 8.322 min Scan# 907

Delta R.T. 0.000 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28




Tgt Ion: 45 Resp: 813047


Ion Ratio Lower Upper

45 100

77 15.8 12.1 18.1


79 12.5 9.7 14.5





#16
Acetophenone
Concen: 27.612 ng/ul
RT: 8.616 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion: 105 Resp: 1002144
Ion Ratio Lower Upper
105 100
77 74.5 60.3 90.5
51 24.1 19.9 29.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#17
N-Nitroso-di-n-propylamine
Concen: 28.844 ng/ul
RT: 8.605 min Scan# 955
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28


Tgt Ion: 70 Resp: 503974
Ion Ratio Lower Upper
70 100
42 50.9 41.7 62.5
101 10.3 7.8 11.8
130 22.5 17.4 26.0





#18
4-Methylphenol
Concen: 28.038 ng/ul
RT: 8.569 min Scan# 949
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion:108 Resp: 723985
Ion Ratio Lower Upper
108 100
107 112.2 89.1 133.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#19
Hexachloroethane
Concen: 26.810 ng/ul
RT: 8.857 min Scan# 998
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28


Tgt Ion:117 Resp: 242180
Ion Ratio Lower Upper
117 100
201 103.2 78.5 117.7
199 64.9 49.9 74.9





#22
Nitrobenzene
Concen: 28.529 ng/ul
RT: 8.999 min Scan# 1022
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597




Tgt Ion: 77 Resp: 747432
Ion Ratio Lower Upper
77 100
123 48.2 37.4 56.2
65 13.0 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#23
Isophorone
Concen: 27.258 ng/ul
RT: 9.522 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Tgt Ion: 82 Resp: 1494409
Ion Ratio Lower Upper
82 100
95 7.0 5.2 7.8
138 18.0 14.2 21.2





#25
2-Nitrophenol
Concen: 32.225 ng/ul
RT: 9.704 min Scan# 1143
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597




Tgt Ion:139 Resp: 402721
Ion Ratio Lower Upper
139 100
65 48.0 37.8 56.8
109 29.2 21.4 32.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#26
2,4-Dimethylphenol
Concen: 27.354 ng/ul
RT: 9.769 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28



Tgt Ion:107 Resp: 750611
Ion Ratio Lower Upper
107 100
121 51.8 40.9 61.3
122 88.1 70.6 105.8









#27
Bis(2-Chloroethoxy)methane
Concen: 27.037 ng/ul
RT: 10.004 min Scan# 1194
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Instrument :
BNA_P
ClientSampleId :
SLCS597

Tgt Ion: 93 Resp: 94853
Ion Ratio Lower Upper
93 100
95 33.0 26.0 39.0
123 11.3 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#29
2,4-Dichlorophenol
Concen: 29.374 ng/ul
RT: 10.240 min Scan# 1233
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28


Tgt Ion:162 Resp: 688755
Ion Ratio Lower Upper
162 100
164 65.3 51.9 77.9
98 35.9 28.6 43.0





#30
Naphthalene
Concen: 26.342 ng/ul
RT: 10.634 min Scan# 1301
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion:128 Resp: 2161788
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.0 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#32
4-Chloroaniline
Concen: 21.615 ng/ul
RT: 10.757 min Scan# 1321
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:127 Resp: 748009
Ion Ratio Lower Upper
127 100
129 31.8 26.0 39.0









#33
Hexachlorobutadiene
Concen: 27.654 ng/ul
RT: 10.910 min Scan# 1347
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Instrument :
BNA_P
ClientSampleId :
SLCS597

Tgt Ion:225 Resp: 403417
Ion Ratio Lower Upper
225 100
223 61.3 49.0 73.6
227 63.3 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#34
Caprolactam
Concen: 30.748 ng/ul m
RT: 11.569 min Scan# 1459
Delta R.T. 0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28


Tgt Ion:113 Resp: 245322
Ion Ratio Lower Upper
113 100
55 158.4 135.6 203.4
56 115.6 98.9 148.3





#35
4-Chloro-3-methylphenol
Concen: 30.115 ng/ul
RT: 11.893 min Scan# 1514
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion:107 Resp: 775888
Ion Ratio Lower Upper
107 100
144 26.3 21.0 31.4
142 83.2 65.4 98.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#36
2-Methylnaphthalene
Concen: 26.755 ng/ul
RT: 12.251 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28


Tgt Ion:142 Resp: 1505305
Ion Ratio Lower Upper
142 100
141 88.6 70.6 106.0





#37
1-Methylnaphthalene
Concen: 26.781 ng/ul
RT: 12.469 min Scan# 1612
Delta R.T. -0.005 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion:142 Resp: 1504762
Ion Ratio Lower Upper
142 100
141 92.4 73.1 109.7
116 3.8 3.0 4.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 26.840 ng/ul
RT: 12.616 min Scan# 1637
Delta R.T. -0.005 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:216 Resp: 811858
Ion Ratio Lower Upper
216 100
214 78.7 62.9 94.3
179 20.6 16.6 24.8
108 16.8 13.2 19.8









#40
Hexachlorocyclopentadiene
Concen: 22.616 ng/ul
RT: 12.587 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Instrument :
BNA_P
ClientSampleId :
SLCS597

Tgt Ion:237 Resp: 316494
Ion Ratio Lower Upper
237 100
235 62.4 50.1 75.1
272 13.8 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#41
2,4,6-Trichlorophenol
Concen: 29.409 ng/ul
RT: 12.869 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:196 Resp: 568622
Ion Ratio Lower Upper
196 100
198 95.6 77.2 115.8
200 31.2 25.1 37.7





#42

2,4,5-Trichlorophenol

Concen: 29.746 ng/ul

RT: 12.940 min Scan# 1693

Delta R.T. -0.005 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS597

Tgt Ion:196 Resp: 627289

Ion Ratio Lower Upper

196 100





198 95.2 77.8 116.8

200 30.2 25.0 37.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#43

1,1'-Biphenyl

Concen: 26.778 ng/ul

RT: 13.269 min Scan# 1748

Delta R.T. 0.000 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28




Tgt Ion:154 Resp: 2027704


Ion Ratio Lower Upper

154 100

153 40.2 31.9 47.9


76 12.3 10.6 16.0





#44
2-Chloronaphthalene
Concen: 26.643 ng/ul
RT: 13.310 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597




Tgt Ion:162 Resp: 158535
Ion Ratio Lower Upper
162 100
164 32.9 26.6 39.8
127 38.9 30.9 46.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#45
2-Nitroaniline
Concen: 35.212 ng/ul
RT: 13.528 min Scan# 1792
Delta R.T. -0.005 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Tgt Ion: 65 Resp: 485621
Ion Ratio Lower Upper
65 100
92 70.8 58.5 87.7
138 110.3 91.1 136.7





#47
Dimethylphthalate
Concen: 28.377 ng/ul
RT: 13.904 min Scan# 1856
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597




Tgt Ion:163 Resp: 2107629
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 9.9 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#48
2,6-Dinitrotoluene
Concen: 36.234 ng/ul
RT: 14.028 min Scan# 1877
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Tgt Ion:165 Resp: 446854
Ion Ratio Lower Upper
165 100
89 55.5 43.4 65.2
121 20.9 17.1 25.7





#50
Acenaphthylene
Concen: 27.561 ng/ul
RT: 14.157 min Scan# 1899
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion:152 Resp: 2490044
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1
153 13.2 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#51
3-Nitroaniline
Concen: 33.175 ng/ul
RT: 14.363 min Scan# 1934
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:138 Resp: 500948
Ion Ratio Lower Upper
138 100
108 10.3 8.3 12.5
92 113.5 90.6 135.8





#52

Acenaphthene

Concen: 26.862 ng/ul

RT: 14.498 min Scan# 1958

Delta R.T. -0.005 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS597

Tgt Ion:153 Resp: 1705492

Ion Ratio Lower Upper

153 100





152 48.1 38.6 58.0

154 89.0 71.6 107.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#53

2,4-Dinitrophenol

Concen: 26.881 ng/ul

RT: 14.575 min Scan# 1970

Delta R.T. 0.000 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28




Tgt Ion:184 Resp: 223872


Ion Ratio Lower Upper

184 100

63 58.9 47.9 71.9


154 61.1 46.9 70.3





#55
4-Nitrophenol
Concen: 30.852 ng/ul
RT: 14.681 min Scan# 1988
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion:109 Resp: 311462
Ion Ratio Lower Upper
109 100
139 139.8 123.1 184.7
65 126.4 107.9 161.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#56
Dibenzofuran
Concen: 27.383 ng/ul
RT: 14.840 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:168 Resp: 2373540
Ion Ratio Lower Upper
168 100
139 37.6 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 34.823 ng/ul
RT: 14.822 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Instrument :
BNA_P
ClientSampleId :
SLCS597

Tgt Ion:165 Resp: 646848
Ion Ratio Lower Upper
165 100
63 43.8 33.1 49.7
182 2.5 2.1 3.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 30.287 ng/ul
RT: 15.069 min Scan# 2054
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28


Tgt Ion:232 Resp: 542094
Ion Ratio Lower Upper
232 100
131 44.5 33.6 50.4
130 2.4 1.8 2.6
166 30.0 24.2 36.2





#59
Diethylphthalate
Concen: 29.878 ng/ul
RT: 15.269 min Scan# 2088
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion:149 Resp: 2159320
Ion Ratio Lower Upper
149 100
177 22.0 18.2 27.2
150 12.5 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#61
Fluorene
Concen: 27.676 ng/ul
RT: 15.492 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28


Tgt Ion:166 Resp: 1906099
Ion Ratio Lower Upper
166 100
165 97.9 77.8 116.8
167 13.5 10.6 15.8





#62
4-Chlorophenyl-phenylether
Concen: 28.125 ng/ul
RT: 15.487 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion:204 Resp: 950128
Ion Ratio Lower Upper
204 100
206 33.2 26.5 39.7
141 59.4 47.3 70.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#63
4-Nitroaniline
Concen: 38.914 ng/ul m
RT: 15.534 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:138 Resp: 530893
Ion Ratio Lower Upper
138 100
92 49.3 40.6 61.0
108 74.9 56.6 85.0









#66
4,6-Dinitro-2-methylphenol
Concen: 29.564 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Instrument :
BNA_P
ClientSampleId :
SLCS597

Tgt Ion:198 Resp: 371958
Ion Ratio Lower Upper
198 100
182 4.0 3.8 5.6
77 24.0 19.5 29.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#67
N-Nitrosodiphenylamine
Concen: 26.940 ng/ul
RT: 15.704 min Scan# 2162
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:169 Resp: 1720920
Ion Ratio Lower Upper
169 100
168 66.4 52.9 79.3
167 36.0 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 28.972 ng/ul
RT: 16.387 min Scan# 2278
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Instrument :
BNA_P
ClientSampleId :
SLCS597

Tgt Ion:248 Resp: 660028
Ion Ratio Lower Upper
248 100
250 98.5 77.4 116.2
141 68.3 56.2 84.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#69
Hexachlorobenzene
Concen: 29.159 ng/ul
RT: 16.492 min Scan# 2296
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:284 Resp: 796731
Ion Ratio Lower Upper
284 100
142 33.0 27.1 40.7
249 30.5 24.5 36.7





#70

Atrazine

Concen: 13.948 ng/ul

RT: 16.669 min Scan# 2326

Delta R.T. 0.000 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS597

Tgt Ion:200 Resp: 30865

Ion Ratio Lower Upper

200 100





173 26.1 21.8 32.6

215 47.5 39.0 58.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#71

Pentachlorophenol

Concen: 27.952 ng/ul

RT: 16.851 min Scan# 2357

Delta R.T. 0.000 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28




Tgt Ion:266 Resp: 469057


Ion Ratio Lower Upper

266 100

264 63.2 50.3 75.5


268 63.1 51.4 77.0





#72
Phenanthrene
Concen: 27.459 ng/ul
RT: 17.239 min Scan# 2424
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion:178 Resp: 3290149
Ion Ratio Lower Upper
178 100
179 15.4 12.6 18.8
176 19.5 15.8 23.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#74
Anthracene
Concen: 27.225 ng/ul
RT: 17.334 min Scan# 2439
Delta R.T. -0.005 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:178 Resp: 3240732
Ion Ratio Lower Upper
178 100
179 15.6 12.5 18.7
176 19.1 15.3 22.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 26.519 ng/uL
RT: 13.228 min Scan# 11
Delta R.T. -0.005 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Instrument :
BNA_P
ClientSampleId :
SLCS597

Tgt Ion:216 Resp: 860390
Ion Ratio Lower Upper
216 100
214 78.7 62.6 94.0
218 47.8 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#76
Pentachlorobenzene
Concen: 24.903 ng/uL
RT: 14.751 min Scan# 2000
Delta R.T. -0.005 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28


Tgt Ion:250 Resp: 853204
Ion Ratio Lower Upper
250 100
248 63.4 50.6 75.8
252 63.2 50.7 76.1





#77
 Carbazole
 Concen: 29.381 ng/ul
 RT: 17.610 min Scan# 2486
 Delta R.T. -0.006 min
 Lab File: BP012472.D
 Acq: 03 Nov 2022 03:28




Instrument :
 BNA_P
 ClientSampleId :
 SLCS597




Tgt Ion:167 Resp: 3114846
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.1 25.7
 139 13.3 10.4 15.6

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
 Supervised By :mohammad ahmed 11/03/2022




#78
 Di-n-butylphthalate
 Concen: 31.609 ng/ul
 RT: 18.157 min Scan# 2579
 Delta R.T. -0.006 min
 Lab File: BP012472.D
 Acq: 03 Nov 2022 03:28




Tgt Ion:149 Resp: 3893153
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.6 11.4
 104 4.9 4.0 6.0





#80
Fluoranthene
Concen: 30.511 ng/ul
RT: 19.216 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597




Tgt Ion:202 Resp: 4036643
Ion Ratio Lower Upper
202 100
101 11.1 9.9 14.9
100 8.2 7.4 11.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#82
Pyrene
Concen: 29.581 ng/ul
RT: 19.569 min Scan# 2819
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28



Tgt Ion:202 Resp: 4053684
Ion Ratio Lower Upper
202 100
101 13.2 11.8 17.6
100 10.6 9.3 13.9





#83

Butylbenzylphthalate

Concen: 36.665 ng/ul

RT: 20.445 min Scan# 2968

Delta R.T. 0.000 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS597

Tgt Ion:149 Resp: 1873217

Ion Ratio Lower Upper

149 100





91 65.9 56.9 85.3

206 21.1 16.2 24.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#84

3,3'-Dichlorobenzidine

Concen: 28.531 ng/ul

RT: 21.216 min Scan# 3099

Delta R.T. -0.005 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28




Tgt Ion:252 Resp: 1322790


Ion Ratio Lower Upper

252 100

254 64.5 51.8 77.8


126 11.1 9.0 13.4





#85
Benzo(a)anthracene
Concen: 28.460 ng/ul
RT: 21.280 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597



Tgt Ion:228 Resp: 3875178
Ion Ratio Lower Upper
228 100
229 19.9 16.2 24.2
226 27.8 22.6 33.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 34.534 ng/ul
RT: 21.198 min Scan# 3096
Delta R.T. -0.005 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28

Tgt Ion:149 Resp: 2697821
Ion Ratio Lower Upper
149 100
167 27.3 22.1 33.1
279 4.5 3.7 5.5





#87

Chrysene

Concen: 28.826 ng/ul

RT: 21.333 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS597

Tgt Ion:228 Resp: 3669405

Ion Ratio Lower Upper

228 100





226 30.5 24.9 37.3

229 20.0 16.0 24.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#89

Di-n-octyl phthalate

Concen: 37.853 ng/ul




RT: 22.116 min Scan# 3252


Delta R.T. -0.005 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28


Tgt Ion:149 Resp: 4540071





#90
Benzo(b)fluoranthene
Concen: 30.263 ng/ul
RT: 22.951 min Scan# 3395
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS597




Tgt Ion:252 Resp: 3853327
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.8
125 9.9 8.7 13.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#91
Benzo(k)fluoranthene
Concen: 30.549 ng/ul
RT: 22.998 min Scan# 3402
Delta R.T. -0.006 min
Lab File: BP012472.D
Acq: 03 Nov 2022 03:28



Tgt Ion:252 Resp: 3707026
Ion Ratio Lower Upper
252 100
253 21.9 18.0 27.0
125 9.6 8.7 13.1





#93

Benzo(a)pyrene

Concen: 29.135 ng/ul

RT: 23.574 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS597

Tgt Ion:252 Resp: 323960

Ion Ratio Lower Upper

252 100





253 21.9 18.0 27.0

125 10.3 9.4 14.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 27.435 ng/ul

RT: 26.151 min Scan# 3938

Delta R.T. -0.011 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28




Tgt Ion:276 Resp: 3801251


Ion Ratio Lower Upper

276 100

138 22.8 20.7 31.1

277 24.9 20.2 30.4





#95

Dibenzo(a,h)anthracene

Concen: 26.495 ng/ul

RT: 26.168 min Scan# 3942

Delta R.T. -0.006 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS597

Tgt Ion:278 Resp: 328711

Ion Ratio Lower Upper

278 100

139 15.8 13.2 19.8





279 24.2 19.4 29.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022

Supervised By :mohammad ahmed 11/03/2022



#96

Benzo(g,h,i)perylene

Concen: 25.715 ng/ul

RT: 26.915 min Scan# 4068

Delta R.T. -0.006 min

Lab File: BP012472.D

Acq: 03 Nov 2022 03:28

Tgt Ion:276 Resp: 3149626

Ion Ratio Lower Upper

276 100

138 20.6 17.8 26.8

277 24.0 19.5 29.3

