

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110222\
 Data File : BP012473.D
 Acq On : 03 Nov 2022 04:02
 Operator : CG/JU
 Sample : PB148675BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS675

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
 Supervised By :mohammad ahmed 11/03/2022

Quant Time: Nov 03 05:20:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 03 00:09:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	329718	20.000	ng/u1	0.00
20) Naphthalene-d8	10.581	136	1532485	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.434	164	1005475	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.198	188	2216959	20.000	ng/u1	0.00
79) Chrysene-d12	21.298	240	2060127	20.000	ng/u1	0.00
88) Perylene-d12	23.674	264	1851372	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	39126	4.778	ng/uL	0.00
4) Pyridine-d5	3.652	84	509912	21.677	ng/u1	0.00
7) Phenol-d5	6.958	99	814394	27.538	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.122	67	456798	25.875	ng/u1	0.00
11) 2-Chlorophenol-d4	7.317	132	611026	28.293	ng/u1	0.00
15) 4-Methylphenol-d8	8.505	113	655195	28.012	ng/u1	0.00
21) Nitrobenzene-d5	8.952	128	318236	32.023	ng/u1	0.00
24) 2-Nitrophenol-d4	9.675	143	367805	35.658	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.210	165	671508	30.040	ng/u1	0.00
31) 4-Chloroaniline-d4	10.734	131	866403	26.030	ng/u1	0.00
46) Dimethylphthalate-d6	13.851	166	1982579	28.804	ng/u1	0.00
49) Acenaphthylene-d8	14.128	160	2232447	28.382	ng/u1	0.00
54) 4-Nitrophenol-d4	14.663	143	389837	29.926	ng/u1	0.00
60) Fluorene-d10	15.434	176	1693657	28.740	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	330115	28.600	ng/u1	0.00
73) Anthracene-d10	17.298	188	2646602	27.706	ng/u1	0.00
81) Pyrene-d10	19.539	212	3192261	30.863	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.522	264	2678600	29.611	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.270	88	76583	8.797	ng/uL	95
5) Pyridine	3.670	79	476494	20.612	ng/u1	96
6) Benzaldehyde	6.934	77	423359	30.354	ng/u1	98
8) Phenol	6.987	94	771090	25.077	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.217	93	595273	24.121	ng/u1	99
12) 2-Chlorophenol	7.346	128	600958	25.562	ng/u1	100
13) 2-Methylphenol	8.234	108	591181	25.001	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.322	45	748290	22.610	ng/u1	99
16) Acetophenone	8.616	105	935895	25.763	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.605	70	463093	26.479	ng/u1	97
18) 4-Methylphenol	8.569	108	670326	25.936	ng/u1	99
19) Hexachloroethane	8.858	117	226623	25.064	ng/u1	94
22) Nitrobenzene	8.993	77	684403	26.396	ng/u1	99
23) Isophorone	9.522	82	1365798	25.172	ng/u1	99
25) 2-Nitrophenol	9.705	139	366566	29.638	ng/u1	98
26) 2,4-Dimethylphenol	9.769	107	605555	22.298	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.005	93	862849	24.851	ng/u1	100
29) 2,4-Dichlorophenol	10.240	162	633409	27.295	ng/u1	99
30) Naphthalene	10.634	128	1993175	24.541	ng/u1	100
32) 4-Chloroaniline	10.758	127	851322	24.857	ng/u1	99
33) Hexachlorobutadiene	10.910	225	371440	25.728	ng/u1	98
34) Caprolactam	11.563	113	224317m	28.408	ng/u1	
35) 4-Chloro-3-methylphenol	11.893	107	712982	27.962	ng/u1	99
36) 2-Methylnaphthalene	12.252	142	1391569	24.991	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110222\
 Data File : BP012473.D
 Acq On : 03 Nov 2022 04:02
 Operator : CG/JU
 Sample : PB148675BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS675


Quant Time: Nov 03 05:20:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 03 00:09:04 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
 Supervised By :mohammad ahmed 11/03/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.469	142	1388509	24.969	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.616	216	746343	25.271	ng/u1	99
40) Hexachlorocyclopentadiene	12.587	237	256721	18.788	ng/u1	100
41) 2,4,6-Trichlorophenol	12.869	196	519053	27.495	ng/u1	99
42) 2,4,5-Trichlorophenol	12.940	196	579357	28.138	ng/u1	100
43) 1,1'-Biphenyl	13.263	154	1856357	25.108	ng/u1	100
44) 2-Chloronaphthalene	13.310	162	1457683	25.090	ng/u1	100
45) 2-Nitroaniline	13.528	65	438543	32.568	ng/u1	99
47) Dimethylphthalate	13.899	163	1903306	26.246	ng/u1	100
48) 2,6-Dinitrotoluene	14.028	165	401899	33.377	ng/u1	99
50) Acenaphthylene	14.157	152	2271282	25.748	ng/u1	100
51) 3-Nitroaniline	14.357	138	447956	30.383	ng/u1	100
52) Acenaphthene	14.498	153	1567311	25.282	ng/u1	99
53) 2,4-Dinitrophenol	14.575	184	165592	20.364	ng/u1	98
55) 4-Nitrophenol	14.681	109	276769	28.079	ng/u1	92
56) Dibenzofuran	14.840	168	2175783	25.709	ng/u1	99
57) 2,4-Dinitrotoluene	14.822	165	579786	31.968	ng/u1	96
58) 2,3,4,6-Tetrachlorophenol	15.069	232	500589	28.644	ng/u1	98
59) Diethylphthalate	15.269	149	1937057	27.451	ng/u1	99
61) Fluorene	15.493	166	1742611	25.914	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.487	204	868024	26.316	ng/u1	99
63) 4-Nitroaniline	15.534	138	471565m	35.402	ng/u1	
66) 4,6-Dinitro-2-methylph...	15.587	198	318430	25.893	ng/u1	99
67) N-Nitrosodiphenylamine	15.704	169	1571344	25.166	ng/u1	99
68) 4-Bromophenyl-phenylether	16.387	248	592375	26.602	ng/u1	98
69) Hexachlorobenzene	16.493	284	718016	26.884	ng/u1	98
70) Atrazine	16.669	200	501213	23.172	ng/u1	99
71) Pentachlorophenol	16.851	266	420354	25.627	ng/u1	99
72) Phenanthrene	17.240	178	2992896	25.554	ng/u1	100
74) Anthracene	17.334	178	2949667	25.351	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.228	216	785996	24.784	ng/uL	100
76) Pentachlorobenzene	14.751	250	783441	23.394	ng/uL	100
77) Carbazole	17.610	167	2816306	27.177	ng/u1	100
78) Di-n-butylphthalate	18.157	149	3483723	28.937	ng/u1	100
80) Fluoranthene	19.216	202	3641253	28.513	ng/u1	97
82) Pyrene	19.569	202	3668131	27.732	ng/u1	97
83) Butylbenzylphthalate	20.439	149	1635108	33.157	ng/u1	100
84) 3,3'-Dichlorobenzidine	21.216	252	1103735	24.664	ng/u1	100
85) Benzo(a)anthracene	21.280	228	3416511	25.995	ng/u1	100
86) Bis(2-ethylhexyl)phtha...	21.198	149	2360482	31.303	ng/u1	100
87) Chrysene	21.333	228	3238252	26.355	ng/u1	99
89) Di-n-octyl phthalate	22.116	149	3862568	34.978	ng/u1	100
90) Benzo(b)fluoranthene	22.951	252	3308378	28.220	ng/u1	99
91) Benzo(k)fluoranthene	22.998	252	3160320	28.286	ng/u1	98
93) Benzo(a)pyrene	23.574	252	2732969	26.695	ng/u1	98
94) Indeno(1,2,3-cd)pyrene	26.151	276	3224242	25.274	ng/u1	97
95) Dibenzo(a,h)anthracene	26.168	278	2802573	24.535	ng/u1	98
96) Benzo(g,h,i)perylene	26.915	276	2701467	23.956	ng/u1	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.797 ng/uL
RT: 3.270 min Scan# 48
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion: 88 Resp: 76583
Ion Ratio Lower Upper
88 100
43 23.4 20.6 31.0
58 68.7 58.0 87.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#5
Pyridine
Concen: 20.612 ng/uL
RT: 3.670 min Scan# 116
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion: 79 Resp: 476494
Ion Ratio Lower Upper
79 100
52 75.6 57.4 86.0
51 32.8 25.7 38.5





#6
Benzaldehyde
Concen: 30.354 ng/ul
RT: 6.934 min Scan# 61
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion: 77 Resp: 423359
Ion Ratio Lower Upper
77 100
105 101.1 81.0 121.6
106 96.0 79.2 118.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#8
Phenol
Concen: 25.077 ng/ul
RT: 6.987 min Scan# 680
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion: 94 Resp: 771090
Ion Ratio Lower Upper
94 100
65 26.8 20.8 31.2
66 35.0 26.2 39.2





#10
Bis(2-Chloroethyl)ether
Concen: 24.121 ng/ul
RT: 7.217 min Scan# 719
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion: 93 Resp: 595273
Ion Ratio Lower Upper
93 100
63 74.1 59.9 89.9
95 32.9 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#12
2-Chlorophenol
Concen: 25.562 ng/ul
RT: 7.346 min Scan# 741
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Tgt Ion:128 Resp: 600958
Ion Ratio Lower Upper
128 100
64 44.1 35.1 52.7
130 31.8 25.8 38.6





#13

2-Methylphenol

Concen: 25.001 ng/ul

RT: 8.234 min Scan# 893

Delta R.T. -0.005 min

Lab File: BP012473.D

Acq: 03 Nov 2022 04:02

Instrument :

BNA_P

ClientSampleId :

SLCS675

Tgt Ion:108 Resp: 59118

Ion Ratio Lower Upper





108 100

107 87.2 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 22.610 ng/ul

RT: 8.322 min Scan# 907

Delta R.T. 0.000 min

Lab File: BP012473.D

Acq: 03 Nov 2022 04:02




Tgt Ion: 45 Resp: 748290


Ion Ratio Lower Upper

45 100

77 15.5 12.1 18.1


79 12.3 9.7 14.5





#16
Acetophenone
Concen: 25.763 ng/ul
RT: 8.616 min Scan# 91
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion: 105 Resp: 935895
Ion Ratio Lower Upper
105 100
77 74.3 60.3 90.5
51 23.5 19.9 29.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#17
N-Nitroso-di-n-propylamine
Concen: 26.479 ng/ul
RT: 8.605 min Scan# 955
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion: 70 Resp: 463093
Ion Ratio Lower Upper
70 100
42 49.5 41.7 62.5
101 10.4 7.8 11.8
130 22.6 17.4 26.0





#18
4-Methylphenol
Concen: 25.936 ng/ul
RT: 8.569 min Scan# 949
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion:108 Resp: 670326
Ion Ratio Lower Upper
108 100
107 110.4 89.1 133.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#19
Hexachloroethane
Concen: 25.064 ng/ul
RT: 8.858 min Scan# 998
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:117 Resp: 226623
Ion Ratio Lower Upper
117 100
201 105.4 78.5 117.7
199 65.1 49.9 74.9





#22
Nitrobenzene
Concen: 26.396 ng/ul
RT: 8.993 min Scan# 1022
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675




Tgt Ion: 77 Resp: 684403
Ion Ratio Lower Upper
77 100
123 47.7 37.4 56.2
65 12.9 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#23
Isophorone
Concen: 25.172 ng/ul
RT: 9.522 min Scan# 1111
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Tgt Ion: 82 Resp: 1365798
Ion Ratio Lower Upper
82 100
95 7.0 5.2 7.8
138 17.9 14.2 21.2





#25
2-Nitrophenol
Concen: 29.638 ng/ul
RT: 9.705 min Scan# 1143
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675




Tgt Ion:139 Resp: 366566
Ion Ratio Lower Upper
139 100
65 48.2 37.8 56.8
109 28.6 21.4 32.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#26
2,4-Dimethylphenol
Concen: 22.298 ng/ul
RT: 9.769 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Tgt Ion:107 Resp: 605555
Ion Ratio Lower Upper
107 100
121 51.3 40.9 61.3
122 88.0 70.6 105.8





#27
Bis(2-Chloroethoxy)methane
Concen: 24.851 ng/ul
RT: 10.005 min Scan# 1194
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion: 93 Resp: 862849
Ion Ratio Lower Upper
93 100
95 32.4 26.0 39.0
123 11.1 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#29
2,4-Dichlorophenol
Concen: 27.295 ng/ul
RT: 10.240 min Scan# 1233
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Tgt Ion:162 Resp: 633409
Ion Ratio Lower Upper
162 100
164 65.6 51.9 77.9
98 34.5 28.6 43.0









#30
Naphthalene
Concen: 24.541 ng/ul
RT: 10.634 min Scan# 1301
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Instrument :
BNA_P
ClientSampleId :
SLCS675

Tgt Ion:128 Resp: 1993175
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.2 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#32
4-Chloroaniline
Concen: 24.857 ng/ul
RT: 10.758 min Scan# 1321
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:127 Resp: 851322
Ion Ratio Lower Upper
127 100
129 32.0 26.0 39.0





#33
Hexachlorobutadiene
Concen: 25.728 ng/ul
RT: 10.910 min Scan# 1348
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion:225 Resp: 371440
Ion Ratio Lower Upper
225 100
223 61.8 49.0 73.6
227 62.8 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#34
Caprolactam
Concen: 28.408 ng/ul m
RT: 11.563 min Scan# 1458
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Tgt Ion:113 Resp: 224317
Ion Ratio Lower Upper
113 100
55 158.5 135.6 203.4
56 120.8 98.9 148.3









#35
4-Chloro-3-methylphenol
Concen: 27.962 ng/ul
RT: 11.893 min Scan# 1514
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Instrument :
BNA_P
ClientSampleId :
SLCS675

Tgt Ion:107 Resp: 712982
Ion Ratio Lower Upper
107 100
144 26.6 21.0 31.4
142 82.6 65.4 98.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#36
2-Methylnaphthalene
Concen: 24.991 ng/ul
RT: 12.252 min Scan# 1575
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:142 Resp: 1391569
Ion Ratio Lower Upper
142 100
141 87.4 70.6 106.0





#37
1-Methylnaphthalene
Concen: 24.969 ng/ul
RT: 12.469 min Scan# 1612
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675




Tgt Ion:142 Resp: 1388509
Ion Ratio Lower Upper
142 100
141 92.5 73.1 109.7
116 3.7 3.0 4.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 25.271 ng/ul
RT: 12.616 min Scan# 1637
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02



Tgt Ion:216 Resp: 746343
Ion Ratio Lower Upper
216 100
214 78.1 62.9 94.3
179 20.8 16.6 24.8
108 17.2 13.2 19.8









#40
Hexachlorocyclopentadiene
Concen: 18.788 ng/ul
RT: 12.587 min Scan# 1632
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Instrument :
BNA_P
ClientSampleId :
SLCS675

Tgt Ion:237 Resp: 25672
Ion Ratio Lower Upper
237 100
235 62.3 50.1 75.1
272 13.2 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#41
2,4,6-Trichlorophenol
Concen: 27.495 ng/ul
RT: 12.869 min Scan# 1680
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:196 Resp: 519053
Ion Ratio Lower Upper
196 100
198 94.8 77.2 115.8
200 31.1 25.1 37.7





#42
2,4,5-Trichlorophenol
Concen: 28.138 ng/ul
RT: 12.940 min Scan# 1693
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion:196 Resp: 579352
Ion Ratio Lower Upper
196 100
198 97.2 77.8 116.8
200 30.7 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#43
1,1'-Biphenyl
Concen: 25.108 ng/ul
RT: 13.263 min Scan# 1747
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:154 Resp: 1856357
Ion Ratio Lower Upper
154 100
153 39.7 31.9 47.9
76 13.0 10.6 16.0





#44
2-Chloronaphthalene
Concen: 25.090 ng/ul
RT: 13.310 min Scan# 1755
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675




Tgt Ion:162 Resp: 1457683
Ion Ratio Lower Upper
162 100
164 33.0 26.6 39.8
127 39.0 30.9 46.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#45
2-Nitroaniline
Concen: 32.568 ng/ul
RT: 13.528 min Scan# 1792
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Tgt Ion: 65 Resp: 438543
Ion Ratio Lower Upper
65 100
92 72.2 58.5 87.7
138 114.1 91.1 136.7





#47
Dimethylphthalate
Concen: 26.246 ng/ul
RT: 13.899 min Scan# 1855
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675




Tgt Ion:163 Resp: 1903306
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.2 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#48
2,6-Dinitrotoluene
Concen: 33.377 ng/ul
RT: 14.028 min Scan# 1877
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02



Tgt Ion:165 Resp: 401899
Ion Ratio Lower Upper
165 100
89 55.1 43.4 65.2
121 22.1 17.1 25.7









#50
Acenaphthylene
Concen: 25.748 ng/ul
RT: 14.157 min Scan# 1899
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Instrument :
BNA_P
ClientSampleId :
SLCS675

Tgt Ion:152 Resp: 227128
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1
153 13.1 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#51
3-Nitroaniline
Concen: 30.383 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:138 Resp: 447956
Ion Ratio Lower Upper
138 100
108 10.0 8.3 12.5
92 113.3 90.6 135.8





#52
Acenaphthene
Concen: 25.282 ng/ul
RT: 14.498 min Scan# 1958
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675




Tgt Ion:153 Resp: 1567311
Ion Ratio Lower Upper
153 100
152 47.5 38.6 58.0
154 89.4 71.6 107.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#53
2,4-Dinitrophenol
Concen: 20.364 ng/ul
RT: 14.575 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Tgt Ion:184 Resp: 165592
Ion Ratio Lower Upper
184 100
63 60.6 47.9 71.9
154 61.1 46.9 70.3





#55
4-Nitrophenol
Concen: 28.079 ng/ul
RT: 14.681 min Scan# 1988
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion:109 Resp: 276769
Ion Ratio Lower Upper
109 100
139 142.6 123.1 184.7
65 125.5 107.9 161.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#56
Dibenzofuran
Concen: 25.709 ng/ul
RT: 14.840 min Scan# 2015
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Tgt Ion:168 Resp: 2175783
Ion Ratio Lower Upper
168 100
139 36.8 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 31.968 ng/ul
RT: 14.822 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Instrument :
BNA_P
ClientSampleId :
SLCS675

Tgt Ion:165 Resp: 579786
Ion Ratio Lower Upper
165 100
63 44.4 33.1 49.7
182 2.5 2.1 3.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 28.644 ng/ul
RT: 15.069 min Scan# 2054
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:232 Resp: 500589
Ion Ratio Lower Upper
232 100
131 44.3 33.6 50.4
130 2.3 1.8 2.6
166 30.7 24.2 36.2





#59
Diethylphthalate
Concen: 27.451 ng/ul
RT: 15.269 min Scan# 2088
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion:149 Resp: 1937052
Ion Ratio Lower Upper
149 100
177 22.2 18.2 27.2
150 12.5 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#61
Fluorene
Concen: 25.914 ng/ul
RT: 15.493 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:166 Resp: 1742611
Ion Ratio Lower Upper
166 100
165 98.2 77.8 116.8
167 13.3 10.6 15.8





#62
4-Chlorophenyl-phenylether
Concen: 26.316 ng/ul
RT: 15.487 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675




Tgt Ion:204 Resp: 868024
Ion Ratio Lower Upper
204 100
206 32.8 26.5 39.7
141 57.7 47.3 70.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#63
4-Nitroaniline
Concen: 35.402 ng/ul m
RT: 15.534 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02



Tgt Ion:138 Resp: 471565
Ion Ratio Lower Upper
138 100
92 51.1 40.6 61.0
108 75.9 56.6 85.0









#66
4,6-Dinitro-2-methylphenol
Concen: 25.893 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Instrument :
BNA_P
ClientSampleId :
SLCS675

Tgt Ion:198 Resp: 318430
Ion Ratio Lower Upper
198 100
182 4.6 3.8 5.6
77 23.7 19.5 29.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#67
N-Nitrosodiphenylamine
Concen: 25.166 ng/ul
RT: 15.704 min Scan# 2162
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Tgt Ion:169 Resp: 1571344
Ion Ratio Lower Upper
169 100
168 66.9 52.9 79.3
167 36.7 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 26.602 ng/ul
RT: 16.387 min Scan# 2278
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Instrument :
BNA_P
ClientSampleId :
SLCS675

Tgt Ion:248 Resp: 592375
Ion Ratio Lower Upper
248 100
250 98.8 77.4 116.2
141 69.1 56.2 84.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#69
Hexachlorobenzene
Concen: 26.884 ng/ul
RT: 16.493 min Scan# 2296
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Tgt Ion:284 Resp: 718016
Ion Ratio Lower Upper
284 100
142 32.3 27.1 40.7
249 29.6 24.5 36.7





#70

Atrazine

Concen: 23.172 ng/ul

RT: 16.669 min Scan# 2326

Delta R.T. 0.000 min

Lab File: BP012473.D

Acq: 03 Nov 2022 04:02

Instrument :

BNA_P

ClientSampleId :





SLCS675

Tgt Ion:	200	Resp:	501213
Ion Ratio	Lower	Upper	
200	100		
173	26.3	21.8	32.6
215	48.1	39.0	58.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#71

Pentachlorophenol

Concen: 25.627 ng/ul




RT: 16.851 min Scan# 2357


Delta R.T. 0.000 min

Lab File: BP012473.D

Acq: 03 Nov 2022 04:02


Tgt Ion:	266	Resp:	420354
Ion Ratio	Lower	Upper	
266	100		
264	63.6	50.3	75.5
268	64.0	51.4	77.0





#72
Phenanthrene
Concen: 25.554 ng/ul
RT: 17.240 min Scan# 2424
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion:178 Resp: 2992896
Ion Ratio Lower Upper
178 100
179 15.5 12.6 18.8
176 19.5 15.8 23.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#74
Anthracene
Concen: 25.351 ng/ul
RT: 17.334 min Scan# 2439
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Tgt Ion:178 Resp: 2949667
Ion Ratio Lower Upper
178 100
179 15.5 12.5 18.7
176 18.5 15.3 22.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 24.784 ng/uL
RT: 13.228 min Scan# 11
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02

Instrument :
BNA_P
ClientSampleId :
SLCS675

Tgt Ion:216 Resp: 785996
Ion Ratio Lower Upper
216 100
214 77.8 62.6 94.0
218 48.4 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#76
Pentachlorobenzene
Concen: 23.394 ng/uL
RT: 14.751 min Scan# 2000
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:250 Resp: 783441
Ion Ratio Lower Upper
250 100
248 63.8 50.6 75.8
252 63.4 50.7 76.1





#77
 Carbazole
 Concen: 27.177 ng/ul
 RT: 17.610 min Scan# 2487
 Delta R.T. -0.005 min
 Lab File: BP012473.D
 Acq: 03 Nov 2022 04:02




Instrument :
 BNA_P
 ClientSampleId :
 SLCS675




Tgt Ion:167 Resp: 2816306
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.1 25.7
 139 13.4 10.4 15.6

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
 Supervised By :mohammad ahmed 11/03/2022




#78
 Di-n-butylphthalate
 Concen: 28.937 ng/ul
 RT: 18.157 min Scan# 2579
 Delta R.T. -0.005 min
 Lab File: BP012473.D
 Acq: 03 Nov 2022 04:02




Tgt Ion:149 Resp: 3483723
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.6 11.4
 104 5.0 4.0 6.0





#80
Fluoranthene
Concen: 28.513 ng/u1
RT: 19.216 min Scan# 21
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion:202 Resp: 3641253
Ion Ratio Lower Upper
202 100
101 11.1 9.9 14.9
100 8.2 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#82
Pyrene
Concen: 27.732 ng/u1
RT: 19.569 min Scan# 2819
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:202 Resp: 3668131
Ion Ratio Lower Upper
202 100
101 13.1 11.8 17.6
100 10.9 9.3 13.9





#83
Butylbenzylphthalate
Concen: 33.157 ng/ul
RT: 20.439 min Scan# 2968
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675




Tgt Ion:149 Resp: 1635108
Ion Ratio Lower Upper
149 100
91 71.5 56.9 85.3
206 20.3 16.2 24.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#84
3,3'-Dichlorobenzidine
Concen: 24.664 ng/ul
RT: 21.216 min Scan# 3099
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Tgt Ion:252 Resp: 1103735
Ion Ratio Lower Upper
252 100
254 64.6 51.8 77.8
126 10.8 9.0 13.4





#85
Benzo(a)anthracene
Concen: 25.995 ng/ul
RT: 21.280 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675




Tgt Ion:228 Resp: 3416511
Ion Ratio Lower Upper
228 100
229 20.1 16.2 24.2
226 27.9 22.6 33.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 31.303 ng/ul
RT: 21.198 min Scan# 3096
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Tgt Ion:149 Resp: 2360482
Ion Ratio Lower Upper
149 100
167 27.5 22.1 33.1
279 4.6 3.7 5.5





#87
Chrysene
Concen: 26.355 ng/ul
RT: 21.333 min Scan# 3119
Delta R.T. 0.001 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675



Tgt Ion:228 Resp: 3238252
Ion Ratio Lower Upper
228 100
226 30.4 24.9 37.3
229 19.8 16.0 24.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#89
Di-n-octyl phthalate
Concen: 34.978 ng/ul
RT: 22.116 min Scan# 3252
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02


Tgt Ion:149 Resp: 3862568





#90
Benzo(b)fluoranthene
Concen: 28.220 ng/ul
RT: 22.951 min Scan# 3395
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02




Instrument :
BNA_P
ClientSampleId :
SLCS675




Tgt Ion:252 Resp: 3308378
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 9.8 8.7 13.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022




#91
Benzo(k)fluoranthene
Concen: 28.286 ng/ul
RT: 22.998 min Scan# 3402
Delta R.T. -0.005 min
Lab File: BP012473.D
Acq: 03 Nov 2022 04:02



Tgt Ion:252 Resp: 3160320
Ion Ratio Lower Upper
252 100
253 21.9 18.0 27.0
125 9.5 8.7 13.1





#93

Benzo(a)pyrene

Concen: 26.695 ng/ul

RT: 23.574 min Scan# 3938

Delta R.T. -0.005 min

Lab File: BP012473.D

Acq: 03 Nov 2022 04:02

Instrument :

BNA_P

ClientSampleId :

SLCS675

Tgt Ion:252 Resp: 2732969

Ion Ratio Lower Upper

252 100





253 21.9 18.0 27.0

125 10.5 9.4 14.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022
Supervised By :mohammad ahmed 11/03/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 25.274 ng/ul

RT: 26.151 min Scan# 3938

Delta R.T. -0.011 min

Lab File: BP012473.D

Acq: 03 Nov 2022 04:02




Tgt Ion:276 Resp: 3224242


Ion Ratio Lower Upper

276 100

138 23.6 20.7 31.1

277 25.6 20.2 30.4





#95

Dibenzo(a,h)anthracene

Concen: 24.535 ng/ul

RT: 26.168 min Scan# 3942

Delta R.T. -0.005 min

Lab File: BP012473.D

Acq: 03 Nov 2022 04:02

Instrument :

BNA_P

ClientSampleId :

SLCS675

Tgt Ion:278 Resp: 280257

Ion Ratio Lower Upper

278 100

139 15.0 13.2 19.8





279 24.2 19.4 29.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2022

Supervised By :mohammad ahmed 11/03/2022



#96

Benzo(g,h,i)perylene

Concen: 23.956 ng/ul

RT: 26.915 min Scan# 4068

Delta R.T. -0.005 min

Lab File: BP012473.D

Acq: 03 Nov 2022 04:02

Tgt Ion:276 Resp: 2701467

Ion Ratio Lower Upper

276 100

138 20.6 17.8 26.8

277 24.0 19.5 29.3

