

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110222\
 Data File : BP012505.D
 Acq On : 04 Nov 2022 00:09
 Operator : CG/JU
 Sample : PB148619BS
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS619

Quant Time: Nov 04 01:04:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 03 00:09:04 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 11/05/2022
 Supervised By :mohammad ahmed 11/06/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	285859	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.581	136	1331693	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.434	164	888113	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	1955977	20.000	ng/u1	-0.01
79) Chrysene-d12	21.292	240	1755702	20.000	ng/u1	0.00
88) Perylene-d12	23.668	264	1550167	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	45460	6.403	ng/uL	0.00
4) Pyridine-d5	3.652	84	392491	19.245	ng/u1	0.00
7) Phenol-d5	6.957	99	985139	38.423	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.116	67	546452	35.702	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.310	132	738002	39.416	ng/u1	-0.01
15) 4-Methylphenol-d8	8.504	113	796765	39.291	ng/u1	0.00
21) Nitrobenzene-d5	8.951	128	384951	44.576	ng/u1	0.00
24) 2-Nitrophenol-d4	9.669	143	435923	48.634	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.210	165	793658	40.857	ng/u1	0.00
31) 4-Chloroaniline-d4	10.734	131	661026	22.854	ng/u1	0.00
46) Dimethylphthalate-d6	13.851	166	2375552	39.074	ng/u1	0.00
49) Acenaphthylene-d8	14.128	160	2678017	38.545	ng/u1	0.00
54) 4-Nitrophenol-d4	14.663	143	451339	39.225	ng/u1	0.00
60) Fluorene-d10	15.433	176	2019583	38.799	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	408642	40.127	ng/u1	0.00
73) Anthracene-d10	17.292	188	3195618	37.917	ng/u1	-0.01
81) Pyrene-d10	19.533	212	3665177	41.580	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.515	264	3041381	40.154	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.264	88	94121	12.470	ng/uL	98
5) Pyridine	3.669	79	585294	29.203	ng/u1	99
6) Benzaldehyde	6.928	77	472415	39.068	ng/u1	98
8) Phenol	6.987	94	969133	36.353	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.210	93	731352	34.182	ng/u1	99
12) 2-Chlorophenol	7.346	128	742568	36.432	ng/u1	100
13) 2-Methylphenol	8.234	108	743344	36.258	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.316	45	927283	32.317	ng/u1	99
16) Acetophenone	8.616	105	1152682	36.599	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.599	70	574762	37.907	ng/u1	99
18) 4-Methylphenol	8.563	108	824558	36.798	ng/u1	98
19) Hexachloroethane	8.851	117	279266	35.625	ng/u1	93
22) Nitrobenzene	8.993	77	859231	38.135	ng/u1	99
23) Isophorone	9.522	82	1695186	35.954	ng/u1	99
25) 2-Nitrophenol	9.698	139	453345	42.181	ng/u1	97
26) 2,4-Dimethylphenol	9.763	107	873441	37.011	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.004	93	1058259	35.074	ng/u1	100
29) 2,4-Dichlorophenol	10.234	162	761001	37.738	ng/u1	100
30) Naphthalene	10.628	128	2424739	34.356	ng/u1	100
32) 4-Chloroaniline	10.757	127	936631	31.472	ng/u1	99
33) Hexachlorobutadiene	10.904	225	442277	35.253	ng/u1	98
34) Caprolactam	11.563	113	270248m	39.386	ng/u1	
35) 4-Chloro-3-methylphenol	11.887	107	876421	39.554	ng/u1	99
36) 2-Methylnaphthalene	12.245	142	1690278	34.933	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110222\
 Data File : BP012505.D
 Acq On : 04 Nov 2022 00:09
 Operator : CG/JU
 Sample : PB148619BS
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS619


Quant Time: Nov 04 01:04:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 03 00:09:04 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/05/2022
 Supervised By :mohammad ahmed 11/06/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.463	142	1700388	35.188	ng/u1	100
39) 1,2,4,5-Tetrachloroben...	12.616	216	905715	34.720	ng/u1	100
40) Hexachlorocyclopentadiene	12.587	237	379308	31.429	ng/u1	100
41) 2,4,6-Trichlorophenol	12.863	196	637038	38.204	ng/u1	99
42) 2,4,5-Trichlorophenol	12.940	196	710491	39.066	ng/u1	99
43) 1,1'-Biphenyl	13.263	154	2280143	34.916	ng/u1	99
44) 2-Chloronaphthalene	13.304	162	1792136	34.923	ng/u1	99
45) 2-Nitroaniline	13.528	65	545897	45.898	ng/u1	99
47) Dimethylphthalate	13.898	163	2357206	36.800	ng/u1	100
48) 2,6-Dinitrotoluene	14.028	165	503491	47.340	ng/u1	99
50) Acenaphthylene	14.157	152	2818451	36.173	ng/u1	99
51) 3-Nitroaniline	14.357	138	528040	40.547	ng/u1	100
52) Acenaphthene	14.498	153	1932183	35.287	ng/u1	100
53) 2,4-Dinitrophenol	14.569	184	229692	31.979	ng/u1	96
55) 4-Nitrophenol	14.675	109	325836	37.425	ng/u1	91
56) Dibenzofuran	14.834	168	2677915	35.823	ng/u1	98
57) 2,4-Dinitrotoluene	14.816	165	709647	44.299	ng/u1	95
58) 2,3,4,6-Tetrachlorophenol	15.063	232	611837	39.636	ng/u1	96
59) Diethylphthalate	15.263	149	2376897	38.136	ng/u1	99
61) Fluorene	15.486	166	2115767	35.621	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.481	204	1055409	36.226	ng/u1	99
63) 4-Nitroaniline	15.528	138	552715m	46.977	ng/u1	
66) 4,6-Dinitro-2-methylph...	15.586	198	399345	36.805	ng/u1	96
67) N-Nitrosodiphenylamine	15.704	169	1939836	35.212	ng/u1	99
68) 4-Bromophenyl-phenylether	16.381	248	733673	37.343	ng/u1	99
69) Hexachlorobenzene	16.492	284	869241	36.889	ng/u1	96
70) Atrazine	16.663	200	535247	28.047	ng/u1	99
71) Pentachlorophenol	16.845	266	503383	34.784	ng/u1	99
72) Phenanthrene	17.239	178	3638214	35.209	ng/u1	100
74) Anthracene	17.328	178	3614780	35.213	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.222	216	959874	34.305	ng/uL	99
76) Pentachlorobenzene	14.751	250	953203	32.261	ng/uL	99
77) Carbazole	17.610	167	3372039	36.881	ng/u1	100
78) Di-n-butylphthalate	18.157	149	4260284	40.109	ng/u1	99
80) Fluoranthene	19.210	202	4318184	39.677	ng/u1	98
82) Pyrene	19.563	202	4380303	38.858	ng/u1	99
83) Butylbenzylphthalate	20.439	149	1954795	46.513	ng/u1	99
84) 3,3'-Dichlorobenzidine	21.216	252	1307489	34.283	ng/u1	99
85) Benzo(a)anthracene	21.274	228	4027256	35.955	ng/u1	99
86) Bis(2-ethylhexyl)phtha...	21.192	149	2835625	44.125	ng/u1	99
87) Chrysene	21.327	228	3758655	35.895	ng/u1	99
89) Di-n-octyl phthalate	22.110	149	4630005	50.074	ng/u1	100
90) Benzo(b)fluoranthene	22.945	252	3834755	39.066	ng/u1	99
91) Benzo(k)fluoranthene	22.992	252	3685544	39.397	ng/u1	99
93) Benzo(a)pyrene	23.568	252	3189321	37.206	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.145	276	3809327	35.663	ng/u1	98
95) Dibenzo(a,h)anthracene	26.156	278	3280798	34.303	ng/u1	99
96) Benzo(g,h,i)perylene	26.903	276	3194793	33.835	ng/u1	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.470 ng/uL
RT: 3.264 min Scan# 47
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion: 88 Resp: 9412
Ion Ratio Lower Upper
88 100
43 25.5 20.6 31.0
58 74.6 58.0 87.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#5
Pyridine
Concen: 29.203 ng/uL
RT: 3.669 min Scan# 116
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Tgt Ion: 79 Resp: 585294
Ion Ratio Lower Upper
79 100
52 70.4 57.4 86.0
51 32.7 25.7 38.5









#6
Benzaldehyde
Concen: 39.068 ng/ul
RT: 6.928 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Instrument :
BNA_P
ClientSampleId :
SLCS619

Tgt Ion: 77 Resp: 472415
Ion Ratio Lower Upper
77 100
105 99.9 81.0 121.6
106 95.8 79.2 118.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#8
Phenol
Concen: 36.353 ng/ul
RT: 6.987 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion: 94 Resp: 969133
Ion Ratio Lower Upper
94 100
65 26.1 20.8 31.2
66 34.0 26.2 39.2





#10
Bis(2-Chloroethyl)ether
Concen: 34.182 ng/ul
RT: 7.210 min Scan# 719
Delta R.T. -0.011 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion: 93 Resp: 731352
Ion Ratio Lower Upper
93 100
63 74.2 59.9 89.9
95 32.9 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#12
2-Chlorophenol
Concen: 36.432 ng/ul
RT: 7.346 min Scan# 741
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Tgt Ion:128 Resp: 742568
Ion Ratio Lower Upper
128 100
64 44.2 35.1 52.7
130 32.1 25.8 38.6





#13

2-Methylphenol

Concen: 36.258 ng/ul

RT: 8.234 min Scan# 892

Delta R.T. -0.006 min

Lab File: BP012505.D

Acq: 04 Nov 2022 00:09

Instrument :

BNA_P

ClientSampleId :

SLCS619

Tgt Ion:108 Resp: 743344

Ion Ratio Lower Upper

108 100





107 88.6 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 11/05/2022

Supervised By :mohammad ahmed 11/06/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 32.317 ng/ul

RT: 8.316 min Scan# 906

Delta R.T. -0.006 min

Lab File: BP012505.D

Acq: 04 Nov 2022 00:09




Tgt Ion: 45 Resp: 927283


Ion Ratio Lower Upper

45 100

77 15.7 12.1 18.1


79 12.1 9.7 14.5





#16
Acetophenone
Concen: 36.599 ng/ul
RT: 8.616 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619




Tgt Ion:105 Resp: 1152682
Ion Ratio Lower Upper
105 100
77 74.3 60.3 90.5
51 23.3 19.9 29.9

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022




#17
N-Nitroso-di-n-propylamine
Concen: 37.907 ng/ul
RT: 8.599 min Scan# 954
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09



Tgt Ion: 70 Resp: 574762
Ion Ratio Lower Upper
70 100
42 51.1 41.7 62.5
101 10.1 7.8 11.8
130 21.9 17.4 26.0









#18
4-Methylphenol
Concen: 36.798 ng/ul
RT: 8.563 min Scan# 949
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Instrument :
BNA_P
ClientSampleId :
SLCS619

Tgt Ion:108 Resp: 824558
Ion Ratio Lower Upper
108 100
107 113.5 89.1 133.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#19
Hexachloroethane
Concen: 35.625 ng/ul
RT: 8.851 min Scan# 997
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:117 Resp: 279266
Ion Ratio Lower Upper
117 100
201 106.3 78.5 117.7
199 67.2 49.9 74.9





#22
Nitrobenzene
Concen: 38.135 ng/ul
RT: 8.993 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion: 77 Resp: 859231
Ion Ratio Lower Upper
77 100
123 47.2 37.4 56.2
65 13.0 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#23
Isophorone
Concen: 35.954 ng/ul
RT: 9.522 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion: 82 Resp: 1695186
Ion Ratio Lower Upper
82 100
95 7.0 5.2 7.8
138 18.2 14.2 21.2





#25
2-Nitrophenol
Concen: 42.181 ng/ul
RT: 9.698 min Scan# 1141
Delta R.T. -0.012 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:139 Resp: 453345
Ion Ratio Lower Upper
139 100
65 47.9 37.8 56.8
109 29.5 21.4 32.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#26
2,4-Dimethylphenol
Concen: 37.011 ng/ul
RT: 9.763 min Scan# 1152
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:107 Resp: 873441
Ion Ratio Lower Upper
107 100
121 50.8 40.9 61.3
122 86.5 70.6 105.8





#27
Bis(2-Chloroethoxy)methane
Concen: 35.074 ng/ul
RT: 10.004 min Scan# 1193
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion: 93 Resp: 1058259
Ion Ratio Lower Upper
93 100
95 32.7 26.0 39.0
123 11.5 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#29
2,4-Dichlorophenol
Concen: 37.738 ng/ul
RT: 10.234 min Scan# 1232
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:162 Resp: 761001
Ion Ratio Lower Upper
162 100
164 64.6 51.9 77.9
98 36.2 28.6 43.0





#30
Naphthalene
Concen: 34.356 ng/ul
RT: 10.628 min Scan# 1299
Delta R.T. -0.012 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:128 Resp: 2424739
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 12.9 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#32
4-Chloroaniline
Concen: 31.472 ng/ul
RT: 10.757 min Scan# 1321
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:127 Resp: 936631
Ion Ratio Lower Upper
127 100
129 32.2 26.0 39.0





#33
Hexachlorobutadiene
Concen: 35.253 ng/ul
RT: 10.904 min Scan# 1347
Delta R.T. -0.012 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:225 Resp: 442272
Ion Ratio Lower Upper
225 100
223 62.2 49.0 73.6
227 63.0 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#34
Caprolactam
Concen: 39.386 ng/ul m
RT: 11.563 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:113 Resp: 270248
Ion Ratio Lower Upper
113 100
55 155.4 135.6 203.4
56 118.3 98.9 148.3





#35
4-Chloro-3-methylphenol
Concen: 39.554 ng/ul
RT: 11.887 min Scan# 1513
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:107 Resp: 876421
Ion Ratio Lower Upper
107 100
144 25.9 21.0 31.4
142 80.7 65.4 98.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#36
2-Methylnaphthalene
Concen: 34.933 ng/ul
RT: 12.245 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:142 Resp: 1690278
Ion Ratio Lower Upper
142 100
141 87.5 70.6 106.0





#37
1-Methylnaphthalene
Concen: 35.188 ng/ul
RT: 12.463 min Scan# 1611
Delta R.T. -0.012 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:142 Resp: 1700388
Ion Ratio Lower Upper
142 100
141 91.1 73.1 109.7
116 3.8 3.0 4.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.720 ng/ul
RT: 12.616 min Scan# 1637
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:216 Resp: 905715
Ion Ratio Lower Upper
216 100
214 78.4 62.9 94.3
179 20.6 16.6 24.8
108 16.3 13.2 19.8





#40
Hexachlorocyclopentadiene
Concen: 31.429 ng/ul
RT: 12.587 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619




Tgt Ion:237 Resp: 379308
Ion Ratio Lower Upper
237 100
235 62.2 50.1 75.1
272 13.3 10.6 15.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022




#41
2,4,6-Trichlorophenol
Concen: 38.204 ng/ul
RT: 12.863 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Tgt Ion:196 Resp: 637038
Ion Ratio Lower Upper
196 100
198 96.0 77.2 115.8
200 30.3 25.1 37.7





#42
2,4,5-Trichlorophenol
Concen: 39.066 ng/ul
RT: 12.940 min Scan# 1692
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:196 Resp: 710491
Ion Ratio Lower Upper
196 100
198 96.3 77.8 116.8
200 30.7 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#43
1,1'-Biphenyl
Concen: 34.916 ng/ul
RT: 13.263 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:154 Resp: 2280143
Ion Ratio Lower Upper
154 100
153 39.5 31.9 47.9
76 12.7 10.6 16.0





#44
2-Chloronaphthalene
Concen: 34.923 ng/ul
RT: 13.304 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:162 Resp: 1792136
Ion Ratio Lower Upper
162 100
164 32.6 26.6 39.8
127 39.5 30.9 46.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#45
2-Nitroaniline
Concen: 45.898 ng/ul
RT: 13.528 min Scan# 1792
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion: 65 Resp: 545897
Ion Ratio Lower Upper
65 100
92 73.5 58.5 87.7
138 115.2 91.1 136.7





#47
Dimethylphthalate
Concen: 36.800 ng/ul
RT: 13.898 min Scan# 1855
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619




Tgt Ion:163 Resp: 2357206
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.2 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022




#48
2,6-Dinitrotoluene
Concen: 47.340 ng/ul
RT: 14.028 min Scan# 1877
Delta R.T. 0.000 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Tgt Ion:165 Resp: 503491
Ion Ratio Lower Upper
165 100
89 53.3 43.4 65.2
121 20.7 17.1 25.7





#50
Acenaphthylene
Concen: 36.173 ng/ul
RT: 14.157 min Scan# 1899
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument : BNA_P
ClientSampleId : SLCS619




Tgt Ion:152 Resp: 281845
Ion Ratio Lower Upper
152 100
151 19.7 16.1 24.1
153 12.9 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022




#51
3-Nitroaniline
Concen: 40.547 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Tgt Ion:138 Resp: 528040
Ion Ratio Lower Upper
138 100
108 10.4 8.3 12.5
92 113.7 90.6 135.8





#52
Acenaphthene
Concen: 35.287 ng/ul
RT: 14.498 min Scan# 1957
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619




Tgt Ion:153 Resp: 1932183
Ion Ratio Lower Upper
153 100
152 48.0 38.6 58.0
154 89.7 71.6 107.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022




#53
2,4-Dinitrophenol
Concen: 31.979 ng/ul
RT: 14.569 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Tgt Ion:184 Resp: 229692
Ion Ratio Lower Upper
184 100
63 61.9 47.9 71.9
154 62.0 46.9 70.3





#55
4-Nitrophenol
Concen: 37.425 ng/ul
RT: 14.675 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:109 Resp: 325836
Ion Ratio Lower Upper
109 100
139 137.9 123.1 184.7
65 129.9 107.9 161.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#56
Dibenzofuran
Concen: 35.823 ng/ul
RT: 14.834 min Scan# 2014
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:168 Resp: 2677915
Ion Ratio Lower Upper
168 100
139 37.5 29.1 43.7





#57
2,4-Dinitrotoluene
Concen: 44.299 ng/ul
RT: 14.816 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:165 Resp: 70964
Ion Ratio Lower Upper
165 100
63 45.0 33.1 49.7
182 2.7 2.1 3.1




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 39.636 ng/ul
RT: 15.063 min Scan# 2053
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Tgt Ion:232 Resp: 611837
Ion Ratio Lower Upper
232 100
131 45.4 33.6 50.4
130 2.3 1.8 2.6
166 30.9 24.2 36.2









#59
Diethylphthalate
Concen: 38.136 ng/ul
RT: 15.263 min Scan# 2088
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Instrument :
BNA_P
ClientSampleId :
SLCS619

Tgt Ion:149 Resp: 237689
Ion Ratio Lower Upper
149 100
177 22.3 18.2 27.2
150 12.7 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#61
Fluorene
Concen: 35.621 ng/ul
RT: 15.486 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:166 Resp: 2115767
Ion Ratio Lower Upper
166 100
165 97.9 77.8 116.8
167 13.7 10.6 15.8





#62
4-Chlorophenyl-phenylether
Concen: 36.226 ng/ul
RT: 15.481 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619




Tgt Ion:204 Resp: 1055409
Ion Ratio Lower Upper
204 100
206 32.8 26.5 39.7
141 60.2 47.3 70.9

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022




#63
4-Nitroaniline
Concen: 46.977 ng/ul m
RT: 15.528 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Tgt Ion:138 Resp: 552715
Ion Ratio Lower Upper
138 100
92 51.0 40.6 61.0
108 76.8 56.6 85.0





#66
4,6-Dinitro-2-methylphenol
Concen: 36.805 ng/ul
RT: 15.586 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:198 Resp: 399345
Ion Ratio Lower Upper
198 100
182 4.2 3.8 5.6
77 22.5 19.5 29.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#67
N-Nitrosodiphenylamine
Concen: 35.212 ng/ul
RT: 15.704 min Scan# 2162
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:169 Resp: 1939836
Ion Ratio Lower Upper
169 100
168 66.3 52.9 79.3
167 37.0 29.1 43.7





#68
4-Bromophenyl-phenylether
Concen: 37.343 ng/ul
RT: 16.381 min Scan# 2277
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:248 Resp: 733673
Ion Ratio Lower Upper
248 100
250 96.9 77.4 116.2
141 67.9 56.2 84.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#69
Hexachlorobenzene
Concen: 36.889 ng/ul
RT: 16.492 min Scan# 2296
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Tgt Ion:284 Resp: 869241
Ion Ratio Lower Upper
284 100
142 30.8 27.1 40.7
249 29.7 24.5 36.7





#70

Atrazine

Concen: 28.047 ng/ul

RT: 16.663 min Scan# 2326

Delta R.T. -0.006 min

Lab File: BP012505.D

Acq: 04 Nov 2022 00:09

Instrument :

BNA_P

ClientSampleId :

SLCS619





Tgt Ion:200 Resp: 53524
 Ion Ratio Lower Upper
 200 100
 173 26.4 21.8 32.6
 215 48.0 39.0 58.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 11/05/2022

Supervised By :mohammad ahmed 11/06/2022



#71

Pentachlorophenol

Concen: 34.784 ng/ul




RT: 16.845 min Scan# 2356


Delta R.T. -0.006 min

Lab File: BP012505.D

Acq: 04 Nov 2022 00:09


Tgt Ion:266 Resp: 503383
 Ion Ratio Lower Upper
 266 100
 264 62.8 50.3 75.5
 268 62.8 51.4 77.0





#72
Phenanthrene
Concen: 35.209 ng/ul
RT: 17.239 min Scan# 2423
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619




Tgt Ion:178 Resp: 3638214
Ion Ratio Lower Upper
178 100
179 15.5 12.6 18.8
176 19.5 15.8 23.6

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022




#74
Anthracene
Concen: 35.213 ng/ul
RT: 17.328 min Scan# 2438
Delta R.T. -0.012 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09



Tgt Ion:178 Resp: 3614780
Ion Ratio Lower Upper
178 100
179 15.9 12.5 18.7
176 18.8 15.3 22.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 34.305 ng/uL
RT: 13.222 min Scan# 1741
Delta R.T. -0.012 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Instrument :
BNA_P
ClientSampleId :
SLCS619

Tgt Ion:216 Resp: 959874
Ion Ratio Lower Upper
216 100
214 77.9 62.6 94.0
218 47.7 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#76
Pentachlorobenzene
Concen: 32.261 ng/uL
RT: 14.751 min Scan# 2000
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:250 Resp: 953203
Ion Ratio Lower Upper
250 100
248 61.9 50.6 75.8
252 64.3 50.7 76.1





#77
 Carbazole
 Concen: 36.881 ng/ul
 RT: 17.610 min Scan# 2486
 Delta R.T. -0.006 min
 Lab File: BP012505.D
 Acq: 04 Nov 2022 00:09




Instrument :
 BNA_P
 ClientSampleId :
 SLCS619




Tgt Ion:167 Resp: 3372039
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.1 25.7
 139 13.2 10.4 15.6

Manual Integrations
 APPROVED



Reviewed By :Yogesh Patel 11/05/2022
 Supervised By :mohammad ahmed 11/06/2022




#78
 Di-n-butylphthalate
 Concen: 40.109 ng/ul
 RT: 18.157 min Scan# 2579
 Delta R.T. -0.006 min
 Lab File: BP012505.D
 Acq: 04 Nov 2022 00:09




Tgt Ion:149 Resp: 4260284
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.6 11.4
 104 4.7 4.0 6.0





#80
Fluoranthene
Concen: 39.677 ng/ul
RT: 19.210 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:202 Resp: 4318184
Ion Ratio Lower Upper
202 100
101 11.7 9.9 14.9
100 8.9 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#82
Pyrene
Concen: 38.858 ng/ul
RT: 19.563 min Scan# 2818
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:202 Resp: 4380303
Ion Ratio Lower Upper
202 100
101 14.1 11.8 17.6
100 11.2 9.3 13.9





#83
Butylbenzylphthalate
Concen: 46.513 ng/ul
RT: 20.439 min Scan# 2968
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619




Tgt Ion:149 Resp: 1954795
Ion Ratio Lower Upper
149 100
91 70.2 56.9 85.3
206 21.0 16.2 24.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022




#84
3,3'-Dichlorobenzidine
Concen: 34.283 ng/ul
RT: 21.216 min Scan# 3099
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Tgt Ion:252 Resp: 1307489
Ion Ratio Lower Upper
252 100
254 64.1 51.8 77.8
126 10.1 9.0 13.4





#85
Benzo(a)anthracene
Concen: 35.955 ng/ul
RT: 21.274 min Scan# 3110
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:228 Resp: 4027256
Ion Ratio Lower Upper
228 100
229 20.1 16.2 24.2
226 27.6 22.6 33.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 44.125 ng/ul
RT: 21.192 min Scan# 3095
Delta R.T. -0.011 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:149 Resp: 2835625
Ion Ratio Lower Upper
149 100
167 27.1 22.1 33.1
279 4.4 3.7 5.5





#87
Chrysene
Concen: 35.895 ng/ul
RT: 21.327 min Scan# 3119
Delta R.T. -0.006 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:228 Resp: 3758655
Ion Ratio Lower Upper
228 100
226 30.6 24.9 37.3
229 20.0 16.0 24.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#89
Di-n-octyl phthalate
Concen: 50.074 ng/ul
RT: 22.110 min Scan# 3251
Delta R.T. -0.012 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:149 Resp: 4630005





#90
Benzo(b)fluoranthene
Concen: 39.066 ng/ul
RT: 22.945 min Scan# 3394
Delta R.T. -0.012 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion:252 Resp: 3834755
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.8
125 10.2 8.7 13.1




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#91
Benzo(k)fluoranthene
Concen: 39.397 ng/ul
RT: 22.992 min Scan# 3401
Delta R.T. -0.011 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Tgt Ion:252 Resp: 3685544
Ion Ratio Lower Upper
252 100
253 22.4 18.0 27.0
125 10.0 8.7 13.1









#93
Benzo(a)pyrene
Concen: 37.206 ng/ul
RT: 23.568 min Scan# 3499
Delta R.T. -0.012 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Instrument :
BNA_P
ClientSampleId :
SLCS619

Tgt Ion:252 Resp: 318932
Ion Ratio Lower Upper
252 100
253 22.0 18.0 27.0
125 10.9 9.4 14.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/05/2022
Supervised By :mohammad ahmed 11/06/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 35.663 ng/ul
RT: 26.145 min Scan# 3937
Delta R.T. -0.017 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09


Tgt Ion:276 Resp: 3809327
Ion Ratio Lower Upper
276 100
138 24.2 20.7 31.1
277 25.3 20.2 30.4





#95
Dibenzo(a,h)anthracene
Concen: 34.303 ng/ul
RT: 26.156 min Scan# 3942
Delta R.T. -0.017 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09




Instrument :
BNA_P
ClientSampleId :
SLCS619



Tgt Ion: 278 Resp: 3280798
Ion Ratio Lower Upper
278 100
139 15.8 13.2 19.8
279 24.1 19.4 29.0

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 11/05/2022
Supervised By : mohammad ahmed 11/06/2022



#96
Benzo(g,h,i)perylene
Concen: 33.835 ng/ul
RT: 26.903 min Scan# 4066
Delta R.T. -0.017 min
Lab File: BP012505.D
Acq: 04 Nov 2022 00:09

Tgt Ion: 276 Resp: 3194793
Ion Ratio Lower Upper
276 100
138 21.1 17.8 26.8
277 23.9 19.5 29.3

