

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111020\
 Data File : BP003897.D
 Acq On : 10 Nov 2020 12:04
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 10 13:42:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 13:41:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.517	0.440	14.9	82	0.00
3	Pyridine	1.516	1.315	13.3	82	0.00
4	n-Nitrosodimethylamine	0.698	0.630	9.7	85	0.00
5 S	2-Fluorophenol	1.198	1.121	6.4	88	0.00
6	Aniline	1.953	1.786	8.6	85	0.00
7 S	Phenol-d6	1.720	1.580	8.1	86	0.00
8	2-Chlorophenol	1.272	1.197	5.9	88	0.00
9	Benzaldehyde	0.873	0.732	16.2	90	0.00
10 C	Phenol	1.660	1.512	8.9	86	0.00
11	bis(2-Chloroethyl)ether	1.334	1.199	10.1	85	0.00
12	1,3-Dichlorobenzene	1.559	1.448	7.1	89	0.00
13 C	1,4-Dichlorobenzene	1.572	1.458	7.3	88	0.00
14	1,2-Dichlorobenzene	1.501	1.399	6.8	88	0.00
15	Benzyl Alcohol	1.191	1.163	2.4	89	0.00
16	2,2'-oxybis(1-Chloropropane	2.247	1.915	14.8	81	0.00
17	2-Methylphenol	1.090	1.019	6.5	87	0.00
18	Hexachloroethane	0.559	0.551	1.4	91	0.00
19 P	n-Nitroso-di-n-propylamine	1.028	0.968	5.8	86	0.00
20	3+4-Methylphenols	1.455	1.371	5.8	87	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
22	Acetophenone	0.537	0.497	7.4	86	0.00
23 S	Nitrobenzene-d5	0.408	0.375	8.1	85	0.00
24	Nitrobenzene	0.406	0.380	6.4	87	0.00
25	Isophorone	0.694	0.661	4.8	87	0.00
26 C	2-Nitrophenol	0.158	0.175	-10.8	97	0.00
27	2,4-Dimethylphenol	0.264	0.252	4.5	89	0.00
28	bis(2-Chloroethoxy)methane	0.476	0.431	9.5	85	0.00
29 C	2,4-Dichlorophenol	0.319	0.317	0.6	91	0.00
30	1,2,4-Trichlorobenzene	0.388	0.370	4.6	90	0.00
31	Naphthalene	1.073	0.986	8.1	87	0.00
32	Benzoic acid	0.161	0.155	3.7	85	0.00
33	4-Chloroaniline	0.456	0.429	5.9	87	0.00
34 C	Hexachlorobutadiene	0.253	0.248	2.0	93	0.00
35	Caprolactam	0.098	0.103	-5.1	92	0.00
36 C	4-Chloro-3-methylphenol	0.344	0.333	3.2	89	0.00
37	2-Methylnaphthalene	0.768	0.712	7.3	87	0.00
38	1-Methylnaphthalene	0.733	0.673	8.2	87	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	0.662	0.623	5.9	89	0.00
41 P	Hexachlorocyclopentadiene	0.334	0.303	9.3	83	0.00
42 S	2,4,6-Tribromophenol	0.250	0.261	-4.4	95	0.00
43 C	2,4,6-Trichlorophenol	0.409	0.414	-1.2	93	0.00
44	2,4,5-Trichlorophenol	0.431	0.419	2.8	90	0.00
45 S	2-Fluorobiphenyl	1.431	1.327	7.3	89	0.00
46	1,1'-Biphenyl	1.553	1.423	8.4	88	0.00
47	2-Chloronaphthalene	1.274	1.163	8.7	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111020\
 Data File : BP003897.D
 Acq On : 10 Nov 2020 12:04
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 10 13:42:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 13:41:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.380	0.387	-1.8	91	0.00
49	Acenaphthylene	1.871	1.736	7.2	88	0.00
50	Dimethylphthalate	1.563	1.455	6.9	88	0.00
51	2,6-Dinitrotoluene	0.321	0.314	2.2	90	0.00
52 C	Acenaphthene	1.246	1.162	6.7	89	0.00
53	3-Nitroaniline	0.355	0.349	1.7	91	0.00
54 P	2,4-Dinitrophenol	0.138	0.137	0.7	91	0.00
55	Dibenzofuran	1.851	1.692	8.6	88	0.00
56 P	4-Nitrophenol	0.256	0.256	0.0	89	0.00
57	2,4-Dinitrotoluene	0.431	0.438	-1.6	92	0.00
58	Fluorene	1.482	1.363	8.0	88	0.00
59	2,3,4,6-Tetrachlorophenol	0.388	0.380	2.1	90	0.00
60	Diethylphthalate	1.531	1.449	5.4	90	0.00
61	4-Chlorophenyl-phenylether	0.810	0.757	6.5	90	0.00
62	4-Nitroaniline	0.370	0.376	-1.6	93	0.00
63	Azobenzene	1.445	1.296	10.3	84	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	0.098	0.099	-1.0	93	0.00
66 c	n-Nitrosodiphenylamine	0.612	0.571	6.7	89	0.00
67	4-Bromophenyl-phenylether	0.234	0.225	3.8	92	0.00
68	Hexachlorobenzene	0.267	0.256	4.1	93	0.00
69	Atrazine	0.201	0.200	0.5	92	0.00
70 C	Pentachlorophenol	0.128	0.115	10.2	82	0.00
71	Phenanthrene	1.122	1.028	8.4	89	0.00
72	Anthracene	1.084	1.006	7.2	90	0.00
73	Carbazole	1.006	0.941	6.5	90	0.00
74	Di-n-butylphthalate	1.164	1.176	-1.0	93	0.00
75 C	Fluoranthene	1.295	1.204	7.0	89	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
77	Benzydine	0.472	0.533	-12.9	96	0.00
78	Pyrene	1.357	1.315	3.1	88	0.00
79 S	Terphenyl-d14	1.035	1.011	2.3	91	0.00
80	Butylbenzylphthalate	0.506	0.567	-12.1	96	0.00
81	Benzo(a)anthracene	1.319	1.239	6.1	87	0.00
82	3,3'-Dichlorobenzidine	0.455	0.463	-1.8	92	0.00
83	Chrysene	1.267	1.189	6.2	87	0.00
84	Bis(2-ethylhexyl)phthalate	0.745	0.791	-6.2	93	0.00
85 c	Di-n-octyl phthalate	1.242	1.358	-9.3	95	0.00
86 I	Perylene-d12	1.000	1.000	0.0	90	0.00
87	Indeno(1,2,3-cd)pyrene	1.443	1.299	10.0	90	0.00
88	Benzo(b)fluoranthene	1.309	1.191	9.0	93	0.00
89	Benzo(k)fluoranthene	1.292	1.091	15.6	83	0.00
90 C	Benzo(a)pyrene	1.212	1.084	10.6	89	0.00
91	Dibenzo(a,h)anthracene	1.203	1.085	9.8	89	0.00
92	Benzo(g,h,i)perylene	1.164	1.057	9.2	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111020\
Data File : BP003897.D
Acq On : 10 Nov 2020 12:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 10 13:42:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 10 13:41:24 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0