

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111324\
 Data File : BP023096.D
 Acq On : 14 Nov 2024 00:57
 Operator : RC/JU
 Sample : PB164783BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Nov 14 01:38:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 13 04:49:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.581	152	71054	20.000	ng/ul	0.00
20) Naphthalene-d8	10.328	136	282585	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.193	164	237040	20.000	ng/ul	-0.01
74) Phenanthrene-d10	17.004	188	591523	20.000	ng/ul	#-0.01
79) Chrysene-d12	21.433	240	683875	20.000	ng/ul	-0.01
88) Perylene-d12	24.639	264	803376	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.134	96	10208	6.906	ng/uL	0.00
4) Pyridine-d5	3.546	84	153196	34.882	ng/ul	0.00
7) Phenol-d5	6.787	99	201772	38.409	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.928	67	112389	36.415	ng/ul	0.00
11) 2-Chlorophenol-d4	7.128	132	149261	38.796	ng/ul	0.00
15) 4-Methylphenol-d8	8.293	113	162408	37.569	ng/ul	0.00
21) Nitrobenzene-d5	8.728	128	73410	37.315	ng/ul	0.00
24) 2-Nitrophenol-d4	9.440	143	88387	37.686	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.975	165	187441	38.163	ng/ul	0.00
31) 4-Chloroaniline-d4	10.481	131	190634	32.534	ng/ul	0.00
46) Dimethylphthalate-d6	13.604	166	617928	36.170	ng/ul	0.00
49) Acenaphthylene-d8	13.881	160	665424	37.329	ng/ul	-0.02
54) 4-Nitrophenol-d4	14.451	143	94124	36.661	ng/ul	0.00
60) Fluorene-d10	15.216	176	554121	36.994	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.351	200	127590	36.007	ng/ul	-0.01
73) Anthracene-d10	17.092	188	949167	37.897	ng/ul	-0.02
81) Pyrene-d10	19.486	212	1254879	39.838	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.427	264	1477123	38.600	ng/ul	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.170	88	22423	13.128	ng/uL	94
5) Pyridine	3.564	79	151650	34.383	ng/ul	98
6) Benzaldehyde	6.752	77	52318	17.307	ng/ul	96
8) Phenol	6.811	94	210156	38.310	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.022	93	151518	36.756	ng/ul	96
12) 2-Chlorophenol	7.163	128	154063	38.570	ng/ul	96
13) 2-Methylphenol	8.028	108	150050	36.644	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.093	45	176198	37.758	ng/ul	95
16) Acetophenone	8.399	105	258148	35.537	ng/ul	96
17) N-Nitroso-di-n-propyla...	8.381	70	134856	37.179	ng/ul	98
18) 4-Methylphenol	8.358	108	168842	37.331	ng/ul	97
19) Hexachloroethane	8.628	117	67873	35.134	ng/ul	95
22) Nitrobenzene	8.769	77	210314	37.257	ng/ul	100
23) Isophorone	9.287	82	381109	37.758	ng/ul	98
25) 2-Nitrophenol	9.469	139	93191	37.645	ng/ul	91
26) 2,4-Dimethylphenol	9.534	107	170094	31.820	ng/ul	100
27) Bis(2-Chloroethoxy)met...	9.757	93	207921	37.438	ng/ul	97
29) 2,4-Dichlorophenol	9.999	162	185301	38.089	ng/ul	96
30) Naphthalene	10.381	128	496833	36.190	ng/ul	98
32) 4-Chloroaniline	10.504	127	128546	22.644	ng/ul	93
33) Hexachlorobutadiene	10.652	225	155368	35.295	ng/ul	98
34) Caprolactam	11.293	113	53055	36.664	ng/ul	94
35) 4-Chloro-3-methylphenol	11.646	107	205060	39.181	ng/ul	94
36) 2-Methylnaphthalene	11.993	142	378340	36.503	ng/ul	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111324\
 Data File : BP023096.D
 Acq On : 14 Nov 2024 00:57
 Operator : RC/JU
 Sample : PB164783BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Nov 14 01:38:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 13 04:49:01 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.210	142	376171	35.503	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.369	216	290372	36.285	ng/ul	97
40) Hexachlorocyclopentadiene	12.340	237	112395	26.740	ng/ul	97
41) 2,4,6-Trichlorophenol	12.622	196	182271	37.405	ng/ul	97
42) 2,4,5-Trichlorophenol	12.710	196	200049	38.080	ng/ul	98
43) 1,1'-Biphenyl	13.022	154	554762	36.524	ng/ul	100
44) 2-Chloronaphthalene	13.063	162	451357	36.352	ng/ul	98
45) 2-Nitroaniline	13.281	65	138834	39.823	ng/ul	90
47) Dimethylphthalate	13.663	163	608876	36.138	ng/ul	99
48) 2,6-Dinitrotoluene	13.793	165	128902	39.331	ng/ul	96
50) Acenaphthylene	13.910	152	668499	37.099	ng/ul	99
51) 3-Nitroaniline	14.134	138	111038	39.336	ng/ul	89
52) Acenaphthene	14.257	153	481083	36.775	ng/ul	97
53) 2,4-Dinitrophenol	14.357	184	79283	34.094	ng/ul	97
55) 4-Nitrophenol	14.469	109	111128	36.996	ng/ul	92
56) Dibenzofuran	14.598	168	732263	36.229	ng/ul	97
57) 2,4-Dinitrotoluene	14.592	165	201968	38.666	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	14.851	232	187848	36.581	ng/ul	98
59) Diethylphthalate	15.040	149	602614	36.164	ng/ul	99
61) Fluorene	15.263	166	608713	36.796	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.257	204	355696	36.516	ng/ul	98
63) 4-Nitroaniline	15.304	138	117277	44.649	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.369	198	137137	36.229	ng/ul	99
67) N-Nitrosodiphenylamine	15.487	169	528748	38.684	ng/ul	99
68) 4-Bromophenyl-phenylether	16.175	248	247987	37.022	ng/ul	96
69) Hexachlorobenzene	16.292	284	313538	37.182	ng/ul	99
70) Atrazine	16.463	200	181481	27.467	ng/ul	99
71) Pentachlorophenol	16.657	266	184760	35.291	ng/ul	98
72) Phenanthrene	17.039	178	1092254	39.109	ng/ul	100
74) Anthracene	17.128	178	1071713	38.167	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	12.987	216	282548	36.178	ng/uL	96
76) Pentachlorobenzene	14.528	250	318648	35.733	ng/uL	98
77) Carbazole	17.422	167	951580	38.843	ng/ul	99
78) Di-n-butylphthalate	18.010	149	1093376	40.219	ng/ul	100
80) Fluoranthene	19.145	202	1459622	40.222	ng/ul	99
82) Pyrene	19.522	202	1565317	40.152	ng/ul	99
83) Butylbenzylphthalate	20.469	149	538172	41.929	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.339	252	568747	37.050	ng/ul	99
85) Benzo(a)anthracene	21.416	228	1617354	38.731	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.351	149	757031	41.504	ng/ul	97
87) Chrysene	21.480	228	1498903	38.148	ng/ul	99
89) Di-n-octyl phthalate	22.563	149	1274887	39.115	ng/ul	100
90) Benzo(b)fluoranthene	23.639	252	1726818	38.618	ng/ul	100
91) Benzo(k)fluoranthene	23.698	252	1714185	38.942	ng/ul#	99
93) Benzo(a)pyrene	24.486	252	1550487	38.357	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	28.274	276	1177582	20.910	ng/ul	97
95) Dibenzo(a,h)anthracene	28.345	278	1654234	36.176	ng/ul	99
96) Benzo(g,h,i)perylene	29.421	276	1603507	37.486	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111324\
Data File : BP023096.D
Acq On : 14 Nov 2024 00:57
Operator : RC/JU
Sample : PB164783BS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :


Quant Time: Nov 14 01:38:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 13 04:49:01 2024
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 13.128 ng/uL
RT: 3.170 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:	88	Resp:	22423
Ion	Ratio	Lower	Upper
88	100		
43	33.1	24.2	36.4
58	68.4	59.2	88.8



#5
Pyridine
Concen: 34.383 ng/uL
RT: 3.564 min Scan# 98
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion:	79	Resp:	151650
Ion	Ratio	Lower	Upper
79	100		
52	75.7	59.5	89.3
51	36.3	30.4	45.6









#6
Benzaldehyde
Concen: 17.307 ng/ul
RT: 6.752 min Scan# 640
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 52318
Ion Ratio Lower Upper
77 100
105 83.3 72.5 108.7
106 82.0 65.5 98.3



#8
Phenol
Concen: 38.310 ng/ul
RT: 6.811 min Scan# 650
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57


Tgt Ion: 94 Resp: 210156
Ion Ratio Lower Upper
94 100
65 31.6 26.2 39.4
66 48.5 40.2 60.4








#10
Bis(2-Chloroethyl)ether
Concen: 36.756 ng/ul
RT: 7.022 min Scan# 686
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Instrument :
BNA_P
ClientSampleId :







Tgt Ion: 93 Resp: 151518
Ion Ratio Lower Upper
93 100
63 71.7 60.6 91.0
95 34.2 29.0 43.4



#12
2-Chlorophenol
Concen: 38.570 ng/ul
RT: 7.163 min Scan# 710
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion: 128 Resp: 154063
Ion Ratio Lower Upper
128 100
64 51.0 39.0 58.6
130 30.2 26.4 39.6









#13
2-Methylphenol
Concen: 36.644 ng/ul
RT: 8.028 min Scan# 858
Delta R.T. -0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 150050
Ion Ratio Lower Upper
108 100
107 91.6 73.3 109.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 37.758 ng/ul
RT: 8.093 min Scan# 868
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion: 45 Resp: 176198
Ion Ratio Lower Upper
45 100
77 19.0 18.2 27.2
79 14.4 11.7 17.5









#16
Acetophenone
Concen: 35.537 ng/ul
RT: 8.399 min Scan# 916
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 258148
Ion Ratio Lower Upper
105 100
77 83.3 69.4 104.0
51 26.8 23.2 34.8



#17
N-Nitroso-di-n-propylamine
Concen: 37.179 ng/ul
RT: 8.381 min Scan# 917
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion: 70 Resp: 134856
Ion Ratio Lower Upper
70 100
42 51.7 42.3 63.5
101 9.7 7.4 11.2
130 21.3 15.4 23.0





#18

4-Methylphenol

Concen: 37.331 ng/ul

RT: 8.358 min Scan# 913

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 168842

Ion Ratio Lower Upper

108 100

107 111.4 91.5 137.3



#19

Hexachloroethane

Concen: 35.134 ng/ul

RT: 8.628 min Scan# 959

Delta R.T. -0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:117 Resp: 67873


Ion Ratio Lower Upper

117 100

201 138.6 105.0 157.6

199 85.2 65.9 98.9









#22
Nitrobenzene
Concen: 37.257 ng/ul
RT: 8.769 min Scan# 983
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 210314
Ion Ratio Lower Upper
77 100
123 40.4 32.2 48.4
65 13.6 10.9 16.3



#23
Isophorone
Concen: 37.758 ng/ul
RT: 9.287 min Scan# 1071
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion: 82 Resp: 381109
Ion Ratio Lower Upper
82 100
95 9.2 7.6 11.4
138 15.8 13.4 20.2









#25
2-Nitrophenol
Concen: 37.645 ng/ul
RT: 9.469 min Scan# 1102
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 93191
Ion Ratio Lower Upper
139 100
65 56.0 52.8 79.2
109 46.4 38.9 58.3



#26
2,4-Dimethylphenol
Concen: 31.820 ng/ul
RT: 9.534 min Scan# 1113
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion:107 Resp: 170094
Ion Ratio Lower Upper
107 100
121 43.4 34.4 51.6
122 70.8 56.7 85.1





#27

Bis(2-Chloroethoxy)methane

Concen: 37.438 ng/ul

RT: 9.757 min Scan# 1151

Delta R.T. 0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion: 93 Resp: 207921

Ion Ratio Lower Upper

93 100

95 32.1 27.0 40.6

123 9.9 8.2 12.4



#29

2,4-Dichlorophenol

Concen: 38.089 ng/ul

RT: 9.999 min Scan# 1192

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:162 Resp: 185301


Ion Ratio Lower Upper

162 100

164 66.9 51.4 77.0

98 36.6 27.0 40.6









#30
Naphthalene
Concen: 36.190 ng/ul
RT: 10.381 min Scan# 1257
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Instrument :
BNA_P
ClientSampleId :

Tgt Ion:128 Resp: 496833




Ion	Ratio	Lower	Upper
128	100		
129	11.4	8.7	13.1
127	13.6	10.2	15.2




#32
4-Chloroaniline
Concen: 22.644 ng/ul
RT: 10.504 min Scan# 1278
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion:127 Resp: 128546

Ion	Ratio	Lower	Upper
127	100		
129	28.9	26.4	39.6





#33

Hexachlorobutadiene

Concen: 35.295 ng/ul

RT: 10.652 min Scan# 1412

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:225 Resp: 155368

Ion Ratio Lower Upper

225 100

223 62.8 49.7 74.5

227 61.5 51.4 77.0



#34

Caprolactam

Concen: 36.664 ng/ul

RT: 11.293 min Scan# 1412

Delta R.T. 0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:113 Resp: 53055


Ion Ratio Lower Upper

113 100

55 142.5 123.0 184.6

56 123.1 99.8 149.8









#35
4-Chloro-3-methylphenol
Concen: 39.181 ng/ul
RT: 11.646 min Scan# 1472
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 205060
Ion Ratio Lower Upper
107 100
144 24.0 18.8 28.2
142 68.9 59.8 89.8



#36
2-Methylnaphthalene
Concen: 36.503 ng/ul
RT: 11.993 min Scan# 1531
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion:142 Resp: 378340
Ion Ratio Lower Upper
142 100
141 87.9 73.1 109.7





#37

1-Methylnaphthalene

Concen: 35.503 ng/ul

RT: 12.210 min Scan# 1569

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:142 Resp: 376171

Ion Ratio Lower Upper

142 100

141 92.5 73.8 110.8

116 4.7 4.2 6.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 36.285 ng/ul

RT: 12.369 min Scan# 1595

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Tgt Ion:216 Resp: 290372




Ion Ratio Lower Upper


216 100

214 78.4 59.4 89.2

179 21.6 17.3 25.9

108 15.2 11.8 17.6





#40

Hexachlorocyclopentadiene

Concen: 26.740 ng/ul

RT: 12.340 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:237 Resp: 112395

Ion Ratio Lower Upper

237 100

235 63.1 48.6 73.0

272 13.3 10.2 15.2



#41

2,4,6-Trichlorophenol

Concen: 37.405 ng/ul

RT: 12.622 min Scan# 1638

Delta R.T. 0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:196 Resp: 182271


Ion Ratio Lower Upper

196 100

198 97.4 80.5 120.7

200 30.8 25.2 37.8





#42

2,4,5-Trichlorophenol

Concen: 38.080 ng/ul

RT: 12.710 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:196 Resp: 200049

Ion Ratio Lower Upper

196 100

198 97.5 79.6 119.4

200 32.3 25.0 37.6



#43

1,1'-Biphenyl

Concen: 36.524 ng/ul

RT: 13.022 min Scan# 1706

Delta R.T. 0.012 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:154 Resp: 554762


Ion Ratio Lower Upper

154 100

153 40.0 32.0 48.0

76 11.0 8.4 12.6





#44

2-Chloronaphthalene

Concen: 36.352 ng/ul

RT: 13.063 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:162 Resp: 451357

Ion Ratio Lower Upper

162 100

164 33.7 26.6 40.0

127 36.0 30.2 45.2



#45

2-Nitroaniline

Concen: 39.823 ng/ul

RT: 13.281 min Scan# 1750

Delta R.T. -0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion: 65 Resp: 138834


Ion Ratio Lower Upper

65 100

92 68.9 54.2 81.2

138 97.3 65.1 97.7









#47
Dimethylphthalate
Concen: 36.138 ng/ul
RT: 13.663 min Scan# 1815
Delta R.T. 0.012 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 608876
Ion Ratio Lower Upper
163 100
194 4.4 3.8 5.6
164 10.5 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 39.331 ng/ul
RT: 13.793 min Scan# 1837
Delta R.T. 0.012 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion:165 Resp: 128902
Ion Ratio Lower Upper
165 100
89 59.7 50.6 75.8
121 19.3 15.8 23.8









#50
Acenaphthylene
Concen: 37.099 ng/ul
RT: 13.910 min Scan# 1857
Delta R.T. -0.012 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:	152	Resp:	668499
Ion Ratio	Lower	Upper	
152	100		
151	20.1	15.9	23.9
153	13.1	10.2	15.2



#51
3-Nitroaniline
Concen: 39.336 ng/ul
RT: 14.134 min Scan# 1895
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57


Tgt Ion:	138	Resp:	111038
Ion Ratio	Lower	Upper	
138	100		
108	12.9	12.7	19.1
92	120.9	107.8	161.8








#52
Acenaphthene
Concen: 36.775 ng/ul
RT: 14.257 min Scan# 1916
Delta R.T. -0.012 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57


Instrument :
BNA_P
ClientSampleId :





Tgt Ion:153 Resp: 481083
Ion Ratio Lower Upper
153 100
152 46.7 39.6 59.4
154 86.3 70.9 106.3




#53
2,4-Dinitrophenol
Concen: 34.094 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. 0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57



Tgt Ion:184 Resp: 79283
Ion Ratio Lower Upper
184 100
63 60.0 50.1 75.1
154 70.5 58.2 87.4









#55
4-Nitrophenol
Concen: 36.996 ng/ul
RT: 14.469 min Scan# 1952
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 111128
Ion Ratio Lower Upper
109 100
139 89.4 64.6 96.8
65 106.8 80.6 120.8



#56
Dibenzofuran
Concen: 36.229 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. -0.012 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion:168 Resp: 732263
Ion Ratio Lower Upper
168 100
139 38.6 29.7 44.5





#57

2,4-Dinitrotoluene

Concen: 38.666 ng/ul

RT: 14.592 min Scan# 1

Delta R.T. -0.018 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:165 Resp: 201968

Ion Ratio Lower Upper

165 100

63 59.5 41.3 61.9

182 3.0 2.5 3.7



#58

2,3,4,6-Tetrachlorophenol

Concen: 36.581 ng/ul

RT: 14.851 min Scan# 2017

Delta R.T. 0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Tgt Ion:232 Resp: 187848




Ion Ratio Lower Upper


232 100

131 34.6 27.0 40.6

130 2.0 1.9 2.9

166 25.8 21.8 32.6





#59

Diethylphthalate

Concen: 36.164 ng/ul

RT: 15.040 min Scan# 2049

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 602614

Ion Ratio Lower Upper

149 100

177 24.1 19.7 29.5

150 12.7 9.8 14.6



#61

Fluorene

Concen: 36.796 ng/ul

RT: 15.263 min Scan# 2087

Delta R.T. -0.012 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:166 Resp: 608713


Ion Ratio Lower Upper

166 100

165 98.4 79.5 119.3

167 13.3 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 36.516 ng/ul
RT: 15.257 min Scan# 2086
Delta R.T. -0.018 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 355696
Ion Ratio Lower Upper
204 100
206 33.1 27.2 40.8
141 53.7 44.6 66.8



#63
4-Nitroaniline
Concen: 44.649 ng/ul
RT: 15.304 min Scan# 2094
Delta R.T. -0.018 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion:138 Resp: 117277
Ion Ratio Lower Upper
138 100
92 60.3 53.9 80.9
108 126.3 106.6 159.8









#66
4,6-Dinitro-2-methylphenol
Concen: 36.229 ng/ul
RT: 15.369 min Scan# 2105
Delta R.T. -0.012 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 137137
Ion Ratio Lower Upper
198 100
182 4.6 3.8 5.8
77 24.2 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 38.684 ng/ul
RT: 15.487 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion:169 Resp: 528748
Ion Ratio Lower Upper
169 100
168 67.7 53.4 80.0
167 36.2 29.4 44.2





#68

4-Bromophenyl-phenylether

Concen: 37.022 ng/ul

RT: 16.175 min Scan# 2242

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:248 Resp: 247987

Ion Ratio Lower Upper

248 100

250 101.3 77.1 115.7

141 62.4 50.7 76.1



#69

Hexachlorobenzene

Concen: 37.182 ng/ul

RT: 16.292 min Scan# 2262

Delta R.T. -0.012 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:284 Resp: 313538


Ion Ratio Lower Upper

284 100

142 27.1 21.5 32.3

249 31.3 25.6 38.4





#70

Atrazine

Concen: 27.467 ng/ul

RT: 16.463 min Scan# 2292

Delta R.T. -0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 181481

Ion Ratio Lower Upper

200 100

173 24.1 18.6 27.8

215 47.6 38.3 57.5



#71

Pentachlorophenol

Concen: 35.291 ng/ul

RT: 16.657 min Scan# 2324

Delta R.T. -0.012 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:266 Resp: 184760


Ion Ratio Lower Upper

266 100

264 62.3 50.7 76.1

268 63.3 51.9 77.9





#72

Phenanthrene

Concen: 39.109 ng/ul

RT: 17.039 min Scan# 2403

Delta R.T. -0.018 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:178 Resp: 1092254

Ion Ratio Lower Upper

178 100

179 15.8 12.7 19.1

176 19.4 15.4 23.2



#74

Anthracene

Concen: 38.167 ng/ul

RT: 17.128 min Scan# 2404

Delta R.T. -0.024 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:178 Resp: 1071713


Ion Ratio Lower Upper

178 100

179 15.6 12.4 18.6

176 18.3 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 36.178 ng/uL

RT: 12.987 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:216 Resp: 282548

Ion Ratio Lower Upper

216 100

214 78.8 60.2 90.2

218 50.3 38.1 57.1



#76

Pentachlorobenzene

Concen: 35.733 ng/uL

RT: 14.528 min Scan# 1962

Delta R.T. -0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:250 Resp: 318648


Ion Ratio Lower Upper

250 100

248 64.4 52.2 78.2

252 64.2 50.1 75.1









#77
 Carbazole
 Concen: 38.843 ng/ul
 RT: 17.422 min Scan# 2455
 Delta R.T. -0.018 min
 Lab File: BP023096.D
 Acq: 14 Nov 2024 00:57




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:167 Resp: 951580
 Ion Ratio Lower Upper
 167 100
 166 21.1 17.5 26.3
 139 12.4 9.6 14.4



#78
 Di-n-butylphthalate
 Concen: 40.219 ng/ul
 RT: 18.010 min Scan# 2554
 Delta R.T. -0.006 min
 Lab File: BP023096.D
 Acq: 14 Nov 2024 00:57

Tgt Ion:149 Resp: 1093376
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.2 10.8
 104 6.1 5.0 7.4









#80
Fluoranthene
Concen: 40.222 ng/ul
RT: 19.145 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:202 Resp: 1459622
Ion Ratio Lower Upper
202 100
101 6.3 5.2 7.8
100 4.6 4.1 6.1



#82
Pyrene
Concen: 40.152 ng/ul
RT: 19.522 min Scan# 2811
Delta R.T. -0.006 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion:202 Resp: 1565317
Ion Ratio Lower Upper
202 100
101 7.5 6.2 9.2
100 6.5 4.9 7.3





#83

Butylbenzylphthalate

Concen: 41.929 ng/ul

RT: 20.469 min Scan# 20

Delta R.T. -0.012 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 538172

Ion Ratio Lower Upper

149 100

91 71.1 55.6 83.4

206 25.4 20.2 30.2



#84

3,3'-Dichlorobenzidine

Concen: 37.050 ng/ul

RT: 21.339 min Scan# 3120

Delta R.T. -0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:252 Resp: 568747


Ion Ratio Lower Upper

252 100

254 64.1 52.0 78.0

126 6.2 5.0 7.6





#85

Benzo(a)anthracene

Concen: 38.731 ng/ul

RT: 21.416 min Scan# 3133

Delta R.T. -0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 1617354

Ion Ratio Lower Upper

228 100

229 19.1 15.5 23.3

226 26.8 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 41.504 ng/ul

RT: 21.351 min Scan# 3122

Delta R.T. 0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:149 Resp: 757031


Ion Ratio Lower Upper

149 100

167 29.1 24.5 36.7

279 5.9 5.5 8.3





#87

Chrysene

Concen: 38.148 ng/ul

RT: 21.480 min Scan# 3144

Delta R.T. -0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 1498903

Ion Ratio Lower Upper

228 100

226 30.3 24.2 36.2

229 19.5 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 39.115 ng/ul




RT: 22.563 min Scan# 3328


Delta R.T. -0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Tgt Ion:149 Resp: 1274887





#90

Benzo(b)fluoranthene

Concen: 38.618 ng/ul

RT: 23.639 min Scan# 3511

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 1726818

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.2

125 4.8 3.7 5.5



#91

Benzo(k)fluoranthene

Concen: 38.942 ng/ul

RT: 23.698 min Scan# 3521

Delta R.T. -0.012 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57




Tgt Ion:252 Resp: 1714185


Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.8


125 4.6 3.0 4.4#








#93
Benzo(a)pyrene
Concen: 38.357 ng/ul
RT: 24.486 min Scan# 3655
Delta R.T. -0.024 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:252 Resp: 1550487
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 5.4 4.2 6.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.910 ng/ul
RT: 28.274 min Scan# 4299
Delta R.T. -0.029 min
Lab File: BP023096.D
Acq: 14 Nov 2024 00:57

Tgt Ion:276 Resp: 1177582
Ion Ratio Lower Upper
276 100
138 9.7 7.0 10.4
277 23.5 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 36.176 ng/ul

RT: 28.345 min Scan# 4411

Delta R.T. 0.000 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 1654234

Ion Ratio Lower Upper

278 100

139 7.8 5.4 8.0

279 23.8 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 37.486 ng/ul

RT: 29.421 min Scan# 4494

Delta R.T. -0.006 min

Lab File: BP023096.D

Acq: 14 Nov 2024 00:57

Tgt Ion:276 Resp: 1603507

Ion Ratio Lower Upper

276 100

138 9.9 7.1 10.7

277 23.9 19.4 29.2

