

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111324\
 Data File : BP023091.D
 Acq On : 13 Nov 2024 21:35
 Operator : RC/JU
 Sample : P4724-23
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH9E2

Quant Time: Nov 14 01:36:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 13 04:49:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.581	152	83113	20.000	ng/u1	0.00
20) Naphthalene-d8	10.322	136	319325	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.198	164	255794	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.004	188	634841	20.000	ng/u1	#-0.01
79) Chrysene-d12	21.427	240	732471	20.000	ng/u1	-0.02
88) Perylene-d12	24.633	264	857963	20.000	ng/u1	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	6695	3.872	ng/uL	0.00
4) Pyridine-d5	3.552	84	85812	16.704	ng/u1	0.00
7) Phenol-d5	6.787	99	123314	20.068	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.928	67	73563	20.377	ng/u1	0.00
11) 2-Chlorophenol-d4	7.128	132	94935	21.095	ng/u1	0.00
15) 4-Methylphenol-d8	8.293	113	100619	19.899	ng/u1	0.00
21) Nitrobenzene-d5	8.722	128	46299	20.827	ng/u1	0.00
24) 2-Nitrophenol-d4	9.440	143	55060	20.775	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.975	165	114454	20.622	ng/u1	0.00
31) 4-Chloroaniline-d4	10.481	131	127646	19.278	ng/u1	0.00
46) Dimethylphthalate-d6	13.599	166	423233	22.957	ng/u1	-0.01
49) Acenaphthylene-d8	13.881	160	438772	22.809	ng/u1	-0.02
54) 4-Nitrophenol-d4	14.481	143	49067	17.710	ng/u1	0.03
60) Fluorene-d10	15.216	176	380719	23.554	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.357	200	51123	13.443	ng/u1	0.00
73) Anthracene-d10	17.098	188	629878	23.433	ng/u1	-0.02
81) Pyrene-d10	19.486	212	825829	24.478	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.416	264	988036	24.176	ng/u1	-0.03

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111324\
Data File : BP023091.D
Acq On : 13 Nov 2024 21:35
Operator : RC/JU
Sample : P4724-23
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BH9E2

Quant Time: Nov 14 01:36:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 13 04:49:01 2024
Response via : Initial Calibration

