

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112822\
 Data File : BP012809.D
 Acq On : 28 Nov 2022 10:45
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020681

Quant Time: Nov 28 22:36:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Nov 27 22:55:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.016	152	419036	20.000	ng/u1	0.00
20) Naphthalene-d8	10.834	136	1874024	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	1260555	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.404	188	2730941	20.000	ng/u1	0.00
79) Chrysene-d12	21.498	240	2884027	20.000	ng/u1	0.00
88) Perylene-d12	24.010	264	2856162	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	79575	8.138	ng/uL	0.00
4) Pyridine-d5	3.817	84	605294	20.684	ng/u1	0.00
7) Phenol-d5	7.163	99	706438	20.516	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.340	67	452168	21.311	ng/u1	0.00
11) 2-Chlorophenol-d4	7.540	132	541808	20.650	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	569390	20.524	ng/u1	0.00
21) Nitrobenzene-d5	9.181	128	247876	21.611	ng/u1	0.00
24) 2-Nitrophenol-d4	9.904	143	265030	21.761	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.446	165	593990	20.902	ng/u1	0.00
31) 4-Chloroaniline-d4	10.963	131	830216	21.335	ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	1805520	20.809	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	2101323	20.756	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	315691	20.408	ng/u1	0.00
60) Fluorene-d10	15.645	176	1627512	21.418	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	251005	19.230	ng/u1	0.00
73) Anthracene-d10	17.504	188	2546908	21.187	ng/u1	0.00
81) Pyrene-d10	19.733	212	3042597	18.156	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.845	264	2878909	20.521	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	86092	8.198	ng/uL	98
5) Pyridine	3.840	79	602269	21.033	ng/u1	99
6) Benzaldehyde	7.146	77	384305	22.858	ng/u1	97
8) Phenol	7.193	94	758793	21.016	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.434	93	618924	20.922	ng/u1	99
12) 2-Chlorophenol	7.575	128	601708	21.082	ng/u1	98
13) 2-Methylphenol	8.452	108	570744	20.339	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.552	45	900260	21.098	ng/u1	99
16) Acetophenone	8.840	105	973638	22.142	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.828	70	470993	21.744	ng/u1	98
18) 4-Methylphenol	8.781	108	642430	20.787	ng/u1	99
19) Hexachloroethane	9.105	117	222921	20.368	ng/u1	100
22) Nitrobenzene	9.222	77	657322	21.850	ng/u1	96
23) Isophorone	9.752	82	1335805	20.142	ng/u1	100
25) 2-Nitrophenol	9.940	139	321525	21.994	ng/u1	98
26) 2,4-Dimethylphenol	9.993	107	690546	21.258	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.234	93	836350	20.569	ng/u1	100
29) 2,4-Dichlorophenol	10.469	162	625871	21.260	ng/u1	99
30) Naphthalene	10.881	128	2085729	21.255	ng/u1	99
32) 4-Chloroaniline	10.987	127	872554	21.833	ng/u1	100
33) Hexachlorobutadiene	11.169	225	403854	20.597	ng/u1	99
34) Caprolactam	11.751	113	116469	12.884	ng/u1	98
35) 4-Chloro-3-methylphenol	12.098	107	643573	21.589	ng/u1	99
36) 2-Methylnaphthalene	12.487	142	1432560	21.200	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112822\
 Data File : BP012809.D
 Acq On : 28 Nov 2022 10:45
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020681

Quant Time: Nov 28 22:36:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Nov 27 22:55:08 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.704	142	1445655	21.346	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.845	216	821176	20.429	ng/ul	99
40) Hexachlorocyclopentadiene	12.828	237	325437	16.802	ng/ul	99
41) 2,4,6-Trichlorophenol	13.081	196	517969	20.594	ng/ul	99
42) 2,4,5-Trichlorophenol	13.151	196	565029	20.898	ng/ul	99
43) 1,1'-Biphenyl	13.487	154	1920454	20.914	ng/ul	100
44) 2-Chloronaphthalene	13.528	162	1534391	20.887	ng/ul	99
45) 2-Nitroaniline	13.728	65	337696	23.495	ng/ul	94
47) Dimethylphthalate	14.104	163	1918058	21.089	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	313104	21.997	ng/ul	96
50) Acenaphthylene	14.375	152	2353530	21.260	ng/ul	99
51) 3-Nitroaniline	14.551	138	319921	19.407	ng/ul	98
52) Acenaphthene	14.716	153	1654354	21.227	ng/ul	99
53) 2,4-Dinitrophenol	14.751	184	148604	16.247	ng/ul	97
55) 4-Nitrophenol	14.845	109	243005	21.622	ng/ul	99
56) Dibenzofuran	15.051	168	2303943	21.447	ng/ul	99
57) 2,4-Dinitrotoluene	15.004	165	473805	22.557	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.269	232	497848	21.363	ng/ul	96
59) Diethylphthalate	15.469	149	1857154	21.094	ng/ul	100
61) Fluorene	15.698	166	1920195	22.289	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.692	204	967524	21.646	ng/ul	99
63) 4-Nitroaniline	15.710	138	309579	22.034	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.769	198	279335	19.938	ng/ul	96
67) N-Nitrosodiphenylamine	15.904	169	1608472	20.724	ng/ul	100
68) 4-Bromophenyl-phenylether	16.592	248	471634	17.547	ng/ul	99
69) Hexachlorobenzene	16.704	284	693024	20.230	ng/ul	98
70) Atrazine	16.857	200	526921	19.787	ng/ul	98
71) Pentachlorophenol	17.051	266	403553	19.182	ng/ul	99
72) Phenanthrene	17.451	178	3111066	21.375	ng/ul	100
74) Anthracene	17.539	178	3121691	21.679	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.451	216	825453	19.551	ng/ul	99
76) Pentachlorobenzene	14.969	250	879704	20.044	ng/ul	99
77) Carbazole	17.804	167	2801949	23.002	ng/ul	100
78) Di-n-butylphthalate	18.351	149	3023888	20.970	ng/ul	99
80) Fluoranthene	19.410	202	3652947	17.525	ng/ul	99
82) Pyrene	19.763	202	3886577	18.028	ng/ul	99
83) Butylbenzylphthalate	20.633	149	1149271	22.781	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.404	252	1054096	19.903	ng/ul	99
85) Benzo(a)anthracene	21.480	228	3898563	19.544	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.392	149	1741461	22.572	ng/ul	98
87) Chrysene	21.533	228	3662378	18.989	ng/ul	100
89) Di-n-octyl phthalate	22.363	149	2718367	16.666	ng/ul	100
90) Benzo(b)fluoranthene	23.239	252	3786520	17.903	ng/ul	100
91) Benzo(k)fluoranthene	23.292	252	3923553	18.585	ng/ul	99
93) Benzo(a)pyrene	23.898	252	3392540	19.506	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.657	276	3698062	22.039	ng/ul	99
95) Dibenzo(a,h)anthracene	26.668	278	3299125	21.668	ng/ul	99
96) Benzo(g,h,i)perylene	27.456	276	3230870	22.386	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112822\
 Data File : BP012809.D
 Acq On : 28 Nov 2022 10:45
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020681

Quant Time: Nov 28 22:36:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Nov 27 22:55:08 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.198 ng/uL
RT: 3.428 min Scan# 41
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion: 88 Resp: 86092
Ion Ratio Lower Upper
88 100
43 28.7 22.6 34.0
58 77.1 59.6 89.4



#5
Pyridine
Concen: 21.033 ng/uL
RT: 3.840 min Scan# 111
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion: 79 Resp: 602269
Ion Ratio Lower Upper
79 100
52 74.0 58.9 88.3
51 35.3 27.3 40.9









#6
Benzaldehyde
Concen: 22.858 ng/ul
RT: 7.146 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion: 77 Resp: 384305
Ion Ratio Lower Upper
77 100
105 97.8 81.4 122.2
106 94.5 77.4 116.0



#8
Phenol
Concen: 21.016 ng/ul
RT: 7.193 min Scan# 681
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion: 94 Resp: 758793
Ion Ratio Lower Upper
94 100
65 26.3 20.9 31.3
66 33.7 26.0 39.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.922 ng/ul
RT: 7.434 min Scan# 722
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion: 93 Resp: 618924
Ion Ratio Lower Upper
93 100
63 75.6 59.7 89.5
95 32.3 25.8 38.6



#12
2-Chlorophenol
Concen: 21.082 ng/ul
RT: 7.575 min Scan# 746
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:128 Resp: 601708
Ion Ratio Lower Upper
128 100
64 45.8 35.6 53.4
130 32.1 26.0 39.0









#13
2-Methylphenol
Concen: 20.339 ng/ul
RT: 8.452 min Scan# 895
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:108 Resp: 570744
Ion Ratio Lower Upper
108 100
107 88.0 69.3 103.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.098 ng/ul
RT: 8.552 min Scan# 912
Delta R.T. 0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion: 45 Resp: 900260
Ion Ratio Lower Upper
45 100
77 14.0 11.3 16.9
79 11.0 8.6 12.8









#16
 Acetophenone
 Concen: 22.142 ng/ul
 RT: 8.840 min Scan# 961
 Delta R.T. -0.006 min
 Lab File: BP012809.D
 Acq: 28 Nov 2022 10:45




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020681


Tgt Ion:105 Resp: 973638
 Ion Ratio Lower Upper
 105 100
 77 75.2 59.7 89.5
 51 26.2 20.9 31.3



#17
 N-Nitroso-di-n-propylamine
 Concen: 21.744 ng/ul
 RT: 8.828 min Scan# 959
 Delta R.T. -0.006 min
 Lab File: BP012809.D
 Acq: 28 Nov 2022 10:45

Tgt Ion: 70 Resp: 470993
 Ion Ratio Lower Upper
 70 100
 42 56.0 46.1 69.1
 101 9.4 7.4 11.2
 130 20.4 15.8 23.6









#18
4-Methylphenol
Concen: 20.787 ng/ul
RT: 8.781 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:108 Resp: 642430
Ion Ratio Lower Upper
108 100
107 111.6 90.2 135.2



#19
Hexachloroethane
Concen: 20.368 ng/ul
RT: 9.105 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:117 Resp: 222921
Ion Ratio Lower Upper
117 100
201 106.6 85.7 128.5
199 65.9 52.9 79.3









#22
Nitrobenzene
Concen: 21.850 ng/ul
RT: 9.222 min Scan# 1026
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion: 77 Resp: 657322
Ion Ratio Lower Upper
77 100
123 44.5 38.2 57.2
65 13.1 10.2 15.4



#23
Isophorone
Concen: 20.142 ng/ul
RT: 9.752 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion: 82 Resp: 1335805
Ion Ratio Lower Upper
82 100
95 6.6 5.2 7.8
138 18.2 14.5 21.7









#25
2-Nitrophenol
Concen: 21.994 ng/ul
RT: 9.940 min Scan# 1148
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:139 Resp: 321525
Ion Ratio Lower Upper
139 100
65 47.8 39.5 59.3
109 26.2 21.4 32.0



#26
2,4-Dimethylphenol
Concen: 21.258 ng/ul
RT: 9.993 min Scan# 1157
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:107 Resp: 690546
Ion Ratio Lower Upper
107 100
121 51.4 41.8 62.6
122 88.7 72.3 108.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.569 ng/ul
 RT: 10.234 min Scan# 1198
 Delta R.T. -0.006 min
 Lab File: BP012809.D
 Acq: 28 Nov 2022 10:45




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020681


Tgt Ion: 93 Resp: 836350
 Ion Ratio Lower Upper
 93 100
 95 32.7 26.1 39.1
 123 11.5 9.0 13.6



#29
 2,4-Dichlorophenol
 Concen: 21.260 ng/ul
 RT: 10.469 min Scan# 1238
 Delta R.T. -0.006 min
 Lab File: BP012809.D
 Acq: 28 Nov 2022 10:45

Tgt Ion: 162 Resp: 625871
 Ion Ratio Lower Upper
 162 100
 164 64.8 52.9 79.3
 98 33.6 26.6 40.0









#30
Naphthalene
Concen: 21.255 ng/ul
RT: 10.881 min Scan# 1308
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:128 Resp: 2085729
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 12.8 10.4 15.6



#32
4-Chloroaniline
Concen: 21.833 ng/ul
RT: 10.987 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:127 Resp: 872554
Ion Ratio Lower Upper
127 100
129 32.3 25.7 38.5









#33
Hexachlorobutadiene
Concen: 20.597 ng/ul
RT: 11.169 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:225 Resp: 403854
Ion Ratio Lower Upper
225 100
223 63.1 49.6 74.4
227 63.4 51.0 76.6



#34
Caprolactam
Concen: 12.884 ng/ul
RT: 11.751 min Scan# 1456
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:113 Resp: 116469
Ion Ratio Lower Upper
113 100
55 177.5 145.9 218.9
56 133.6 107.1 160.7









#35
4-Chloro-3-methylphenol
Concen: 21.589 ng/ul
RT: 12.098 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:107 Resp: 643573
Ion Ratio Lower Upper
107 100
144 27.0 21.4 32.2
142 84.3 66.6 100.0



#36
2-Methylnaphthalene
Concen: 21.200 ng/ul
RT: 12.487 min Scan# 1581
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:142 Resp: 1432560
Ion Ratio Lower Upper
142 100
141 90.3 70.2 105.2









#37
1-Methylnaphthalene
Concen: 21.346 ng/ul
RT: 12.704 min Scan# 1618
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:142 Resp: 1445655
Ion Ratio Lower Upper
142 100
141 91.9 73.3 109.9
116 3.5 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.429 ng/ul
RT: 12.845 min Scan# 1642
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:216 Resp: 821176
Ion Ratio Lower Upper
216 100
214 77.8 63.0 94.6
179 19.6 16.2 24.4
108 15.5 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 16.802 ng/ul
RT: 12.828 min Scan# 1639
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:237 Resp: 325437
Ion Ratio Lower Upper
237 100
235 63.9 50.2 75.4
272 12.2 10.2 15.2



#41
2,4,6-Trichlorophenol
Concen: 20.594 ng/ul
RT: 13.081 min Scan# 1682
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:196 Resp: 517969
Ion Ratio Lower Upper
196 100
198 96.1 76.5 114.7
200 30.0 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 20.898 ng/ul
RT: 13.151 min Scan# 1694
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:196 Resp: 565029
Ion Ratio Lower Upper
196 100
198 95.7 77.6 116.4
200 31.2 24.2 36.4



#43
1,1'-Biphenyl
Concen: 20.914 ng/ul
RT: 13.487 min Scan# 1751
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:154 Resp: 1920454
Ion Ratio Lower Upper
154 100
153 40.1 32.0 48.0
76 12.7 10.2 15.4









#44
2-Chloronaphthalene
Concen: 20.887 ng/ul
RT: 13.528 min Scan# 1758
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:162 Resp: 1534391
Ion Ratio Lower Upper
162 100
164 32.4 26.5 39.7
127 36.6 28.8 43.2



#45
2-Nitroaniline
Concen: 23.495 ng/ul
RT: 13.728 min Scan# 1792
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion: 65 Resp: 337696
Ion Ratio Lower Upper
65 100
92 70.5 57.8 86.8
138 109.9 94.6 142.0





#47

Dimethylphthalate

Concen: 21.089 ng/ul

RT: 14.104 min Scan# 1856

Delta R.T. -0.006 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45

Instrument :

BNA_P

ClientSampleId :

SSTD020681





Tgt Ion:163 Resp: 1918058

Ion Ratio Lower Upper

163 100

194 4.1 3.4 5.0

164 10.1 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 21.997 ng/ul

RT: 14.222 min Scan# 1876

Delta R.T. -0.006 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45




Tgt Ion:165 Resp: 313104


Ion Ratio Lower Upper

165 100

89 57.3 42.7 64.1

121 21.3 16.9 25.3









#50
Acenaphthylene
Concen: 21.260 ng/ul
RT: 14.375 min Scan# 1902
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:152 Resp: 2353530
Ion Ratio Lower Upper
152 100
151 19.7 15.9 23.9
153 12.8 10.7 16.1



#51
3-Nitroaniline
Concen: 19.407 ng/ul
RT: 14.551 min Scan# 1932
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:138 Resp: 319921
Ion Ratio Lower Upper
138 100
108 11.5 9.9 14.9
92 114.8 90.2 135.2









#52
Acenaphthene
Concen: 21.227 ng/ul
RT: 14.716 min Scan# 1960
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:153 Resp: 1654354
Ion Ratio Lower Upper
153 100
152 48.0 38.2 57.4
154 87.8 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 16.247 ng/ul
RT: 14.751 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:184 Resp: 148604
Ion Ratio Lower Upper
184 100
63 64.0 48.1 72.1
154 58.4 47.4 71.2









#55
4-Nitrophenol
Concen: 21.622 ng/ul
RT: 14.845 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:109 Resp: 243005
Ion Ratio Lower Upper
109 100
139 148.1 119.2 178.8
65 139.0 110.7 166.1



#56
Dibenzofuran
Concen: 21.447 ng/ul
RT: 15.051 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:168 Resp: 2303943
Ion Ratio Lower Upper
168 100
139 34.1 27.7 41.5









#57
2,4-Dinitrotoluene
Concen: 22.557 ng/ul
RT: 15.004 min Scan# 2009
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:165 Resp: 473805
Ion Ratio Lower Upper
165 100
63 40.7 31.0 46.4
182 2.5 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.363 ng/ul
RT: 15.269 min Scan# 2054
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:232 Resp: 497848
Ion Ratio Lower Upper
232 100
131 43.5 32.0 48.0
130 2.1 1.5 2.3
166 30.6 23.2 34.8









#59
Diethylphthalate
Concen: 21.094 ng/ul
RT: 15.469 min Scan# 2088
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:149 Resp: 1857154
Ion Ratio Lower Upper
149 100
177 22.7 18.1 27.1
150 12.4 9.8 14.8



#61
Fluorene
Concen: 22.289 ng/ul
RT: 15.698 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:166 Resp: 1920195
Ion Ratio Lower Upper
166 100
165 96.9 78.2 117.2
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 21.646 ng/ul
RT: 15.692 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:204 Resp: 967524
Ion Ratio Lower Upper
204 100
206 32.7 26.0 39.0
141 51.4 40.7 61.1



#63
4-Nitroaniline
Concen: 22.034 ng/ul
RT: 15.710 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:138 Resp: 309579
Ion Ratio Lower Upper
138 100
92 50.1 38.6 57.8
108 70.7 56.4 84.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.938 ng/ul
 RT: 15.769 min Scan# 2139
 Delta R.T. -0.006 min
 Lab File: BP012809.D
 Acq: 28 Nov 2022 10:45




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020681


Tgt Ion:198 Resp: 279335
 Ion Ratio Lower Upper
 198 100
 182 4.8 4.1 6.1
 77 21.4 18.9 28.3



#67
 N-Nitrosodiphenylamine
 Concen: 20.724 ng/ul
 RT: 15.904 min Scan# 2162
 Delta R.T. -0.006 min
 Lab File: BP012809.D
 Acq: 28 Nov 2022 10:45

Tgt Ion:169 Resp: 1608472
 Ion Ratio Lower Upper
 169 100
 168 66.9 53.2 79.8
 167 36.8 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 17.547 ng/ul
RT: 16.592 min Scan# 211
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:248 Resp: 471634
Ion Ratio Lower Upper
248 100
250 97.6 78.6 118.0
141 61.8 48.7 73.1



#69
Hexachlorobenzene
Concen: 20.230 ng/ul
RT: 16.704 min Scan# 2298
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:284 Resp: 693024
Ion Ratio Lower Upper
284 100
142 33.6 25.0 37.6
249 30.6 24.6 36.8





#70

Atrazine

Concen: 19.787 ng/ul

RT: 16.857 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45

Instrument :

BNA_P

ClientSampleId :

SSTD020681





Tgt Ion:200 Resp: 526921

Ion Ratio Lower Upper

200 100

173 26.2 20.3 30.5

215 48.9 40.1 60.1



#71

Pentachlorophenol

Concen: 19.182 ng/ul

RT: 17.051 min Scan# 2357

Delta R.T. -0.000 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45




Tgt Ion:266 Resp: 403553


Ion Ratio Lower Upper

266 100

264 62.6 50.7 76.1

268 64.6 51.8 77.8





#72

Phenanthrene

Concen: 21.375 ng/ul

RT: 17.451 min Scan# 2425

Delta R.T. -0.000 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45

Instrument :

BNA_P

ClientSampleId :

SSTD020681





Tgt Ion:178 Resp: 3111066

Ion Ratio Lower Upper

178 100

179 15.3 12.3 18.5

176 19.4 15.4 23.0



#74

Anthracene

Concen: 21.679 ng/ul

RT: 17.539 min Scan# 2440

Delta R.T. -0.006 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45




Tgt Ion:178 Resp: 3121691


Ion Ratio Lower Upper

178 100

179 15.7 12.4 18.6

176 18.4 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.551 ng/uL
RT: 13.451 min Scan# 1745
Delta R.T. -0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:216 Resp: 825453
Ion Ratio Lower Upper
216 100
214 78.4 62.3 93.5
218 47.7 39.1 58.7



#76
Pentachlorobenzene
Concen: 20.044 ng/uL
RT: 14.969 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:250 Resp: 879704
Ion Ratio Lower Upper
250 100
248 62.5 50.6 76.0
252 64.0 50.5 75.7





#77

Carbazole

Concen: 23.002 ng/ul

RT: 17.804 min Scan# 2485

Delta R.T. -0.006 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45

Instrument :





BNA_P

ClientSampleId :

SSTD020681

Tgt Ion:167 Resp: 2801949

Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.1	25.7
139	12.5	9.8	14.8



#78

Di-n-butylphthalate

Concen: 20.970 ng/ul

RT: 18.351 min Scan# 2578




Delta R.T. -0.006 min


Lab File: BP012809.D

Acq: 28 Nov 2022 10:45

Tgt Ion:149 Resp: 3023888

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.2	10.8
104	4.9	3.7	5.5









#80
Fluoranthene
Concen: 17.525 ng/ul
RT: 19.410 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:202 Resp: 3652947
Ion Ratio Lower Upper
202 100
101 10.7 8.7 13.1
100 8.2 6.9 10.3



#82
Pyrene
Concen: 18.028 ng/ul
RT: 19.763 min Scan# 2818
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:202 Resp: 3886577
Ion Ratio Lower Upper
202 100
101 12.4 10.3 15.5
100 10.3 8.6 13.0





#83

Butylbenzylphthalate

Concen: 22.781 ng/ul

RT: 20.633 min Scan# 2966

Delta R.T. -0.000 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45

Instrument :

BNA_P

ClientSampleId :

SSTD020681





Tgt Ion:149 Resp: 1149271

Ion Ratio Lower Upper

149 100

91 66.8 58.2 87.2

206 22.0 16.5 24.7



#84

3,3'-Dichlorobenzidine

Concen: 19.903 ng/ul

RT: 21.404 min Scan# 3097

Delta R.T. -0.000 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45




Tgt Ion:252 Resp: 1054096


Ion Ratio Lower Upper

252 100

254 65.6 52.3 78.5

126 10.5 7.6 11.4





#85

Benzo(a)anthracene

Concen: 19.544 ng/ul

RT: 21.480 min Scan# 3110

Delta R.T. -0.000 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45

Instrument :

BNA_P

ClientSampleId :

SSTD020681





Tgt Ion:228 Resp: 3898563

Ion Ratio Lower Upper

228 100

229 19.7 16.2 24.2

226 27.4 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.572 ng/ul

RT: 21.392 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45




Tgt Ion:149 Resp: 1741461


Ion Ratio Lower Upper

149 100

167 27.2 22.7 34.1

279 3.9 3.8 5.6





#87

Chrysene

Concen: 18.989 ng/ul

RT: 21.533 min Scan# 3119

Delta R.T. -0.000 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45

Instrument :

BNA_P

ClientSampleId :

SSTD020681





Tgt Ion:228 Resp: 3662378

Ion Ratio Lower Upper

228 100

226 30.1 24.3 36.5

229 20.0 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 16.666 ng/ul




RT: 22.363 min Scan# 3260


Delta R.T. -0.000 min

Lab File: BP012809.D

Acq: 28 Nov 2022 10:45

Tgt Ion:149 Resp: 2718367









#90
Benzo(b)fluoranthene
Concen: 17.903 ng/ul
RT: 23.239 min Scan# 3409
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:252 Resp: 3786520
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 9.3 7.4 11.2



#91
Benzo(k)fluoranthene
Concen: 18.585 ng/ul
RT: 23.292 min Scan# 3418
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:252 Resp: 3923553
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.8
125 9.0 7.8 11.8









#93
Benzo(a)pyrene
Concen: 19.506 ng/ul
RT: 23.898 min Scan# 3521
Delta R.T. -0.000 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45




Instrument :
BNA_P
ClientSampleId :
SSTD020681


Tgt Ion:252 Resp: 3392540
Ion Ratio Lower Upper
252 100
253 21.6 17.6 26.4
125 10.5 8.5 12.7



#94
Indeno(1,2,3-cd)pyrene
Concen: 22.039 ng/ul
RT: 26.657 min Scan# 3990
Delta R.T. 0.006 min
Lab File: BP012809.D
Acq: 28 Nov 2022 10:45

Tgt Ion:276 Resp: 3698062
Ion Ratio Lower Upper
276 100
138 23.8 18.3 27.5
277 25.2 19.8 29.8









#95
 Dibenzo(a,h)anthracene
 Concen: 21.668 ng/ul
 RT: 26.668 min Scan# 3992
 Delta R.T. -0.006 min
 Lab File: BP012809.D
 Acq: 28 Nov 2022 10:45

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020681

Tgt Ion:278 Resp: 3299125
 Ion Ratio Lower Upper
 278 100
 139 14.7 12.6 18.8
 279 24.1 19.4 29.2



#96
 Benzo(g,h,i)perylene
 Concen: 22.386 ng/ul
 RT: 27.456 min Scan# 4126
 Delta R.T. -0.000 min
 Lab File: BP012809.D
 Acq: 28 Nov 2022 10:45

Tgt Ion:276 Resp: 3230870
 Ion Ratio Lower Upper
 276 100
 138 20.3 16.7 25.1
 277 23.5 19.1 28.7

