

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012885.D
 Acq On : 30 Nov 2022 18:56
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020688

Quant Time: Nov 30 21:56:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	571246	20.000	ng/ul	0.00
20) Naphthalene-d8	10.828	136	2544308	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.645	164	1731002	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.398	188	3647658	20.000	ng/ul	0.00
79) Chrysene-d12	21.486	240	2922464	20.000	ng/ul	0.00
88) Perylene-d12	23.998	264	2226459	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	104034	7.804	ng/uL	0.00
4) Pyridine-d5	3.817	84	752090	18.852	ng/ul	0.00
7) Phenol-d5	7.158	99	895429	19.076	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.334	67	564229	19.507	ng/ul	0.00
11) 2-Chlorophenol-d4	7.534	132	718546	20.089	ng/ul	0.00
15) 4-Methylphenol-d8	8.716	113	741764	19.613	ng/ul	0.00
21) Nitrobenzene-d5	9.175	128	353422	22.695	ng/ul	0.00
24) 2-Nitrophenol-d4	9.899	143	387794	23.453	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	800975	20.761	ng/ul	0.00
31) 4-Chloroaniline-d4	10.957	131	1076608	20.378	ng/ul	0.00
46) Dimethylphthalate-d6	14.051	166	2422967	20.336	ng/ul	0.00
49) Acenaphthylene-d8	14.340	160	2777643	19.980	ng/ul	0.00
54) 4-Nitrophenol-d4	14.828	143	392948	18.498	ng/ul	0.00
60) Fluorene-d10	15.640	176	2125201	20.367	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	357279	20.493	ng/ul	0.00
73) Anthracene-d10	17.498	188	3223196	20.074	ng/ul	0.00
81) Pyrene-d10	19.728	212	3450995	20.322	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.833	264	2133910	19.512	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	109740	7.665	ng/uL	99
5) Pyridine	3.840	79	756129	19.370	ng/ul	99
6) Benzaldehyde	7.146	77	476352	20.784	ng/ul	99
8) Phenol	7.187	94	952494	19.351	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.428	93	794590	19.703	ng/ul	99
12) 2-Chlorophenol	7.569	128	780340	20.056	ng/ul	98
13) 2-Methylphenol	8.446	108	747164	19.531	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.546	45	1137214	19.550	ng/ul	100
16) Acetophenone	8.840	105	1231796	20.549	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.822	70	616435	20.875	ng/ul	99
18) 4-Methylphenol	8.775	108	829206	19.682	ng/ul	100
19) Hexachloroethane	9.099	117	301647	20.217	ng/ul	99
22) Nitrobenzene	9.216	77	867940	21.251	ng/ul	98
23) Isophorone	9.746	82	1773784	19.700	ng/ul	99
25) 2-Nitrophenol	9.934	139	450797	22.713	ng/ul	97
26) 2,4-Dimethylphenol	9.987	107	903881	20.494	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.228	93	1116355	20.223	ng/ul	100
29) 2,4-Dichlorophenol	10.463	162	825877	20.664	ng/ul	99
30) Naphthalene	10.875	128	2661148	19.974	ng/ul	100
32) 4-Chloroaniline	10.981	127	1122290	20.684	ng/ul	99
33) Hexachlorobutadiene	11.163	225	546707	20.537	ng/ul	99
34) Caprolactam	11.757	113	161731	13.178	ng/ul	86
35) 4-Chloro-3-methylphenol	12.093	107	842632	20.820	ng/ul	99
36) 2-Methylnaphthalene	12.475	142	1889800	20.599	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012885.D
 Acq On : 30 Nov 2022 18:56
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020688

Quant Time: Nov 30 21:56:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.698	142	1890335	20.559	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.840	216	1103667	19.994	ng/ul	99
40) Hexachlorocyclopentadiene	12.822	237	535815	20.146	ng/ul	100
41) 2,4,6-Trichlorophenol	13.075	196	705486	20.426	ng/ul	99
42) 2,4,5-Trichlorophenol	13.146	196	769870	20.735	ng/ul	97
43) 1,1'-Biphenyl	13.481	154	2512357	19.924	ng/ul	100
44) 2-Chloronaphthalene	13.522	162	2003415	19.860	ng/ul	99
45) 2-Nitroaniline	13.722	65	476421	24.138	ng/ul	96
47) Dimethylphthalate	14.098	163	2521907	20.192	ng/ul	99
48) 2,6-Dinitrotoluene	14.216	165	462411	23.658	ng/ul	99
50) Acenaphthylene	14.369	152	3066769	20.174	ng/ul	99
51) 3-Nitroaniline	14.545	138	447331	19.761	ng/ul	97
52) Acenaphthene	14.710	153	2136677	19.965	ng/ul	98
53) 2,4-Dinitrophenol	14.745	184	226924	18.067	ng/ul	99
55) 4-Nitrophenol	14.840	109	287635	18.637	ng/ul	97
56) Dibenzofuran	15.045	168	2991096	20.276	ng/ul	99
57) 2,4-Dinitrotoluene	14.998	165	658740	22.838	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.263	232	663492	20.733	ng/ul	100
59) Diethylphthalate	15.463	149	2444928	20.223	ng/ul	99
61) Fluorene	15.692	166	2441257	20.636	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.687	204	1289366	21.006	ng/ul	99
63) 4-Nitroaniline	15.710	138	386348	20.025	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.763	198	391381	20.914	ng/ul	96
67) N-Nitrosodiphenylamine	15.898	169	2092218	20.182	ng/ul	99
68) 4-Bromophenyl-phenylether	16.587	248	741323	20.649	ng/ul	98
69) Hexachlorobenzene	16.698	284	917766	20.057	ng/ul	100
70) Atrazine	16.851	200	700226	19.687	ng/ul	99
71) Pentachlorophenol	17.045	266	522382	18.590	ng/ul	98
72) Phenanthrene	17.445	178	3842525	19.765	ng/ul	100
74) Anthracene	17.534	178	3868803	20.116	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.445	216	1113422	19.744	ng/ul	99
76) Pentachlorobenzene	14.963	250	1185067	20.216	ng/ul	99
77) Carbazole	17.798	167	3290664	20.225	ng/ul	100
78) Di-n-butylphthalate	18.345	149	3930756	20.408	ng/ul	99
80) Fluoranthene	19.404	202	4259091	20.164	ng/ul	98
82) Pyrene	19.757	202	4392103	20.105	ng/ul	98
83) Butylbenzylphthalate	20.622	149	1392054	27.230	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.398	252	1022868	19.059	ng/ul	100
85) Benzo(a)anthracene	21.469	228	3697605	18.293	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.386	149	1899626	24.298	ng/ul	100
87) Chrysene	21.527	228	3494579	17.881	ng/ul	100
89) Di-n-octyl phthalate	22.345	149	2611214	20.537	ng/ul	100
90) Benzo(b)fluoranthene	23.227	252	2981358	18.083	ng/ul	100
91) Benzo(k)fluoranthene	23.274	252	3015660	18.324	ng/ul	100
93) Benzo(a)pyrene	23.886	252	2538766	18.726	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.627	276	2725708	20.839	ng/ul	99
95) Dibenzo(a,h)anthracene	26.651	278	2474790	20.851	ng/ul	99
96) Benzo(g,h,i)perylene	27.439	276	2414206	21.459	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012885.D
 Acq On : 30 Nov 2022 18:56
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020688

Quant Time: Nov 30 21:56:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.665 ng/uL
RT: 3.428 min Scan# 41
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion: 88 Resp: 109740
Ion Ratio Lower Upper
88 100
43 29.0 22.6 34.0
58 73.6 59.6 89.4



#5
Pyridine
Concen: 19.370 ng/uL
RT: 3.840 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion: 79 Resp: 756129
Ion Ratio Lower Upper
79 100
52 72.3 58.9 88.3
51 33.8 27.3 40.9









#6
Benzaldehyde
Concen: 20.784 ng/ul
RT: 7.146 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion: 77 Resp: 476352
Ion Ratio Lower Upper
77 100
105 102.7 81.4 122.2
106 98.5 77.4 116.0



#8
Phenol
Concen: 19.351 ng/ul
RT: 7.187 min Scan# 680
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion: 94 Resp: 952494
Ion Ratio Lower Upper
94 100
65 26.6 20.9 31.3
66 33.5 26.0 39.0









#10
Bis(2-Chloroethyl)ether
Concen: 19.703 ng/ul
RT: 7.428 min Scan# 721
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion: 93 Resp: 794590
Ion Ratio Lower Upper
93 100
63 73.5 59.7 89.5
95 32.7 25.8 38.6



#12
2-Chlorophenol
Concen: 20.056 ng/ul
RT: 7.569 min Scan# 745
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:128 Resp: 780340
Ion Ratio Lower Upper
128 100
64 43.3 35.6 53.4
130 31.2 26.0 39.0





#13

2-Methylphenol

Concen: 19.531 ng/ul

RT: 8.446 min Scan# 894

Delta R.T. -0.006 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Instrument :

BNA_P

ClientSampleId :





SSTD020688

Tgt Ion:108 Resp: 747164

Ion Ratio Lower Upper

108 100

107 86.9 69.3 103.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.550 ng/ul

RT: 8.546 min Scan# 911

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56




Tgt Ion: 45 Resp: 1137214


Ion Ratio Lower Upper

45 100

77 14.1 11.3 16.9

79 10.9 8.6 12.8









#16
Acetophenone
Concen: 20.549 ng/ul
RT: 8.840 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:105 Resp: 1231796
Ion Ratio Lower Upper
105 100
77 75.2 59.7 89.5
51 25.3 20.9 31.3



#17
N-Nitroso-di-n-propylamine
Concen: 20.875 ng/ul
RT: 8.822 min Scan# 958
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion: 70 Resp: 616435
Ion Ratio Lower Upper
70 100
42 57.0 46.1 69.1
101 9.6 7.4 11.2
130 21.3 15.8 23.6









#18
4-Methylphenol
Concen: 19.682 ng/ul
RT: 8.775 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:108 Resp: 829206
Ion Ratio Lower Upper
108 100
107 112.4 90.2 135.2



#19
Hexachloroethane
Concen: 20.217 ng/ul
RT: 9.099 min Scan# 1005
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:117 Resp: 301647
Ion Ratio Lower Upper
117 100
201 108.1 85.7 128.5
199 67.8 52.9 79.3









#22
Nitrobenzene
Concen: 21.251 ng/ul
RT: 9.216 min Scan# 1025
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion: 77 Resp: 867940
Ion Ratio Lower Upper
77 100
123 46.5 38.2 57.2
65 13.0 10.2 15.4



#23
Isophorone
Concen: 19.700 ng/ul
RT: 9.746 min Scan# 1115
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion: 82 Resp: 1773784
Ion Ratio Lower Upper
82 100
95 6.6 5.2 7.8
138 18.4 14.5 21.7









#25
2-Nitrophenol
Concen: 22.713 ng/ul
RT: 9.934 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:139 Resp: 450797
Ion Ratio Lower Upper
139 100
65 46.7 39.5 59.3
109 25.8 21.4 32.0



#26
2,4-Dimethylphenol
Concen: 20.494 ng/ul
RT: 9.987 min Scan# 1156
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:107 Resp: 903881
Ion Ratio Lower Upper
107 100
121 52.3 41.8 62.6
122 89.1 72.3 108.5





#27

Bis(2-Chloroethoxy)methane

Concen: 20.223 ng/ul

RT: 10.228 min Scan# 1197

Delta R.T. -0.006 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion: 93 Resp: 1116355

Ion Ratio Lower Upper

93 100

95 32.8 26.1 39.1

123 11.3 9.0 13.6



#29

2,4-Dichlorophenol

Concen: 20.664 ng/ul

RT: 10.463 min Scan# 1237

Delta R.T. -0.006 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56




Tgt Ion:162 Resp: 825877


Ion Ratio Lower Upper

162 100

164 66.9 52.9 79.3

98 34.1 26.6 40.0









#30
Naphthalene
Concen: 19.974 ng/ul
RT: 10.875 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:128 Resp: 2661148
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 20.684 ng/ul
RT: 10.981 min Scan# 1325
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:127 Resp: 1122290
Ion Ratio Lower Upper
127 100
129 32.4 25.7 38.5









#33
Hexachlorobutadiene
Concen: 20.537 ng/ul
RT: 11.163 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:225 Resp: 546707
Ion Ratio Lower Upper
225 100
223 62.5 49.6 74.4
227 64.5 51.0 76.6



#34
Caprolactam
Concen: 13.178 ng/ul
RT: 11.757 min Scan# 1457
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:113 Resp: 161731
Ion Ratio Lower Upper
113 100
55 158.5 145.9 218.9
56 120.4 107.1 160.7









#35
4-Chloro-3-methylphenol
Concen: 20.820 ng/ul
RT: 12.093 min Scan# 1514
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:107 Resp: 842632
Ion Ratio Lower Upper
107 100
144 27.4 21.4 32.2
142 83.6 66.6 100.0



#36
2-Methylnaphthalene
Concen: 20.599 ng/ul
RT: 12.475 min Scan# 1579
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:142 Resp: 1889800
Ion Ratio Lower Upper
142 100
141 87.4 70.2 105.2





#37

1-Methylnaphthalene

Concen: 20.559 ng/ul

RT: 12.698 min Scan# 1617

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:142 Resp: 1890335

Ion Ratio Lower Upper

142 100

141 93.1 73.3 109.9

116 3.7 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.994 ng/ul

RT: 12.840 min Scan# 1641

Delta R.T. -0.006 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Tgt Ion:216 Resp: 1103667




Ion Ratio Lower Upper


216 100

214 79.3 63.0 94.6

179 20.1 16.2 24.4

108 14.9 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 20.146 ng/ul
RT: 12.822 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:237 Resp: 535815
Ion Ratio Lower Upper
237 100
235 62.8 50.2 75.4
272 13.4 10.2 15.2



#41
2,4,6-Trichlorophenol
Concen: 20.426 ng/ul
RT: 13.075 min Scan# 1681
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:196 Resp: 705486
Ion Ratio Lower Upper
196 100
198 94.3 76.5 114.7
200 30.9 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 20.735 ng/ul

RT: 13.146 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:196 Resp: 769870

Ion Ratio Lower Upper

196 100

198 93.0 77.6 116.4

200 30.8 24.2 36.4



#43

1,1'-Biphenyl

Concen: 19.924 ng/ul

RT: 13.481 min Scan# 1750

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56




Tgt Ion:154 Resp: 2512357


Ion Ratio Lower Upper

154 100

153 39.9 32.0 48.0

76 12.2 10.2 15.4









#44
2-Chloronaphthalene
Concen: 19.860 ng/ul
RT: 13.522 min Scan# 1757
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:162 Resp: 2003415
Ion Ratio Lower Upper
162 100
164 33.7 26.5 39.7
127 36.3 28.8 43.2



#45
2-Nitroaniline
Concen: 24.138 ng/ul
RT: 13.722 min Scan# 1791
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion: 65 Resp: 476421
Ion Ratio Lower Upper
65 100
92 70.4 57.8 86.8
138 113.4 94.6 142.0









#47
Dimethylphthalate
Concen: 20.192 ng/ul
RT: 14.098 min Scan# 1855
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:163 Resp: 2521907
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 10.2 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 23.658 ng/ul
RT: 14.216 min Scan# 1875
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:165 Resp: 462411
Ion Ratio Lower Upper
165 100
89 53.1 42.7 64.1
121 20.5 16.9 25.3





#50

Acenaphthylene

Concen: 20.174 ng/ul

RT: 14.369 min Scan# 1901

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:152 Resp: 3066769

Ion Ratio Lower Upper

152 100

151 19.7 15.9 23.9

153 13.1 10.7 16.1



#51

3-Nitroaniline

Concen: 19.761 ng/ul

RT: 14.545 min Scan# 1931

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56




Tgt Ion:138 Resp: 447331


Ion Ratio Lower Upper

138 100

108 11.2 9.9 14.9

92 109.2 90.2 135.2









#52
Acenaphthene
Concen: 19.965 ng/ul
RT: 14.710 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:153 Resp: 2136677
Ion Ratio Lower Upper
153 100
152 47.9 38.2 57.4
154 89.4 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 18.067 ng/ul
RT: 14.745 min Scan# 1965
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:184 Resp: 226924
Ion Ratio Lower Upper
184 100
63 60.3 48.1 72.1
154 58.0 47.4 71.2









#55
4-Nitrophenol
Concen: 18.637 ng/ul
RT: 14.840 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:109 Resp: 287635
Ion Ratio Lower Upper
109 100
139 154.1 119.2 178.8
65 141.5 110.7 166.1



#56
Dibenzofuran
Concen: 20.276 ng/ul
RT: 15.045 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:168 Resp: 2991096
Ion Ratio Lower Upper
168 100
139 34.3 27.7 41.5









#57
2,4-Dinitrotoluene
Concen: 22.838 ng/ul
RT: 14.998 min Scan# 2008
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:165 Resp: 658740
Ion Ratio Lower Upper
165 100
63 40.0 31.0 46.4
182 2.8 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.733 ng/ul
RT: 15.263 min Scan# 2053
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:232 Resp: 663492
Ion Ratio Lower Upper
232 100
131 40.3 32.0 48.0
130 2.0 1.5 2.3
166 28.9 23.2 34.8









#59
Diethylphthalate
Concen: 20.223 ng/ul
RT: 15.463 min Scan# 2087
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:149 Resp: 2444928
Ion Ratio Lower Upper
149 100
177 23.2 18.1 27.1
150 12.3 9.8 14.8



#61
Fluorene
Concen: 20.636 ng/ul
RT: 15.692 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:166 Resp: 2441257
Ion Ratio Lower Upper
166 100
165 97.7 78.2 117.2
167 13.5 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 21.006 ng/ul
RT: 15.687 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:204 Resp: 1289366
Ion Ratio Lower Upper
204 100
206 32.9 26.0 39.0
141 49.9 40.7 61.1



#63
4-Nitroaniline
Concen: 20.025 ng/ul
RT: 15.710 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:138 Resp: 386348
Ion Ratio Lower Upper
138 100
92 48.9 38.6 57.8
108 70.4 56.4 84.6









#66
4,6-Dinitro-2-methylphenol
Concen: 20.914 ng/ul
RT: 15.763 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:198 Resp: 391381
Ion Ratio Lower Upper
198 100
182 4.3 4.1 6.1
77 21.7 18.9 28.3



#67
N-Nitrosodiphenylamine
Concen: 20.182 ng/ul
RT: 15.898 min Scan# 2161
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:169 Resp: 2092218
Ion Ratio Lower Upper
169 100
168 65.5 53.2 79.8
167 36.4 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 20.649 ng/ul
RT: 16.587 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:248 Resp: 741323
Ion Ratio Lower Upper
248 100
250 96.8 78.6 118.0
141 58.4 48.7 73.1



#69
Hexachlorobenzene
Concen: 20.057 ng/ul
RT: 16.698 min Scan# 2297
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:284 Resp: 917766
Ion Ratio Lower Upper
284 100
142 31.1 25.0 37.6
249 30.5 24.6 36.8









#70
Atrazine
Concen: 19.687 ng/ul
RT: 16.851 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:200 Resp: 700226
Ion Ratio Lower Upper
200 100
173 25.8 20.3 30.5
215 49.5 40.1 60.1



#71
Pentachlorophenol
Concen: 18.590 ng/ul
RT: 17.045 min Scan# 2356
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:266 Resp: 522382
Ion Ratio Lower Upper
266 100
264 65.0 50.7 76.1
268 63.0 51.8 77.8





#72

Phenanthrene

Concen: 19.765 ng/ul

RT: 17.445 min Scan# 2424

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:178 Resp: 3842525

Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 19.1 15.4 23.0



#74

Anthracene

Concen: 20.116 ng/ul

RT: 17.534 min Scan# 2439

Delta R.T. -0.006 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56




Tgt Ion:178 Resp: 3868803


Ion Ratio Lower Upper

178 100

179 15.7 12.4 18.6

176 18.7 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.744 ng/uL
RT: 13.445 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:216 Resp: 1113422
Ion Ratio Lower Upper
216 100
214 78.6 62.3 93.5
218 48.7 39.1 58.7



#76
Pentachlorobenzene
Concen: 20.216 ng/uL
RT: 14.963 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:250 Resp: 1185067
Ion Ratio Lower Upper
250 100
248 62.7 50.6 76.0
252 64.1 50.5 75.7









#77
Carbazole
Concen: 20.225 ng/ul
RT: 17.798 min Scan# 2484
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:167 Resp: 3290664
Ion Ratio Lower Upper
167 100
166 21.4 17.1 25.7
139 12.3 9.8 14.8



#78
Di-n-butylphthalate
Concen: 20.408 ng/ul
RT: 18.345 min Scan# 2577
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:149 Resp: 3930756
Ion Ratio Lower Upper
149 100
150 9.3 7.2 10.8
104 4.7 3.7 5.5









#80
Fluoranthene
Concen: 20.164 ng/ul
RT: 19.404 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:202 Resp: 4259091
Ion Ratio Lower Upper
202 100
101 10.0 8.7 13.1
100 7.7 6.9 10.3



#82
Pyrene
Concen: 20.105 ng/ul
RT: 19.757 min Scan# 2817
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:202 Resp: 4392103
Ion Ratio Lower Upper
202 100
101 12.1 10.3 15.5
100 10.0 8.6 13.0





#83

Butylbenzylphthalate

Concen: 27.230 ng/ul

RT: 20.622 min Scan# 2964

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:149 Resp: 1392054

Ion Ratio Lower Upper

149 100

91 69.7 58.2 87.2

206 22.1 16.5 24.7



#84

3,3'-Dichlorobenzidine

Concen: 19.059 ng/ul

RT: 21.398 min Scan# 3096

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56




Tgt Ion:252 Resp: 1022868


Ion Ratio Lower Upper

252 100

254 65.4 52.3 78.5

126 9.6 7.6 11.4









#85
Benzo(a)anthracene
Concen: 18.293 ng/ul
RT: 21.469 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:228 Resp: 3697605
Ion Ratio Lower Upper
228 100
229 20.0 16.2 24.2
226 27.5 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 24.298 ng/ul
RT: 21.386 min Scan# 3094
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:149 Resp: 1899626
Ion Ratio Lower Upper
149 100
167 28.5 22.7 34.1
279 4.8 3.8 5.6





#87

Chrysene

Concen: 17.881 ng/ul

RT: 21.527 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:228 Resp: 3494579

Ion Ratio Lower Upper

228 100

226 30.5 24.3 36.5

229 19.8 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 20.537 ng/ul




RT: 22.345 min Scan# 3257


Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Tgt Ion:149 Resp: 2611214









#90
Benzo(b)fluoranthene
Concen: 18.083 ng/ul
RT: 23.227 min Scan# 3407
Delta R.T. 0.000 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:252 Resp: 2981358
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 9.4 7.4 11.2



#91
Benzo(k)fluoranthene
Concen: 18.324 ng/ul
RT: 23.274 min Scan# 3415
Delta R.T. -0.006 min
Lab File: BP012885.D
Acq: 30 Nov 2022 18:56

Tgt Ion:252 Resp: 3015660
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 9.7 7.8 11.8





#93

Benzo(a)pyrene

Concen: 18.726 ng/ul

RT: 23.886 min Scan# 3519

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:252 Resp: 2538766

Ion Ratio Lower Upper

252 100

253 22.7 17.6 26.4

125 10.5 8.5 12.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.839 ng/ul

RT: 26.627 min Scan# 3985

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56




Tgt Ion:276 Resp: 2725708


Ion Ratio Lower Upper

276 100

138 22.0 18.3 27.5

277 24.9 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 20.851 ng/ul

RT: 26.651 min Scan# 3989

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:278 Resp: 2474790

Ion Ratio Lower Upper

278 100

139 15.1 12.6 18.8

279 24.2 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 21.459 ng/ul

RT: 27.439 min Scan# 4123

Delta R.T. 0.000 min

Lab File: BP012885.D

Acq: 30 Nov 2022 18:56

Tgt Ion:276 Resp: 2414206

Ion Ratio Lower Upper

276 100

138 20.7 16.7 25.1

277 23.8 19.1 28.7

