

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012913.D
 Acq On : 01 Dec 2022 12:01
 Operator : CG/JU
 Sample : N5737-02MS
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A4345MS

Quant Time: Dec 01 21:36:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	407922	20.000	ng/ul	0.00
20) Naphthalene-d8	10.822	136	1705328	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.645	164	1076035	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.398	188	2336717	20.000	ng/ul	0.00
79) Chrysene-d12	21.486	240	2389079	20.000	ng/ul	0.00
88) Perylene-d12	23.992	264	2385300	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	57278	6.017	ng/uL	0.00
4) Pyridine-d5	3.822	84	382263	13.418	ng/ul	0.00
7) Phenol-d5	7.158	99	295326	8.810	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	668467	32.364	ng/ul	0.00
11) 2-Chlorophenol-d4	7.534	132	701910	27.481	ng/ul	0.00
15) 4-Methylphenol-d8	8.710	113	496775	18.394	ng/ul	0.00
21) Nitrobenzene-d5	9.175	128	407581	39.049	ng/ul	0.00
24) 2-Nitrophenol-d4	9.899	143	432491	39.024	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	870045	33.646	ng/ul	0.00
31) 4-Chloroaniline-d4	10.957	131	1028498	29.045	ng/ul	0.00
46) Dimethylphthalate-d6	14.051	166	2783148	37.577	ng/ul	0.00
49) Acenaphthylene-d8	14.340	160	3114027	36.034	ng/ul	0.00
54) 4-Nitrophenol-d4	14.828	143	172903	13.094	ng/ul	0.00
60) Fluorene-d10	15.634	176	2462990	37.971	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	481807	43.140	ng/ul	0.00
73) Anthracene-d10	17.498	188	3938293	38.288	ng/ul	0.00
81) Pyrene-d10	19.722	212	4715202	33.967	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.833	264	4637299	39.580	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	728874	71.294	ng/uL	100
5) Pyridine	3.840	79	412458	14.797	ng/ul	98
6) Benzaldehyde	7.140	77	618660	37.800	ng/ul	99
8) Phenol	7.187	94	315967	8.990	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.428	93	897684	31.171	ng/ul	100
12) 2-Chlorophenol	7.569	128	734337	26.430	ng/ul	98
13) 2-Methylphenol	8.446	108	559121	20.468	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.540	45	1239417	29.838	ng/ul	100
16) Acetophenone	8.834	105	1353608	31.622	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.816	70	645162	30.596	ng/ul	98
18) 4-Methylphenol	8.775	108	549822	18.276	ng/ul	97
19) Hexachloroethane	9.099	117	337963	31.720	ng/ul	94
22) Nitrobenzene	9.216	77	977427	35.705	ng/ul	99
23) Isophorone	9.746	82	1863058	30.871	ng/ul	99
25) 2-Nitrophenol	9.934	139	468319	35.204	ng/ul	97
26) 2,4-Dimethylphenol	9.987	107	846844	28.648	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.228	93	1178139	31.842	ng/ul	100
29) 2,4-Dichlorophenol	10.463	162	846580	31.603	ng/ul	98
30) Naphthalene	10.875	128	2860723	32.036	ng/ul	99
32) 4-Chloroaniline	10.981	127	1007534	27.704	ng/ul	100
33) Hexachlorobutadiene	11.157	225	557588	31.250	ng/ul	100
34) Caprolactam	11.769	113	49569	6.026	ng/ul	96
35) 4-Chloro-3-methylphenol	12.093	107	811244	29.905	ng/ul	97
36) 2-Methylnaphthalene	12.475	142	1971730	32.065	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012913.D
 Acq On : 01 Dec 2022 12:01
 Operator : CG/JU
 Sample : N5737-02MS
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A4345MS

Quant Time: Dec 01 21:36:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.693	142	1995254	32.376	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.840	216	1129560	32.919	ng/ul	99
40) Hexachlorocyclopentadiene	12.822	237	504763	30.530	ng/ul	99
41) 2,4,6-Trichlorophenol	13.075	196	753085	35.077	ng/ul	99
42) 2,4,5-Trichlorophenol	13.145	196	827427	35.851	ng/ul	98
43) 1,1'-Biphenyl	13.481	154	2653425	33.851	ng/ul	99
44) 2-Chloronaphthalene	13.522	162	2119498	33.799	ng/ul	100
45) 2-Nitroaniline	13.722	65	577899	47.101	ng/ul	95
47) Dimethylphthalate	14.098	163	2807538	36.162	ng/ul	99
48) 2,6-Dinitrotoluene	14.216	165	544847	44.842	ng/ul	99
50) Acenaphthylene	14.369	152	3304845	34.973	ng/ul	100
51) 3-Nitroaniline	14.545	138	529877	37.655	ng/ul	97
52) Acenaphthene	14.710	153	2325638	34.957	ng/ul	98
53) 2,4-Dinitrophenol	14.745	184	287572	36.832	ng/ul	99
55) 4-Nitrophenol	14.845	109	101561	10.586	ng/ul	99
56) Dibenzofuran	15.039	168	3307721	36.071	ng/ul	100
57) 2,4-Dinitrotoluene	14.998	165	774596	43.201	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.263	232	742341	37.317	ng/ul	99
59) Diethylphthalate	15.457	149	2797198	37.220	ng/ul	99
61) Fluorene	15.692	166	2742491	37.293	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.687	204	1416941	37.136	ng/ul	98
63) 4-Nitroaniline	15.710	138	547305	45.634	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.763	198	461730	38.516	ng/ul	98
67) N-Nitrosodiphenylamine	15.898	169	2371896	35.716	ng/ul	99
68) 4-Bromophenyl-phenylether	16.581	248	853956	37.131	ng/ul	99
69) Hexachlorobenzene	16.698	284	1051248	35.863	ng/ul	98
70) Atrazine	16.851	200	797630	35.007	ng/ul	98
71) Pentachlorophenol	17.039	266	630927	35.050	ng/ul	98
72) Phenanthrene	17.439	178	4513773	36.244	ng/ul	99
74) Anthracene	17.533	178	4540214	36.850	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.445	216	1147393	31.762	ng/ul	100
76) Pentachlorobenzene	14.957	250	1148950	30.596	ng/ul	99
77) Carbazole	17.798	167	4208343	40.376	ng/ul	100
78) Di-n-butylphthalate	18.345	149	4961844	40.214	ng/ul	100
80) Fluoranthene	19.398	202	5509315	31.907	ng/ul	99
82) Pyrene	19.757	202	5650223	31.638	ng/ul	96
83) Butylbenzylphthalate	20.622	149	2227718	53.306	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.392	252	463724	10.570	ng/ul	99
85) Benzo(a)anthracene	21.469	228	5646585	34.172	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.380	149	3215347	50.310	ng/ul	100
87) Chrysene	21.521	228	5311652	33.246	ng/ul	99
89) Di-n-octyl phthalate	22.345	149	5346610	39.251	ng/ul	100
90) Benzo(b)fluoranthene	23.227	252	5885467	33.320	ng/ul	99
91) Benzo(k)fluoranthene	23.274	252	5640027	31.989	ng/ul	99
93) Benzo(a)pyrene	23.886	252	5094692	35.076	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.627	276	6316067	45.072	ng/ul	98
95) Dibenzo(a,h)anthracene	26.645	278	5551915	43.663	ng/ul	99
96) Benzo(g,h,i)perylene	27.433	276	5404802	44.841	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012913.D
 Acq On : 01 Dec 2022 12:01
 Operator : CG/JU
 Sample : N5737-02MS
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A4345MS

Quant Time: Dec 01 21:36:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 71.294 ng/uL
RT: 3.428 min Scan# 41
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion: 88 Resp: 728874
Ion Ratio Lower Upper
88 100
43 28.3 22.6 34.0
58 75.1 59.6 89.4



#5
Pyridine
Concen: 14.797 ng/uL
RT: 3.840 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion: 79 Resp: 412458
Ion Ratio Lower Upper
79 100
52 75.4 58.9 88.3
51 34.9 27.3 40.9









#6
Benzaldehyde
Concen: 37.800 ng/ul
RT: 7.140 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion: 77 Resp: 618660
Ion Ratio Lower Upper
77 100
105 100.3 81.4 122.2
106 96.2 77.4 116.0



#8
Phenol
Concen: 8.990 ng/ul
RT: 7.187 min Scan# 680
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion: 94 Resp: 315967
Ion Ratio Lower Upper
94 100
65 26.9 20.9 31.3
66 34.1 26.0 39.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 31.171 ng/ul
 RT: 7.428 min Scan# 720
 Delta R.T. 0.000 min
 Lab File: BP012913.D
 Acq: 01 Dec 2022 12:01

Instrument :
 BNA_P
 ClientSampleId :
 A4345MS

Tgt Ion: 93 Resp: 897684




Ion	Ratio	Lower	Upper
93	100		
63	74.8	59.7	89.5
95	32.7	25.8	38.6




#12
 2-Chlorophenol
 Concen: 26.430 ng/ul
 RT: 7.569 min Scan# 745
 Delta R.T. 0.000 min
 Lab File: BP012913.D
 Acq: 01 Dec 2022 12:01

Tgt Ion:128 Resp: 734337

Ion	Ratio	Lower	Upper
128	100		
64	43.8	35.6	53.4
130	31.3	26.0	39.0





#13

2-Methylphenol

Concen: 20.468 ng/ul

RT: 8.446 min Scan# 894

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :





A4345MS

Tgt Ion:108 Resp: 559121

Ion Ratio Lower Upper

108 100

107 88.2 69.3 103.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 29.838 ng/ul

RT: 8.540 min Scan# 910

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion: 45 Resp: 1239417


Ion Ratio Lower Upper

45 100

77 14.1 11.3 16.9

79 11.0 8.6 12.8









#16
Acetophenone
Concen: 31.622 ng/ul
RT: 8.834 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:105 Resp: 1353608
Ion Ratio Lower Upper
105 100
77 73.3 59.7 89.5
51 25.2 20.9 31.3



#17
N-Nitroso-di-n-propylamine
Concen: 30.596 ng/ul
RT: 8.816 min Scan# 957
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion: 70 Resp: 645162
Ion Ratio Lower Upper
70 100
42 56.9 46.1 69.1
101 10.0 7.4 11.2
130 21.5 15.8 23.6





#18

4-Methylphenol

Concen: 18.276 ng/ul

RT: 8.775 min Scan# 91

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :





A4345MS

Tgt Ion:108 Resp: 549822

Ion Ratio Lower Upper

108 100

107 109.3 90.2 135.2



#19

Hexachloroethane

Concen: 31.720 ng/ul

RT: 9.099 min Scan# 1005

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:117 Resp: 337963


Ion Ratio Lower Upper

117 100

201 113.4 85.7 128.5

199 70.7 52.9 79.3









#22
Nitrobenzene
Concen: 35.705 ng/ul
RT: 9.216 min Scan# 1025
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion: 77 Resp: 977427
Ion Ratio Lower Upper
77 100
123 47.4 38.2 57.2
65 13.1 10.2 15.4



#23
Isophorone
Concen: 30.871 ng/ul
RT: 9.746 min Scan# 1115
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion: 82 Resp: 1863058
Ion Ratio Lower Upper
82 100
95 6.6 5.2 7.8
138 18.6 14.5 21.7









#25
2-Nitrophenol
Concen: 35.204 ng/ul
RT: 9.934 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:139 Resp: 468319
Ion Ratio Lower Upper
139 100
65 46.4 39.5 59.3
109 26.5 21.4 32.0



#26
2,4-Dimethylphenol
Concen: 28.648 ng/ul
RT: 9.987 min Scan# 1156
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:107 Resp: 846844
Ion Ratio Lower Upper
107 100
121 51.6 41.8 62.6
122 90.3 72.3 108.5





#27

Bis(2-Chloroethoxy)methane

Concen: 31.842 ng/ul

RT: 10.228 min Scan# 1197

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion: 93 Resp: 1178139

Ion Ratio Lower Upper

93 100

95 32.6 26.1 39.1

123 11.6 9.0 13.6



#29

2,4-Dichlorophenol

Concen: 31.603 ng/ul

RT: 10.463 min Scan# 1237

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:162 Resp: 846580


Ion Ratio Lower Upper

162 100

164 64.2 52.9 79.3

98 33.4 26.6 40.0









#30
Naphthalene
Concen: 32.036 ng/ul
RT: 10.875 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:128 Resp: 2860723
Ion Ratio Lower Upper
128 100
129 11.5 8.7 13.1
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 27.704 ng/ul
RT: 10.981 min Scan# 1325
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:127 Resp: 1007534
Ion Ratio Lower Upper
127 100
129 32.4 25.7 38.5





#33

Hexachlorobutadiene

Concen: 31.250 ng/ul

RT: 11.157 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:225 Resp: 557588

Ion Ratio Lower Upper

225 100

223 62.4 49.6 74.4

227 63.5 51.0 76.6



#34

Caprolactam

Concen: 6.026 ng/ul

RT: 11.769 min Scan# 1459

Delta R.T. 0.012 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:113 Resp: 49569


Ion Ratio Lower Upper

113 100

55 189.0 145.9 218.9

56 129.7 107.1 160.7









#35
4-Chloro-3-methylphenol
Concen: 29.905 ng/ul
RT: 12.093 min Scan# 1514
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:107 Resp: 811244
Ion Ratio Lower Upper
107 100
144 28.0 21.4 32.2
142 86.1 66.6 100.0



#36
2-Methylnaphthalene
Concen: 32.065 ng/ul
RT: 12.475 min Scan# 1579
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:142 Resp: 1971730
Ion Ratio Lower Upper
142 100
141 88.0 70.2 105.2





#37

1-Methylnaphthalene

Concen: 32.376 ng/ul

RT: 12.693 min Scan# 1616

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:142 Resp: 1995254

Ion Ratio Lower Upper

142 100

141 91.4 73.3 109.9

116 3.3 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 32.919 ng/ul

RT: 12.840 min Scan# 1641

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Tgt Ion:216 Resp: 1129560




Ion Ratio Lower Upper


216 100

214 78.4 63.0 94.6

179 19.2 16.2 24.4

108 14.4 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 30.530 ng/ul
RT: 12.822 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:237 Resp: 504763
Ion Ratio Lower Upper
237 100
235 62.6 50.2 75.4
272 13.5 10.2 15.2



#41
2,4,6-Trichlorophenol
Concen: 35.077 ng/ul
RT: 13.075 min Scan# 1681
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:196 Resp: 753085
Ion Ratio Lower Upper
196 100
198 94.8 76.5 114.7
200 31.2 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 35.851 ng/ul

RT: 13.145 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:196 Resp: 827427

Ion Ratio Lower Upper

196 100

198 94.9 77.6 116.4

200 31.2 24.2 36.4



#43

1,1'-Biphenyl

Concen: 33.851 ng/ul

RT: 13.481 min Scan# 1750

Delta R.T. 0.000 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:154 Resp: 2653425


Ion Ratio Lower Upper

154 100

153 39.2 32.0 48.0

76 12.0 10.2 15.4









#44
2-Chloronaphthalene
Concen: 33.799 ng/ul
RT: 13.522 min Scan# 1757
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:162 Resp: 2119498
Ion Ratio Lower Upper
162 100
164 33.4 26.5 39.7
127 35.7 28.8 43.2



#45
2-Nitroaniline
Concen: 47.101 ng/ul
RT: 13.722 min Scan# 1791
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion: 65 Resp: 577899
Ion Ratio Lower Upper
65 100
92 71.2 57.8 86.8
138 110.4 94.6 142.0





#47

Dimethylphthalate

Concen: 36.162 ng/ul

RT: 14.098 min Scan# 1854

Delta R.T. 0.000 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:163 Resp: 2807538

Ion Ratio Lower Upper

163 100

194 4.1 3.4 5.0

164 10.2 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 44.842 ng/ul

RT: 14.216 min Scan# 1875

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:165 Resp: 544847


Ion Ratio Lower Upper

165 100

89 52.6 42.7 64.1

121 20.3 16.9 25.3









#50
Acenaphthylene
Concen: 34.973 ng/ul
RT: 14.369 min Scan# 1900
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:152 Resp: 3304845
Ion Ratio Lower Upper
152 100
151 19.8 15.9 23.9
153 13.1 10.7 16.1



#51
3-Nitroaniline
Concen: 37.655 ng/ul
RT: 14.545 min Scan# 1931
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:138 Resp: 529877
Ion Ratio Lower Upper
138 100
108 11.3 9.9 14.9
92 109.8 90.2 135.2





#52

Acenaphthene

Concen: 34.957 ng/ul

RT: 14.710 min Scan# 1959

Delta R.T. 0.000 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:153 Resp: 2325638

Ion Ratio Lower Upper

153 100

152 47.9 38.2 57.4

154 89.7 69.8 104.6



#53

2,4-Dinitrophenol

Concen: 36.832 ng/ul

RT: 14.745 min Scan# 1965

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:184 Resp: 287572


Ion Ratio Lower Upper

184 100

63 60.4 48.1 72.1

154 58.3 47.4 71.2









#55
4-Nitrophenol
Concen: 10.586 ng/ul
RT: 14.845 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:109 Resp: 101561
Ion Ratio Lower Upper
109 100
139 149.9 119.2 178.8
65 141.0 110.7 166.1



#56
Dibenzofuran
Concen: 36.071 ng/ul
RT: 15.039 min Scan# 2015
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:168 Resp: 3307721
Ion Ratio Lower Upper
168 100
139 34.9 27.7 41.5









#57
2,4-Dinitrotoluene
Concen: 43.201 ng/ul
RT: 14.998 min Scan# 2008
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:165 Resp: 774596
Ion Ratio Lower Upper
165 100
63 39.9 31.0 46.4
182 2.4 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 37.317 ng/ul
RT: 15.263 min Scan# 2053
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:232 Resp: 742341
Ion Ratio Lower Upper
232 100
131 39.2 32.0 48.0
130 2.0 1.5 2.3
166 28.2 23.2 34.8









#59
Diethylphthalate
Concen: 37.220 ng/ul
RT: 15.457 min Scan# 2086
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:149 Resp: 2797198
Ion Ratio Lower Upper
149 100
177 23.2 18.1 27.1
150 12.5 9.8 14.8



#61
Fluorene
Concen: 37.293 ng/ul
RT: 15.692 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:166 Resp: 2742491
Ion Ratio Lower Upper
166 100
165 96.4 78.2 117.2
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 37.136 ng/ul
RT: 15.687 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:204 Resp: 1416941
Ion Ratio Lower Upper
204 100
206 32.8 26.0 39.0
141 49.1 40.7 61.1



#63
4-Nitroaniline
Concen: 45.634 ng/ul
RT: 15.710 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:138 Resp: 547305
Ion Ratio Lower Upper
138 100
92 48.0 38.6 57.8
108 70.5 56.4 84.6









#66
4,6-Dinitro-2-methylphenol
Concen: 38.516 ng/ul
RT: 15.763 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:198 Resp: 461730
Ion Ratio Lower Upper
198 100
182 4.3 4.1 6.1
77 22.6 18.9 28.3



#67
N-Nitrosodiphenylamine
Concen: 35.716 ng/ul
RT: 15.898 min Scan# 2161
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:169 Resp: 2371896
Ion Ratio Lower Upper
169 100
168 66.1 53.2 79.8
167 36.2 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 37.131 ng/ul
RT: 16.581 min Scan# 2277
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:248 Resp: 853956
Ion Ratio Lower Upper
248 100
250 97.6 78.6 118.0
141 62.6 48.7 73.1



#69
Hexachlorobenzene
Concen: 35.863 ng/ul
RT: 16.698 min Scan# 2297
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:284 Resp: 1051248
Ion Ratio Lower Upper
284 100
142 29.3 25.0 37.6
249 30.1 24.6 36.8





#70

Atrazine

Concen: 35.007 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:200 Resp: 797630

Ion Ratio Lower Upper

200 100

173 27.2 20.3 30.5

215 49.7 40.1 60.1



#71

Pentachlorophenol

Concen: 35.050 ng/ul

RT: 17.039 min Scan# 2355

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:266 Resp: 630927


Ion Ratio Lower Upper

266 100

264 62.1 50.7 76.1

268 62.5 51.8 77.8





#72

Phenanthrene

Concen: 36.244 ng/ul

RT: 17.439 min Scan# 2423

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:178 Resp: 4513773

Ion Ratio Lower Upper

178 100

179 15.9 12.3 18.5

176 19.4 15.4 23.0



#74

Anthracene

Concen: 36.850 ng/ul

RT: 17.533 min Scan# 2439

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:178 Resp: 4540214


Ion Ratio Lower Upper

178 100

179 15.6 12.4 18.6

176 18.6 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.762 ng/uL
RT: 13.445 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:216 Resp: 1147393
Ion Ratio Lower Upper
216 100
214 77.7 62.3 93.5
218 48.3 39.1 58.7



#76
Pentachlorobenzene
Concen: 30.596 ng/uL
RT: 14.957 min Scan# 2001
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:250 Resp: 1148950
Ion Ratio Lower Upper
250 100
248 62.4 50.6 76.0
252 63.3 50.5 75.7





#77

Carbazole

Concen: 40.376 ng/ul

RT: 17.798 min Scan# 2484

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:167 Resp: 4208343

Ion Ratio Lower Upper

167 100

166 21.5 17.1 25.7

139 12.2 9.8 14.8



#78

Di-n-butylphthalate

Concen: 40.214 ng/ul

RT: 18.345 min Scan# 2577

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:149 Resp: 4961844


Ion Ratio Lower Upper

149 100

150 9.2 7.2 10.8

104 4.6 3.7 5.5





#80

Fluoranthene

Concen: 31.907 ng/ul

RT: 19.398 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:202 Resp: 5509315

Ion Ratio Lower Upper

202 100

101 10.6 8.7 13.1

100 7.9 6.9 10.3



#82

Pyrene

Concen: 31.638 ng/ul

RT: 19.757 min Scan# 2817

Delta R.T. 0.000 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:202 Resp: 5650223


Ion Ratio Lower Upper

202 100

101 11.4 10.3 15.5

100 9.5 8.6 13.0





#83

Butylbenzylphthalate

Concen: 53.306 ng/ul

RT: 20.622 min Scan# 2964

Delta R.T. 0.000 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:149 Resp: 2227718

Ion Ratio Lower Upper

149 100

91 65.0 58.2 87.2

206 23.0 16.5 24.7



#84

3,3'-Dichlorobenzidine

Concen: 10.570 ng/ul

RT: 21.392 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:252 Resp: 463724


Ion Ratio Lower Upper

252 100

254 64.7 52.3 78.5

126 10.7 7.6 11.4





#85

Benzo(a)anthracene

Concen: 34.172 ng/ul

RT: 21.469 min Scan# 3108

Delta R.T. 0.000 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:228 Resp: 5646585

Ion Ratio Lower Upper

228 100

229 20.2 16.2 24.2

226 27.6 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 50.310 ng/ul

RT: 21.380 min Scan# 3093

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01




Tgt Ion:149 Resp: 3215347


Ion Ratio Lower Upper

149 100

167 28.5 22.7 34.1

279 4.8 3.8 5.6









#87
Chrysene
Concen: 33.246 ng/ul
RT: 21.521 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:228 Resp: 5311652
Ion Ratio Lower Upper
228 100
226 30.6 24.3 36.5
229 20.3 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 39.251 ng/ul
RT: 22.345 min Scan# 3257
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:149 Resp: 5346610









#90
Benzo(b)fluoranthene
Concen: 33.320 ng/ul
RT: 23.227 min Scan# 3406
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:252 Resp: 5885467
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 8.9 7.4 11.2



#91
Benzo(k)fluoranthene
Concen: 31.989 ng/ul
RT: 23.274 min Scan# 3415
Delta R.T. -0.006 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:252 Resp: 5640027
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 9.1 7.8 11.8









#93
Benzo(a)pyrene
Concen: 35.076 ng/ul
RT: 23.886 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:252 Resp: 5094692
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 9.6 8.5 12.7



#94
Indeno(1,2,3-cd)pyrene
Concen: 45.072 ng/ul
RT: 26.627 min Scan# 3985
Delta R.T. 0.000 min
Lab File: BP012913.D
Acq: 01 Dec 2022 12:01

Tgt Ion:276 Resp: 6316067
Ion Ratio Lower Upper
276 100
138 22.0 18.3 27.5
277 25.5 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 43.663 ng/ul

RT: 26.645 min Scan# 3988

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:278 Resp: 5551915

Ion Ratio Lower Upper

278 100

139 14.7 12.6 18.8

279 24.3 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 44.841 ng/ul

RT: 27.433 min Scan# 4122

Delta R.T. -0.006 min

Lab File: BP012913.D

Acq: 01 Dec 2022 12:01

Tgt Ion:276 Resp: 5404802

Ion Ratio Lower Upper

276 100

138 19.8 16.7 25.1

277 24.0 19.1 28.7

