

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012895.D
 Acq On : 01 Dec 2022 00:39
 Operator : CG/JU
 Sample : N5737-02MS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A4345MS

Quant Time: Dec 01 02:19:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.011	152	549005	20.000	ng/u1	0.00
20) Naphthalene-d8	10.828	136	2221775	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.646	164	1336808	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	2738555	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	2808122	20.000	ng/u1	0.00
88) Perylene-d12	23.998	264	2899187	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	57200	4.465	ng/uL	0.00
4) Pyridine-d5	3.823	84	389235	10.152	ng/u1	0.00
7) Phenol-d5	7.158	99	310854	6.891	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.334	67	691350	24.870	ng/u1	0.00
11) 2-Chlorophenol-d4	7.534	132	732714	21.315	ng/u1	0.00
15) 4-Methylphenol-d8	8.711	113	502168	13.816	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	412952	30.367	ng/u1	0.00
24) 2-Nitrophenol-d4	9.899	143	435744	30.178	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	862023	25.587	ng/u1	0.00
31) 4-Chloroaniline-d4	10.958	131	994777	21.562	ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	2618721	28.460	ng/u1	0.00
49) Acenaphthylene-d8	14.340	160	2987209	27.824	ng/u1	0.00
54) 4-Nitrophenol-d4	14.828	143	157506	9.601	ng/u1	0.00
60) Fluorene-d10	15.640	176	2322967	28.827	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	428026	32.701	ng/u1	0.00
73) Anthracene-d10	17.498	188	3560784	29.538	ng/u1	0.00
81) Pyrene-d10	19.728	212	4209183	25.797	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	4267951	29.970	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.429	88	752538	54.693	ng/uL	99
5) Pyridine	3.840	79	420380	11.206	ng/u1	99
6) Benzaldehyde	7.140	77	656310	29.796	ng/u1	98
8) Phenol	7.187	94	329741	6.971	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.428	93	938270	24.208	ng/u1	99
12) 2-Chlorophenol	7.569	128	765286	20.466	ng/u1	99
13) 2-Methylphenol	8.446	108	569478	15.490	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.540	45	1296946	23.199	ng/u1	99
16) Acetophenone	8.834	105	1376429	23.892	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.822	70	659279	23.231	ng/u1	98
18) 4-Methylphenol	8.775	108	550666	13.600	ng/u1	100
19) Hexachloroethane	9.099	117	352957	24.614	ng/u1	95
22) Nitrobenzene	9.216	77	994643	27.888	ng/u1	100
23) Isophorone	9.746	82	1875982	23.860	ng/u1	99
25) 2-Nitrophenol	9.934	139	481994	27.810	ng/u1	96
26) 2,4-Dimethylphenol	9.987	107	834287	21.663	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.228	93	1212769	25.159	ng/u1	99
29) 2,4-Dichlorophenol	10.463	162	837825	24.006	ng/u1	99
30) Naphthalene	10.875	128	2929628	25.182	ng/u1	100
32) 4-Chloroaniline	10.981	127	968917	20.450	ng/u1	99
33) Hexachlorobutadiene	11.163	225	580810	24.985	ng/u1	99
34) Caprolactam	11.763	113	48334	4.510	ng/u1	98
35) 4-Chloro-3-methylphenol	12.093	107	776497	21.971	ng/u1	99
36) 2-Methylnaphthalene	12.475	142	1991591	24.859	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012895.D
 Acq On : 01 Dec 2022 00:39
 Operator : CG/JU
 Sample : N5737-02MS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A4345MS

Quant Time: Dec 01 02:19:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.693	142	1988547	24.767	ng/u1	100
39) 1,2,4,5-Tetrachloroben...	12.840	216	1139746	26.737	ng/u1	98
40) Hexachlorocyclopentadiene	12.822	237	507692	24.717	ng/u1	98
41) 2,4,6-Trichlorophenol	13.075	196	725339	27.194	ng/u1	99
42) 2,4,5-Trichlorophenol	13.146	196	774793	27.021	ng/u1	99
43) 1,1'-Biphenyl	13.481	154	2616049	26.864	ng/u1	99
44) 2-Chloronaphthalene	13.522	162	2090412	26.832	ng/u1	99
45) 2-Nitroaniline	13.722	65	527630	34.615	ng/u1	96
47) Dimethylphthalate	14.099	163	2630442	27.271	ng/u1	99
48) 2,6-Dinitrotoluene	14.216	165	495946	32.855	ng/u1	99
50) Acenaphthylene	14.369	152	3177014	27.062	ng/u1	100
51) 3-Nitroaniline	14.546	138	472794	27.045	ng/u1	97
52) Acenaphthene	14.710	153	2254148	27.273	ng/u1	99
53) 2,4-Dinitrophenol	14.746	184	251143	25.891	ng/u1	98
55) 4-Nitrophenol	14.840	109	87640	7.353	ng/u1	97
56) Dibenzofuran	15.046	168	3154277	27.688	ng/u1	99
57) 2,4-Dinitrotoluene	14.998	165	702410	31.533	ng/u1	97
58) 2,3,4,6-Tetrachlorophenol	15.263	232	670881	27.146	ng/u1	100
59) Diethylphthalate	15.463	149	2574460	27.574	ng/u1	98
61) Fluorene	15.693	166	2580353	28.244	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.687	204	1330888	28.077	ng/u1	98
63) 4-Nitroaniline	15.710	138	485281	32.569	ng/u1	99
66) 4,6-Dinitro-2-methylph...	15.763	198	415398	29.567	ng/u1	98
67) N-Nitrosodiphenylamine	15.898	169	2197089	28.229	ng/u1	99
68) 4-Bromophenyl-phenylether	16.581	248	835594	31.002	ng/u1	97
69) Hexachlorobenzene	16.698	284	975258	28.389	ng/u1	98
70) Atrazine	16.857	200	711264	26.636	ng/u1	99
71) Pentachlorophenol	17.045	266	549335	26.039	ng/u1	99
72) Phenanthrene	17.445	178	4132261	28.312	ng/u1	99
74) Anthracene	17.534	178	4116075	28.506	ng/u1	100
75) 1,2,3,4-Tetrachloroben...	13.446	216	1137409	26.865	ng/uL	99
76) Pentachlorobenzene	14.963	250	1102907	25.060	ng/uL	100
77) Carbazole	17.798	167	3780072	30.946	ng/u1	100
78) Di-n-butylphthalate	18.345	149	4487200	31.031	ng/u1	100
80) Fluoranthene	19.398	202	4913498	24.210	ng/u1	100
82) Pyrene	19.757	202	5093372	24.264	ng/u1	96
83) Butylbenzylphthalate	20.622	149	1995811	40.630	ng/u1	92
84) 3,3'-Dichlorobenzidine	21.398	252	429607	8.331	ng/u1	99
85) Benzo(a)anthracene	21.469	228	5107773	26.298	ng/u1	100
86) Bis(2-ethylhexyl)phtha...	21.380	149	2882799	38.375	ng/u1	100
87) Chrysene	21.528	228	4836985	25.757	ng/u1	100
89) Di-n-octyl phthalate	22.345	149	4822555	29.129	ng/u1	100
90) Benzo(b)fluoranthene	23.227	252	5301303	24.693	ng/u1	99
91) Benzo(k)fluoranthene	23.280	252	5297504	24.720	ng/u1	99
93) Benzo(a)pyrene	23.886	252	4682853	26.526	ng/u1	100
94) Indeno(1,2,3-cd)pyrene	26.633	276	5885969	34.558	ng/u1	99
95) Dibenzo(a,h)anthracene	26.651	278	5177132	33.498	ng/u1	98
96) Benzo(g,h,i)perylene	27.439	276	5049790	34.470	ng/u1	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 54.693 ng/uL
RT: 3.429 min Scan# 41
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion: 88 Resp: 752538
Ion Ratio Lower Upper
88 100
43 27.3 22.6 34.0
58 74.2 59.6 89.4



#5
Pyridine
Concen: 11.206 ng/uL
RT: 3.840 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion: 79 Resp: 420380
Ion Ratio Lower Upper
79 100
52 73.8 58.9 88.3
51 34.8 27.3 40.9









#6
Benzaldehyde
Concen: 29.796 ng/ul
RT: 7.140 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion: 77 Resp: 656310
Ion Ratio Lower Upper
77 100
105 98.5 81.4 122.2
106 95.4 77.4 116.0



#8
Phenol
Concen: 6.971 ng/ul
RT: 7.187 min Scan# 680
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion: 94 Resp: 329741
Ion Ratio Lower Upper
94 100
65 26.1 20.9 31.3
66 32.1 26.0 39.0









#10
Bis(2-Chloroethyl)ether
Concen: 24.208 ng/ul
RT: 7.428 min Scan# 721
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion: 93 Resp: 938270
Ion Ratio Lower Upper
93 100
63 74.4 59.7 89.5
95 32.8 25.8 38.6



#12
2-Chlorophenol
Concen: 20.466 ng/ul
RT: 7.569 min Scan# 745
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:128 Resp: 765286
Ion Ratio Lower Upper
128 100
64 45.1 35.6 53.4
130 32.1 26.0 39.0





#13

2-Methylphenol

Concen: 15.490 ng/ul

RT: 8.446 min Scan# 894

Delta R.T. -0.006 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39

Instrument :

BNA_P

ClientSampleId :





A4345MS

Tgt Ion:108 Resp: 569478

Ion Ratio Lower Upper

108 100

107 86.9 69.3 103.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 23.199 ng/ul

RT: 8.540 min Scan# 910

Delta R.T. -0.006 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39




Tgt Ion: 45 Resp: 1296946


Ion Ratio Lower Upper

45 100

77 13.5 11.3 16.9

79 10.7 8.6 12.8









#16
Acetophenone
Concen: 23.892 ng/ul
RT: 8.834 min Scan# 961
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:105 Resp: 1376429
Ion Ratio Lower Upper
105 100
77 74.0 59.7 89.5
51 25.6 20.9 31.3



#17
N-Nitroso-di-n-propylamine
Concen: 23.231 ng/ul
RT: 8.822 min Scan# 958
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion: 70 Resp: 659279
Ion Ratio Lower Upper
70 100
42 57.2 46.1 69.1
101 9.8 7.4 11.2
130 22.1 15.8 23.6









#18
4-Methylphenol
Concen: 13.600 ng/ul
RT: 8.775 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:108 Resp: 550666
Ion Ratio Lower Upper
108 100
107 112.3 90.2 135.2



#19
Hexachloroethane
Concen: 24.614 ng/ul
RT: 9.099 min Scan# 1005
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:117 Resp: 352957
Ion Ratio Lower Upper
117 100
201 112.1 85.7 128.5
199 71.1 52.9 79.3









#22
Nitrobenzene
Concen: 27.888 ng/ul
RT: 9.216 min Scan# 1025
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion: 77 Resp: 994643
Ion Ratio Lower Upper
77 100
123 47.5 38.2 57.2
65 12.6 10.2 15.4



#23
Isophorone
Concen: 23.860 ng/ul
RT: 9.746 min Scan# 1115
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion: 82 Resp: 1875982
Ion Ratio Lower Upper
82 100
95 6.6 5.2 7.8
138 18.4 14.5 21.7









#25
2-Nitrophenol
Concen: 27.810 ng/ul
RT: 9.934 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:139 Resp: 481994
Ion Ratio Lower Upper
139 100
65 45.4 39.5 59.3
109 26.2 21.4 32.0



#26
2,4-Dimethylphenol
Concen: 21.663 ng/ul
RT: 9.987 min Scan# 1156
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:107 Resp: 834287
Ion Ratio Lower Upper
107 100
121 53.0 41.8 62.6
122 88.2 72.3 108.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 25.159 ng/ul
 RT: 10.228 min Scan# 1197
 Delta R.T. -0.006 min
 Lab File: BP012895.D
 Acq: 01 Dec 2022 00:39

Instrument :
 BNA_P
 ClientSampleId :
 A4345MS

Tgt Ion: 93 Resp: 1212769




Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.1	39.1
123	11.6	9.0	13.6




#29
 2,4-Dichlorophenol
 Concen: 24.006 ng/ul
 RT: 10.463 min Scan# 1237
 Delta R.T. -0.006 min
 Lab File: BP012895.D
 Acq: 01 Dec 2022 00:39

Tgt Ion:162 Resp: 837825

Ion	Ratio	Lower	Upper
162	100		
164	65.8	52.9	79.3
98	34.0	26.6	40.0









#30
Naphthalene
Concen: 25.182 ng/ul
RT: 10.875 min Scan# 1307
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:128 Resp: 2929628
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 20.450 ng/ul
RT: 10.981 min Scan# 1325
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:127 Resp: 968917
Ion Ratio Lower Upper
127 100
129 32.5 25.7 38.5









#33
Hexachlorobutadiene
Concen: 24.985 ng/ul
RT: 11.163 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:225 Resp: 580810
Ion Ratio Lower Upper
225 100
223 62.9 49.6 74.4
227 64.9 51.0 76.6



#34
Caprolactam
Concen: 4.510 ng/ul
RT: 11.763 min Scan# 1458
Delta R.T. 0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:113 Resp: 48334
Ion Ratio Lower Upper
113 100
55 180.8 145.9 218.9
56 130.9 107.1 160.7









#35
4-Chloro-3-methylphenol
Concen: 21.971 ng/ul
RT: 12.093 min Scan# 1514
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:107 Resp: 776497
Ion Ratio Lower Upper
107 100
144 27.4 21.4 32.2
142 83.9 66.6 100.0



#36
2-Methylnaphthalene
Concen: 24.859 ng/ul
RT: 12.475 min Scan# 1579
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:142 Resp: 1991591
Ion Ratio Lower Upper
142 100
141 87.2 70.2 105.2









#37
1-Methylnaphthalene
Concen: 24.767 ng/ul
RT: 12.693 min Scan# 1616
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:142 Resp: 1988547
Ion Ratio Lower Upper
142 100
141 91.2 73.3 109.9
116 3.5 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 26.737 ng/ul
RT: 12.840 min Scan# 1641
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:216 Resp: 1139746
Ion Ratio Lower Upper
216 100
214 77.9 63.0 94.6
179 19.4 16.2 24.4
108 14.6 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 24.717 ng/ul
RT: 12.822 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:237 Resp: 507692
Ion Ratio Lower Upper
237 100
235 64.4 50.2 75.4
272 12.9 10.2 15.2



#41
2,4,6-Trichlorophenol
Concen: 27.194 ng/ul
RT: 13.075 min Scan# 1681
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:196 Resp: 725339
Ion Ratio Lower Upper
196 100
198 94.9 76.5 114.7
200 31.1 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 27.021 ng/ul
RT: 13.146 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:196 Resp: 774793
Ion Ratio Lower Upper
196 100
198 96.3 77.6 116.4
200 31.9 24.2 36.4



#43
1,1'-Biphenyl
Concen: 26.864 ng/ul
RT: 13.481 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:154 Resp: 2616049
Ion Ratio Lower Upper
154 100
153 39.5 32.0 48.0
76 12.4 10.2 15.4





#44

2-Chloronaphthalene

Concen: 26.832 ng/ul

RT: 13.522 min Scan# 1757

Delta R.T. -0.006 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:162 Resp: 2090412

Ion Ratio Lower Upper

162 100

164 32.7 26.5 39.7

127 36.2 28.8 43.2



#45

2-Nitroaniline

Concen: 34.615 ng/ul

RT: 13.722 min Scan# 1791

Delta R.T. -0.006 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39




Tgt Ion: 65 Resp: 527630


Ion Ratio Lower Upper

65 100

92 71.1 57.8 86.8

138 111.6 94.6 142.0





#47

Dimethylphthalate

Concen: 27.271 ng/ul

RT: 14.099 min Scan# 1855

Delta R.T. 0.000 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:163 Resp: 2630442

Ion Ratio Lower Upper

163 100

194 4.1 3.4 5.0

164 10.1 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 32.855 ng/ul

RT: 14.216 min Scan# 1875

Delta R.T. -0.006 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39




Tgt Ion:165 Resp: 495946


Ion Ratio Lower Upper

165 100

89 52.3 42.7 64.1

121 20.7 16.9 25.3









#50
Acenaphthylene
Concen: 27.062 ng/ul
RT: 14.369 min Scan# 1901
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:152 Resp: 3177014
Ion Ratio Lower Upper
152 100
151 20.1 15.9 23.9
153 13.1 10.7 16.1



#51
3-Nitroaniline
Concen: 27.045 ng/ul
RT: 14.546 min Scan# 1931
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:138 Resp: 472794
Ion Ratio Lower Upper
138 100
108 10.9 9.9 14.9
92 109.3 90.2 135.2









#52
Acenaphthene
Concen: 27.273 ng/ul
RT: 14.710 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:153 Resp: 2254148
Ion Ratio Lower Upper
153 100
152 47.5 38.2 57.4
154 88.6 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 25.891 ng/ul
RT: 14.746 min Scan# 1965
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:184 Resp: 251143
Ion Ratio Lower Upper
184 100
63 60.4 48.1 72.1
154 57.1 47.4 71.2









#55
4-Nitrophenol
Concen: 7.353 ng/ul
RT: 14.840 min Scan# 1981
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:109 Resp: 87640
Ion Ratio Lower Upper
109 100
139 149.6 119.2 178.8
65 145.1 110.7 166.1



#56
Dibenzofuran
Concen: 27.688 ng/ul
RT: 15.046 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:168 Resp: 3154277
Ion Ratio Lower Upper
168 100
139 33.9 27.7 41.5









#57
2,4-Dinitrotoluene
Concen: 31.533 ng/ul
RT: 14.998 min Scan# 2008
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:165 Resp: 702410
Ion Ratio Lower Upper
165 100
63 40.5 31.0 46.4
182 2.6 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 27.146 ng/ul
RT: 15.263 min Scan# 2053
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:232 Resp: 670881
Ion Ratio Lower Upper
232 100
131 40.2 32.0 48.0
130 2.2 1.5 2.3
166 29.3 23.2 34.8









#59
Diethylphthalate
Concen: 27.574 ng/ul
RT: 15.463 min Scan# 2087
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:149 Resp: 2574460
Ion Ratio Lower Upper
149 100
177 23.7 18.1 27.1
150 12.2 9.8 14.8



#61
Fluorene
Concen: 28.244 ng/ul
RT: 15.693 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:166 Resp: 2580353
Ion Ratio Lower Upper
166 100
165 96.6 78.2 117.2
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 28.077 ng/ul
RT: 15.687 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:204 Resp: 1330888
Ion Ratio Lower Upper
204 100
206 33.1 26.0 39.0
141 49.5 40.7 61.1



#63
4-Nitroaniline
Concen: 32.569 ng/ul
RT: 15.710 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:138 Resp: 485281
Ion Ratio Lower Upper
138 100
92 49.4 38.6 57.8
108 70.6 56.4 84.6









#66
4,6-Dinitro-2-methylphenol
Concen: 29.567 ng/ul
RT: 15.763 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:198 Resp: 415398
Ion Ratio Lower Upper
198 100
182 4.4 4.1 6.1
77 22.8 18.9 28.3



#67
N-Nitrosodiphenylamine
Concen: 28.229 ng/ul
RT: 15.898 min Scan# 2161
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:169 Resp: 2197089
Ion Ratio Lower Upper
169 100
168 66.9 53.2 79.8
167 36.4 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 31.002 ng/ul
RT: 16.581 min Scan# 2277
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:248 Resp: 835594
Ion Ratio Lower Upper
248 100
250 99.8 78.6 118.0
141 64.2 48.7 73.1



#69
Hexachlorobenzene
Concen: 28.389 ng/ul
RT: 16.698 min Scan# 2297
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:284 Resp: 975258
Ion Ratio Lower Upper
284 100
142 30.0 25.0 37.6
249 30.2 24.6 36.8





#70

Atrazine

Concen: 26.636 ng/ul

RT: 16.857 min Scan# 2324

Delta R.T. 0.006 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:200 Resp: 711264

Ion Ratio Lower Upper

200 100

173 25.4 20.3 30.5

215 49.4 40.1 60.1



#71

Pentachlorophenol

Concen: 26.039 ng/ul

RT: 17.045 min Scan# 2356

Delta R.T. 0.000 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39




Tgt Ion:266 Resp: 549335


Ion Ratio Lower Upper

266 100

264 64.3 50.7 76.1

268 64.7 51.8 77.8









#72
Phenanthrene
Concen: 28.312 ng/ul
RT: 17.445 min Scan# 2424
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:178 Resp: 4132261
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 19.6 15.4 23.0



#74
Anthracene
Concen: 28.506 ng/ul
RT: 17.534 min Scan# 2439
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:178 Resp: 4116075
Ion Ratio Lower Upper
178 100
179 15.6 12.4 18.6
176 18.6 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 26.865 ng/uL
RT: 13.446 min Scan# 1744
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:216 Resp: 1137409
Ion Ratio Lower Upper
216 100
214 78.5 62.3 93.5
218 48.7 39.1 58.7



#76
Pentachlorobenzene
Concen: 25.060 ng/uL
RT: 14.963 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:250 Resp: 1102907
Ion Ratio Lower Upper
250 100
248 63.4 50.6 76.0
252 63.4 50.5 75.7





#77

Carbazole

Concen: 30.946 ng/ul

RT: 17.798 min Scan# 2484

Delta R.T. -0.006 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:167 Resp: 3780072

Ion Ratio Lower Upper

167 100

166 21.4 17.1 25.7

139 12.2 9.8 14.8



#78

Di-n-butylphthalate

Concen: 31.031 ng/ul

RT: 18.345 min Scan# 2577

Delta R.T. -0.006 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39




Tgt Ion:149 Resp: 4487200


Ion Ratio Lower Upper

149 100

150 9.2 7.2 10.8

104 4.5 3.7 5.5





#80

Fluoranthene

Concen: 24.210 ng/ul

RT: 19.398 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:202 Resp: 4913498

Ion Ratio Lower Upper

202 100

101 10.9 8.7 13.1

100 8.3 6.9 10.3



#82

Pyrene

Concen: 24.264 ng/ul

RT: 19.757 min Scan# 2817

Delta R.T. 0.000 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39




Tgt Ion:202 Resp: 5093372


Ion Ratio Lower Upper

202 100

101 11.5 10.3 15.5

100 9.5 8.6 13.0





#83

Butylbenzylphthalate

Concen: 40.630 ng/ul

RT: 20.622 min Scan# 2964

Delta R.T. 0.000 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:149 Resp: 1995811

Ion Ratio Lower Upper

149 100

91 65.3 58.2 87.2

206 22.7 16.5 24.7



#84

3,3'-Dichlorobenzidine

Concen: 8.331 ng/ul

RT: 21.398 min Scan# 3096

Delta R.T. 0.000 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39




Tgt Ion:252 Resp: 429607


Ion Ratio Lower Upper

252 100

254 64.7 52.3 78.5

126 9.9 7.6 11.4









#85
Benzo(a)anthracene
Concen: 26.298 ng/ul
RT: 21.469 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:228 Resp: 5107773
Ion Ratio Lower Upper
228 100
229 20.3 16.2 24.2
226 27.7 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 38.375 ng/ul
RT: 21.380 min Scan# 3093
Delta R.T. -0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:149 Resp: 2882799
Ion Ratio Lower Upper
149 100
167 28.2 22.7 34.1
279 4.5 3.8 5.6





#87

Chrysene

Concen: 25.757 ng/ul

RT: 21.528 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39

Instrument :

BNA_P

ClientSampleId :

A4345MS





Tgt Ion:228 Resp: 4836985

Ion Ratio Lower Upper

228 100

226 30.3 24.3 36.5

229 19.7 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 29.129 ng/ul




RT: 22.345 min Scan# 3257


Delta R.T. 0.000 min

Lab File: BP012895.D

Acq: 01 Dec 2022 00:39

Tgt Ion:149 Resp: 4822555









#90
Benzo(b)fluoranthene
Concen: 24.693 ng/ul
RT: 23.227 min Scan# 3407
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:252 Resp: 5301303
Ion Ratio Lower Upper
252 100
253 22.5 17.6 26.4
125 9.1 7.4 11.2



#91
Benzo(k)fluoranthene
Concen: 24.720 ng/ul
RT: 23.280 min Scan# 3416
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:252 Resp: 5297504
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 9.0 7.8 11.8









#93
Benzo(a)pyrene
Concen: 26.526 ng/ul
RT: 23.886 min Scan# 3519
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39




Instrument :
BNA_P
ClientSampleId :
A4345MS


Tgt Ion:252 Resp: 4682853
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 10.3 8.5 12.7



#94
Indeno(1,2,3-cd)pyrene
Concen: 34.558 ng/ul
RT: 26.633 min Scan# 3986
Delta R.T. 0.006 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion:276 Resp: 5885969
Ion Ratio Lower Upper
276 100
138 22.1 18.3 27.5
277 25.1 19.8 29.8









#95
Dibenzo(a,h)anthracene
Concen: 33.498 ng/ul
RT: 26.651 min Scan# 3989
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Instrument :
BNA_P
ClientSampleId :
A4345MS

Tgt Ion: 278 Resp: 5177132

Ion	Ratio	Lower	Upper
278	100		
139	14.6	12.6	18.8
279	23.9	19.4	29.2



#96
Benzo(g,h,i)perylene
Concen: 34.470 ng/ul
RT: 27.439 min Scan# 4123
Delta R.T. 0.000 min
Lab File: BP012895.D
Acq: 01 Dec 2022 00:39

Tgt Ion: 276 Resp: 5049790

Ion	Ratio	Lower	Upper
276	100		
138	20.0	16.7	25.1
277	24.1	19.1	28.7

