

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012898.D
 Acq On : 01 Dec 2022 02:21
 Operator : CG/JU
 Sample : N5737-05
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A4347

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/01/2022
 Supervised By :mohammad ahmed 12/02/2022

Quant Time: Dec 01 02:59:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	456831	20.000	ng/u1	0.00
20) Naphthalene-d8	10.822	136	1835255	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.645	164	1136398	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	2440706	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	2504645	20.000	ng/u1	0.00
88) Perylene-d12	23.992	264	2618798	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	52413	4.916	ng/uL	0.00
4) Pyridine-d5	3.822	84	367973	11.534	ng/u1	0.00
7) Phenol-d5	7.157	99	251228	6.692	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.334	67	644512	27.863	ng/u1	0.00
11) 2-Chlorophenol-d4	7.534	132	662730	23.169	ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	442969	14.646	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	396186	35.270	ng/u1	0.00
24) 2-Nitrophenol-d4	9.899	143	395269	33.140	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	748794	26.907	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	915254	24.017	ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	2499159	31.950	ng/u1	0.00
49) Acenaphthylene-d8	14.339	160	2821380	30.913	ng/u1	0.00
54) 4-Nitrophenol-d4	14.828	143	69802m	5.005	ng/u1	0.00
60) Fluorene-d10	15.634	176	2220051	32.408	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	288614	24.741	ng/u1	0.00
73) Anthracene-d10	17.498	188	3547977	33.024	ng/u1	0.00
81) Pyrene-d10	19.727	212	4211613	28.939	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	4252539	33.059	ng/u1	0.00
Target Compounds						
2) 1,4-Dioxane	3.428	88	549486	47.993	ng/uL	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
Data File : BP012898.D
Acq On : 01 Dec 2022 02:21
Operator : CG/JU
Sample : N5737-05
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
BNA_P
ClientSampleId :
A4347

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022


Quant Time: Dec 01 02:59:08 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 29 22:53:34 2022
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 47.993 ng/uL
 RT: 3.428 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: BP012898.D
 Acq: 01 Dec 2022 02:21

Instrument :
 BNA_P
 ClientSampleId :
 A4347



Tgt Ion: 88 Resp: 549486
 Ion Ratio Lower Upper
 88 100
 43 27.0 22.6 34.0
 58 73.4 59.6 89.4

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/01/2022
 Supervised By :mohammad ahmed 12/02/2022

