

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012901.D
 Acq On : 01 Dec 2022 04:03
 Operator : CG/JU
 Sample : PB149294BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS294

Quant Time: Dec 01 04:57:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 12/01/2022
 Supervised By :mohammad ahmed 12/02/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.011	152	654156	20.000	ng/u1	0.00
20) Naphthalene-d8	10.822	136	3006386	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.646	164	1970197	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	3815544	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	2481457	20.000	ng/u1	0.00
88) Perylene-d12	23.992	264	2348138	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	85746	5.617	ng/uL	0.00
4) Pyridine-d5	3.817	84	1125076	24.627	ng/u1	0.00
7) Phenol-d5	7.158	99	1700732	31.639	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.334	67	1005313	30.351	ng/u1	0.00
11) 2-Chlorophenol-d4	7.534	132	1323446	32.311	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	1366664	31.556	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	659951	35.865	ng/u1	0.00
24) 2-Nitrophenol-d4	9.899	143	748658	38.318	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	1458609	31.995	ng/u1	0.00
31) 4-Chloroaniline-d4	10.958	131	1841618	29.500	ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	4142507	30.547	ng/u1	0.00
49) Acenaphthylene-d8	14.340	160	4787246	30.255	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	673016	27.836	ng/u1	0.00
60) Fluorene-d10	15.640	176	3578199	30.128	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	639839	35.086	ng/u1	0.00
73) Anthracene-d10	17.498	188	5164061	30.746	ng/u1	0.00
81) Pyrene-d10	19.728	212	4912232	34.069	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	3640331	31.562	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.429	88	177525	10.828	ng/uL	100
5) Pyridine	3.840	79	1187228	26.559	ng/u1	98
6) Benzaldehyde	7.140	77	921653m	35.116	ng/u1	
8) Phenol	7.187	94	1742397	30.913	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.428	93	1387209	30.038	ng/u1	100
12) 2-Chlorophenol	7.569	128	1383104	31.043	ng/u1	99
13) 2-Methylphenol	8.446	108	1347249	30.754	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.534	45	1945438	29.205	ng/u1	99
16) Acetophenone	8.840	105	2111557	30.761	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.822	70	1049936	31.049	ng/u1	98
18) 4-Methylphenol	8.781	108	1482791	30.734	ng/u1	97
19) Hexachloroethane	9.099	117	535553	31.344	ng/u1	94
22) Nitrobenzene	9.216	77	1533660	31.779	ng/u1	100
23) Isophorone	9.752	82	3089943	29.043	ng/u1	99
25) 2-Nitrophenol	9.934	139	827543	35.286	ng/u1	96
26) 2,4-Dimethylphenol	9.987	107	1498706	28.759	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.228	93	1941247	29.761	ng/u1	100
29) 2,4-Dichlorophenol	10.463	162	1447281	30.646	ng/u1	99
30) Naphthalene	10.875	128	4553882	28.927	ng/u1	100
32) 4-Chloroaniline	10.981	127	1777704	27.728	ng/u1	99
33) Hexachlorobutadiene	11.163	225	946258	30.082	ng/u1	100
34) Caprolactam	11.763	113	428533m	29.550	ng/u1	
35) 4-Chloro-3-methylphenol	12.099	107	1469842	30.735	ng/u1	96
36) 2-Methylnaphthalene	12.475	142	3205101	29.566	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012901.D
 Acq On : 01 Dec 2022 04:03
 Operator : CG/JU
 Sample : PB149294BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS294

Quant Time: Dec 01 04:57:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 12/01/2022
 Supervised By :mohammad ahmed 12/02/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.693	142	3195969	29.416	ng/u1	100
39) 1,2,4,5-Tetrachloroben...	12.840	216	1861460	29.629	ng/u1	99
40) Hexachlorocyclopentadiene	12.822	237	956040	31.581	ng/u1	99
41) 2,4,6-Trichlorophenol	13.075	196	1198206	30.481	ng/u1	99
42) 2,4,5-Trichlorophenol	13.146	196	1301617	30.801	ng/u1	100
43) 1,1'-Biphenyl	13.481	154	4226340	29.447	ng/u1	99
44) 2-Chloronaphthalene	13.522	162	3369913	29.350	ng/u1	99
45) 2-Nitroaniline	13.722	65	859490	38.259	ng/u1	97
47) Dimethylphthalate	14.099	163	4215844	29.657	ng/u1	99
48) 2,6-Dinitrotoluene	14.222	165	836455	37.599	ng/u1	95
50) Acenaphthylene	14.369	152	5086827	29.400	ng/u1	100
51) 3-Nitroaniline	14.546	138	820922	31.862	ng/u1	98
52) Acenaphthene	14.710	153	3535661	29.026	ng/u1	99
53) 2,4-Dinitrophenol	14.751	184	421576	29.490	ng/u1	93
55) 4-Nitrophenol	14.846	109	465541	26.503	ng/u1	98
56) Dibenzofuran	15.046	168	4941599	29.432	ng/u1	99
57) 2,4-Dinitrotoluene	15.004	165	1146408	34.920	ng/u1	97
58) 2,3,4,6-Tetrachlorophenol	15.263	232	1064162	29.217	ng/u1	100
59) Diethylphthalate	15.463	149	4055102	29.469	ng/u1	99
61) Fluorene	15.693	166	3908570	29.028	ng/u1	100
62) 4-Chlorophenyl-phenyle...	15.687	204	2108734	30.185	ng/u1	98
63) 4-Nitroaniline	15.710	138	737035	33.563	ng/u1	98
66) 4,6-Dinitro-2-methylph...	15.769	198	667734	34.112	ng/u1#	92
67) N-Nitrosodiphenylamine	15.898	169	3404153	31.392	ng/u1	100
68) 4-Bromophenyl-phenylether	16.581	248	1311363	34.920	ng/u1	98
69) Hexachlorobenzene	16.698	284	1502006	31.381	ng/u1	98
70) Atrazine	16.851	200	609304	16.377	ng/u1	100
71) Pentachlorophenol	17.045	266	752662	25.607	ng/u1	99
72) Phenanthrene	17.445	178	6027120	29.638	ng/u1	100
74) Anthracene	17.534	178	5959507	29.623	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.446	216	1851809	31.393	ng/uL	100
76) Pentachlorobenzene	14.963	250	1786657	29.137	ng/uL	99
77) Carbazole	17.798	167	4979538	29.258	ng/u1	100
78) Di-n-butylphthalate	18.345	149	6081573	30.185	ng/u1	99
80) Fluoranthene	19.404	202	6075570	33.876	ng/u1	97
82) Pyrene	19.757	202	6077787	32.766	ng/u1	97
83) Butylbenzylphthalate	20.622	149	1948309	44.885	ng/u1	93
84) 3,3'-Dichlorobenzidine	21.398	252	1428599	31.350	ng/u1	98
85) Benzo(a)anthracene	21.469	228	4828301	28.132	ng/u1	99
86) Bis(2-ethylhexyl)phtha...	21.380	149	2619414	39.460	ng/u1	100
87) Chrysene	21.522	228	4501830	27.128	ng/u1	99
89) Di-n-octyl phthalate	22.345	149	4024445	30.012	ng/u1	100
90) Benzo(b)fluoranthene	23.227	252	4612155	26.524	ng/u1	99
91) Benzo(k)fluoranthene	23.275	252	4524739	26.069	ng/u1	100
93) Benzo(a)pyrene	23.886	252	4112406	28.761	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.627	276	5418330	39.278	ng/u1	99
95) Dibenzo(a,h)anthracene	26.651	278	4769704	38.105	ng/u1	99
96) Benzo(g,h,i)perylene	27.433	276	4721179	39.789	ng/u1	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP113022\
 Data File : BP012901.D
 Acq On : 01 Dec 2022 04:03
 Operator : CG/JU
 Sample : PB149294BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS294

Quant Time: Dec 01 04:57:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
 Supervised By :mohammad ahmed 12/02/2022





#2
1,4-Dioxane
Concen: 10.828 ng/uL
RT: 3.429 min Scan# 41
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion: 88 Resp: 177525
Ion Ratio Lower Upper
88 100
43 28.7 22.6 34.0
58 74.5 59.6 89.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#5
Pyridine
Concen: 26.559 ng/uL
RT: 3.840 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03



Tgt Ion: 79 Resp: 1187228
Ion Ratio Lower Upper
79 100
52 71.7 58.9 88.3
51 34.0 27.3 40.9









#6
Benzaldehyde
Concen: 35.116 ng/ul m
RT: 7.140 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03

Instrument :
BNA_P
ClientSampleId :
SLCS294

Tgt Ion: 77 Resp: 921653
Ion Ratio Lower Upper
77 100
105 100.2 81.4 122.2
106 97.1 77.4 116.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#8
Phenol
Concen: 30.913 ng/ul
RT: 7.187 min Scan# 680
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion: 94 Resp: 1742397
Ion Ratio Lower Upper
94 100
65 26.4 20.9 31.3
66 33.0 26.0 39.0





#10
Bis(2-Chloroethyl)ether
Concen: 30.038 ng/ul
RT: 7.428 min Scan# 721
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion: 93 Resp: 1387209
Ion Ratio Lower Upper
93 100
63 74.2 59.7 89.5
95 32.4 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#12
2-Chlorophenol
Concen: 31.043 ng/ul
RT: 7.569 min Scan# 745
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03

Tgt Ion:128 Resp: 1383104
Ion Ratio Lower Upper
128 100
64 43.5 35.6 53.4
130 32.2 26.0 39.0









#13
2-Methylphenol
Concen: 30.754 ng/ul
RT: 8.446 min Scan# 894
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03

Instrument :
BNA_P
ClientSampleId :
SLCS294

Tgt Ion:108 Resp: 1347249
Ion Ratio Lower Upper
108 100
107 86.3 69.3 103.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 29.205 ng/ul
RT: 8.534 min Scan# 909
Delta R.T. -0.012 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion: 45 Resp: 1945438
Ion Ratio Lower Upper
45 100
77 14.5 11.3 16.9
79 10.6 8.6 12.8





#16
Acetophenone
Concen: 30.761 ng/ul
RT: 8.840 min Scan# 961
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion: 105 Resp: 211155
Ion Ratio Lower Upper
105 100
77 74.1 59.7 89.5
51 25.0 20.9 31.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#17
N-Nitroso-di-n-propylamine
Concen: 31.049 ng/ul
RT: 8.822 min Scan# 958
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion: 70 Resp: 1049936
Ion Ratio Lower Upper
70 100
42 56.9 46.1 69.1
101 9.8 7.4 11.2
130 22.2 15.8 23.6





#18
4-Methylphenol
Concen: 30.734 ng/ul
RT: 8.781 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion:108 Resp: 1482791
Ion Ratio Lower Upper
108 100
107 109.8 90.2 135.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#19
Hexachloroethane
Concen: 31.344 ng/ul
RT: 9.099 min Scan# 1005
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion:117 Resp: 535553
Ion Ratio Lower Upper
117 100
201 114.0 85.7 128.5
199 71.0 52.9 79.3





#22
Nitrobenzene
Concen: 31.779 ng/ul
RT: 9.216 min Scan# 1025
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion: 77 Resp: 1533660
Ion Ratio Lower Upper
77 100
123 47.6 38.2 57.2
65 12.8 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#23
Isophorone
Concen: 29.043 ng/ul
RT: 9.752 min Scan# 1116
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03



Tgt Ion: 82 Resp: 3089943
Ion Ratio Lower Upper
82 100
95 6.6 5.2 7.8
138 18.8 14.5 21.7









#25
2-Nitrophenol
Concen: 35.286 ng/ul
RT: 9.934 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03

Instrument :
BNA_P
ClientSampleId :
SLCS294

Tgt Ion:139 Resp: 827543
Ion Ratio Lower Upper
139 100
65 45.7 39.5 59.3
109 25.3 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#26
2,4-Dimethylphenol
Concen: 28.759 ng/ul
RT: 9.987 min Scan# 1156
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion:107 Resp: 1498706
Ion Ratio Lower Upper
107 100
121 52.4 41.8 62.6
122 89.1 72.3 108.5





#27
Bis(2-Chloroethoxy)methane
Concen: 29.761 ng/ul
RT: 10.228 min Scan# 1197
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion: 93 Resp: 1941247
Ion Ratio Lower Upper
93 100
95 32.5 26.1 39.1
123 11.4 9.0 13.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#29
2,4-Dichlorophenol
Concen: 30.646 ng/ul
RT: 10.463 min Scan# 1237
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:162 Resp: 1447281
Ion Ratio Lower Upper
162 100
164 66.4 52.9 79.3
98 33.8 26.6 40.0





#30
Naphthalene
Concen: 28.927 ng/ul
RT: 10.875 min Scan# 1307
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion:128 Resp: 4553882
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.0 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#32
4-Chloroaniline
Concen: 27.728 ng/ul
RT: 10.981 min Scan# 1325
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion:127 Resp: 1777704
Ion Ratio Lower Upper
127 100
129 32.5 25.7 38.5





#33
Hexachlorobutadiene
Concen: 30.082 ng/ul
RT: 11.163 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:225 Resp: 946258
Ion Ratio Lower Upper
225 100
223 62.2 49.6 74.4
227 63.9 51.0 76.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#34
Caprolactam
Concen: 29.550 ng/ul m
RT: 11.763 min Scan# 1458
Delta R.T. 0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:113 Resp: 428533
Ion Ratio Lower Upper
113 100
55 170.2 145.9 218.9
56 122.0 107.1 160.7





#35
4-Chloro-3-methylphenol
Concen: 30.735 ng/ul
RT: 12.099 min Scan# 1514
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion:107 Resp: 1469842
Ion Ratio Lower Upper
107 100
144 28.4 21.4 32.2
142 86.7 66.6 100.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#36
2-Methylnaphthalene
Concen: 29.566 ng/ul
RT: 12.475 min Scan# 1579
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion:142 Resp: 3205101
Ion Ratio Lower Upper
142 100
141 87.7 70.2 105.2





#37
1-Methylnaphthalene
Concen: 29.416 ng/ul
RT: 12.693 min Scan# 1617
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:142 Resp: 3195969
Ion Ratio Lower Upper
142 100
141 91.2 73.3 109.9
116 3.6 2.9 4.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 29.629 ng/ul
RT: 12.840 min Scan# 1641
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:216 Resp: 1861460
Ion Ratio Lower Upper
216 100
214 79.0 63.0 94.6
179 19.7 16.2 24.4
108 14.8 12.7 19.1





#40
Hexachlorocyclopentadiene
Concen: 31.581 ng/ul
RT: 12.822 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:237 Resp: 956040
Ion Ratio Lower Upper
237 100
235 62.3 50.2 75.4
272 13.2 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#41
2,4,6-Trichlorophenol
Concen: 30.481 ng/ul
RT: 13.075 min Scan# 1681
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:196 Resp: 1198206
Ion Ratio Lower Upper
196 100
198 96.5 76.5 114.7
200 30.8 25.3 37.9





#42
2,4,5-Trichlorophenol
Concen: 30.801 ng/ul
RT: 13.146 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:196 Resp: 1301617
Ion Ratio Lower Upper
196 100
198 96.9 77.6 116.4
200 30.8 24.2 36.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#43
1,1'-Biphenyl
Concen: 29.447 ng/ul
RT: 13.481 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:154 Resp: 4226340
Ion Ratio Lower Upper
154 100
153 40.2 32.0 48.0
76 12.0 10.2 15.4





#44
2-Chloronaphthalene
Concen: 29.350 ng/ul
RT: 13.522 min Scan# 1757
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion:162 Resp: 3369913
Ion Ratio Lower Upper
162 100
164 32.8 26.5 39.7
127 35.7 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#45
2-Nitroaniline
Concen: 38.259 ng/ul
RT: 13.722 min Scan# 1791
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion: 65 Resp: 859490
Ion Ratio Lower Upper
65 100
92 72.0 57.8 86.8
138 113.6 94.6 142.0





#47
Dimethylphthalate
Concen: 29.657 ng/ul
RT: 14.099 min Scan# 1855
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:163 Resp: 421584
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.0
164 10.1 7.8 11.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#48
2,6-Dinitrotoluene
Concen: 37.599 ng/ul
RT: 14.222 min Scan# 1876
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:165 Resp: 836455
Ion Ratio Lower Upper
165 100
89 49.2 42.7 64.1
121 19.2 16.9 25.3





#50
Acenaphthylene
Concen: 29.400 ng/ul
RT: 14.369 min Scan# 1901
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:152 Resp: 508682
Ion Ratio Lower Upper
152 100
151 19.9 15.9 23.9
153 13.3 10.7 16.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#51
3-Nitroaniline
Concen: 31.862 ng/ul
RT: 14.546 min Scan# 1931
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:138 Resp: 820922
Ion Ratio Lower Upper
138 100
108 10.6 9.9 14.9
92 111.0 90.2 135.2





#52
Acenaphthene
Concen: 29.026 ng/ul
RT: 14.710 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:153 Resp: 353566
Ion Ratio Lower Upper
153 100
152 47.5 38.2 57.4
154 88.5 69.8 104.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#53
2,4-Dinitrophenol
Concen: 29.490 ng/ul
RT: 14.751 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:184 Resp: 421576
Ion Ratio Lower Upper
184 100
63 54.5 48.1 72.1
154 54.0 47.4 71.2





#55
4-Nitrophenol
Concen: 26.503 ng/ul
RT: 14.846 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion:109 Resp: 46554
Ion Ratio Lower Upper
109 100
139 152.4 119.2 178.8
65 138.7 110.7 166.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#56
Dibenzofuran
Concen: 29.432 ng/ul
RT: 15.046 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion:168 Resp: 4941599
Ion Ratio Lower Upper
168 100
139 34.1 27.7 41.5





#57
2,4-Dinitrotoluene
Concen: 34.920 ng/ul
RT: 15.004 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:165 Resp: 1146408
Ion Ratio Lower Upper
165 100
63 36.8 31.0 46.4
182 2.7 2.0 3.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 29.217 ng/ul
RT: 15.263 min Scan# 2053
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:232 Resp: 1064162
Ion Ratio Lower Upper
232 100
131 39.7 32.0 48.0
130 1.8 1.5 2.3
166 29.1 23.2 34.8





#59
Diethylphthalate
Concen: 29.469 ng/ul
RT: 15.463 min Scan# 2087
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:149 Resp: 4055102
Ion Ratio Lower Upper
149 100
177 23.7 18.1 27.1
150 12.3 9.8 14.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#61
Fluorene
Concen: 29.028 ng/ul
RT: 15.693 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:166 Resp: 3908570
Ion Ratio Lower Upper
166 100
165 98.0 78.2 117.2
167 13.8 10.9 16.3





#62
4-Chlorophenyl-phenylether
Concen: 30.185 ng/ul
RT: 15.687 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:204 Resp: 2108734
Ion Ratio Lower Upper
204 100
206 33.2 26.0 39.0
141 48.9 40.7 61.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#63
4-Nitroaniline
Concen: 33.563 ng/ul
RT: 15.710 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:138 Resp: 737035
Ion Ratio Lower Upper
138 100
92 48.1 38.6 57.8
108 67.7 56.4 84.6





#66
4,6-Dinitro-2-methylphenol
Concen: 34.112 ng/ul
RT: 15.769 min Scan# 2138
Delta R.T. 0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion:198 Resp: 667734
Ion Ratio Lower Upper
198 100
182 3.9 4.1 6.1
77 19.2 18.9 28.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#67
N-Nitrosodiphenylamine
Concen: 31.392 ng/ul
RT: 15.898 min Scan# 2161
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion:169 Resp: 3404153
Ion Ratio Lower Upper
169 100
168 66.4 53.2 79.8
167 36.5 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 34.920 ng/ul
RT: 16.581 min Scan# 2277
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion:248 Resp: 1311363
Ion Ratio Lower Upper
248 100
250 96.7 78.6 118.0
141 63.8 48.7 73.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#69
Hexachlorobenzene
Concen: 31.381 ng/ul
RT: 16.698 min Scan# 2297
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03

Tgt Ion:284 Resp: 1502006
Ion Ratio Lower Upper
284 100
142 29.5 25.0 37.6
249 30.0 24.6 36.8





#70

Atrazine

Concen: 16.377 ng/ul

RT: 16.851 min Scan# 2323

Delta R.T. 0.000 min

Lab File: BP012901.D

Acq: 01 Dec 2022 04:03

Instrument :

BNA_P

ClientSampleId :





SLCS294

Tgt Ion:200 Resp: 609304
 Ion Ratio Lower Upper
 200 100
 173 25.4 20.3 30.5
 215 49.7 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 12/01/2022
 Supervised By :mohammad ahmed 12/02/2022



#71

Pentachlorophenol

Concen: 25.607 ng/ul




RT: 17.045 min Scan# 2356


Delta R.T. 0.000 min

Lab File: BP012901.D

Acq: 01 Dec 2022 04:03


Tgt Ion:266 Resp: 752662
 Ion Ratio Lower Upper
 266 100
 264 62.9 50.7 76.1
 268 64.4 51.8 77.8





#72
Phenanthrene
Concen: 29.638 ng/ul
RT: 17.445 min Scan# 2424
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion:178 Resp: 6027120
Ion Ratio Lower Upper
178 100
179 15.7 12.3 18.5
176 19.3 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#74
Anthracene
Concen: 29.623 ng/ul
RT: 17.534 min Scan# 2439
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03

Tgt Ion:178 Resp: 5959507
Ion Ratio Lower Upper
178 100
179 15.8 12.4 18.6
176 18.9 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 31.393 ng/uL

RT: 13.446 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP012901.D

Acq: 01 Dec 2022 04:03

Instrument :

BNA_P

ClientSampleId :

SLCS294

Tgt Ion:216 Resp: 1851809

Ion Ratio Lower Upper

216 100

214 78.2 62.3 93.5





218 48.5 39.1 58.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 12/01/2022

Supervised By :mohammad ahmed 12/02/2022



#76

Pentachlorobenzene

Concen: 29.137 ng/uL

RT: 14.963 min Scan# 2002

Delta R.T. 0.000 min

Lab File: BP012901.D

Acq: 01 Dec 2022 04:03




Tgt Ion:250 Resp: 1786657


Ion Ratio Lower Upper

250 100

248 63.3 50.6 76.0


252 64.0 50.5 75.7





#77
 Carbazole
 Concen: 29.258 ng/ul
 RT: 17.798 min Scan# 2484
 Delta R.T. -0.006 min
 Lab File: BP012901.D
 Acq: 01 Dec 2022 04:03




Instrument : BNA_P
 ClientSampleId : SLCS294




Tgt Ion:167 Resp: 4979538
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.1 25.7
 139 12.1 9.8 14.8

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
 Supervised By :mohammad ahmed 12/02/2022




#78
 Di-n-butylphthalate
 Concen: 30.185 ng/ul
 RT: 18.345 min Scan# 2577
 Delta R.T. -0.006 min
 Lab File: BP012901.D
 Acq: 01 Dec 2022 04:03




Tgt Ion:149 Resp: 6081573
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.2 10.8
 104 4.6 3.7 5.5





#80
Fluoranthene
Concen: 33.876 ng/ul
RT: 19.404 min Scan# 2757
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:202 Resp: 6075576
Ion Ratio Lower Upper
202 100
101 9.8 8.7 13.1
100 7.5 6.9 10.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#82
Pyrene
Concen: 32.766 ng/ul
RT: 19.757 min Scan# 2817
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:202 Resp: 6077787
Ion Ratio Lower Upper
202 100
101 11.8 10.3 15.5
100 9.5 8.6 13.0





#83
Butylbenzylphthalate
Concen: 44.885 ng/ul
RT: 20.622 min Scan# 2964
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:149 Resp: 1948309
Ion Ratio Lower Upper
149 100
91 66.4 58.2 87.2
206 23.2 16.5 24.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#84
3,3'-Dichlorobenzidine
Concen: 31.350 ng/ul
RT: 21.398 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Tgt Ion:252 Resp: 1428599
Ion Ratio Lower Upper
252 100
254 63.9 52.3 78.5
126 9.5 7.6 11.4





#85
Benzo(a)anthracene
Concen: 28.132 ng/ul
RT: 21.469 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
Client Sample Id :
SLCS294



Tgt Ion: 228 Resp: 4828301
Ion Ratio Lower Upper
228 100
229 19.8 16.2 24.2
226 27.4 22.2 33.2




Manual Integrations
APPROVED


Reviewed By : Jagrut Upadhyay 12/01/2022
Supervised By : mohammad ahmed 12/02/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 39.460 ng/ul
RT: 21.380 min Scan# 3093
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion: 149 Resp: 2619414
Ion Ratio Lower Upper
149 100
167 28.6 22.7 34.1
279 4.5 3.8 5.6





#87
Chrysene
Concen: 27.128 ng/ul
RT: 21.522 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion:228 Resp: 4501830
Ion Ratio Lower Upper
228 100
226 30.1 24.3 36.5
229 20.3 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#89
Di-n-octyl phthalate
Concen: 30.012 ng/ul
RT: 22.345 min Scan# 3257
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03


Tgt Ion:149 Resp: 4024445





#90
Benzo(b)fluoranthene
Concen: 26.524 ng/ul
RT: 23.227 min Scan# 3407
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294




Tgt Ion:252 Resp: 4612155
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 8.8 7.4 11.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022




#91
Benzo(k)fluoranthene
Concen: 26.069 ng/ul
RT: 23.275 min Scan# 3415
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03



Tgt Ion:252 Resp: 4524739
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 9.3 7.8 11.8





#93

Benzo(a)pyrene

Concen: 28.761 ng/ul

RT: 23.886 min Scan# 3519

Delta R.T. 0.000 min

Lab File: BP012901.D

Acq: 01 Dec 2022 04:03

Instrument :

BNA_P

ClientSampleId :

SLCS294

Tgt Ion:252 Resp: 4112406

Ion Ratio Lower Upper

252 100





253 22.2 17.6 26.4

125 10.2 8.5 12.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 39.278 ng/ul

RT: 26.627 min Scan# 3985

Delta R.T. 0.000 min

Lab File: BP012901.D

Acq: 01 Dec 2022 04:03




Tgt Ion:276 Resp: 5418330


Ion Ratio Lower Upper

276 100

138 22.2 18.3 27.5


277 25.0 19.8 29.8





#95
Dibenzo(a,h)anthracene
Concen: 38.105 ng/ul
RT: 26.651 min Scan# 3989
Delta R.T. 0.000 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03




Instrument :
BNA_P
ClientSampleId :
SLCS294



Tgt Ion: 278 Resp: 4769704
Ion Ratio Lower Upper
278 100
139 15.1 12.6 18.8
279 24.0 19.4 29.2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/01/2022
Supervised By :mohammad ahmed 12/02/2022



#96
Benzo(g,h,i)perylene
Concen: 39.789 ng/ul
RT: 27.433 min Scan# 4122
Delta R.T. -0.006 min
Lab File: BP012901.D
Acq: 01 Dec 2022 04:03

Tgt Ion: 276 Resp: 4721179
Ion Ratio Lower Upper
276 100
138 20.1 16.7 25.1
277 23.8 19.1 28.7

