

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120822\
 Data File : BP012986.D
 Acq On : 08 Dec 2022 15:57
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Dec 09 01:55:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 08 00:35:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	553501	20.000	ng/u1	0.00
20) Naphthalene-d8	10.793	136	2403680	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	1629184	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.375	188	3712783	20.000	ng/u1	0.00
79) Chrysene-d12	21.457	240	3509866	20.000	ng/u1	0.00
88) Perylene-d12	23.951	264	3537766	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	100250	7.744	ng/uL	0.00
4) Pyridine-d5	3.781	84	698881	19.005	ng/u1	0.00
7) Phenol-d5	7.128	99	856292	18.993	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	544024	20.004	ng/u1	0.00
11) 2-Chlorophenol-d4	7.505	132	702140	19.784	ng/u1	0.00
15) 4-Methylphenol-d8	8.687	113	716487	19.396	ng/u1	0.00
21) Nitrobenzene-d5	9.146	128	349714	19.802	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	397874	20.011	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.404	165	781855	20.246	ng/u1	0.00
31) 4-Chloroaniline-d4	10.928	131	1104810	20.265	ng/u1	0.00
46) Dimethylphthalate-d6	14.022	166	2507446	20.026	ng/u1	0.00
49) Acenaphthylene-d8	14.310	160	2769073	20.126	ng/u1	0.00
54) 4-Nitrophenol-d4	14.804	143	440146	19.542	ng/u1	0.00
60) Fluorene-d10	15.610	176	2152263	20.318	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.728	200	439436	18.392	ng/u1	0.00
73) Anthracene-d10	17.469	188	3388107	20.227	ng/u1	0.00
81) Pyrene-d10	19.704	212	4017783	19.943	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.786	264	3514544	19.931	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	105343	7.795	ng/uL	98
5) Pyridine	3.805	79	706652	19.335	ng/u1	97
6) Benzaldehyde	7.110	77	337175	19.100	ng/u1	97
8) Phenol	7.157	94	879686	19.298	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.393	93	738604	19.742	ng/u1	99
12) 2-Chlorophenol	7.534	128	718099	19.731	ng/u1	98
13) 2-Methylphenol	8.416	108	690453	19.429	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.510	45	1030251	20.067	ng/u1	98
16) Acetophenone	8.804	105	1174664	20.739	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.787	70	584072	20.506	ng/u1	99
18) 4-Methylphenol	8.746	108	773245	19.644	ng/u1	99
19) Hexachloroethane	9.063	117	286169	19.676	ng/u1	98
22) Nitrobenzene	9.187	77	832562	20.090	ng/u1	99
23) Isophorone	9.716	82	1660955	19.821	ng/u1	99
25) 2-Nitrophenol	9.899	139	432281	20.066	ng/u1	96
26) 2,4-Dimethylphenol	9.957	107	840257	19.990	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.199	93	1067863	19.930	ng/u1	99
29) 2,4-Dichlorophenol	10.434	162	764696	20.214	ng/u1	100
30) Naphthalene	10.840	128	2513565	19.995	ng/u1	99
32) 4-Chloroaniline	10.951	127	1105427	20.489	ng/u1	99
33) Hexachlorobutadiene	11.128	225	515058	19.830	ng/u1	99
34) Caprolactam	11.734	113	148219	11.491	ng/u1	93
35) 4-Chloro-3-methylphenol	12.063	107	798965	20.695	ng/u1	100
36) 2-Methylnaphthalene	12.445	142	1753660	20.334	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120822\
 Data File : BP012986.D
 Acq On : 08 Dec 2022 15:57
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Dec 09 01:55:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 08 00:35:07 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	1803456	20.334	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.810	216	1044312	19.907	ng/ul	98
40) Hexachlorocyclopentadiene	12.793	237	620770	18.184	ng/ul	99
41) 2,4,6-Trichlorophenol	13.045	196	672020	19.933	ng/ul	99
42) 2,4,5-Trichlorophenol	13.116	196	732717	20.233	ng/ul	99
43) 1,1'-Biphenyl	13.451	154	2442059	19.958	ng/ul	99
44) 2-Chloronaphthalene	13.492	162	1915696	19.893	ng/ul	100
45) 2-Nitroaniline	13.698	65	480798	20.592	ng/ul	99
47) Dimethylphthalate	14.069	163	2493618	20.094	ng/ul	99
48) 2,6-Dinitrotoluene	14.192	165	474361	20.568	ng/ul	96
50) Acenaphthylene	14.339	152	2993654	20.239	ng/ul	100
51) 3-Nitroaniline	14.522	138	394125	18.098	ng/ul	98
52) Acenaphthene	14.681	153	2041596	19.918	ng/ul	99
53) 2,4-Dinitrophenol	14.722	184	261192	16.568	ng/ul	98
55) 4-Nitrophenol	14.822	109	306871	19.968	ng/ul	99
56) Dibenzofuran	15.016	168	2942929	20.194	ng/ul	99
57) 2,4-Dinitrotoluene	14.975	165	693617	20.705	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.239	232	669780	20.266	ng/ul	97
59) Diethylphthalate	15.434	149	2421191	20.070	ng/ul	99
61) Fluorene	15.663	166	2390037	20.449	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.657	204	1264812	20.463	ng/ul	99
63) 4-Nitroaniline	15.681	138	360829	18.829	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.739	198	442893	18.969	ng/ul	99
67) N-Nitrosodiphenylamine	15.869	169	2064031	20.052	ng/ul	100
68) 4-Bromophenyl-phenylether	16.557	248	794493	19.705	ng/ul	98
69) Hexachlorobenzene	16.675	284	945416	19.747	ng/ul	98
70) Atrazine	16.828	200	797861	19.891	ng/ul	99
71) Pentachlorophenol	17.016	266	535051	18.454	ng/ul	99
72) Phenanthrene	17.416	178	3907822	20.207	ng/ul	100
74) Anthracene	17.504	178	3927256	20.339	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.416	216	1047287	19.771	ng/ul	99
76) Pentachlorobenzene	14.934	250	1069845	19.809	ng/ul	100
77) Carbazole	17.775	167	3467454	21.323	ng/ul	100
78) Di-n-butylphthalate	18.316	149	4062031	20.516	ng/ul	100
80) Fluoranthene	19.375	202	4661393	20.386	ng/ul	99
82) Pyrene	19.727	202	4834220	20.495	ng/ul	100
83) Butylbenzylphthalate	20.592	149	1802391	19.780	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.369	252	1493363	18.834	ng/ul	100
85) Benzo(a)anthracene	21.439	228	4710517	20.217	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.351	149	2601796	20.241	ng/ul	99
87) Chrysene	21.498	228	4474767	20.308	ng/ul	100
89) Di-n-octyl phthalate	22.304	149	4234237	21.364	ng/ul	100
90) Benzo(b)fluoranthene	23.186	252	4659795	20.656	ng/ul	100
91) Benzo(k)fluoranthene	23.233	252	4467862	19.972	ng/ul	100
93) Benzo(a)pyrene	23.839	252	3946254	20.090	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.556	276	4600252	18.801	ng/ul	99
95) Dibenzo(a,h)anthracene	26.574	278	3830572	18.909	ng/ul	99
96) Benzo(g,h,i)perylene	27.356	276	3551451	18.080	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120822\
Data File : BP012986.D
Acq On : 08 Dec 2022 15:57
Operator : CG/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Dec 09 01:55:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 08 00:35:07 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.795 ng/uL
RT: 3.393 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 88 Resp: 105343
Ion Ratio Lower Upper
88 100
43 27.5 22.4 33.6
58 71.3 58.2 87.4



#5
Pyridine
Concen: 19.335 ng/uL
RT: 3.805 min Scan# 105
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion: 79 Resp: 706652
Ion Ratio Lower Upper
79 100
52 72.9 60.7 91.1
51 34.5 28.9 43.3









#6
Benzaldehyde
Concen: 19.100 ng/ul
RT: 7.110 min Scan# 60
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 337175
Ion Ratio Lower Upper
77 100
105 103.0 81.5 122.3
106 100.5 76.9 115.3



#8
Phenol
Concen: 19.298 ng/ul
RT: 7.157 min Scan# 675
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57


Tgt Ion: 94 Resp: 879686
Ion Ratio Lower Upper
94 100
65 26.3 21.1 31.7
66 32.9 27.0 40.6








#10
Bis(2-Chloroethyl)ether
Concen: 19.742 ng/ul
RT: 7.393 min Scan# 716
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Instrument :
BNA_P
ClientSampleId :







Tgt Ion: 93 Resp: 738604
Ion Ratio Lower Upper
93 100
63 75.6 59.8 89.8
95 33.3 26.4 39.6



#12
2-Chlorophenol
Concen: 19.731 ng/ul
RT: 7.534 min Scan# 739
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:128 Resp: 718099
Ion Ratio Lower Upper
128 100
64 45.5 34.9 52.3
130 32.1 25.8 38.6









#13
2-Methylphenol
Concen: 19.429 ng/ul
RT: 8.416 min Scan# 890
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 690453
Ion Ratio Lower Upper
108 100
107 86.7 69.1 103.7



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.067 ng/ul
RT: 8.510 min Scan# 905
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion: 45 Resp: 1030251
Ion Ratio Lower Upper
45 100
77 14.3 11.0 16.6
79 10.8 9.3 13.9









#16
Acetophenone
Concen: 20.739 ng/ul
RT: 8.804 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 1174664
Ion Ratio Lower Upper
105 100
77 75.3 59.3 88.9
51 26.2 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 20.506 ng/ul
RT: 8.787 min Scan# 952
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion: 70 Resp: 584072
Ion Ratio Lower Upper
70 100
42 56.2 45.5 68.3
101 9.5 7.9 11.9
130 21.4 17.4 26.2









#18
4-Methylphenol
Concen: 19.644 ng/ul
RT: 8.746 min Scan# 946
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 773245
Ion Ratio Lower Upper
108 100
107 110.7 87.4 131.2



#19
Hexachloroethane
Concen: 19.676 ng/ul
RT: 9.063 min Scan# 999
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:117 Resp: 286169
Ion Ratio Lower Upper
117 100
201 116.4 93.7 140.5
199 70.8 58.6 88.0









#22
Nitrobenzene
Concen: 20.090 ng/ul
RT: 9.187 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 832562
Ion Ratio Lower Upper
77 100
123 47.5 38.3 57.5
65 13.0 10.3 15.5



#23
Isophorone
Concen: 19.821 ng/ul
RT: 9.716 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion: 82 Resp: 1660955
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.2
138 18.6 15.3 22.9









#25
2-Nitrophenol
Concen: 20.066 ng/ul
RT: 9.899 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 432281
Ion Ratio Lower Upper
139 100
65 48.8 36.6 54.8
109 28.0 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 19.990 ng/ul
RT: 9.957 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:107 Resp: 840257
Ion Ratio Lower Upper
107 100
121 53.1 41.8 62.6
122 90.2 72.6 108.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.930 ng/ul
 RT: 10.199 min Scan# 1192
 Delta R.T. 0.000 min
 Lab File: BP012986.D
 Acq: 08 Dec 2022 15:57

Instrument :
 BNA_P
 ClientSampleId :

Tgt Ion: 93 Resp: 1067863




Ion	Ratio	Lower	Upper
93	100		
95	31.9	26.0	39.0
123	11.6	9.2	13.8




#29
 2,4-Dichlorophenol
 Concen: 20.214 ng/ul
 RT: 10.434 min Scan# 1232
 Delta R.T. -0.006 min
 Lab File: BP012986.D
 Acq: 08 Dec 2022 15:57

Tgt Ion:162 Resp: 764696

Ion	Ratio	Lower	Upper
162	100		
164	65.8	52.5	78.7
98	33.7	26.6	39.8









#30
Naphthalene
Concen: 19.995 ng/ul
RT: 10.840 min Scan# 1302
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 2513565
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.2 10.4 15.6



#32
4-Chloroaniline
Concen: 20.489 ng/ul
RT: 10.951 min Scan# 1320
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:127 Resp: 1105427
Ion Ratio Lower Upper
127 100
129 32.5 25.6 38.4









#33
Hexachlorobutadiene
Concen: 19.830 ng/ul
RT: 11.128 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:225 Resp: 515058
Ion Ratio Lower Upper
225 100
223 61.6 49.2 73.8
227 64.0 50.2 75.2



#34
Caprolactam
Concen: 11.491 ng/ul
RT: 11.734 min Scan# 1453
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:113 Resp: 148219
Ion Ratio Lower Upper
113 100
55 163.0 141.9 212.9
56 122.8 99.6 149.4









#35
4-Chloro-3-methylphenol
Concen: 20.695 ng/ul
RT: 12.063 min Scan# 1510
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 798965
Ion Ratio Lower Upper
107 100
144 26.9 21.4 32.2
142 81.8 65.8 98.6



#36
2-Methylnaphthalene
Concen: 20.334 ng/ul
RT: 12.445 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:142 Resp: 1753660
Ion Ratio Lower Upper
142 100
141 87.9 70.8 106.2









#37
1-Methylnaphthalene
Concen: 20.334 ng/ul
RT: 12.663 min Scan# 1612
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 1803456
Ion Ratio Lower Upper
142 100
141 91.7 72.7 109.1
116 3.7 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.907 ng/ul
RT: 12.810 min Scan# 1636
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:216 Resp: 1044312
Ion Ratio Lower Upper
216 100
214 78.5 61.9 92.9
179 19.6 15.0 22.6
108 14.7 11.0 16.4





#40

Hexachlorocyclopentadiene

Concen: 18.184 ng/ul

RT: 12.793 min Scan# 1633

Delta R.T. 0.000 min

Lab File: BP012986.D

Acq: 08 Dec 2022 15:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:237 Resp: 620770

Ion Ratio Lower Upper

237 100

235 61.3 49.7 74.5

272 13.4 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 19.933 ng/ul

RT: 13.045 min Scan# 1676

Delta R.T. -0.006 min

Lab File: BP012986.D

Acq: 08 Dec 2022 15:57




Tgt Ion:196 Resp: 672020


Ion Ratio Lower Upper

196 100

198 97.0 77.0 115.6


200 31.1 25.3 37.9








#42
2,4,5-Trichlorophenol
Concen: 20.233 ng/ul
RT: 13.116 min Scan# 1688
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:196 Resp: 732717
Ion Ratio Lower Upper
196 100
198 95.5 77.1 115.7
200 31.0 25.4 38.2



#43
1,1'-Biphenyl
Concen: 19.958 ng/ul
RT: 13.451 min Scan# 1745
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:154 Resp: 2442059
Ion Ratio Lower Upper
154 100
153 40.1 31.5 47.3
76 11.7 9.6 14.4









#44
2-Chloronaphthalene
Concen: 19.893 ng/ul
RT: 13.492 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:162 Resp: 1915696
Ion Ratio Lower Upper
162 100
164 32.9 26.3 39.5
127 36.2 29.3 43.9



#45
2-Nitroaniline
Concen: 20.592 ng/ul
RT: 13.698 min Scan# 1787
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion: 65 Resp: 480798
Ion Ratio Lower Upper
65 100
92 69.8 56.4 84.6
138 114.8 90.3 135.5









#47
Dimethylphthalate
Concen: 20.094 ng/ul
RT: 14.069 min Scan# 1850
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 2493618
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.1 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 20.568 ng/ul
RT: 14.192 min Scan# 1871
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:165 Resp: 474361
Ion Ratio Lower Upper
165 100
89 50.4 43.0 64.4
121 19.6 16.1 24.1









#50
Acenaphthylene
Concen: 20.239 ng/ul
RT: 14.339 min Scan# 1896
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 2993654
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.3 10.6 16.0



#51
3-Nitroaniline
Concen: 18.098 ng/ul
RT: 14.522 min Scan# 1927
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:138 Resp: 394125
Ion Ratio Lower Upper
138 100
108 10.1 8.6 13.0
92 108.6 88.4 132.6









#52
Acenaphthene
Concen: 19.918 ng/ul
RT: 14.681 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 2041596
Ion Ratio Lower Upper
153 100
152 47.9 38.2 57.4
154 89.3 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 16.568 ng/ul
RT: 14.722 min Scan# 1961
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:184 Resp: 261192
Ion Ratio Lower Upper
184 100
63 58.2 48.4 72.6
154 58.1 45.5 68.3









#55
4-Nitrophenol
Concen: 19.968 ng/ul
RT: 14.822 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 306871
Ion Ratio Lower Upper
109 100
139 149.5 120.8 181.2
65 136.1 107.9 161.9



#56
Dibenzofuran
Concen: 20.194 ng/ul
RT: 15.016 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:168 Resp: 2942929
Ion Ratio Lower Upper
168 100
139 34.4 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 20.705 ng/ul
RT: 14.975 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 693617
Ion Ratio Lower Upper
165 100
63 39.2 31.8 47.6
182 2.7 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.266 ng/ul
RT: 15.239 min Scan# 2049
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:232 Resp: 669780
Ion Ratio Lower Upper
232 100
131 37.6 31.8 47.6
130 2.1 1.8 2.6
166 27.1 22.9 34.3









#59
Diethylphthalate
Concen: 20.070 ng/ul
RT: 15.434 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 2421191
Ion Ratio Lower Upper
149 100
177 23.6 18.4 27.6
150 12.5 9.9 14.9



#61
Fluorene
Concen: 20.449 ng/ul
RT: 15.663 min Scan# 2121
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:166 Resp: 2390037
Ion Ratio Lower Upper
166 100
165 97.4 77.4 116.2
167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.463 ng/ul
RT: 15.657 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 1264812
Ion Ratio Lower Upper
204 100
206 33.7 26.3 39.5
141 50.3 40.4 60.6



#63
4-Nitroaniline
Concen: 18.829 ng/ul
RT: 15.681 min Scan# 2124
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:138 Resp: 360829
Ion Ratio Lower Upper
138 100
92 50.7 39.0 58.4
108 73.0 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 18.969 ng/ul
RT: 15.739 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 442893
Ion Ratio Lower Upper
198 100
182 4.1 3.3 4.9
77 21.3 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 20.052 ng/ul
RT: 15.869 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:169 Resp: 2064031
Ion Ratio Lower Upper
169 100
168 66.5 53.0 79.6
167 36.1 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 19.705 ng/ul
RT: 16.557 min Scan# 2273
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 794493
Ion Ratio Lower Upper
248 100
250 97.9 77.4 116.0
141 58.1 48.3 72.5



#69
Hexachlorobenzene
Concen: 19.747 ng/ul
RT: 16.675 min Scan# 2293
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:284 Resp: 945416
Ion Ratio Lower Upper
284 100
142 27.3 22.6 34.0
249 29.4 24.6 37.0









#70
Atrazine
Concen: 19.891 ng/ul
RT: 16.828 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:200 Resp: 797861
Ion Ratio Lower Upper
200 100
173 25.2 20.5 30.7
215 49.0 39.9 59.9



#71
Pentachlorophenol
Concen: 18.454 ng/ul
RT: 17.016 min Scan# 2351
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:266 Resp: 535051
Ion Ratio Lower Upper
266 100
264 64.5 50.9 76.3
268 64.1 51.4 77.2









#72
Phenanthrene
Concen: 20.207 ng/ul
RT: 17.416 min Scan# 2419
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:178 Resp: 3907822
Ion Ratio Lower Upper
178 100
179 15.4 12.5 18.7
176 19.1 15.4 23.0



#74
Anthracene
Concen: 20.339 ng/ul
RT: 17.504 min Scan# 2434
Delta R.T. -0.006 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:178 Resp: 3927256
Ion Ratio Lower Upper
178 100
179 15.7 12.5 18.7
176 18.2 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.771 ng/uL
RT: 13.416 min Scan# 1739
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 1047287
Ion Ratio Lower Upper
216 100
214 77.3 62.2 93.2
218 47.7 38.9 58.3



#76
Pentachlorobenzene
Concen: 19.809 ng/uL
RT: 14.934 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57


Tgt Ion:250 Resp: 1069845
Ion Ratio Lower Upper
250 100
248 63.2 50.6 76.0
252 64.4 51.1 76.7








#77
 Carbazole
 Concen: 21.323 ng/ul
 RT: 17.775 min Scan# 2480
 Delta R.T. 0.000 min
 Lab File: BP012986.D
 Acq: 08 Dec 2022 15:57

Instrument :
 BNA_P
 ClientSampleId :







Tgt Ion:167 Resp: 3467454
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.2 25.8
 139 12.1 9.4 14.2



#78
 Di-n-butylphthalate
 Concen: 20.516 ng/ul
 RT: 18.316 min Scan# 2572
 Delta R.T. 0.000 min
 Lab File: BP012986.D
 Acq: 08 Dec 2022 15:57

Tgt Ion:149 Resp: 4062031
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.5 11.3
 104 4.7 3.8 5.6









#80
Fluoranthene
Concen: 20.386 ng/u1
RT: 19.375 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:202 Resp: 4661393
Ion Ratio Lower Upper
202 100
101 10.0 8.2 12.2
100 7.5 6.1 9.1



#82
Pyrene
Concen: 20.495 ng/u1
RT: 19.727 min Scan# 2812
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:202 Resp: 4834220
Ion Ratio Lower Upper
202 100
101 12.1 9.4 14.2
100 9.9 7.8 11.8





#83

Butylbenzylphthalate

Concen: 19.780 ng/ul

RT: 20.592 min Scan# 2959

Delta R.T. 0.000 min

Lab File: BP012986.D

Acq: 08 Dec 2022 15:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 1802391

Ion Ratio Lower Upper

149 100

91 67.8 53.4 80.0

206 22.7 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 18.834 ng/ul

RT: 21.369 min Scan# 3091

Delta R.T. 0.000 min

Lab File: BP012986.D

Acq: 08 Dec 2022 15:57




Tgt Ion:252 Resp: 1493363


Ion Ratio Lower Upper

252 100

254 65.0 51.8 77.8

126 9.5 7.4 11.2





#85

Benzo(a)anthracene

Concen: 20.217 ng/ul

RT: 21.439 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BP012986.D

Acq: 08 Dec 2022 15:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 4710517

Ion Ratio Lower Upper

228 100

229 19.9 16.1 24.1

226 27.1 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.241 ng/ul

RT: 21.351 min Scan# 3088

Delta R.T. 0.000 min

Lab File: BP012986.D

Acq: 08 Dec 2022 15:57




Tgt Ion:149 Resp: 2601796


Ion Ratio Lower Upper

149 100

167 28.7 22.5 33.7

279 4.6 3.8 5.8





#87

Chrysene

Concen: 20.308 ng/ul

RT: 21.498 min Scan# 3113

Delta R.T. 0.000 min

Lab File: BP012986.D

Acq: 08 Dec 2022 15:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 4474767

Ion Ratio Lower Upper

228 100

226 30.3 24.5 36.7

229 19.8 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 21.364 ng/ul




RT: 22.304 min Scan# 3250


Delta R.T. 0.000 min

Lab File: BP012986.D

Acq: 08 Dec 2022 15:57

Tgt Ion:149 Resp: 4234237









#90
Benzo(b)fluoranthene
Concen: 20.656 ng/ul
RT: 23.186 min Scan# 3400
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 4659795
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 8.5 6.8 10.2



#91
Benzo(k)fluoranthene
Concen: 19.972 ng/ul
RT: 23.233 min Scan# 3408
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:252 Resp: 4467862
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 8.6 6.8 10.2









#93
Benzo(a)pyrene
Concen: 20.090 ng/ul
RT: 23.839 min Scan# 3511
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 3946254
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 9.5 7.5 11.3



#94
Indeno(1,2,3-cd)pyrene
Concen: 18.801 ng/ul
RT: 26.556 min Scan# 3973
Delta R.T. 0.000 min
Lab File: BP012986.D
Acq: 08 Dec 2022 15:57

Tgt Ion:276 Resp: 4600252
Ion Ratio Lower Upper
276 100
138 20.8 17.1 25.7
277 25.2 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 18.909 ng/ul

RT: 26.574 min Scan# 3975

Delta R.T. 0.006 min

Lab File: BP012986.D

Acq: 08 Dec 2022 15:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 3830572

Ion Ratio Lower Upper

278 100

139 14.5 11.1 16.7

279 24.1 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 18.080 ng/ul

RT: 27.356 min Scan# 4109

Delta R.T. 0.006 min

Lab File: BP012986.D

Acq: 08 Dec 2022 15:57

Tgt Ion:276 Resp: 3551451

Ion Ratio Lower Upper

276 100

138 17.9 15.4 23.2

277 24.2 19.1 28.7

