

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120822\
 Data File : BP013028.D
 Acq On : 09 Dec 2022 20:49
 Operator : CG/JU
 Sample : PB149458BS
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS458

Quant Time: Dec 09 21:59:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 08 00:35:07 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	580299	20.000	ng/u1	0.00
20) Naphthalene-d8	10.787	136	2489730	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	1633274	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.369	188	3703574	20.000	ng/u1	0.00
79) Chrysene-d12	21.457	240	3282012	20.000	ng/u1	0.00
88) Perylene-d12	23.945	264	3035681	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	73662	5.427	ng/uL	0.00
4) Pyridine-d5	3.781	84	995871	25.830	ng/u1	0.00
7) Phenol-d5	7.128	99	1435773	30.376	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	836138	29.325	ng/u1	0.00
11) 2-Chlorophenol-d4	7.505	132	1145481	30.785	ng/u1	0.00
15) 4-Methylphenol-d8	8.687	113	1161549	29.992	ng/u1	0.00
21) Nitrobenzene-d5	9.146	128	574945	31.430	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	675036	32.778	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.404	165	1256688	31.416	ng/u1	0.00
31) 4-Chloroaniline-d4	10.928	131	1520841	26.931	ng/u1	0.00
46) Dimethylphthalate-d6	14.022	166	3703028	29.500	ng/u1	0.00
49) Acenaphthylene-d8	14.310	160	4195102	30.415	ng/u1	0.00
54) 4-Nitrophenol-d4	14.810	143	712624	31.561	ng/u1	0.00
60) Fluorene-d10	15.610	176	3188088	30.021	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.728	200	706367	29.638	ng/u1	0.00
73) Anthracene-d10	17.469	188	5065505	30.317	ng/u1	0.00
81) Pyrene-d10	19.698	212	5908680	31.365	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.786	264	4705169	31.096	ng/u1	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.387	88	150787	10.643	ng/uL	97
5) Pyridine	3.799	79	1009988	26.358	ng/u1	99
6) Benzaldehyde	7.110	77	704940	38.088	ng/u1	99
8) Phenol	7.157	94	1478253	30.932	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.393	93	1144338	29.174	ng/u1	100
12) 2-Chlorophenol	7.534	128	1172517	30.729	ng/u1	99
13) 2-Methylphenol	8.416	108	1129701	30.321	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.504	45	1629779	30.279	ng/u1	99
16) Acetophenone	8.804	105	1746657	29.413	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.793	70	883681	29.592	ng/u1	99
18) 4-Methylphenol	8.752	108	1234888	29.923	ng/u1	100
19) Hexachloroethane	9.057	117	449650	29.489	ng/u1	93
22) Nitrobenzene	9.187	77	1284188	29.917	ng/u1	98
23) Isophorone	9.716	82	2579798	29.722	ng/u1	100
25) 2-Nitrophenol	9.899	139	695217	31.156	ng/u1	98
26) 2,4-Dimethylphenol	9.957	107	1281710	29.439	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.193	93	1569255	28.276	ng/u1	99
29) 2,4-Dichlorophenol	10.434	162	1217770	31.078	ng/u1	99
30) Naphthalene	10.840	128	3799295	29.178	ng/u1	100
32) 4-Chloroaniline	10.951	127	1287896	23.047	ng/u1	99
33) Hexachlorobutadiene	11.122	225	805743	29.949	ng/u1	99
34) Caprolactam	11.734	113	399062m	29.869	ng/u1	
35) 4-Chloro-3-methylphenol	12.069	107	1255447	31.395	ng/u1	97
36) 2-Methylnaphthalene	12.445	142	2657912	29.754	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120822\
 Data File : BP013028.D
 Acq On : 09 Dec 2022 20:49
 Operator : CG/JU
 Sample : PB149458BS
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS458


Quant Time: Dec 09 21:59:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 08 00:35:07 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	2656728	28.920	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.810	216	1589699	30.227	ng/ul	100
40) Hexachlorocyclopentadiene	12.787	237	713161	20.838	ng/ul	100
41) 2,4,6-Trichlorophenol	13.045	196	1042942	30.857	ng/ul	99
42) 2,4,5-Trichlorophenol	13.116	196	1150416	31.687	ng/ul	98
43) 1,1'-Biphenyl	13.445	154	3527601	28.758	ng/ul	99
44) 2-Chloronaphthalene	13.492	162	2843592	29.454	ng/ul	99
45) 2-Nitroaniline	13.698	65	801043	34.222	ng/ul	97
47) Dimethylphthalate	14.069	163	3655475	29.382	ng/ul	99
48) 2,6-Dinitrotoluene	14.187	165	760831	32.907	ng/ul	99
50) Acenaphthylene	14.334	152	4366557	29.447	ng/ul	100
51) 3-Nitroaniline	14.522	138	789250	36.151	ng/ul	97
52) Acenaphthene	14.681	153	3026735	29.455	ng/ul	100
53) 2,4-Dinitrophenol	14.722	184	462014	29.233	ng/ul	98
55) 4-Nitrophenol	14.822	109	487868	31.666	ng/ul	99
56) Dibenzofuran	15.010	168	4292546	29.381	ng/ul	100
57) 2,4-Dinitrotoluene	14.975	165	1105009	32.902	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.239	232	1008286	30.433	ng/ul	96
59) Diethylphthalate	15.434	149	3646165	30.148	ng/ul	99
61) Fluorene	15.663	166	3486371	29.754	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.651	204	1845889	29.790	ng/ul	98
63) 4-Nitroaniline	15.686	138	843711	43.918	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.739	198	689586	29.608	ng/ul	98
67) N-Nitrosodiphenylamine	15.869	169	3076552	29.963	ng/ul	100
68) 4-Bromophenyl-phenylether	16.551	248	1194342	29.695	ng/ul	99
69) Hexachlorobenzene	16.669	284	1422882	29.794	ng/ul	98
70) Atrazine	16.822	200	720965	18.018	ng/ul	100
71) Pentachlorophenol	17.016	266	846032	29.253	ng/ul	99
72) Phenanthrene	17.410	178	5789473	30.012	ng/ul	100
74) Anthracene	17.504	178	5841284	30.326	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.416	216	1590600	30.102	ng/ul	100
76) Pentachlorobenzene	14.934	250	1571010	29.161	ng/ul	99
77) Carbazole	17.775	167	5327704	32.844	ng/ul	100
78) Di-n-butylphthalate	18.316	149	6463738	32.727	ng/ul	99
80) Fluoranthene	19.374	202	6916799	32.349	ng/ul	99
82) Pyrene	19.727	202	7091000	32.150	ng/ul	100
83) Butylbenzylphthalate	20.592	149	2926077	34.340	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.369	252	2270671	30.625	ng/ul	100
85) Benzo(a)anthracene	21.439	228	6765458	31.052	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.345	149	4273349	35.553	ng/ul	99
87) Chrysene	21.498	228	6319335	30.670	ng/ul	100
89) Di-n-octyl phthalate	22.304	149	7261507	42.698	ng/ul	100
90) Benzo(b)fluoranthene	23.180	252	6254332	32.309	ng/ul	99
91) Benzo(k)fluoranthene	23.233	252	5880988	30.638	ng/ul	99
93) Benzo(a)pyrene	23.839	252	5162316	30.627	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.556	276	5611876	26.729	ng/ul	99
95) Dibenzo(a,h)anthracene	26.568	278	4999867	28.763	ng/ul	99
96) Benzo(g,h,i)perylene	27.356	276	4514294	26.783	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.643 ng/uL
RT: 3.387 min Scan# 34
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion: 88 Resp: 15078
Ion Ratio Lower Upper
88 100
43 30.2 22.4 33.6
58 75.1 58.2 87.4


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#5
Pyridine
Concen: 26.358 ng/uL
RT: 3.799 min Scan# 104
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion: 79 Resp: 1009988
Ion Ratio Lower Upper
79 100
52 76.9 60.7 91.1
51 35.6 28.9 43.3





#6
Benzaldehyde
Concen: 38.088 ng/ul
RT: 7.110 min Scan# 60
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion: 77 Resp: 704940
Ion Ratio Lower Upper
77 100
105 101.8 81.5 122.3
106 97.6 76.9 115.3


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#8
Phenol
Concen: 30.932 ng/ul
RT: 7.157 min Scan# 675
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion: 94 Resp: 1478253
Ion Ratio Lower Upper
94 100
65 26.5 21.1 31.7
66 33.6 27.0 40.6





#10
Bis(2-Chloroethyl)ether
Concen: 29.174 ng/ul
RT: 7.393 min Scan# 715
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion: 93 Resp: 1144338
Ion Ratio Lower Upper
93 100
63 74.9 59.8 89.8
95 32.4 26.4 39.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#12
2-Chlorophenol
Concen: 30.729 ng/ul
RT: 7.534 min Scan# 739
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49

Tgt Ion:128 Resp: 1172517
Ion Ratio Lower Upper
128 100
64 44.6 34.9 52.3
130 32.3 25.8 38.6





#13

2-Methylphenol

Concen: 30.321 ng/ul

RT: 8.416 min Scan# 889

Delta R.T. -0.006 min

Lab File: BP013028.D

Acq: 09 Dec 2022 20:49

Instrument :

BNA_P

ClientSampleId :

SLCS458

Tgt Ion:108 Resp: 112970

Ion Ratio Lower Upper



108 100

107 88.0 69.1 103.7

Manual Integrations



APPROVED

Reviewed By :Jagrut
Upadhyay



12/10/2022
Supervised By :mohammad
ahmed

12/10/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 30.279 ng/ul

RT: 8.504 min Scan# 904

Delta R.T. -0.012 min

Lab File: BP013028.D

Acq: 09 Dec 2022 20:49




Tgt Ion: 45 Resp: 1629779


Ion Ratio Lower Upper

45 100

77 14.1 11.0 16.6


79 10.8 9.3 13.9





#16
Acetophenone
Concen: 29.413 ng/ul
RT: 8.804 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion: 105 Resp: 174665
Ion Ratio Lower Upper
105 100
77 73.9 59.3 88.9
51 25.8 19.6 29.4


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#17
N-Nitroso-di-n-propylamine
Concen: 29.592 ng/ul
RT: 8.793 min Scan# 953
Delta R.T. 0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Tgt Ion: 70 Resp: 883681
Ion Ratio Lower Upper
70 100
42 57.4 45.5 68.3
101 10.0 7.9 11.9
130 23.2 17.4 26.2







#18
4-Methylphenol
Concen: 29.923 ng/ul
RT: 8.752 min Scan# 945
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49

Instrument :
BNA_P
ClientSampleId :
SLCS458

Tgt Ion:108 Resp: 123488
Ion Ratio Lower Upper
108 100
107 108.9 87.4 131.2



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#19
Hexachloroethane
Concen: 29.489 ng/ul
RT: 9.057 min Scan# 998
Delta R.T. -0.012 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:117 Resp: 449650
Ion Ratio Lower Upper
117 100
201 109.4 93.7 140.5
199 67.5 58.6 88.0





#22
Nitrobenzene
Concen: 29.917 ng/u1
RT: 9.187 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion: 77 Resp: 1284188
Ion Ratio Lower Upper
77 100
123 49.1 38.3 57.5
65 13.5 10.3 15.5


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#23
Isophorone
Concen: 29.722 ng/u1
RT: 9.716 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion: 82 Resp: 2579798
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 18.9 15.3 22.9





#25
2-Nitrophenol
Concen: 31.156 ng/ul
RT: 9.899 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:139 Resp: 69521
Ion Ratio Lower Upper
139 100
65 46.7 36.6 54.8
109 26.3 21.6 32.4


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#26
2,4-Dimethylphenol
Concen: 29.439 ng/ul
RT: 9.957 min Scan# 1151
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion:107 Resp: 1281710
Ion Ratio Lower Upper
107 100
121 52.7 41.8 62.6
122 90.5 72.6 108.8





#27
Bis(2-Chloroethoxy)methane
Concen: 28.276 ng/ul
RT: 10.193 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion: 93 Resp: 156925
Ion Ratio Lower Upper
93 100
95 32.0 26.0 39.0
123 11.4 9.2 13.8


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#29
2,4-Dichlorophenol
Concen: 31.078 ng/ul
RT: 10.434 min Scan# 1232
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:162 Resp: 1217770
Ion Ratio Lower Upper
162 100
164 66.3 52.5 78.7
98 33.4 26.6 39.8





#30
Naphthalene
Concen: 29.178 ng/ul
RT: 10.840 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:128 Resp: 379929
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.0 10.4 15.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#32
4-Chloroaniline
Concen: 23.047 ng/ul
RT: 10.951 min Scan# 1320
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:127 Resp: 1287896
Ion Ratio Lower Upper
127 100
129 32.4 25.6 38.4





#33
Hexachlorobutadiene
Concen: 29.949 ng/ul
RT: 11.122 min Scan# 1349
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:225 Resp: 80574
Ion Ratio Lower Upper
225 100
223 62.2 49.2 73.8
227 63.9 50.2 75.2


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#34
Caprolactam
Concen: 29.869 ng/ul m
RT: 11.734 min Scan# 1453
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:113 Resp: 399062
Ion Ratio Lower Upper
113 100
55 166.9 141.9 212.9
56 122.0 99.6 149.4





#35
4-Chloro-3-methylphenol
Concen: 31.395 ng/ul
RT: 12.069 min Scan# 1509
Delta R.T. 0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:107 Resp: 125544
Ion Ratio Lower Upper
107 100
144 27.6 21.4 32.2
142 85.4 65.8 98.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#36
2-Methylnaphthalene
Concen: 29.754 ng/ul
RT: 12.445 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:142 Resp: 2657912
Ion Ratio Lower Upper
142 100
141 89.3 70.8 106.2





#37
1-Methylnaphthalene
Concen: 28.920 ng/ul
RT: 12.663 min Scan# 1611
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:142 Resp: 2656728
Ion Ratio Lower Upper
142 100
141 91.5 72.7 109.1
116 3.5 2.7 4.1


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 30.227 ng/ul
RT: 12.810 min Scan# 1636
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion:216 Resp: 1589699
Ion Ratio Lower Upper
216 100
214 77.5 61.9 92.9
179 19.1 15.0 22.6
108 13.7 11.0 16.4





#40
Hexachlorocyclopentadiene
Concen: 20.838 ng/ul
RT: 12.787 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:237 Resp: 71316
Ion Ratio Lower Upper
237 100
235 62.3 49.7 74.5
272 13.1 10.8 16.2



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#41
2,4,6-Trichlorophenol
Concen: 30.857 ng/ul
RT: 13.045 min Scan# 1676
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:196 Resp: 1042942
Ion Ratio Lower Upper
196 100
198 96.9 77.0 115.6
200 31.2 25.3 37.9





#42
2,4,5-Trichlorophenol
Concen: 31.687 ng/ul
RT: 13.116 min Scan# 1688
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:196 Resp: 1150410
Ion Ratio Lower Upper
196 100
198 94.8 77.1 115.7
200 30.8 25.4 38.2


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#43
1,1'-Biphenyl
Concen: 28.758 ng/ul
RT: 13.445 min Scan# 1744
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion:154 Resp: 3527601
Ion Ratio Lower Upper
154 100
153 40.1 31.5 47.3
76 12.3 9.6 14.4





#44
2-Chloronaphthalene
Concen: 29.454 ng/ul
RT: 13.492 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:162 Resp: 2843592
Ion Ratio Lower Upper
162 100
164 33.6 26.3 39.5
127 36.2 29.3 43.9


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#45
2-Nitroaniline
Concen: 34.222 ng/ul
RT: 13.698 min Scan# 1787
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion: 65 Resp: 801043
Ion Ratio Lower Upper
65 100
92 71.5 56.4 84.6
138 117.2 90.3 135.5





#47
Dimethylphthalate
Concen: 29.382 ng/ul
RT: 14.069 min Scan# 1850
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:163 Resp: 365547
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.0
164 10.2 8.3 12.5


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#48
2,6-Dinitrotoluene
Concen: 32.907 ng/ul
RT: 14.187 min Scan# 1870
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion:165 Resp: 760831
Ion Ratio Lower Upper
165 100
89 52.8 43.0 64.4
121 19.4 16.1 24.1





#50
Acenaphthylene
Concen: 29.447 ng/ul
RT: 14.334 min Scan# 1895
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:152 Resp: 436655
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.6
153 13.4 10.6 16.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#51
3-Nitroaniline
Concen: 36.151 ng/ul
RT: 14.522 min Scan# 1927
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:138 Resp: 789250
Ion Ratio Lower Upper
138 100
108 10.1 8.6 13.0
92 107.2 88.4 132.6





#52
Acenaphthene
Concen: 29.455 ng/ul
RT: 14.681 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:153 Resp: 302673
Ion Ratio Lower Upper
153 100
152 47.6 38.2 57.4
154 88.9 70.8 106.2



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#53
2,4-Dinitrophenol
Concen: 29.233 ng/ul
RT: 14.722 min Scan# 1961
Delta R.T. 0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:184 Resp: 462014
Ion Ratio Lower Upper
184 100
63 58.2 48.4 72.6
154 57.3 45.5 68.3





#55
4-Nitrophenol
Concen: 31.666 ng/ul
RT: 14.822 min Scan# 1977
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:109 Resp: 48786
Ion Ratio Lower Upper
109 100
139 151.4 120.8 181.2
65 137.7 107.9 161.9


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#56
Dibenzofuran
Concen: 29.381 ng/ul
RT: 15.010 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:168 Resp: 4292546
Ion Ratio Lower Upper
168 100
139 34.7 27.9 41.9





#57
2,4-Dinitrotoluene
Concen: 32.902 ng/ul
RT: 14.975 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:165 Resp: 1105009
Ion Ratio Lower Upper
165 100
63 38.4 31.8 47.6
182 2.8 2.1 3.1


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 30.433 ng/ul
RT: 15.239 min Scan# 2049
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion:232 Resp: 1008286
Ion Ratio Lower Upper
232 100
131 36.2 31.8 47.6
130 1.8 1.8 2.6
166 27.0 22.9 34.3





#59
Diethylphthalate
Concen: 30.148 ng/ul
RT: 15.434 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:149 Resp: 364616
Ion Ratio Lower Upper
149 100
177 23.8 18.4 27.6
150 12.4 9.9 14.9


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#61
Fluorene
Concen: 29.754 ng/ul
RT: 15.663 min Scan# 2121
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:166 Resp: 3486371
Ion Ratio Lower Upper
166 100
165 97.2 77.4 116.2
167 13.8 10.6 16.0





#62
4-Chlorophenyl-phenylether
Concen: 29.790 ng/ul
RT: 15.651 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:204 Resp: 1845889
Ion Ratio Lower Upper
204 100
206 33.4 26.3 39.5
141 52.2 40.4 60.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#63
4-Nitroaniline
Concen: 43.918 ng/ul
RT: 15.686 min Scan# 2125
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion:138 Resp: 843711
Ion Ratio Lower Upper
138 100
92 48.8 39.0 58.4
108 71.4 55.1 82.7





#66
4,6-Dinitro-2-methylphenol
Concen: 29.608 ng/ul
RT: 15.739 min Scan# 2133
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:198 Resp: 689580
Ion Ratio Lower Upper
198 100
182 3.6 3.3 4.9
77 21.7 16.6 25.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#67
N-Nitrosodiphenylamine
Concen: 29.963 ng/ul
RT: 15.869 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:169 Resp: 3076552
Ion Ratio Lower Upper
169 100
168 66.0 53.0 79.6
167 36.3 29.1 43.7





#68
4-Bromophenyl-phenylether
Concen: 29.695 ng/ul
RT: 16.551 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:248 Resp: 119434
Ion Ratio Lower Upper
248 100
250 97.9 77.4 116.0
141 61.4 48.3 72.5


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#69
Hexachlorobenzene
Concen: 29.794 ng/ul
RT: 16.669 min Scan# 2292
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:284 Resp: 1422882
Ion Ratio Lower Upper
284 100
142 30.1 22.6 34.0
249 30.6 24.6 37.0





#70
Atrazine
Concen: 18.018 ng/ul
RT: 16.822 min Scan# 2318
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:200 Resp: 72096
Ion Ratio Lower Upper
200 100
173 25.6 20.5 30.7
215 50.4 39.9 59.9


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#71
Pentachlorophenol
Concen: 29.253 ng/ul
RT: 17.016 min Scan# 2351
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:266 Resp: 846032
Ion Ratio Lower Upper
266 100
264 63.1 50.9 76.3
268 63.4 51.4 77.2





#72
Phenanthrene
Concen: 30.012 ng/ul
RT: 17.410 min Scan# 2419
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:178 Resp: 578947
Ion Ratio Lower Upper
178 100
179 15.7 12.5 18.7
176 19.3 15.4 23.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#74
Anthracene
Concen: 30.326 ng/ul
RT: 17.504 min Scan# 2434
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:178 Resp: 5841284
Ion Ratio Lower Upper
178 100
179 15.7 12.5 18.7
176 18.7 14.8 22.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 30.102 ng/uL
RT: 13.416 min Scan# 1739
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:216 Resp: 1590600
Ion Ratio Lower Upper
216 100
214 77.9 62.2 93.2
218 48.2 38.9 58.3


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#76
Pentachlorobenzene
Concen: 29.161 ng/uL
RT: 14.934 min Scan# 1997
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49




Tgt Ion:250 Resp: 1571010
Ion Ratio Lower Upper
250 100
248 62.7 50.6 76.0
252 64.9 51.1 76.7





#77
 Carbazole
 Concen: 32.844 ng/ul
 RT: 17.775 min Scan# 2480
 Delta R.T. -0.000 min
 Lab File: BP013028.D
 Acq: 09 Dec 2022 20:49


Instrument :
 BNA_P
 ClientSampleId :
 SLCS458



Tgt Ion:167 Resp: 5327704
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.2 25.8
 139 12.0 9.4 14.2


Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay






12/10/2022
 Supervised By :mohammad
 ahmed


12/10/2022



#78
 Di-n-butylphthalate
 Concen: 32.727 ng/ul
 RT: 18.316 min Scan# 2572
 Delta R.T. -0.000 min
 Lab File: BP013028.D
 Acq: 09 Dec 2022 20:49


Tgt Ion:149 Resp: 6463738
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.5 11.3
 104 4.5 3.8 5.6





#80
Fluoranthene
Concen: 32.349 ng/uI
RT: 19.374 min Scan# 2752
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:202 Resp: 6916799
Ion Ratio Lower Upper
202 100
101 9.9 8.2 12.2
100 7.4 6.1 9.1


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#82
Pyrene
Concen: 32.150 ng/uI
RT: 19.727 min Scan# 2812
Delta R.T. 0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Tgt Ion:202 Resp: 7091000
Ion Ratio Lower Upper
202 100
101 11.8 9.4 14.2
100 9.8 7.8 11.8





#83

Butylbenzylphthalate

Concen: 34.340 ng/ul

RT: 20.592 min Scan# 2959

Delta R.T. -0.000 min

Lab File: BP013028.D

Acq: 09 Dec 2022 20:49

Instrument :

BNA_P

ClientSampleId :



SLCS458

Tgt Ion:149 Resp: 292607

Ion	Ratio	Lower	Upper
149	100		
91	65.2	53.4	80.0
206	24.3	17.9	26.9



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay



12/10/2022
Supervised By :mohammad
ahmed

12/10/2022



#84

3,3'-Dichlorobenzidine

Concen: 30.625 ng/ul

RT: 21.369 min Scan# 3091




Delta R.T. -0.000 min


Lab File: BP013028.D

Acq: 09 Dec 2022 20:49

Tgt Ion:252 Resp: 2270671


Ion	Ratio	Lower	Upper
252	100		
254	64.5	51.8	77.8
126	9.1	7.4	11.2





#85
Benzo(a)anthracene
Concen: 31.052 ng/ul
RT: 21.439 min Scan# 3103
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:228 Resp: 6765458
Ion Ratio Lower Upper
228 100
229 19.9 16.1 24.1
226 27.7 22.0 33.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 35.553 ng/ul
RT: 21.345 min Scan# 3087
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:149 Resp: 4273349
Ion Ratio Lower Upper
149 100
167 28.7 22.5 33.7
279 4.7 3.8 5.8





#87
Chrysene
Concen: 30.670 ng/ul
RT: 21.498 min Scan# 3113
Delta R.T. -0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:228 Resp: 6319335
Ion Ratio Lower Upper
228 100
226 30.8 24.5 36.7
229 19.9 16.0 24.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






12/10/2022
Supervised By :mohammad
ahmed


12/10/2022



#89
Di-n-octyl phthalate
Concen: 42.698 ng/ul
RT: 22.304 min Scan# 3250
Delta R.T. 0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49


Tgt Ion:149 Resp: 7261507





#90
Benzo(b)fluoranthene
Concen: 32.309 ng/u1
RT: 23.180 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:252 Resp: 625433
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 9.1 6.8 10.2


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed



12/10/2022




#91
Benzo(k)fluoranthene
Concen: 30.638 ng/u1
RT: 23.233 min Scan# 3408
Delta R.T. 0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Tgt Ion:252 Resp: 5880988
Ion Ratio Lower Upper
252 100
253 22.4 17.7 26.5
125 8.9 6.8 10.2





#93

Benzo(a)pyrene

Concen: 30.627 ng/ul

RT: 23.839 min Scan# 3

Delta R.T. 0.000 min

Lab File: BP013028.D

Acq: 09 Dec 2022 20:49

Instrument :

BNA_P

ClientSampleId :

SLCS458

Tgt Ion:252 Resp: 5162310

Ion Ratio Lower Upper

252 100

253 21.7 17.5 26.3



125 9.3 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :Jagrut



Upadhyay



12/10/2022

Supervised By :mohammad
ahmed

12/10/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 26.729 ng/ul

RT: 26.556 min Scan# 3973

Delta R.T. 0.000 min

Lab File: BP013028.D

Acq: 09 Dec 2022 20:49




Tgt Ion:276 Resp: 5611876


Ion Ratio Lower Upper

276 100

138 22.2 17.1 25.7


277 25.4 20.0 30.0





#95
Dibenzo(a,h)anthracene
Concen: 28.763 ng/ul
RT: 26.568 min Scan# 3975
Delta R.T. 0.000 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Instrument :
BNA_P
ClientSampleId :
SLCS458



Tgt Ion:278 Resp: 499986
Ion Ratio Lower Upper
278 100
139 14.3 11.1 16.7
279 24.1 19.0 28.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




12/10/2022
Supervised By :mohammad
ahmed

12/10/2022



#96
Benzo(g,h,i)perylene
Concen: 26.783 ng/ul
RT: 27.356 min Scan# 4109
Delta R.T. 0.006 min
Lab File: BP013028.D
Acq: 09 Dec 2022 20:49



Tgt Ion:276 Resp: 4514294
Ion Ratio Lower Upper
276 100
138 18.5 15.4 23.2
277 24.1 19.1 28.7

