

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP122019\
 Data File : BP001341.D
 Acq On : 20 Dec 2019 13:30
 Operator : JU
 Sample : K6300-02MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSI-MW-4MS

Manual Integrations
 APPROVED

mohammad
 12/23/2019 11:56:06 AM

Quant Time: Dec 20 23:28:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP121919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 20 03:20:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.30	152	55237	20.00	ng	0.00
21) Naphthalene-d8	10.34	136	189969	20.00	ng	0.00
39) Acenaphthene-d10	13.20	164	107032	20.00	ng	0.00
64) Phenanthrene-d10	15.60	188	208072	20.00	ng	0.00
76) Chrysene-d12	19.25	240	161843	20.00	ng	0.00
87) Perylene-d12	21.02	264	166738	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	6.21	112	228253	66.79	ng	0.01
7) Phenol-d6	7.75	99	192284	43.12	ng	0.00
23) Nitrobenzene-d5	9.19	82	446077	90.17	ng	0.00
42) 2,4,6-Tribromophenol	14.50	330	184242	137.68	ng	0.01
45) 2-Fluorobiphenyl	12.12	172	772444	91.26	ng	0.00
79) Terphenyl-d14	17.89	244	953819	100.94	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.59	88	25458	18.100	ng	99
3) Pyridine	3.47	79	43034	11.200	ng	95
4) n-Nitrosodimethylamine	3.34	42	49201	20.223	ng	94
6) Aniline	7.77	93	70218	12.508	ng	96
8) 2-Chlorophenol	7.96	128	132877	38.672	ng	96
9) Benzaldehyde	7.60	77	149598	47.819	ng	97
10) Phenol	7.77	94	77650	15.239	ng	93
11) bis(2-Chloroethyl)ether	7.90	93	161274	44.737	ng	99
12) 1,3-Dichlorobenzene	8.21	146	168756	39.431	ng	98
13) 1,4-Dichlorobenzene	8.33	146	173694	39.835	ng	98
14) 1,2-Dichlorobenzene	8.57	146	167138	40.236	ng	99
15) Benzyl Alcohol	8.53	79	122674	29.316	ng	99
16) 2,2'-oxybis(1-Chloropropan	8.77	45	235792	41.300	ng	99
17) 2-Methylphenol	8.73	107	97343	32.060	ng	94
18) Hexachloroethane	9.10	117	70507	38.011	ng	97
19) n-Nitroso-di-n-propylamine	8.97	70	143927	45.461	ng	97
20) 3+4-Methylphenols	8.97	107	114504	28.419	ng	96
22) Acetophenone	8.95	105	261404	43.514	ng	# 97
24) Nitrobenzene	9.22	77	218313	44.838	ng	99
25) Isophorone	9.61	82	369936	46.507	ng	99
26) 2-Nitrophenol	9.73	139	80690	46.475	ng	96
27) 2,4-Dimethylphenol	9.82	122	121007	47.388	ng	98
28) bis(2-Chloroethoxy)methane	9.98	93	200683	42.242	ng	99
29) 2,4-Dichlorophenol	10.12	162	134038	42.392	ng	99
30) 1,2,4-Trichlorobenzene	10.25	180	156098	40.136	ng	98
31) Naphthalene	10.37	128	450425	43.436	ng	100
32) Benzoic acid	9.95	122	20642	13.975	ng	95
33) 4-Chloroaniline	10.46	127	35317	8.328	ng	98
34) Hexachlorobutadiene	10.59	225	102256	37.696	ng	99
35) Caprolactam	11.05	113	6512	6.921	ng	# 82
36) 4-Chloro-3-methylphenol	11.28	107	146202	39.568	ng	96
37) 2-Methylnaphthalene	11.50	142	320251	44.734	ng	99
38) 1-Methylnaphthalene	11.66	142	297432	44.192	ng	96
40) 1,2,4,5-Tetrachlorobenzene	11.78	216	171851	43.424	ng	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP122019\
 Data File : BP001341.D
 Acq On : 20 Dec 2019 13:30
 Operator : JU
 Sample : K6300-02MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSI-MW-4MS


Manual Integrations
 APPROVED

mohammad
 12/23/2019 11:56:06 AM

Quant Time: Dec 20 23:28:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP121919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 20 03:20:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	11.77	237	194027	99.127	nq	100
43) 2,4,6-Trichlorophenol	11.97	196	107506	44.171	nq	99
44) 2,4,5-Trichlorophenol	12.03	196	114095	45.625	nq	98
46) 1,1'-Biphenyl	12.27	154	399330	44.584	nq	99
47) 2-Chloronaphthalene	12.30	162	315864	43.623	nq	99
48) 2-Nitroaniline	12.46	65	125047	43.128	nq	99
49) Acenaphthylene	12.97	152	485934	45.909	nq	100
50) Dimethylphthalate	12.80	163	418600	47.672	nq	98
51) 2,6-Dinitrotoluene	12.88	165	79918	44.823	nq	98
52) Acenaphthene	13.26	154	278942	43.739	nq	98
53) 3-Nitroaniline	13.13	138	25430	13.288	nq	97
54) 2,4-Dinitrophenol	13.32	184	63222	91.829	nq	96
55) Dibenzofuran	13.55	168	452666	45.302	nq	97
56) 4-Nitrophenol	13.44	139	50969	37.247	nq	95
57) 2,4-Dinitrotoluene	13.54	165	112686	45.472	nq	95
58) Fluorene	14.10	166	356650	44.625	nq	97
59) 2,3,4,6-Tetrachlorophenol	13.75	232	94049m	43.806	nq	
60) Diethylphthalate	13.97	149	387866	42.827	nq	99
61) 4-Chlorophenyl-phenylether	14.12	204	185045	44.046	nq	99
62) 4-Nitroaniline	12.47	138	96837	43.697	nq	95
63) Azobenzene	14.39	77	421967	44.256	nq	99
65) 4,6-Dinitro-2-methylphenol	14.20	198	45373	48.690	nq	96
66) n-Nitrosodiphenylamine	14.32	169	298303	45.222	nq	98
67) 4-Bromophenyl-phenylether	14.92	248	110110	44.759	nq	95
68) Hexachlorobenzene	15.00	284	121027	43.215	nq	98
69) Atrazine	15.21	200	114744	48.427	nq	99
70) Pentachlorophenol	15.32	266	137403	95.671	nq	94
71) Phenanthrene	15.64	178	531767	44.795	nq	99
72) Anthracene	15.72	178	549147	46.729	nq	100
73) Carbazole	15.96	167	480592	45.870	nq	99
74) Di-n-butylphthalate	16.52	149	621246	43.208	nq	100
75) Fluoranthene	17.35	202	590882	44.514	nq	100
77) Benzidine	17.54	184	176785	37.517	nq	99
78) Pyrene	17.66	202	599823	47.733	nq	99
80) Butylbenzylphthalate	18.54	149	257659	43.981	nq	95
81) Benzo(a)anthracene	19.24	228	538223	45.638	nq	99
82) 3,3'-Dichlorobenzidine	19.20	252	107176	26.172	nq	97
83) Chrysene	19.29	228	512395	45.014	nq	99
84) Bis(2-ethylhexyl)phthalate	19.29	149	369904	43.029	nq	99
85) Di-n-octyl phthalate	20.07	149	612733	43.005	nq	100
86) Indeno(1,2,3-cd)pyrene	22.67	276	529165	43.043	nq	100
88) Benzo(b)fluoranthene	20.54	252	516386	47.158	nq	99
89) Benzo(k)fluoranthene	20.57	252	465573	44.634	nq	99
90) Benzo(a)pyrene	20.95	252	436695	43.778	nq	99
91) Dibenzo(a,h)anthracene	22.69	278	442931	45.413	nq	99
92) Benzo(g,h,i)perylene	23.16	276	438218	47.020	nq	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 8.30 min Scan# 1056
 Delta R.T. -0.00 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Instrument :
 BNA_P
 ClientSampleId :
 SSI-MW-4MS




Tot Ion	Ratio	Lower	Upper
152	100		
150	154.3	124.0	186.0
115	70.1	59.3	88.9

Manual Integrations
 APPROVED

mohammad
 12/23/2019 11:56:06 AM




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E






#2
 1,4-Dioxane
 Concen: 18.100 ng
 RT: 2.59 min Scan# 86
 Delta R.T. 0.07 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
88	100		
58	87.0	69.6	104.4
43	44.9	33.7	50.5



Abundance Ion 88.00 (87.70 to 88.70): BP0
 Ion 58.00 (57.70 to 58.70): BP0
 Ion 43.00 (42.70 to 43.70): BP0









#3
 Pvridine
 Concen: 11.200 ng
 RT: 3.47 min Scan# 235
 Delta R.T. 0.10 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Instrument :
 BNA_P
 ClientSampleId :
 SSI-MW-4MS

Tot Ion	Ion	Ratio	Lower	Upper
79	79	100		
52	79.1	68.5	102.7	
51	51.1	39.5	59.3	




Manual Integrations
 APPROVED


mohammad
 12/23/2019 11:56:06 AM



#4
 n-Nitrosodimethylamine
 Concen: 20.223 ng
 RT: 3.34 min Scan# 213
 Delta R.T. 0.06 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Tgt Ion	Ion	Ratio	Lower	Upper
42	42	100		
74	84.8	72.2	108.4	
44	6.3	4.5	6.7	









#5
2-Fluorophenol
Concen: 66.785 ng
RT: 6.21 min Scan# 701
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Resp Lower	Upper
112	100		
64	71.7	57.6	86.4
63	44.5	35.1	52.7




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#6
Aniline
Concen: 12.508 ng
RT: 7.77 min Scan# 967
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Resp Lower	Upper
93	100		
66	91.6	69.0	103.6
65	53.2	42.2	63.4





#7

Phenol-d6

Concen: 43.116 ng

RT: 7.75 min Scan# 963

Delta R.T. -0.00 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Instrument :

BNA_P





ClientSampleId :

SSI-MW-4MS

Tot Ion:	99	Resp:	192284
Ion Ratio	Lower	Upper	
99	100		
42	25.9	20.9	31.3
71	49.1	40.0	60.0

Manual Integrations
APPROVED

mohammad
12/23/2019 11:56:06 AM



#8

2-Chlorophenol

Concen: 38.672 ng




RT: 7.96 min Scan# 999


Delta R.T. -0.00 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Tgt Ion:	128	Resp:	132877
Ion Ratio	Lower	Upper	
128	100		
130	31.3	13.5	53.5
64	61.0	38.2	78.2





#9

Benzaldehyde

Concen: 47.819 ng

RT: 7.60 min Scan# 937

Delta R.T. 0.00 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Instrument :

BNA_P





ClientSampleId :

SSI-MW-4MS

Tot Ion:	77	Resp:	149598
Ion Ratio	Lower	Upper	
77	100		
105	89.8	68.6	108.6
106	86.2	62.5	102.5

Manual Integrations
APPROVED

mohammad
12/23/2019 11:56:06 AM



#10

Phenol

Concen: 15.239 ng




RT: 7.77 min Scan# 966


Delta R.T. -0.00 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Tgt Ion:	94	Resp:	77650
Ion Ratio	Lower	Upper	
94	100		
65	45.5	30.4	70.4
66	76.0	62.0	102.0









#11
bis(2-Chloroethyl)ether
Concen: 44.737 ng
RT: 7.90 min Scan# 989
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ion	Ratio	Resp	Lower	Upper
93	100		161274		
63	84.8	66.2	106.2		
95	31.9	12.2	52.2		




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#12
1,3-Dichlorobenzene
Concen: 39.431 ng
RT: 8.21 min Scan# 1041
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
146	100		168756		
148	66.1	51.9	77.9		
75	44.1	36.2	54.4		









#13
1,4-Dichlorobenzene
Concen: 39.835 ng
RT: 8.33 min Scan# 1061
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
146	100		
148	65.2	53.3	79.9
111	48.4	40.2	60.4




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#14
1,2-Dichlorobenzene
Concen: 40.236 ng
RT: 8.57 min Scan# 1102
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
146	100		
148	64.6	50.8	76.2
111	48.8	39.4	59.2









#15
Benzyl Alcohol
Concen: 29.316 ng
RT: 8.53 min Scan# 1096
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
79	100		
108	60.7	49.3	73.9
77	68.2	54.4	81.6




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.300 ng
RT: 8.77 min Scan# 1136
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
45	100		
77	14.1	0.0	34.8
79	11.0	0.0	31.2






#17
2-Methylphenol
Concen: 32.060 ng
RT: 8.73 min Scan# 1129
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
107	100		
108	107.9	90.7	136.1
77	61.4	52.7	79.1
79	57.9	50.4	75.6

Manual Integrations
APPROVED

mohammad
12/23/2019 11:56:06 AM






Abundance

Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E


Ion 77.00 (76.70 to 77.70): BP0

Ion 79.00 (78.70 to 79.70): BP0



#18
Hexachloroethane
Concen: 38.011 ng
RT: 9.10 min Scan# 1193
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
117	100		
119	98.3	75.0	112.4
201	95.0	75.6	113.4











Abundance

Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E


Ion 201.00 (200.70 to 201.70): E





#19
n-Nitroso-di-n-propylamine
Concen: 45.461 ng
RT: 8.97 min Scan# 1171
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tot Ion	Ratio	Lower	Upper
70	100		
42	65.2	54.1	81.1
101	9.6	7.0	10.4
130	16.0	13.0	19.6




Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS


Manual Integrations
APPROVED

mohammad
12/23/2019 11:56:06 AM

#20
3+4-Methylphenols
Concen: 28.419 ng
RT: 8.97 min Scan# 1171
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
107	100		
108	92.3	67.2	107.2
77	47.6	28.0	68.0
79	41.2	18.6	58.6









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.34 min Scan# 1403
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ion	Ratio	Resp	Lower	Upper
136	100		189969		
137	10.8	8.8	13.2		
54	10.5	8.8	13.2		
68	5.2	4.6	6.8		




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#22
Acetophenone
Concen: 43.514 ng
RT: 8.95 min Scan# 1167
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
105	100		261404		
71	1.6	0.8	1.2#		
51	36.7	30.6	45.8		
120	19.3	16.5	24.7		









#23
 Nitrobenzene-d5
 Concen: 90.172 ng
 RT: 9.19 min Scan# 1207
 Delta R.T. -0.00 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Instrument :
 BNA_P
 ClientSampleId :
 SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
82	100		
128	33.8	27.0	40.4
54	62.0	49.0	73.4




Manual Integrations
 APPROVED


mohammad
 12/23/2019 11:56:06 AM



#24
 Nitrobenzene
 Concen: 44.838 ng
 RT: 9.22 min Scan# 1212
 Delta R.T. -0.00 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
77	100		
123	36.4	29.4	44.0
65	16.0	12.4	18.6





#25

Isophorone

Concen: 46.507 ng

RT: 9.61 min Scan# 1279

Delta R.T. -0.00 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Instrument :

BNA_P





ClientSampleId :

SSI-MW-4MS

Tot Ion:	82	Resp:	369936
Ion Ratio	Lower	Upper	
82	100		
95	8.8	7.3	10.9
138	14.5	11.5	17.3

Manual Integrations
APPROVED

mohammad
12/23/2019 11:56:06 AM



#26

2-Nitrophenol

Concen: 46.475 ng




RT: 9.73 min Scan# 1299


Delta R.T. 0.01 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Tgt Ion:	139	Resp:	80690
Ion Ratio	Lower	Upper	
139	100		
109	50.0	39.1	58.7
65	72.1	61.2	91.8









#27
2,4-Dimethylphenol
Concen: 47.388 ng
RT: 9.82 min Scan# 1315
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
122	100		
107	142.9	111.9	167.9
121	59.6	47.9	71.9




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#28
bis(2-Chloroethoxy)methane
Concen: 42.242 ng
RT: 9.98 min Scan# 1342
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
93	100		
95	31.4	25.4	38.0
123	10.4	8.8	13.2









#29
2,4-Dichlorophenol
Concen: 42.392 ng
RT: 10.12 min Scan# 1365
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampled :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
162	100		
164	65.0	45.9	85.9
98	43.0	23.2	63.2




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#30
1,2,4-Trichlorobenzene
Concen: 40.136 ng
RT: 10.25 min Scan# 1388
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
180	100		
182	96.6	75.0	112.6
145	32.3	25.8	38.8









#31
Naphthalene
Concen: 43.436 ng
RT: 10.37 min Scan# 1409
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion:	128	Resp:	450425
Ion Ratio	Lower	Upper	
128	100		
129	11.0	8.9	13.3
127	13.2	10.7	16.1




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#32
Benzoic acid
Concen: 13.975 ng
RT: 9.95 min Scan# 1337
Delta R.T. -0.04 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion:	122	Resp:	20642
Ion Ratio	Lower	Upper	
122	100		
105	130.3	118.5	158.5
77	117.1	101.3	141.3









#33
4-Chloroaniline
Concen: 8.328 ng
RT: 10.46 min Scan# 1424
Delta R.T. -0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
127	100		
129	31.8	25.2	37.8
65	44.6	33.6	50.4
92	23.6	19.8	29.6




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#34
Hexachlorobutadiene
Concen: 37.696 ng
RT: 10.59 min Scan# 1446
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.7	49.2	73.8
227	63.4	50.3	75.5









#35
Caprolactam
Concen: 6.921 ng
RT: 11.05 min Scan# 1523
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
113	100		
55	185.4	191.5	231.5#
56	153.0	160.5	200.5#




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#36
4-Chloro-3-methylphenol
Concen: 39.568 ng
RT: 11.28 min Scan# 1563
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
107	100		
144	23.4	16.9	25.3
142	67.3	55.9	83.9









#37
2-Methylnaphthalene
Concen: 44.734 ng
RT: 11.50 min Scan# 1601
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Resp Lower	Upper
142	100		
141	88.7	70.2	105.4
115	49.4	38.7	58.1




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#38
1-Methylnaphthalene
Concen: 44.192 ng
RT: 11.66 min Scan# 1628
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Resp Lower	Upper
142	100		
141	92.7	71.3	106.9
116	5.4	4.2	6.2









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 13.20 min Scan# 1890
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
164	100		
162	102.7	80.3	120.5
160	47.1	36.6	54.8




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#40
1,2,4,5-Tetrachlorobenzene
Concen: 43.424 ng
RT: 11.78 min Scan# 1648
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
216	100		
214	78.3	62.6	94.0
179	22.9	18.1	27.1
108	25.6	18.2	27.2





#41

Hexachlorocyclopentadiene

Concen: 99.127 ng

RT: 11.77 min Scan# 1647

Delta R.T. -0.00 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Instrument :

BNA_P





ClientSampleId :

SSI-MW-4MS

Tot Ion:	237	Resp:	194027
Ion Ratio	Lower	Upper	
237	100		
235	62.0	41.9	81.9
272	12.0	0.0	32.0

Manual Integrations
APPROVED

mohammad
12/23/2019 11:56:06 AM



#42

2,4,6-Tribromophenol

Concen: 137.685 ng




RT: 14.50 min Scan# 2111


Delta R.T. 0.01 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Tgt Ion:	330	Resp:	184242
Ion Ratio	Lower	Upper	
330	100		
332	97.9	78.7	118.1
141	45.5	36.2	54.4









#43
2,4,6-Trichlorophenol
Concen: 44.171 ng
RT: 11.97 min Scan# 1680
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Resp	Lower	Upper
196	100	107506		
198	97.4	78.6	118.0	
200	31.0	24.3	36.5	




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#44
2,4,5-Trichlorophenol
Concen: 45.625 ng
RT: 12.03 min Scan# 1690
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	114095		
198	95.7	77.7	116.5	
97	64.4	50.7	76.1	
132	27.8	20.7	31.1	









#45
2-Fluorobiphenyl
Concen: 91.257 ng
RT: 12.12 min Scan# 1706
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
172	100		
171	34.4	27.7	41.5
170	23.7	18.6	28.0




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#46
1,1'-Biphenyl
Concen: 44.584 ng
RT: 12.27 min Scan# 1732
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
154	100		
153	39.5	20.2	60.2
76	13.5	0.0	33.1









#47
2-Chloronaphthalene
Concen: 43.623 ng
RT: 12.30 min Scan# 1736
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion:	162	Resp:	315864
Ion Ratio	Lower	Upper	
162	100		
127	42.2	34.2	51.4
164	33.2	26.4	39.6




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#48
2-Nitroaniline
Concen: 43.128 ng
RT: 12.46 min Scan# 1764
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion:	65	Resp:	125047
Ion Ratio	Lower	Upper	
65	100		
92	61.1	48.3	72.5
138	76.7	60.2	90.4









#49
Acenaphthylene
Concen: 45.909 ng
RT: 12.97 min Scan# 1850
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
152	100		
151	20.1	16.2	24.2
153	12.8	10.4	15.6




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#50
Dimethylphthalate
Concen: 47.672 ng
RT: 12.80 min Scan# 1822
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
163	100		
194	3.8	3.4	5.0
164	10.5	7.9	11.9









#51
2,6-Dinitrotoluene
Concen: 44.823 ng
RT: 12.88 min Scan# 1835
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
165	100		
63	88.1	71.3	106.9
89	83.2	68.5	102.7




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#52
Acenaphthene
Concen: 43.739 ng
RT: 13.26 min Scan# 1899
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
154	100		
153	116.2	90.6	135.8
152	54.1	42.4	63.6









#53
3-Nitroaniline
Concen: 13.288 ng
RT: 13.13 min Scan# 1878
Delta R.T. -0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
138	100		
108	13.4	10.1	15.1
92	147.6	114.6	172.0




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#54
2,4-Dinitrophenol
Concen: 91.829 ng
RT: 13.32 min Scan# 1910
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
184	100		
63	93.0	78.6	117.8
154	80.3	62.2	93.2









#55
Dibenzofuran
Concen: 45.302 ng
RT: 13.55 min Scan# 1948
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
168	100		
139	42.3	35.8	53.8
169	13.2	10.5	15.7




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#56
4-Nitrophenol
Concen: 37.247 ng
RT: 13.44 min Scan# 1930
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
139	100		
109	131.2	110.1	150.1
65	140.5	131.3	171.3









#57
2,4-Dinitrotoluene
Concen: 45.472 ng
RT: 13.54 min Scan# 1947
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Resp	Lower	Upper
165	100			
63	89.6	67.6	101.4	
89	106.3	81.3	121.9	
182	2.6	2.2	3.2	




Manual Integrations
APPROVED




mohammad
12/23/2019 11:56:06 AM



#58
Fluorene
Concen: 44.625 ng
RT: 14.10 min Scan# 2043
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Resp	Lower	Upper
166	100			
165	98.9	76.9	115.3	
167	13.5	10.3	15.5	









#59
2,3,4,6-Tetrachlorophenol
Concen: 43.806 ng/ml
RT: 13.75 min Scan# 1983
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ion	Ratio	Resp	Lower	Upper
232	100				
131	45.7	36.1	54.1		
130	2.9	2.5	3.7		
166	27.4	22.5	33.7		


Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#60
Diethylphthalate
Concen: 42.827 ng/ml
RT: 13.97 min Scan# 2020
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
149	100				
177	21.5	16.7	25.1		
150	12.3	10.2	15.4		









#61
4-Chlorophenyl-phenylether
Concen: 44.046 ng
RT: 14.12 min Scan# 2046
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Resp Lower	Upper
204	100		
206	33.5	27.4	41.0
141	71.8	57.3	85.9




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#62
4-Nitroaniline
Concen: 43.697 ng
RT: 12.47 min Scan# 1765
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Resp Lower	Upper
138	100		
92	74.5	60.3	100.3
108	34.0	14.5	54.5









#63
Azobenzene
Concen: 44.256 ng
RT: 14.39 min Scan# 2091
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
77	100		
182	21.4	2.0	42.0
105	18.2	0.0	38.1
51	40.5	20.0	60.0




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 15.60 min Scan# 2298
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
188	100		
94	7.7	6.6	10.0
80	9.0	7.8	11.6









#65
4,6-Dinitro-2-methylphenol
Concen: 48.690 ng
RT: 14.20 min Scan# 2060
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
198	100		
51	70.8	47.4	87.4
105	50.1	33.1	73.1




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#66
n-Nitrosodiphenylamine
Concen: 45.222 ng
RT: 14.32 min Scan# 2080
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
169	100		
168	68.5	53.4	80.2
167	37.3	29.4	44.0









#67
4-Bromophenyl-phenylether
Concen: 44.759 ng
RT: 14.92 min Scan# 2182
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
248	100		
250	95.1	79.3	118.9
141	90.5	77.5	116.3




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#68
Hexachlorobenzene
Concen: 43.215 ng
RT: 15.00 min Scan# 2196
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
284	100		
142	42.9	36.1	54.1
249	33.6	26.3	39.5









#69
Atrazine
Concen: 48.427 ng
RT: 15.21 min Scan# 2231
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
200	100		
173	25.3	6.8	46.8
215	46.3	26.1	66.1




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#70
Pentachlorophenol
Concen: 95.671 ng
RT: 15.32 min Scan# 2250
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
266	100		
268	64.1	55.4	83.2
264	61.5	53.2	79.8









#71
Phenanthrene
Concen: 44.795 ng
RT: 15.64 min Scan# 2304
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.2	22.8
179	15.5	12.0	18.0




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#72
Anthracene
Concen: 46.729 ng
RT: 15.72 min Scan# 2318
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.6
179	15.2	11.9	17.9









#73
 Carbazole
 Concen: 45.870 ng
 RT: 15.96 min Scan# 2359
 Delta R.T. -0.00 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Instrument :
 BNA_P
 ClientSampleId :
 SSI-MW-4MS

Tot Ion:	167	Resp:	480592
Ion Ratio	Lower	Upper	
167	100		
166	21.6	17.6	26.4
139	14.3	11.1	16.7




Manual Integrations
 APPROVED


mohammad
 12/23/2019 11:56:06 AM



#74
 Di-n-butylphthalate
 Concen: 43.208 ng
 RT: 16.52 min Scan# 2453
 Delta R.T. -0.00 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Tgt Ion:	149	Resp:	621246
Ion Ratio	Lower	Upper	
149	100		
150	9.1	7.5	11.3
104	6.8	5.4	8.2





#75

Fluoranthene

Concen: 44.514 ng

RT: 17.35 min Scan# 2595

Delta R.T. 0.01 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Instrument :

BNA_P





ClientSampleId :

SSI-MW-4MS

Tot Ion	Ion	Ratio	Resp	Lower	Upper
202	100		590882		
101	9.5	0.0	29.9		
203	17.6	0.0	37.6		

Manual Integrations
APPROVED

mohammad
12/23/2019 11:56:06 AM



#76

Chrysene-d12

Concen: 20.000 ng




RT: 19.25 min Scan# 2918


Delta R.T. 0.01 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
240	100		161843		
120	8.3	7.6	11.4		
236	26.3	21.0	31.4		





#77

Benzidine

Concen: 37.517 ng

RT: 17.54 min Scan# 2627

Delta R.T. -0.00 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Instrument :

BNA_P





ClientSampleId :

SSI-MW-4MS

Tot Ion:	184	Resp:	176785
Ion Ratio	Lower	Upper	
184	100		
185	14.2	11.7	17.5
183	11.6	9.8	14.8

Manual Integrations
APPROVED

mohammad
12/23/2019 11:56:06 AM



#78

Pyrene

Concen: 47.733 ng




RT: 17.66 min Scan# 2647


Delta R.T. 0.01 min

Lab File: BP001341.D

Acq: 20 Dec 2019 13:30

Tgt Ion:	202	Resp:	599823
Ion Ratio	Lower	Upper	
202	100		
200	21.8	16.8	25.2
203	17.4	13.8	20.8









#79
 Terphenyl-d14
 Concen: 100.942 ng
 RT: 17.89 min Scan# 2687
 Delta R.T. 0.01 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Instrument :
 BNA_P
 ClientSampleId :
 SSI-MW-4MS

Tot Ion	Ratio	Resp Lower	Upper
244	100		
212	8.2	6.6	9.8
122	9.4	7.6	11.4




Manual Integrations
 APPROVED


mohammad
 12/23/2019 11:56:06 AM



#80
 Butylbenzylphthalate
 Concen: 43.981 ng
 RT: 18.54 min Scan# 2797
 Delta R.T. -0.00 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Resp Lower	Upper
149	100		
91	86.9	65.2	97.8
206	18.9	15.6	23.4









#81
Benzo(a)anthracene
Concen: 45.638 ng
RT: 19.24 min Scan# 2916
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
228	100		
226	27.1	21.2	31.8
229	19.3	15.4	23.0




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#82
3,3'-Dichlorobenzidine
Concen: 26.172 ng
RT: 19.20 min Scan# 2910
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
252	100		
254	66.7	51.8	77.8
126	12.1	8.2	12.2









#83
 Chrysene
 Concen: 45.014 ng
 RT: 19.29 min Scan# 2924
 Delta R.T. 0.01 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Instrument :
 BNA_P
 ClientSampleId :
 SSI-MW-4MS

Tot Ion:	228	Resp:	512395
Ion Ratio	Lower	Upper	
228	100		
226	29.2	23.1	34.7
229	18.8	15.4	23.2




Manual Integrations
 APPROVED


mohammad
 12/23/2019 11:56:06 AM



#84
 Bis(2-ethylhexyl)phthalate
 Concen: 43.029 ng
 RT: 19.29 min Scan# 2925
 Delta R.T. 0.01 min
 Lab File: BP001341.D
 Acq: 20 Dec 2019 13:30

Tgt Ion:	149	Resp:	369904
Ion Ratio	Lower	Upper	
149	100		
167	26.8	21.8	32.6
279	4.2	3.0	4.4









#85
Di-n-octyl phthalate
Concen: 43.005 ng
RT: 20.07 min Scan# 3058
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	11.9	9.4	14.2




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#86
Indeno(1,2,3-cd)pyrene
Concen: 43.043 ng
RT: 22.67 min Scan# 3499
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
276	100		
138	19.4	15.8	23.6
277	25.2	20.2	30.4









#87
Pervlene-d12
Concen: 20.000 ng
RT: 21.02 min Scan# 3219
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.2	28.8
265	22.7	17.4	26.2




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#88
Benzo(b)fluoranthene
Concen: 47.158 ng
RT: 20.54 min Scan# 3137
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.8	26.6
125	7.3	6.3	9.5









#89
Benzo(k)fluoranthene
Concen: 44.634 ng
RT: 20.57 min Scan# 3143
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.2	17.4	26.0
125	7.1	5.9	8.9




Manual Integrations
APPROVED


mohammad
12/23/2019 11:56:06 AM



#90
Benzo(a)pyrene
Concen: 43.778 ng
RT: 20.95 min Scan# 3207
Delta R.T. -0.00 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	8.3	6.3	9.5









#91
Dibenzo(a,h)anthracene
Concen: 45.413 ng
RT: 22.69 min Scan# 3503
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Instrument :
BNA_P
ClientSampleId :
SSI-MW-4MS

Tot Ion	Ratio	Lower	Upper
278	100		
139	12.7	10.3	15.5
279	23.6	19.6	29.4

Manual Integrations
APPROVED

mohammad
12/23/2019 11:56:06 AM



#92
Benzo(g,h,i)perylene
Concen: 47.020 ng
RT: 23.16 min Scan# 3582
Delta R.T. 0.01 min
Lab File: BP001341.D
Acq: 20 Dec 2019 13:30

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.0	19.6	29.4
138	16.0	12.6	18.8

