

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122422\
 Data File : BPCCC308.D
 Acq On : 24 Dec 2022 20:20
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Dec 25 00:25:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 00:19:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	622752	20.000	ng/ul	0.00
20) Naphthalene-d8	10.752	136	2561426	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.587	164	1466307	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.345	188	2547348	20.000	ng/ul	0.00
79) Chrysene-d12	21.433	240	2424483	20.000	ng/ul	0.00
88) Perylene-d12	23.916	264	2820238	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.328	96	107298	7.367	ng/uL	0.00
4) Pyridine-d5	3.758	84	826885	19.985	ng/ul	0.00
7) Phenol-d5	7.105	99	1008269	19.877	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.264	67	608330	19.881	ng/ul	0.00
11) 2-Chlorophenol-d4	7.469	132	812090	20.337	ng/ul	0.00
15) 4-Methylphenol-d8	8.652	113	786339	18.919	ng/ul	0.00
21) Nitrobenzene-d5	9.110	128	390151	20.731	ng/ul	0.00
24) 2-Nitrophenol-d4	9.834	143	441958	20.860	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	827480	20.107	ng/ul	0.00
31) 4-Chloroaniline-d4	10.893	131	989155	17.026	ng/ul	0.00
46) Dimethylphthalate-d6	13.993	166	2119875	18.811	ng/ul	0.00
49) Acenaphthylene-d8	14.281	160	2458287	19.852	ng/ul	0.00
54) 4-Nitrophenol-d4	14.787	143	293967	14.502	ng/ul	0.00
60) Fluorene-d10	15.581	176	1696257	17.792	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	247576	15.103	ng/ul	0.00
73) Anthracene-d10	17.445	188	2254697	19.619	ng/ul	0.00
81) Pyrene-d10	19.675	212	2521072	18.116	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.751	264	2723413	19.374	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.364	88	112834	7.421	ng/uL	92
5) Pyridine	3.775	79	828075	20.138	ng/ul	95
6) Benzaldehyde	7.075	77	392316	19.752	ng/ul	96
8) Phenol	7.128	94	1020823	19.904	ng/ul	96
10) Bis(2-Chloroethyl)ether	7.358	93	798738	18.975	ng/ul	95
12) 2-Chlorophenol	7.499	128	821721	20.067	ng/ul	95
13) 2-Methylphenol	8.387	108	772828	19.328	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.469	45	1200032	20.775	ng/ul	98
16) Acetophenone	8.769	105	1238604	19.436	ng/ul#	95
17) N-Nitroso-di-n-propyla...	8.752	70	622990	19.440	ng/ul	94
18) 4-Methylphenol	8.716	108	837754	18.916	ng/ul	97
19) Hexachloroethane	9.022	117	319365	19.517	ng/ul	96
22) Nitrobenzene	9.152	77	926882	20.989	ng/ul	97
23) Isophorone	9.681	82	1690384	18.930	ng/ul	97
25) 2-Nitrophenol	9.863	139	466571	20.324	ng/ul	92
26) 2,4-Dimethylphenol	9.922	107	897409	20.035	ng/ul	95
27) Bis(2-Chloroethoxy)met...	10.157	93	1079696	18.910	ng/ul	99
29) 2,4-Dichlorophenol	10.399	162	807851	20.040	ng/ul	98
30) Naphthalene	10.805	128	2619410	19.554	ng/ul	99
32) 4-Chloroaniline	10.916	127	994964	17.306	ng/ul	100
33) Hexachlorobutadiene	11.087	225	570404	20.608	ng/ul	99
34) Caprolactam	11.704	113	208230	15.149	ng/ul	94
35) 4-Chloro-3-methylphenol	12.040	107	764525	18.584	ng/ul	99
36) 2-Methylnaphthalene	12.410	142	1724525	18.765	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122422\
 Data File : BPCCC308.D
 Acq On : 24 Dec 2022 20:20
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Dec 25 00:25:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 00:19:17 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.628	142	1745554	18.469	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.775	216	1055412	22.353	ng/ul	99
40) Hexachlorocyclopentadiene	12.751	237	232113	7.554	ng/ul	100
41) 2,4,6-Trichlorophenol	13.016	196	639286	21.068	ng/ul	99
42) 2,4,5-Trichlorophenol	13.093	196	666566	20.451	ng/ul	99
43) 1,1'-Biphenyl	13.416	154	2293236	20.824	ng/ul	99
44) 2-Chloronaphthalene	13.463	162	1812439	20.911	ng/ul	99
45) 2-Nitroaniline	13.669	65	458234	21.806	ng/ul	98
47) Dimethylphthalate	14.040	163	2113343	18.921	ng/ul	99
48) 2,6-Dinitrotoluene	14.163	165	407500	19.632	ng/ul	99
50) Acenaphthylene	14.304	152	2639310	19.825	ng/ul	100
51) 3-Nitroaniline	14.493	138	343696	17.535	ng/ul	95
52) Acenaphthene	14.651	153	1799769	19.509	ng/ul	99
53) 2,4-Dinitrophenol	14.698	184	140882	9.929	ng/ul	96
55) 4-Nitrophenol	14.804	109	223961	16.192	ng/ul	94
56) Dibenzofuran	14.987	168	2498867	19.051	ng/ul	100
57) 2,4-Dinitrotoluene	14.951	165	525241	17.420	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.210	232	526289	17.693	ng/ul	98
59) Diethylphthalate	15.404	149	1935786	17.829	ng/ul	99
61) Fluorene	15.634	166	1935191	18.396	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.628	204	1065932	19.161	ng/ul	100
63) 4-Nitroaniline	15.657	138	290717	16.856	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.716	198	242302	15.126	ng/ul	98
67) N-Nitrosodiphenylamine	15.840	169	1594325	22.575	ng/ul	98
68) 4-Bromophenyl-phenylether	16.528	248	609952	22.049	ng/ul	99
69) Hexachlorobenzene	16.645	284	680962	20.731	ng/ul	100
70) Atrazine	16.798	200	546412	19.854	ng/ul	99
71) Pentachlorophenol	16.992	266	365659	18.382	ng/ul	99
72) Phenanthrene	17.387	178	2613238	19.696	ng/ul	100
74) Anthracene	17.481	178	2622422	19.795	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.381	216	1010970	27.817	ng/ul	99
76) Pentachlorobenzene	14.904	250	921884	24.879	ng/ul	99
77) Carbazole	17.751	167	2181218	19.550	ng/ul	99
78) Di-n-butylphthalate	18.286	149	2564152	18.875	ng/ul	100
80) Fluoranthene	19.351	202	2884831	18.264	ng/ul	99
82) Pyrene	19.704	202	3050412	18.722	ng/ul	100
83) Butylbenzylphthalate	20.569	149	1052776	16.725	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.345	252	949183	17.330	ng/ul	99
85) Benzo(a)anthracene	21.416	228	3230546	20.072	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.327	149	1504891	16.949	ng/ul	99
87) Chrysene	21.475	228	3011463	19.785	ng/ul	100
89) Di-n-octyl phthalate	22.275	149	2745164	17.375	ng/ul	100
90) Benzo(b)fluoranthene	23.157	252	3420674	19.021	ng/ul	99
91) Benzo(k)fluoranthene	23.204	252	3302576	18.519	ng/ul	99
93) Benzo(a)pyrene	23.804	252	3027374	19.333	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.510	276	4040930	20.717	ng/ul	98
95) Dibenzo(a,h)anthracene	26.527	278	3376402	20.907	ng/ul	99
96) Benzo(g,h,i)perylene	27.304	276	3301063	21.081	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122422\
 Data File : BPCCC308.D
 Acq On : 24 Dec 2022 20:20
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Dec 25 00:25:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 00:19:17 2022
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.421 ng/uL
 RT: 3.364 min Scan# 30
 Delta R.T. -0.006 min
 Lab File: BPCCC308.D
 Acq: 24 Dec 2022 20:20




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion: 88 Resp: 112834
 Ion Ratio Lower Upper
 88 100
 43 33.4 22.4 33.6
 58 79.1 58.2 87.4



#5
 Pyridine
 Concen: 20.138 ng/uL
 RT: 3.775 min Scan# 100
 Delta R.T. -0.006 min
 Lab File: BPCCC308.D
 Acq: 24 Dec 2022 20:20

Tgt Ion: 79 Resp: 828075
 Ion Ratio Lower Upper
 79 100
 52 80.9 60.7 91.1
 51 38.6 28.9 43.3









#6
Benzaldehyde
Concen: 19.752 ng/ul
RT: 7.075 min Scan# 60
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 392316
Ion Ratio Lower Upper
77 100
105 96.8 81.5 122.3
106 93.9 76.9 115.3



#8
Phenol
Concen: 19.904 ng/ul
RT: 7.128 min Scan# 670
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion: 94 Resp: 1020823
Ion Ratio Lower Upper
94 100
65 28.3 21.1 31.7
66 36.5 27.0 40.6





#10

Bis(2-Chloroethyl)ether

Concen: 18.975 ng/ul

RT: 7.358 min Scan# 709

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion: 93 Resp: 798738

Ion Ratio Lower Upper

93 100

63 81.3 59.8 89.8

95 32.9 26.4 39.6



#12

2-Chlorophenol

Concen: 20.067 ng/ul

RT: 7.499 min Scan# 733

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20




Tgt Ion:128 Resp: 821721


Ion Ratio Lower Upper

128 100

64 49.0 34.9 52.3

130 31.9 25.8 38.6





#13

2-Methylphenol

Concen: 19.328 ng/ul

RT: 8.387 min Scan# 884

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 772828

Ion Ratio Lower Upper

108 100

107 88.3 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.775 ng/ul

RT: 8.469 min Scan# 898

Delta R.T. -0.012 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20




Tgt Ion: 45 Resp: 1200032


Ion Ratio Lower Upper

45 100

77 13.1 11.0 16.6

79 10.4 9.3 13.9









#16
Acetophenone
Concen: 19.436 ng/ul
RT: 8.769 min Scan# 949
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 1238604
Ion Ratio Lower Upper
105 100
77 76.7 59.3 88.9
51 29.5 19.6 29.4#



#17
N-Nitroso-di-n-propylamine
Concen: 19.440 ng/ul
RT: 8.752 min Scan# 946
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion: 70 Resp: 622990
Ion Ratio Lower Upper
70 100
42 62.8 45.5 68.3
101 9.4 7.9 11.9
130 20.1 17.4 26.2









#18
4-Methylphenol
Concen: 18.916 ng/ul
RT: 8.716 min Scan# 940
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 837754
Ion Ratio Lower Upper
108 100
107 113.0 87.4 131.2



#19
Hexachloroethane
Concen: 19.517 ng/ul
RT: 9.022 min Scan# 992
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:117 Resp: 319365
Ion Ratio Lower Upper
117 100
201 112.4 93.7 140.5
199 70.5 58.6 88.0









#22
Nitrobenzene
Concen: 20.989 ng/ul
RT: 9.152 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 926882
Ion Ratio Lower Upper
77 100
123 46.1 38.3 57.5
65 14.1 10.3 15.5



#23
Isophorone
Concen: 18.930 ng/ul
RT: 9.681 min Scan# 1104
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion: 82 Resp: 1690384
Ion Ratio Lower Upper
82 100
95 6.9 5.4 8.2
138 17.6 15.3 22.9









#25
2-Nitrophenol
Concen: 20.324 ng/ul
RT: 9.863 min Scan# 1135
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 466571
Ion Ratio Lower Upper
139 100
65 52.3 36.6 54.8
109 29.0 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 20.035 ng/ul
RT: 9.922 min Scan# 1145
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20


Tgt Ion:107 Resp: 897409
Ion Ratio Lower Upper
107 100
121 49.5 41.8 62.6
122 85.0 72.6 108.8








#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.910 ng/ul
 RT: 10.157 min Scan# 1185
 Delta R.T. -0.006 min
 Lab File: BPCCC308.D
 Acq: 24 Dec 2022 20:20

Instrument :
 BNA_P
 ClientSampleId :







Tgt Ion: 93 Resp: 1079696
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.0 39.0
 123 11.2 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 20.040 ng/ul
 RT: 10.399 min Scan# 1226
 Delta R.T. -0.006 min
 Lab File: BPCCC308.D
 Acq: 24 Dec 2022 20:20

Tgt Ion:162 Resp: 807851
 Ion Ratio Lower Upper
 162 100
 164 66.2 52.5 78.7
 98 35.1 26.6 39.8





#30

Naphthalene

Concen: 19.554 ng/ul

RT: 10.805 min Scan# 1295

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:128 Resp: 2619410

Ion Ratio Lower Upper

128 100

129 10.9 8.7 13.1

127 13.4 10.4 15.6



#32

4-Chloroaniline

Concen: 17.306 ng/ul

RT: 10.916 min Scan# 1314

Delta R.T. -0.006 min

Lab File: BPCCC308.D




Acq: 24 Dec 2022 20:20


Tgt Ion:127 Resp: 994964

Ion Ratio Lower Upper

127 100

129 32.2 25.6 38.4





#33

Hexachlorobutadiene

Concen: 20.608 ng/ul

RT: 11.087 min Scan# 11

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20





Instrument :

BNA_P

ClientSampleId :

Tgt Ion:225 Resp: 570404

Ion	Ratio	Lower	Upper
225	100		
223	61.3	49.2	73.8
227	64.3	50.2	75.2



#34

Caprolactam

Concen: 15.149 ng/ul

RT: 11.704 min Scan# 1448




Delta R.T. -0.006 min


Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Tgt Ion:113 Resp: 208230

Ion	Ratio	Lower	Upper
113	100		
55	183.1	141.9	212.9
56	134.5	99.6	149.4









#35
4-Chloro-3-methylphenol
Concen: 18.584 ng/ul
RT: 12.040 min Scan# 1505
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 764525
Ion Ratio Lower Upper
107 100
144 26.8 21.4 32.2
142 83.6 65.8 98.6



#36
2-Methylnaphthalene
Concen: 18.765 ng/ul
RT: 12.410 min Scan# 1568
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:142 Resp: 1724525
Ion Ratio Lower Upper
142 100
141 88.9 70.8 106.2









#37
1-Methylnaphthalene
Concen: 18.469 ng/ul
RT: 12.628 min Scan# 1605
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 1745554
Ion Ratio Lower Upper
142 100
141 92.5 72.7 109.1
116 3.6 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.353 ng/ul
RT: 12.775 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:216 Resp: 1055412
Ion Ratio Lower Upper
216 100
214 77.1 61.9 92.9
179 19.3 15.0 22.6
108 14.2 11.0 16.4





#40

Hexachlorocyclopentadiene

Concen: 7.554 ng/ul

RT: 12.751 min Scan# 1626

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:237 Resp: 232113

Ion Ratio Lower Upper

237 100

235 61.9 49.7 74.5

272 13.0 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 21.068 ng/ul

RT: 13.016 min Scan# 1671

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20




Tgt Ion:196 Resp: 639286


Ion Ratio Lower Upper

196 100

198 95.2 77.0 115.6


200 30.9 25.3 37.9








#42
2,4,5-Trichlorophenol
Concen: 20.451 ng/ul
RT: 13.093 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:196 Resp: 666566
Ion Ratio Lower Upper
196 100
198 97.6 77.1 115.7
200 32.0 25.4 38.2



#43
1,1'-Biphenyl
Concen: 20.824 ng/ul
RT: 13.416 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:154 Resp: 2293236
Ion Ratio Lower Upper
154 100
153 40.2 31.5 47.3
76 12.8 9.6 14.4





#44

2-Chloronaphthalene

Concen: 20.911 ng/ul

RT: 13.463 min Scan# 1747

Delta R.T. 0.000 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:162 Resp: 1812439

Ion Ratio Lower Upper

162 100

164 32.8 26.3 39.5

127 35.9 29.3 43.9



#45

2-Nitroaniline

Concen: 21.806 ng/ul

RT: 13.669 min Scan# 1782

Delta R.T. 0.000 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20




Tgt Ion: 65 Resp: 458234


Ion Ratio Lower Upper

65 100

92 69.4 56.4 84.6

138 109.7 90.3 135.5









#47
Dimethylphthalate
Concen: 18.921 ng/ul
RT: 14.040 min Scan# 1845
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 2113343
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.0 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 19.632 ng/ul
RT: 14.163 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:165 Resp: 407500
Ion Ratio Lower Upper
165 100
89 54.7 43.0 64.4
121 20.3 16.1 24.1









#50
Acenaphthylene
Concen: 19.825 ng/ul
RT: 14.304 min Scan# 1890
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 2639310
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.4 10.6 16.0



#51
3-Nitroaniline
Concen: 17.535 ng/ul
RT: 14.493 min Scan# 1922
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:138 Resp: 343696
Ion Ratio Lower Upper
138 100
108 10.5 8.6 13.0
92 116.2 88.4 132.6









#52
Acenaphthene
Concen: 19.509 ng/ul
RT: 14.651 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 1799769
Ion Ratio Lower Upper
153 100
152 48.1 38.2 57.4
154 89.1 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 9.929 ng/ul
RT: 14.698 min Scan# 1957
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:184 Resp: 140882
Ion Ratio Lower Upper
184 100
63 63.0 48.4 72.6
154 60.3 45.5 68.3









#55
4-Nitrophenol
Concen: 16.192 ng/ul
RT: 14.804 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 223961
Ion Ratio Lower Upper
109 100
139 135.4 120.8 181.2
65 135.0 107.9 161.9



#56
Dibenzofuran
Concen: 19.051 ng/ul
RT: 14.987 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:168 Resp: 2498867
Ion Ratio Lower Upper
168 100
139 34.8 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 17.420 ng/ul
RT: 14.951 min Scan# 2000
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 525241
Ion Ratio Lower Upper
165 100
63 42.5 31.8 47.6
182 2.9 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 17.693 ng/ul
RT: 15.210 min Scan# 2044
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:232 Resp: 526289
Ion Ratio Lower Upper
232 100
131 38.7 31.8 47.6
130 2.1 1.8 2.6
166 27.0 22.9 34.3









#59
Diethylphthalate
Concen: 17.829 ng/ul
RT: 15.404 min Scan# 2077
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 1935786
Ion Ratio Lower Upper
149 100
177 23.4 18.4 27.6
150 12.1 9.9 14.9



#61
Fluorene
Concen: 18.396 ng/ul
RT: 15.634 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:166 Resp: 1935191
Ion Ratio Lower Upper
166 100
165 97.7 77.4 116.2
167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 19.161 ng/ul
RT: 15.628 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 1065932
Ion Ratio Lower Upper
204 100
206 33.1 26.3 39.5
141 50.2 40.4 60.6



#63
4-Nitroaniline
Concen: 16.856 ng/ul
RT: 15.657 min Scan# 2120
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:138 Resp: 290717
Ion Ratio Lower Upper
138 100
92 51.4 39.0 58.4
108 75.7 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 15.126 ng/ul
RT: 15.716 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 242302
Ion Ratio Lower Upper
198 100
182 4.6 3.3 4.9
77 21.9 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 22.575 ng/ul
RT: 15.840 min Scan# 2151
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:169 Resp: 1594325
Ion Ratio Lower Upper
169 100
168 64.9 53.0 79.6
167 35.6 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 22.049 ng/ul
RT: 16.528 min Scan# 21
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 609952
Ion Ratio Lower Upper
248 100
250 96.5 77.4 116.0
141 58.3 48.3 72.5



#69
Hexachlorobenzene
Concen: 20.731 ng/ul
RT: 16.645 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:284 Resp: 680962
Ion Ratio Lower Upper
284 100
142 27.9 22.6 34.0
249 30.7 24.6 37.0









#70
Atrazine
Concen: 19.854 ng/ul
RT: 16.798 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:200 Resp: 546412
Ion Ratio Lower Upper
200 100
173 25.1 20.5 30.7
215 50.8 39.9 59.9



#71
Pentachlorophenol
Concen: 18.382 ng/ul
RT: 16.992 min Scan# 2347
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:266 Resp: 365659
Ion Ratio Lower Upper
266 100
264 62.5 50.9 76.3
268 64.8 51.4 77.2









#72
Phenanthrene
Concen: 19.696 ng/ul
RT: 17.387 min Scan# 2414
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:178 Resp: 2613238
Ion Ratio Lower Upper
178 100
179 15.4 12.5 18.7
176 19.4 15.4 23.0



#74
Anthracene
Concen: 19.795 ng/ul
RT: 17.481 min Scan# 2430
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:178 Resp: 2622422
Ion Ratio Lower Upper
178 100
179 15.3 12.5 18.7
176 18.5 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 27.817 ng/uL
RT: 13.381 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 1010970
Ion Ratio Lower Upper
216 100
214 77.0 62.2 93.2
218 48.1 38.9 58.3



#76
Pentachlorobenzene
Concen: 24.879 ng/uL
RT: 14.904 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:250 Resp: 921884
Ion Ratio Lower Upper
250 100
248 62.3 50.6 76.0
252 62.7 51.1 76.7





#77

Carbazole

Concen: 19.550 ng/ul

RT: 17.751 min Scan# 2476

Delta R.T. 0.000 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:167 Resp: 2181218

Ion Ratio Lower Upper

167 100

166 21.2 17.2 25.8

139 12.2 9.4 14.2



#78

Di-n-butylphthalate

Concen: 18.875 ng/ul

RT: 18.286 min Scan# 2567

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20




Tgt Ion:149 Resp: 2564152


Ion Ratio Lower Upper

149 100

150 9.2 7.5 11.3

104 4.8 3.8 5.6





#80

Fluoranthene

Concen: 18.264 ng/ul

RT: 19.351 min Scan# 21

Delta R.T. 0.000 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:202 Resp: 2884831

Ion Ratio Lower Upper

202 100

101 9.6 8.2 12.2

100 7.4 6.1 9.1



#82

Pyrene

Concen: 18.722 ng/ul

RT: 19.704 min Scan# 2808

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20




Tgt Ion:202 Resp: 3050412


Ion Ratio Lower Upper

202 100

101 11.9 9.4 14.2

100 9.6 7.8 11.8





#83

Butylbenzylphthalate

Concen: 16.725 ng/ul

RT: 20.569 min Scan# 2956

Delta R.T. 0.000 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 1052776

Ion Ratio Lower Upper

149 100

91 68.7 53.4 80.0

206 22.8 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 17.330 ng/ul

RT: 21.345 min Scan# 3087

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20




Tgt Ion:252 Resp: 949183


Ion Ratio Lower Upper

252 100

254 64.1 51.8 77.8

126 9.6 7.4 11.2





#85

Benzo(a)anthracene

Concen: 20.072 ng/ul

RT: 21.416 min Scan# 3099

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 3230546

Ion Ratio Lower Upper

228 100

229 19.7 16.1 24.1

226 27.8 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.949 ng/ul

RT: 21.327 min Scan# 3084

Delta R.T. 0.000 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20




Tgt Ion:149 Resp: 1504891


Ion Ratio Lower Upper

149 100

167 28.5 22.5 33.7

279 5.0 3.8 5.8





#87

Chrysene

Concen: 19.785 ng/ul

RT: 21.475 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 3011463

Ion Ratio Lower Upper

228 100

226 30.3 24.5 36.7

229 19.9 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 17.375 ng/ul




RT: 22.275 min Scan# 3245


Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Tgt Ion:149 Resp: 2745164









#90
Benzo(b)fluoranthene
Concen: 19.021 ng/ul
RT: 23.157 min Scan# 31
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 3420674
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 7.7 6.8 10.2



#91
Benzo(k)fluoranthene
Concen: 18.519 ng/ul
RT: 23.204 min Scan# 3403
Delta R.T. 0.000 min
Lab File: BPCCC308.D
Acq: 24 Dec 2022 20:20

Tgt Ion:252 Resp: 3302576
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 8.2 6.8 10.2





#93

Benzo(a)pyrene

Concen: 19.333 ng/ul

RT: 23.804 min Scan# 3

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 3027374

Ion Ratio Lower Upper

252 100

253 21.8 17.5 26.3

125 9.4 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.717 ng/ul

RT: 26.510 min Scan# 3965

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20




Tgt Ion:276 Resp: 4040930


Ion Ratio Lower Upper

276 100

138 20.1 17.1 25.7

277 24.7 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 20.907 ng/ul

RT: 26.527 min Scan# 3970

Delta R.T. 0.000 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 3376402

Ion Ratio Lower Upper

278 100

139 13.1 11.1 16.7

279 24.3 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 21.081 ng/ul

RT: 27.304 min Scan# 4100

Delta R.T. -0.006 min

Lab File: BPCCC308.D

Acq: 24 Dec 2022 20:20

Tgt Ion:276 Resp: 3301063

Ion Ratio Lower Upper

276 100

138 18.2 15.4 23.2

277 23.4 19.1 28.7

