

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122422\
 Data File : BP013317.D
 Acq On : 25 Dec 2022 01:24
 Operator : CG/JU
 Sample : N6016-21
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DBYE6

Quant Time: Dec 25 04:22:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 00:19:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	568287	20.000	ng/u1	0.00
20) Naphthalene-d8	10.752	136	2117875	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.581	164	1092439	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.340	188	2297449	20.000	ng/u1	0.00
79) Chrysene-d12	21.433	240	2553167	20.000	ng/u1	0.00
88) Perylene-d12	23.910	264	2961660	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.329	96	52881	3.979	ng/uL	0.00
4) Pyridine-d5	3.758	84	714198	18.916	ng/u1	0.00
7) Phenol-d5	7.099	99	927994	20.048	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.264	67	618165	22.138	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	779513	21.392	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	663980	17.507	ng/u1	0.00
21) Nitrobenzene-d5	9.105	128	361536	23.234	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.828	143	395582	22.581	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.375	165	696346	20.465	ng/u1	0.00
31) 4-Chloroaniline-d4	10.893	131	841322	17.514	ng/u1	0.00
46) Dimethylphthalate-d6	13.987	166	1844256	21.966	ng/u1	-0.01
49) Acenaphthylene-d8	14.275	160	2115171	22.927	ng/u1	0.00
54) 4-Nitrophenol-d4	14.787	143	234753	15.544	ng/u1	0.00
60) Fluorene-d10	15.575	176	1567415	22.067	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	145542	9.844	ng/u1	0.00
73) Anthracene-d10	17.440	188	2374443	22.909	ng/u1	0.00
81) Pyrene-d10	19.675	212	3187206	21.748	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.751	264	3469008	23.499	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122422\
Data File : BP013317.D
Acq On : 25 Dec 2022 01:24
Operator : CG/JU
Sample : N6016-21
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DBYE6

Quant Time: Dec 25 04:22:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Dec 24 00:19:17 2022
Response via : Initial Calibration

