

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP043024\
 Data File : BP020126.D
 Acq On : 30 Apr 2024 19:36
 Operator : MA/JU
 Sample : P2220-01 10X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E0808

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/03/2024
 Supervised By :mohammad ahmed 05/03/2024

Quant Time: May 01 01:20:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 01 00:58:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	104129	20.000	ng/u1	0.00
20) Naphthalene-d8	10.410	136	438907	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.322	164	320807	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.163	188	720314	20.000	ng/u1	0.00
79) Chrysene-d12	21.651	240	569762	20.000	ng/u1	-0.02
88) Perylene-d12	25.027	264	556069	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	6577	2.638	ng/uL	0.00
4) Pyridine-d5	3.628	84	71610	9.749	ng/u1	0.01
7) Phenol-d5	6.822	99	81268	9.204	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	6.987	67	226481	39.816	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	187066	30.175	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	143838	21.397	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	126075	37.405	ng/u1	0.00
24) 2-Nitrophenol-d4	9.510	143	136303	36.439	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	227955	32.382	ng/u1	0.00
31) 4-Chloroaniline-d4	10.569	131	287708	31.341	ng/u1	0.01
46) Dimethylphthalate-d6	13.734	166	945096	41.247	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	1012782	38.780	ng/u1	0.00
54) 4-Nitrophenol-d4	14.610	143	29378m	7.888	ng/u1	0.04
60) Fluorene-d10	15.345	176	796899	39.983	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.498	200	155590	33.782	ng/u1	0.00
73) Anthracene-d10	17.275	188	1333630	41.700	ng/u1	0.01
81) Pyrene-d10	19.663	212	1521218	50.185	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.798	264	1166379	42.989	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP043024\
Data File : BP020126.D
Acq On : 30 Apr 2024 19:36
Operator : MA/JU
Sample : P2220-01 10X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E0808

Quant Time: May 01 01:20:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 01 00:58:38 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/03/2024
Supervised By :mohammad ahmed 05/03/2024

