

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP043024\
 Data File : BP020125.D
 Acq On : 30 Apr 2024 18:51
 Operator : MA/JU
 Sample : PB160589TB 10X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLEB589

Quant Time: May 01 01:13:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 01 00:58:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	76616	20.000	ng/u1	0.00
20) Naphthalene-d8	10.405	136	322137	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.310	164	241031	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.163	188	567506	20.000	ng/u1	0.00
79) Chrysene-d12	21.657	240	599934	20.000	ng/u1	-0.01
88) Perylene-d12	25.033	264	638275	20.000	ng/u1	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.223	96	13181	7.184	ng/uL	0.00
4) Pyridine-d5	3.623	84	182022	33.678	ng/u1	0.00
7) Phenol-d5	6.817	99	228507	35.172	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.987	67	159080	38.010	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	171641	37.629	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	185948	37.594	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	88046	35.591	ng/u1	0.00
24) 2-Nitrophenol-d4	9.511	143	102267	37.250	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	173618	33.604	ng/u1	0.00
31) 4-Chloroaniline-d4	10.569	131	244109	36.230	ng/u1	0.01
46) Dimethylphthalate-d6	13.728	166	709558	41.217	ng/u1	0.00
49) Acenaphthylene-d8	13.999	160	725584	36.979	ng/u1	0.00
54) 4-Nitrophenol-d4	14.569	143	99456	35.541	ng/u1	0.00
60) Fluorene-d10	15.340	176	572278	38.217	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.498	200	113801	31.362	ng/u1	0.00
73) Anthracene-d10	17.263	188	968247	38.427	ng/u1	0.00
81) Pyrene-d10	19.663	212	1232033	38.600	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.816	264	1185374	38.062	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP043024\
Data File : BP020125.D
Acq On : 30 Apr 2024 18:51
Operator : MA/JU
Sample : PB160589TB 10X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SLEB589

Quant Time: May 01 01:13:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 01 00:58:38 2024
Response via : Initial Calibration

