

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092121\
 Data File : BP007082.D
 Acq On : 21 Sep 2021 17:21
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP092121

Quant Time: Sep 21 17:59:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 16:50:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	0.00
2	1,4-Dioxane	40.000	41.021	-2.6	116	0.00
3	Pyridine	40.000	41.533	-3.8	113	0.00
4	n-Nitrosodimethylamine	40.000	41.940	-4.8	115	0.00
5 S	2-Fluorophenol	80.000	79.561	0.5	111	-0.01
6	Aniline	40.000	39.543	1.1	107	-0.01
7 S	Phenol-d6	80.000	79.015	1.2	107	0.00
8	2-Chlorophenol	40.000	39.331	1.7	109	-0.01
9	Benzaldehyde	40.000	40.941	-2.4	109	-0.01
10 C	Phenol	40.000	39.211	2.0	107	-0.01
11	bis(2-Chloroethyl)ether	40.000	39.140	2.1	107	0.00
12	1,3-Dichlorobenzene	40.000	38.725	3.2	108	0.00
13 C	1,4-Dichlorobenzene	40.000	38.737	3.2	108	0.00
14	1,2-Dichlorobenzene	40.000	38.685	3.3	107	0.00
15	Benzyl Alcohol	40.000	39.609	1.0	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.903	2.7	107	0.00
17	2-Methylphenol	40.000	39.073	2.3	106	0.00
18	Hexachloroethane	40.000	38.480	3.8	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.992	2.5	106	0.00
20	3+4-Methylphenols	40.000	39.370	1.6	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	39.058	2.4	106	0.00
23 S	Nitrobenzene-d5	80.000	79.330	0.8	106	-0.01
24	Nitrobenzene	40.000	39.558	1.1	105	-0.01
25	Isophorone	40.000	39.499	1.3	105	-0.01
26 C	2-Nitrophenol	40.000	40.180	-0.4	106	0.00
27	2,4-Dimethylphenol	40.000	39.274	1.8	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.851	2.9	105	0.00
29 C	2,4-Dichlorophenol	40.000	39.511	1.2	105	0.00
30	1,2,4-Trichlorobenzene	40.000	38.636	3.4	105	0.00
31	Naphthalene	40.000	38.540	3.7	104	0.00
32	Benzoic acid	40.000	43.390	-8.5	110	0.00
33	4-Chloroaniline	40.000	38.817	3.0	104	0.00
34 C	Hexachlorobutadiene	40.000	38.859	2.9	106	0.00
35	Caprolactam	40.000	40.563	-1.4	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.245	1.9	105	0.00
37	2-Methylnaphthalene	40.000	38.530	3.7	105	0.00
38	1-Methylnaphthalene	40.000	38.404	4.0	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.693	3.3	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.662	0.8	104	0.00
42 S	2,4,6-Tribromophenol	80.000	79.962	0.0	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.914	0.2	105	0.00
44	2,4,5-Trichlorophenol	40.000	39.410	1.5	103	0.00
45 S	2-Fluorobiphenyl	80.000	77.023	3.7	104	0.00
46	1,1'-Biphenyl	40.000	38.541	3.6	104	0.00
47	2-Chloronaphthalene	40.000	38.666	3.3	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092121\
 Data File : BP007082.D
 Acq On : 21 Sep 2021 17:21
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP092121

Quant Time: Sep 21 17:59:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 16:50:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.994	-2.5	105	0.00
49	Acenaphthylene	40.000	39.150	2.1	103	0.00
50	Dimethylphthalate	40.000	38.634	3.4	104	0.00
51	2,6-Dinitrotoluene	40.000	40.247	-0.6	105	0.00
52 C	Acenaphthene	40.000	38.784	3.0	103	0.00
53	3-Nitroaniline	40.000	40.583	-1.5	104	0.00
54 P	2,4-Dinitrophenol	40.000	43.194	-8.0	107	0.00
55	Dibenzofuran	40.000	38.679	3.3	104	0.00
56 P	4-Nitrophenol	40.000	42.221	-5.6	104	0.00
57	2,4-Dinitrotoluene	40.000	41.140	-2.9	105	0.00
58	Fluorene	40.000	38.886	2.8	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.643	0.9	105	0.00
60	Diethylphthalate	40.000	38.820	2.9	104	0.00
61	4-Chlorophenyl-phenylether	40.000	38.766	3.1	104	0.00
62	4-Nitroaniline	40.000	41.763	-4.4	105	0.00
63	Azobenzene	40.000	39.506	1.2	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.464	-6.2	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.998	2.5	104	0.00
67	4-Bromophenyl-phenylether	40.000	39.005	2.5	105	0.00
68	Hexachlorobenzene	40.000	38.668	3.3	104	0.00
69	Atrazine	40.000	39.635	0.9	103	0.00
70 C	Pentachlorophenol	40.000	40.898	-2.2	105	0.00
71	Phenanthrene	40.000	38.935	2.7	104	0.00
72	Anthracene	40.000	39.203	2.0	104	0.00
73	Carbazole	40.000	39.524	1.2	104	0.00
74	Di-n-butylphthalate	40.000	39.840	0.4	104	0.00
75 C	Fluoranthene	40.000	39.553	1.1	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	42.349	-5.9	111	0.00
78	Pyrene	40.000	38.704	3.2	104	0.00
79 S	Terphenyl-d14	80.000	78.011	2.5	104	0.00
80	Butylbenzylphthalate	40.000	39.942	0.1	104	0.00
81	Benzo(a)anthracene	40.000	38.803	3.0	104	0.00
82	3,3'-Dichlorobenzidine	40.000	39.526	1.2	103	0.00
83	Chrysene	40.000	38.715	3.2	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.206	-0.5	105	0.00
85 c	Di-n-octyl phthalate	40.000	39.762	0.6	104	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	39.202	2.0	104	0.01
88	Benzo(b)fluoranthene	40.000	39.447	1.4	104	0.00
89	Benzo(k)fluoranthene	40.000	38.907	2.7	104	0.00
90 C	Benzo(a)pyrene	40.000	39.139	2.2	104	0.00
91	Dibenzo(a,h)anthracene	40.000	39.150	2.1	104	0.01
92	Benzo(g,h,i)perylene	40.000	38.946	2.6	103	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092121\
Data File : BP007082.D
Acq On : 21 Sep 2021 17:21
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP092121

Quant Time: Sep 21 17:59:42 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 21 16:50:24 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0