



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092121\
 Data File : BP007073.D
 Acq On : 21 Sep 2021 09:06
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Sep 21 16:50:24 2021






AutoFind: Scans 2488, 2489, 2490; Background Corrected with Scan 2478

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.1	157659	PASS
68	69	0.00	2	1.9	3314	PASS
69	198	0.00	100	27.4	172192	PASS
70	69	0.00	2	0.5	810	PASS
127	198	10	80	40.6	255061	PASS
197	198	0.00	2	0.7	4414	PASS
198	198	100	100	100.0	628273	PASS
199	198	5	9	6.8	42883	PASS
275	198	10	60	27.4	172267	PASS
365	198	1	100	3.4	21542	PASS
441	198	0.01	100	14.0	88224	PASS
442	442	100	100	100.0	572203	PASS
443	442	15	24	19.4	110868	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092121\
 Data File : BP007073.D
 Acq On : 21 Sep 2021 09:06
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Sep 21 10:50:45 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 17 10:50:16 2021
 Response via : Initial Calibration



TIC: BP007073.D\data.ms

(70) Pentachlorophenol (C)

16.934min (+ 0.000) 198479.55 ng




response 779134

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	64.68
264.00	62.90	62.81
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092121\
 Data File : BP007073.D
 Acq On : 21 Sep 2021 09:06
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Sep 21 10:50:45 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 17 10:50:16 2021
 Response via : Initial Calibration



TIC: BP007073.D\data.ms

(77) Benzidine

19.469min (+ 0.000) -20.00 ng

response 4091161

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	14.60
183.00	11.90	12.14
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/21/2021	BNA_P	<u>BP007073.D</u>
Compound Name	Response	Retention Time
DDT	2160802	20.67
DDD	26283	20.24
DDE	3104	19.75
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
29387	2190189	1.34