


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092822\
 Data File : BP011828.D
 Acq On : 28 Sep 2022 13:20
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092122.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Sep 28 02:03:33 2022






AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2499

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.0	182796	PASS
68	69	0.00	2	1.5	2927	PASS
69	198	0.00	100	34.0	194206	PASS
70	69	0.00	2	0.6	1073	PASS
127	198	10	80	48.5	277091	PASS
197	198	0.00	2	0.3	1926	PASS
198	198	100	100	100.0	571876	PASS
199	198	5	9	6.7	38581	PASS
275	198	10	60	24.2	138539	PASS
365	198	1	100	2.6	14652	PASS
441	198	0.01	100	8.4	48117	PASS
442	442	100	100	100.0	307051	PASS
443	442	15	24	19.4	59645	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092822\
Data File : BP011828.D
Acq On : 28 Sep 2022 13:20
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Sep 29 05:34:09 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 02:03:33 2022
Response via : Initial Calibration



TIC: BP011828.D\data.ms

(70) Pentachlorophenol (C)

16.951min (-0.006) 261790.70 ng




response 691025

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.10	64.25
264.00	63.20	64.06
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092822\
 Data File : BP011828.D
 Acq On : 28 Sep 2022 13:20
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Sep 29 05:34:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:03:33 2022
 Response via : Initial Calibration



TIC: BP011828.D\data.ms

(77) Benzidine

19.492min (-0.006) 4195346.73 ng

response 4315443

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.60	14.50
183.00	11.60	11.68
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/28/2022	BNA_P	<u>BP011828.D</u>
Compound Name	Response	Retention Time
DDT	1997127	20.692
DDD	15857	20.316
DDE	2407	19.769
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
18264	2015391	0.91

Instrument :
BNA_P
ClientSampleId :
DFTPP