




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010424\
Data File : BP019270.D
Acq On : 04 Jan 2024 18:48
Operator : MA/JU
Sample : 05951-18
Misc :
ALS Vial : 14 Sample Multiplier: 1

Quant Time: Jan 04 23:54:28 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 29 03:48:11 2023
Response via : Initial Calibration



TIC: BP019270.D\data.ms

(54) 4-Nitrophenol-d4 (S)




14.639min (-0.024) 0.04 ng/ul

response 212

Ion	Exp%	Act%
143.00	100.00	100.00
113.00	73.50	17.41#
41.00	33.40	698.12#
42.00	23.00	100.71#

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010424\
Data File : BP019270.D
Acq On : 04 Jan 2024 18:48
Operator : MA/JU
Sample : 05951-18
Misc :
ALS Vial : 14 Sample Multiplier: 1

Quant Time: Jan 04 23:54:28 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 29 03:48:11 2023
Response via : Initial Calibration



TIC: BP019270.D\data.ms

(54) 4-Nitrophenol-d4 (S)**14.692min (+ 0.029) 15.86 ng/ul m****response 94578**

Ion	Exp%	Act%
143.00	100.00	100.00
113.00	73.50	73.58
41.00	33.40	58.13#
42.00	23.00	25.23

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010424\
 Data File : BP019270.D
 Acq On : 04 Jan 2024 18:48
 Operator : MA/JU
 Sample : 05951-18
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Quant Time: Jan 05 00:21:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 03:48:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.798	152	167239	20.000	ng/ul	0.00
20) Naphthalene-d8	10.592	136	627941	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.439	164	399899	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.198	188	955107	20.000	ng/ul	0.00
79) Chrysene-d12	21.304	240	992895	20.000	ng/ul	0.00
88) Perylene-d12	23.674	264	1175508	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	14008	2.861	ng/uL	0.00
4) Pyridine-d5	3.657	84	100272	8.158	ng/ul	0.00
7) Phenol-d5	6.987	99	256727	17.118	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	146173	16.847	ng/ul	0.00
11) 2-Chlorophenol-d4	7.334	132	205561	17.002	ng/ul	0.00
15) 4-Methylphenol-d8	8.528	113	183482	15.714	ng/ul	0.00
21) Nitrobenzene-d5	8.963	128	98854	18.211	ng/ul	0.00
24) 2-Nitrophenol-d4	9.687	143	114497	19.517	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	212129	17.907	ng/ul	0.00
31) 4-Chloroaniline-d4	10.745	131	78104	5.360	ng/ul	0.00
46) Dimethylphthalate-d6	13.863	166	616192	17.445	ng/ul	0.00
49) Acenaphthylene-d8	14.133	160	718875	18.757	ng/ul	0.00
54) 4-Nitrophenol-d4	14.692	143	94578m	15.856	ng/ul	0.03
60) Fluorene-d10	15.439	176	559927	18.349	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.574	200	65517	10.464	ng/ul	0.00
73) Anthracene-d10	17.298	188	854649	17.135	ng/ul	0.00
81) Pyrene-d10	19.545	212	1145319	18.927	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.521	264	1176680	17.415	ng/ul	0.00
Target Compounds						
					Qvalue	
6) Benzaldehyde	6.945	77	23741	3.409	ng/ul	96
71) Pentachlorophenol	16.857	266	68845	8.083	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010424\
Data File : BP019270.D
Acq On : 04 Jan 2024 18:48
Operator : MA/JU
Sample : 05951-18
Misc :
ALS Vial : 14 Sample Multiplier: 1

Quant Time: Jan 05 00:21:32 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 29 03:48:11 2023
Response via : Initial Calibration

