

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010622\
 Data File : BP008704.D
 Acq On : 06 Jan 2022 11:59
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Quant Time: Jan 06 16:18:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 06 16:16:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	430972	20.000	ng	0.00
21) Naphthalene-d8	10.681	136	1832294	20.000	ng	0.00
39) Acenaphthene-d10	14.516	164	1362006	20.000	ng	0.00
64) Phenanthrene-d10	17.263	188	2995807	20.000	ng	0.00
76) Chrysene-d12	21.345	240	3017529	20.000	ng	-0.01
86) Perylene-d12	23.769	264	3286549	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.464	112	355966	15.076	ng	0.00
7) Phenol-d6	7.052	99	447881	13.628	ng	0.00
23) Nitrobenzene-d5	9.034	82	400726	13.010	ng	0.00
42) 2,4,6-Tribromophenol	16.004	330	267520	14.518	ng	0.00
45) 2-Fluorobiphenyl	13.140	172	1273037	13.542	ng	0.00
79) Terphenyl-d14	19.816	244	2285015	15.461	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.364	88	71791	7.526	ng	# 95
3) Pyridine	3.770	79	143024	5.497	ng	# 95
4) n-Nitrosodimethylamine	3.676	42	68281	8.127	ng	# 80
6) Aniline	7.205	93	282504	7.396	ng	99
8) 2-Chlorophenol	7.446	128	192035	6.904	ng	99
9) Benzaldehyde	7.017	77	156633	8.374	ng	99
10) Phenol	7.075	94	230162	6.937	ng	96
11) bis(2-Chloroethyl)ether	7.305	93	183862	6.871	ng	100
12) 1,3-Dichlorobenzene	7.769	146	227276	6.981	ng	98
13) 1,4-Dichlorobenzene	7.916	146	230822	6.970	ng	99
14) 1,2-Dichlorobenzene	8.234	146	221183	6.837	ng	100
15) Benzyl Alcohol	8.122	79	163472	6.944	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.416	45	208906	6.861	ng	99
17) 2-Methylphenol	8.328	107	157132	6.806	ng	97
18) Hexachloroethane	8.963	117	77344	7.164	ng	93
19) n-Nitroso-di-n-propyla...	8.687	70	148124	7.058	ng	95
20) 3+4-Methylphenols	8.658	107	213885	6.594	ng	99
22) Acetophenone	8.705	105	304529	6.931	ng	# 97
24) Nitrobenzene	9.081	77	202022	6.887	ng	99
25) Isophorone	9.611	82	402652	6.998	ng	99
26) 2-Nitrophenol	9.793	139	99787	7.162	ng	93
27) 2,4-Dimethylphenol	9.858	122	192546	7.659	ng	99
28) bis(2-Chloroethoxy)met...	10.093	93	258118	6.479	ng	99
29) 2,4-Dichlorophenol	10.334	162	201533	6.623	ng	98
30) 1,2,4-Trichlorobenzene	10.546	180	228668	6.373	ng	98
31) Naphthalene	10.734	128	641678	6.747	ng	99
32) Benzoic acid	9.952	122	108144	6.810	ng	99
33) 4-Chloroaniline	10.840	127	279651	6.857	ng	99
34) Hexachlorobutadiene	11.028	225	140645	6.457	ng	98
35) Caprolactam	11.610	113	69322	6.931	ng	89
36) 4-Chloro-3-methylphenol	11.969	107	204803	6.733	ng	99
37) 2-Methylnaphthalene	12.340	142	500232	6.988	ng	99
38) 1-Methylnaphthalene	12.557	142	466324	6.663	ng	97
40) 1,2,4,5-Tetrachloroben...	12.710	216	283079	6.597	ng	99
41) Hexachlorocyclopentadiene	12.693	237	156847	6.874	ng	99
43) 2,4,6-Trichlorophenol	12.951	196	179647	6.435	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010622\
 Data File : BP008704.D
 Acq On : 06 Jan 2022 11:59
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Quant Time: Jan 06 16:18:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 06 16:16:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.022	196	197832	6.418	ng	97
46) 1,1'-Biphenyl	13.351	154	682870	6.778	ng	99
47) 2-Chloronaphthalene	13.387	162	515478	6.529	ng	98
48) 2-Nitroaniline	13.587	65	115147	6.653	ng	94
49) Acenaphthylene	14.234	152	805980	6.646	ng	100
50) Dimethylphthalate	13.969	163	690399	6.581	ng	100
51) 2,6-Dinitrotoluene	14.081	165	139377	6.475	ng	90
52) Acenaphthene	14.575	154	498441	6.282	ng	99
53) 3-Nitroaniline	14.410	138	139560	6.229	ng	95
54) 2,4-Dinitrophenol	14.622	184	57703	6.019	ng	# 88
55) Dibenzofuran	14.910	168	831770	6.539	ng	99
56) 4-Nitrophenol	14.728	139	112084	6.353	ng	92
57) 2,4-Dinitrotoluene	14.869	165	183881	6.402	ng	# 87
58) Fluorene	15.563	166	673825	6.798	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.140	232	178147	6.451	ng	# 87
60) Diethylphthalate	15.334	149	678479	6.617	ng	98
61) 4-Chlorophenyl-phenyle...	15.557	204	358871	6.499	ng	98
62) 4-Nitroaniline	15.569	138	144516	6.351	ng	91
63) Azobenzene	15.845	77	536914	6.488	ng	93
65) 4,6-Dinitro-2-methylph...	15.640	198	84525	5.611	ng	95
66) n-Nitrosodiphenylamine	15.769	169	596966	6.603	ng	99
67) 4-Bromophenyl-phenylether	16.451	248	222041	6.774	ng	96
68) Hexachlorobenzene	16.575	284	258511	6.424	ng	99
69) Atrazine	16.722	200	227529	6.728	ng	99
70) Pentachlorophenol	16.916	266	142624	6.508	ng	99
71) Phenanthrene	17.304	178	1083774	6.710	ng	97
72) Anthracene	17.398	178	1091101	6.933	ng	98
73) Carbazole	17.663	167	991762	6.421	ng	98
74) Di-n-butylphthalate	18.222	149	1157500	6.849	ng	98
75) Fluoranthene	19.269	202	1292895	6.959	ng	96
77) Benzidine	19.445	184	784514	8.654	ng	98
78) Pyrene	19.622	202	1342745	6.840	ng	98
80) Butylbenzylphthalate	20.498	149	510565	6.521	ng	96
81) Benzo(a)anthracene	21.333	228	1312602	6.705	ng	96
82) 3,3'-Dichlorobenzidine	21.263	252	560075	7.781	ng	99
83) Chrysene	21.380	228	1262751	6.898	ng	96
84) Bis(2-ethylhexyl)phtha...	21.263	149	780791	6.601	ng	99
85) Di-n-octyl phthalate	22.192	149	1360846	7.433	ng	99
87) Indeno(1,2,3-cd)pyrene	26.286	276	1636971	7.415	ng	98
88) Benzo(b)fluoranthene	23.027	252	1406454	6.654	ng	99
89) Benzo(k)fluoranthene	23.075	252	1350338	6.427	ng	99
90) Benzo(a)pyrene	23.657	252	1344071	7.888	ng	99
91) Dibenzo(a,h)anthracene	26.304	278	1403023	7.534	ng	98
92) Benzo(g,h,i)perylene	27.063	276	1372348	7.881	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 06 16:18:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 06 16:16:23 2022
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.881 min Scan# 81
 Delta R.T. 0.000 min
 Lab File: BP008704.D
 Acq: 06 Jan 2022 11:59




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:152 Resp: 430972
 Ion Ratio Lower Upper
 152 100
 150 154.2 127.0 190.6
 115 53.6 43.7 65.5



#2
 1,4-Dioxane
 Concen: 7.526 ng
 RT: 3.364 min Scan# 64
 Delta R.T. 0.000 min
 Lab File: BP008704.D
 Acq: 06 Jan 2022 11:59

Tgt Ion: 88 Resp: 71791
 Ion Ratio Lower Upper
 88 100
 58 65.5 50.6 75.8
 43 28.6 18.2 27.4#









#3
Pyridine
Concen: 5.497 ng
RT: 3.770 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 79 Resp: 143024
Ion Ratio Lower Upper
79 100
52 57.0 45.1 67.7
51 36.3 23.1 34.7#



#4
n-Nitrosodimethylamine
Concen: 8.127 ng
RT: 3.676 min Scan# 117
Delta R.T. 0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion: 42 Resp: 68281
Ion Ratio Lower Upper
42 100
74 137.0 131.1 196.7
44 16.8 6.2 9.2#









#5
2-Fluorophenol
Concen: 15.076 ng
RT: 5.464 min Scan# 41
Delta R.T. 0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:112 Resp: 355966
Ion Ratio Lower Upper
112 100
64 47.7 39.7 59.5
63 25.9 20.0 30.0



#6
Aniline
Concen: 7.396 ng
RT: 7.205 min Scan# 717
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion: 93 Resp: 282504
Ion Ratio Lower Upper
93 100
66 34.9 27.8 41.6
65 18.2 14.0 21.0









#7
Phenol-d6
Concen: 13.628 ng
RT: 7.052 min Scan# 691
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 99 Resp: 447881
Ion Ratio Lower Upper
99 100
42 16.0 10.8 16.2
71 31.5 23.8 35.6



#8
2-Chlorophenol
Concen: 6.904 ng
RT: 7.446 min Scan# 758
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion: 128 Resp: 192035
Ion Ratio Lower Upper
128 100
130 32.5 12.6 52.6
64 39.9 18.6 58.6









#9
Benzaldehyde
Concen: 8.374 ng
RT: 7.017 min Scan# 685
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 77 Resp: 156633
Ion Ratio Lower Upper
77 100
105 104.0 84.2 124.2
106 101.7 80.6 120.6



#10
Phenol
Concen: 6.937 ng
RT: 7.075 min Scan# 695
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion: 94 Resp: 230162
Ion Ratio Lower Upper
94 100
65 26.5 4.1 44.1
66 34.3 12.4 52.4









#11
bis(2-Chloroethyl)ether
Concen: 6.871 ng
RT: 7.305 min Scan# 734
Delta R.T. 0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 93 Resp: 183862
Ion Ratio Lower Upper
93 100
63 66.1 46.0 86.0
95 31.6 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 6.981 ng
RT: 7.769 min Scan# 813
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion: 146 Resp: 227276
Ion Ratio Lower Upper
146 100
148 63.7 51.9 77.9
75 26.9 20.5 30.7









#13
1,4-Dichlorobenzene
Concen: 6.970 ng
RT: 7.916 min Scan# 838
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:146 Resp: 230822
Ion Ratio Lower Upper
146 100
148 64.4 52.3 78.5
111 37.9 29.8 44.8



#14
1,2-Dichlorobenzene
Concen: 6.837 ng
RT: 8.234 min Scan# 892
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:146 Resp: 221183
Ion Ratio Lower Upper
146 100
148 64.8 52.2 78.2
111 38.9 31.1 46.7









#15
Benzyl Alcohol
Concen: 6.944 ng
RT: 8.122 min Scan# 81
Delta R.T. 0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 79 Resp: 163472
Ion Ratio Lower Upper
79 100
108 85.1 67.8 101.8
77 65.1 50.7 76.1



#16
2,2'-oxybis(1-Chloropropane)
Concen: 6.861 ng
RT: 8.416 min Scan# 923
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion: 45 Resp: 208906
Ion Ratio Lower Upper
45 100
77 17.6 0.0 37.4
79 13.8 0.0 34.1









#17
2-Methylphenol
Concen: 6.806 ng
RT: 8.328 min Scan# 908
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	107	Resp:	157132
Ion	Ratio	Lower	Upper
107	100		
108	112.5	90.0	135.0
77	45.0	32.7	49.1
79	43.7	32.1	48.1



#18
Hexachloroethane
Concen: 7.164 ng
RT: 8.963 min Scan# 1016
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:	117	Resp:	77344
Ion	Ratio	Lower	Upper
117	100		
119	92.1	76.9	115.3
201	106.4	93.7	140.5









#19
n-Nitroso-di-n-propylamine
Concen: 7.058 ng
RT: 8.687 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	70	Resp:	148124
Ion	Ratio	Lower	Upper
70	100		
42	51.0	38.2	57.2
101	11.4	8.7	13.1
130	23.6	21.0	31.6



#20
3+4-Methylphenols
Concen: 6.594 ng
RT: 8.658 min Scan# 964
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:	107	Resp:	213885
Ion	Ratio	Lower	Upper
107	100		
108	90.7	69.8	109.8
77	30.8	10.4	50.4
79	27.7	6.4	46.4









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.681 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:136 Resp: 1832294
Ion Ratio Lower Upper
136 100
137 11.0 9.0 13.4
54 6.1 4.9 7.3
68 5.3 4.3 6.5



#22
Acetophenone
Concen: 6.931 ng
RT: 8.705 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:105 Resp: 304529
Ion Ratio Lower Upper
105 100
71 2.5 5.3 7.9#
51 22.2 16.3 24.5
120 24.2 19.2 28.8









#23
Nitrobenzene-d5
Concen: 13.010 ng
RT: 9.034 min Scan# 1028
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 82 Resp: 400726
Ion Ratio Lower Upper
82 100
128 46.3 39.4 59.2
54 43.2 32.6 48.8



#24
Nitrobenzene
Concen: 6.887 ng
RT: 9.081 min Scan# 1036
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion: 77 Resp: 202022
Ion Ratio Lower Upper
77 100
123 51.3 41.4 62.2
65 13.0 9.8 14.8









#25
Isophorone
Concen: 6.998 ng
RT: 9.611 min Scan# 1126
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 82 Resp: 402652
Ion Ratio Lower Upper
82 100
95 7.2 5.4 8.2
138 20.1 16.6 24.8



#26
2-Nitrophenol
Concen: 7.162 ng
RT: 9.793 min Scan# 1157
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion: 139 Resp: 99787
Ion Ratio Lower Upper
139 100
109 28.6 20.7 31.1
65 44.1 31.6 47.4





#27

2,4-Dimethylphenol

Concen: 7.659 ng

RT: 9.858 min Scan# 1168

Delta R.T. 0.000 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC005





Tgt Ion:122 Resp: 192546

Ion Ratio Lower Upper

122 100

107 110.8 89.6 134.4

121 58.1 47.8 71.8



#28

bis(2-Chloroethoxy)methane

Concen: 6.479 ng

RT: 10.093 min Scan# 1208

Delta R.T. 0.000 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59




Tgt Ion: 93 Resp: 258118


Ion Ratio Lower Upper

93 100

95 31.9 25.8 38.8

123 12.7 9.7 14.5









#29
2,4-Dichlorophenol
Concen: 6.623 ng
RT: 10.334 min Scan# 1249
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:162 Resp: 201533
Ion Ratio Lower Upper
162 100
164 63.7 44.9 84.9
98 29.9 11.2 51.2



#30
1,2,4-Trichlorobenzene
Concen: 6.373 ng
RT: 10.546 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:180 Resp: 228668
Ion Ratio Lower Upper
180 100
182 97.9 76.7 115.1
145 28.8 22.4 33.6









#31
Naphthalene
Concen: 6.747 ng
RT: 10.734 min Scan# 1117
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:128 Resp: 641678
Ion Ratio Lower Upper
128 100
129 10.8 8.8 13.2
127 12.7 10.4 15.6



#32
Benzoic acid
Concen: 6.810 ng
RT: 9.952 min Scan# 1184
Delta R.T. -0.059 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:122 Resp: 108144
Ion Ratio Lower Upper
122 100
105 118.9 100.1 140.1
77 87.4 65.5 105.5









#33
4-Chloroaniline
Concen: 6.857 ng
RT: 10.840 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	127	Resp:	279651
Ion	Ratio	Lower	Upper
127	100		
129	32.8	26.0	39.0
65	26.0	20.5	30.7
92	17.5	14.1	21.1



#34
Hexachlorobutadiene
Concen: 6.457 ng
RT: 11.028 min Scan# 1367
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:	225	Resp:	140645
Ion	Ratio	Lower	Upper
225	100		
223	60.2	50.1	75.1
227	63.3	51.5	77.3









#35
Caprolactam
Concen: 6.931 ng
RT: 11.610 min Scan# 1466
Delta R.T. -0.023 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:113 Resp: 69322
Ion Ratio Lower Upper
113 100
55 140.0 107.1 147.1
56 105.3 74.1 114.1



#36
4-Chloro-3-methylphenol
Concen: 6.733 ng
RT: 11.969 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:107 Resp: 204803
Ion Ratio Lower Upper
107 100
144 28.4 23.2 34.8
142 86.7 70.1 105.1









#37
2-Methylnaphthalene
Concen: 6.988 ng
RT: 12.340 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:142 Resp: 500232
Ion Ratio Lower Upper
142 100
141 88.2 70.5 105.7
115 32.7 24.6 36.8



#38
1-Methylnaphthalene
Concen: 6.663 ng
RT: 12.557 min Scan# 1627
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:142 Resp: 466324
Ion Ratio Lower Upper
142 100
141 89.0 73.2 109.8
116 3.7 2.7 4.1









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.516 min Scan# 1960
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:164 Resp: 1362006
Ion Ratio Lower Upper
164 100
162 100.4 79.9 119.9
160 44.6 35.5 53.3



#40
1,2,4,5-Tetrachlorobenzene
Concen: 6.597 ng
RT: 12.710 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:216 Resp: 283079
Ion Ratio Lower Upper
216 100
214 78.4 63.1 94.7
179 19.6 16.0 24.0
108 15.6 11.9 17.9









#41
Hexachlorocyclopentadiene
Concen: 6.874 ng
RT: 12.693 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:237 Resp: 156847
Ion Ratio Lower Upper
237 100
235 63.3 42.9 82.9
272 13.1 0.0 33.8



#42
2,4,6-Tribromophenol
Concen: 14.518 ng
RT: 16.004 min Scan# 2213
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:330 Resp: 267520
Ion Ratio Lower Upper
330 100
332 96.7 77.0 115.4
141 29.1 21.3 31.9





#43

2,4,6-Trichlorophenol

Concen: 6.435 ng

RT: 12.951 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC005





Tgt Ion:196 Resp: 179647

Ion Ratio Lower Upper

196 100

198 100.4 77.8 116.8

200 32.3 24.7 37.1



#44

2,4,5-Trichlorophenol

Concen: 6.418 ng

RT: 13.022 min Scan# 1706

Delta R.T. 0.000 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59

Tgt Ion:196 Resp: 197832




Ion Ratio Lower Upper


196 100

198 98.0 77.7 116.5

97 41.8 29.6 44.4

132 23.9 18.7 28.1









#45
2-Fluorobiphenyl
Concen: 13.542 ng
RT: 13.140 min Scan# 1726
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:172 Resp: 1273037
Ion Ratio Lower Upper
172 100
171 34.5 29.3 43.9
170 23.1 19.5 29.3



#46
1,1'-Biphenyl
Concen: 6.778 ng
RT: 13.351 min Scan# 1762
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:154 Resp: 682870
Ion Ratio Lower Upper
154 100
153 39.6 20.3 60.3
76 10.6 0.0 31.1









#47
2-Chloronaphthalene
Concen: 6.529 ng
RT: 13.387 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:162 Resp: 515478
Ion Ratio Lower Upper
162 100
127 35.0 28.4 42.6
164 31.9 26.8 40.2



#48
2-Nitroaniline
Concen: 6.653 ng
RT: 13.587 min Scan# 1802
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion: 65 Resp: 115147
Ion Ratio Lower Upper
65 100
92 74.3 62.2 93.4
138 123.8 105.8 158.8





#49

Acenaphthylene

Concen: 6.646 ng

RT: 14.234 min Scan# 1912

Delta R.T. 0.000 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC005





Tgt Ion:152 Resp: 805980

Ion Ratio Lower Upper

152 100

151 19.8 16.1 24.1

153 13.4 10.6 16.0



#50

Dimethylphthalate

Concen: 6.581 ng

RT: 13.969 min Scan# 1867

Delta R.T. -0.006 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59




Tgt Ion:163 Resp: 690399


Ion Ratio Lower Upper

163 100

194 4.4 3.8 5.6

164 10.3 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 6.475 ng
RT: 14.081 min Scan# 1887
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:165 Resp: 139377
Ion Ratio Lower Upper
165 100
63 45.4 31.8 47.8
89 52.6 36.1 54.1



#52
Acenaphthene
Concen: 6.282 ng
RT: 14.575 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:154 Resp: 498441
Ion Ratio Lower Upper
154 100
153 113.0 89.4 134.2
152 53.8 42.6 63.8









#53
3-Nitroaniline
Concen: 6.229 ng
RT: 14.410 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:138 Resp: 139560
Ion Ratio Lower Upper
138 100
108 10.5 8.0 12.0
92 107.9 81.7 122.5



#54
2,4-Dinitrophenol
Concen: 6.019 ng
RT: 14.622 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:184 Resp: 57703
Ion Ratio Lower Upper
184 100
63 55.6 35.6 53.4#
154 60.1 43.4 65.0









#55
Dibenzofuran
Concen: 6.539 ng
RT: 14.910 min Scan# 2028
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:168 Resp: 831770
Ion Ratio Lower Upper
168 100
139 34.6 27.4 41.0
169 13.3 10.9 16.3



#56
4-Nitrophenol
Concen: 6.353 ng
RT: 14.728 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:139 Resp: 112084
Ion Ratio Lower Upper
139 100
109 70.3 42.2 82.2
65 82.9 57.3 97.3









#57
2,4-Dinitrotoluene
Concen: 6.402 ng
RT: 14.869 min Scan# 2021
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:165 Resp: 183881
Ion Ratio Lower Upper
165 100
63 36.9 23.3 34.9#
89 65.0 44.2 66.4
182 3.0 2.6 4.0



#58
Fluorene
Concen: 6.798 ng
RT: 15.563 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:166 Resp: 673825
Ion Ratio Lower Upper
166 100
165 97.5 77.2 115.8
167 13.5 11.2 16.8









#59
2,3,4,6-Tetrachlorophenol
Concen: 6.451 ng
RT: 15.140 min Scan# 2067
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:232 Resp: 178147
Ion Ratio Lower Upper
232 100
131 38.9 25.2 37.8#
130 2.3 1.4 2.2#
166 28.2 17.8 26.6#



#60
Diethylphthalate
Concen: 6.617 ng
RT: 15.334 min Scan# 2099
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:149 Resp: 678479
Ion Ratio Lower Upper
149 100
177 23.7 19.9 29.9
150 12.3 10.3 15.5









#61
4-Chlorophenyl-phenylether
Concen: 6.499 ng
RT: 15.557 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:204 Resp: 358871
Ion Ratio Lower Upper
204 100
206 32.9 26.3 39.5
141 50.6 39.1 58.7



#62
4-Nitroaniline
Concen: 6.351 ng
RT: 15.569 min Scan# 2139
Delta R.T. -0.012 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:138 Resp: 144516
Ion Ratio Lower Upper
138 100
92 47.9 24.3 64.3
108 73.3 44.8 84.8









#63
Azobenzene
Concen: 6.488 ng
RT: 15.845 min Scan# 2187
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 77	Resp: 536914
Ion Ratio	Lower Upper
77 100	
182 29.9	15.3 55.3
105 21.5	3.2 43.2
51 27.4	3.4 43.4



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.263 min Scan# 2427
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:188	Resp: 2995807		
Ion	Ratio	Lower	Upper
188	100		
94	7.5	6.8	10.2
80	7.9	7.2	10.8









#65
4,6-Dinitro-2-methylphenol
Concen: 5.611 ng
RT: 15.640 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:198 Resp: 84525
Ion Ratio Lower Upper
198 100
51 30.7 6.8 46.8
105 35.7 14.4 54.4



#66
n-Nitrosodiphenylamine
Concen: 6.603 ng
RT: 15.769 min Scan# 2173
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:169 Resp: 596966
Ion Ratio Lower Upper
169 100
168 65.7 53.3 79.9
167 36.2 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 6.774 ng
RT: 16.451 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:248 Resp: 222041
Ion Ratio Lower Upper
248 100
250 100.6 77.6 116.4
141 63.0 47.4 71.2



#68
Hexachlorobenzene
Concen: 6.424 ng
RT: 16.575 min Scan# 2310
Delta R.T. 0.000 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:284 Resp: 258511
Ion Ratio Lower Upper
284 100
142 27.9 22.4 33.6
249 29.6 24.4 36.6





#69

Atrazine

Concen: 6.728 ng

RT: 16.722 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC005





Tgt Ion:200 Resp: 227529

Ion Ratio Lower Upper

200 100

173 25.8 6.5 46.5

215 50.1 29.6 69.6



#70

Pentachlorophenol

Concen: 6.508 ng

RT: 16.916 min Scan# 2368

Delta R.T. -0.006 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59




Tgt Ion:266 Resp: 142624


Ion Ratio Lower Upper

266 100

268 64.1 50.9 76.3

264 61.6 50.3 75.5









#71
Phenanthrene
Concen: 6.710 ng
RT: 17.304 min Scan# 2435
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:178 Resp: 1083774
Ion Ratio Lower Upper
178 100
176 18.8 16.0 24.0
179 15.1 13.0 19.4



#72
Anthracene
Concen: 6.933 ng
RT: 17.398 min Scan# 2450
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:178 Resp: 1091101
Ion Ratio Lower Upper
178 100
176 18.4 15.1 22.7
179 15.0 12.8 19.2





#73

Carbazole

Concen: 6.421 ng

RT: 17.663 min Scan# 2496

Delta R.T. -0.006 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC005





Tgt Ion:167 Resp: 991762

Ion Ratio Lower Upper

167 100

166 20.9 17.7 26.5

139 12.1 9.8 14.6



#74

Di-n-butylphthalate

Concen: 6.849 ng

RT: 18.222 min Scan# 2590

Delta R.T. -0.006 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59




Tgt Ion:149 Resp: 1157500


Ion Ratio Lower Upper

149 100

150 9.0 7.9 11.9

104 4.7 3.7 5.5









#75
Fluoranthene
Concen: 6.959 ng
RT: 19.269 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:202 Resp: 1292895
Ion Ratio Lower Upper
202 100
101 8.9 0.0 30.7
203 17.4 0.0 38.9



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.345 min Scan# 3121
Delta R.T. -0.012 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:240 Resp: 3017529
Ion Ratio Lower Upper
240 100
120 9.2 7.7 11.5
236 26.0 21.2 31.8









#77
Benzidine
Concen: 8.654 ng
RT: 19.445 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:184 Resp: 784514
Ion Ratio Lower Upper
184 100
185 14.1 12.2 18.2
183 11.5 9.5 14.3



#78
Pyrene
Concen: 6.840 ng
RT: 19.622 min Scan# 2828
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:202 Resp: 1342745
Ion Ratio Lower Upper
202 100
200 21.7 18.0 27.0
203 17.7 15.0 22.6









#79
 Terphenyl-d14
 Concen: 15.461 ng
 RT: 19.816 min Scan# 2861
 Delta R.T. -0.012 min
 Lab File: BP008704.D
 Acq: 06 Jan 2022 11:59




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:244 Resp: 2285015
 Ion Ratio Lower Upper
 244 100
 212 6.7 6.5 9.7
 122 9.3 9.2 13.8



#80
 Butylbenzylphthalate
 Concen: 6.521 ng
 RT: 20.498 min Scan# 2977
 Delta R.T. -0.006 min
 Lab File: BP008704.D
 Acq: 06 Jan 2022 11:59

Tgt Ion:149 Resp: 510565
 Ion Ratio Lower Upper
 149 100
 91 59.2 50.6 76.0
 206 24.9 19.7 29.5









#81
Benzo(a)anthracene
Concen: 6.705 ng
RT: 21.333 min Scan# 3119
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:228 Resp: 1312602
Ion Ratio Lower Upper
228 100
226 26.5 23.3 34.9
229 19.7 17.0 25.6



#82
3,3'-Dichlorobenzidine
Concen: 7.781 ng
RT: 21.263 min Scan# 3107
Delta R.T. -0.012 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:252 Resp: 560075
Ion Ratio Lower Upper
252 100
254 66.6 52.4 78.6
126 8.6 7.3 10.9









#83
Chrysene
Concen: 6.898 ng
RT: 21.380 min Scan# 3129
Delta R.T. -0.012 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:228 Resp: 1262751
Ion Ratio Lower Upper
228 100
226 29.5 26.0 39.0
229 19.6 16.9 25.3



#84
Bis(2-ethylhexyl)phthalate
Concen: 6.601 ng
RT: 21.263 min Scan# 3107
Delta R.T. -0.006 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:149 Resp: 780791
Ion Ratio Lower Upper
149 100
167 30.4 25.0 37.4
279 6.2 4.8 7.2









#85
Di-n-octyl phthalate
Concen: 7.433 ng
RT: 22.192 min Scan# 31
Delta R.T. -0.012 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 1360846
Ion Ratio Lower Upper
149 100
167 1.7 1.4 2.0
43 6.4 5.4 8.2



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.769 min Scan# 3533
Delta R.T. -0.017 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:264 Resp: 3286549
Ion Ratio Lower Upper
264 100
260 24.6 19.1 28.7
265 22.1 17.1 25.7





#87

Indeno(1,2,3-cd)pyrene

Concen: 7.415 ng

RT: 26.286 min Scan# 3407

Delta R.T. -0.041 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC005





Tgt Ion:276 Resp: 1636971

Ion Ratio Lower Upper

276 100

138 20.2 17.8 26.6

277 25.7 20.4 30.6



#88

Benzo(b)fluoranthene

Concen: 6.654 ng

RT: 23.027 min Scan# 3407

Delta R.T. -0.017 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59




Tgt Ion:252 Resp: 1406454


Ion Ratio Lower Upper

252 100

253 22.1 18.0 27.0

125 8.1 6.9 10.3









#89
Benzo(k)fluoranthene
Concen: 6.427 ng
RT: 23.075 min Scan# 3419
Delta R.T. -0.023 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:252 Resp: 1350338
Ion Ratio Lower Upper
252 100
253 22.4 18.2 27.2
125 8.3 7.3 10.9



#90
Benzo(a)pyrene
Concen: 7.888 ng
RT: 23.657 min Scan# 3514
Delta R.T. -0.023 min
Lab File: BP008704.D
Acq: 06 Jan 2022 11:59

Tgt Ion:252 Resp: 1344071
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.8
125 9.8 7.8 11.6





#91

Dibenzo(a,h)anthracene

Concen: 7.534 ng

RT: 26.304 min Scan# 3964

Delta R.T. -0.041 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC005





Tgt Ion:278 Resp: 1403023

Ion Ratio Lower Upper

278 100

139 13.1 11.5 17.3

279 24.4 19.1 28.7



#92

Benzo(g,h,i)perylene

Concen: 7.881 ng

RT: 27.063 min Scan# 4093

Delta R.T. -0.047 min

Lab File: BP008704.D

Acq: 06 Jan 2022 11:59

Tgt Ion:276 Resp: 1372348

Ion Ratio Lower Upper

276 100

277 23.3 19.2 28.8

138 17.4 15.0 22.4

