

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010622\
 Data File : BP008705.D
 Acq On : 06 Jan 2022 12:33
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Jan 06 16:18:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 06 16:16:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	335615	20.000	ng	0.00
21) Naphthalene-d8	10.681	136	1496984	20.000	ng	0.00
39) Acenaphthene-d10	14.510	164	1122668	20.000	ng	0.00
64) Phenanthrene-d10	17.263	188	2497260	20.000	ng	0.00
76) Chrysene-d12	21.345	240	2598397	20.000	ng	-0.01
86) Perylene-d12	23.769	264	2947329	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.458	112	459386	24.984	ng	0.00
7) Phenol-d6	7.046	99	606317	23.691	ng	0.00
23) Nitrobenzene-d5	9.034	82	555821	22.088	ng	0.00
42) 2,4,6-Tribromophenol	15.998	330	380272	25.036	ng	-0.01
45) 2-Fluorobiphenyl	13.140	172	1761431	22.732	ng	0.00
79) Terphenyl-d14	19.822	244	3285504	25.817	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.364	88	90923	12.241	ng	98
3) Pyridine	3.770	79	187897	9.273	ng	# 96
4) n-Nitrosodimethylamine	3.670	42	89244	13.640	ng	# 84
6) Aniline	7.205	93	384726	12.934	ng	100
8) 2-Chlorophenol	7.446	128	256024	11.821	ng	100
9) Benzaldehyde	7.016	77	209350	14.372	ng	98
10) Phenol	7.075	94	311436	12.054	ng	99
11) bis(2-Chloroethyl)ether	7.299	93	243762	11.698	ng	98
12) 1,3-Dichlorobenzene	7.769	146	297440	11.732	ng	99
13) 1,4-Dichlorobenzene	7.916	146	307086	11.908	ng	99
14) 1,2-Dichlorobenzene	8.234	146	296498	11.769	ng	99
15) Benzyl Alcohol	8.122	79	225371	12.294	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.416	45	287259	12.115	ng	100
17) 2-Methylphenol	8.328	107	212757	11.833	ng	96
18) Hexachloroethane	8.963	117	100620	11.968	ng	96
19) n-Nitroso-di-n-propyla...	8.687	70	202590	12.396	ng	94
20) 3+4-Methylphenols	8.652	107	300188	11.884	ng	97
22) Acetophenone	8.699	105	424357	11.822	ng	# 96
24) Nitrobenzene	9.075	77	284088	11.854	ng	96
25) Isophorone	9.610	82	571310	12.153	ng	99
26) 2-Nitrophenol	9.793	139	141464	12.428	ng	94
27) 2,4-Dimethylphenol	9.857	122	267564	13.026	ng	98
28) bis(2-Chloroethoxy)met...	10.093	93	354191	10.883	ng	99
29) 2,4-Dichlorophenol	10.334	162	281784	11.335	ng	99
30) 1,2,4-Trichlorobenzene	10.546	180	314568	10.731	ng	99
31) Naphthalene	10.728	128	887786	11.426	ng	99
32) Benzoic acid	9.963	122	154357	11.898	ng	97
33) 4-Chloroaniline	10.840	127	394546	11.842	ng	99
34) Hexachlorobutadiene	11.022	225	191194	10.744	ng	98
35) Caprolactam	11.610	113	99827	12.217	ng	93
36) 4-Chloro-3-methylphenol	11.969	107	292089	11.754	ng	96
37) 2-Methylnaphthalene	12.340	142	695188	11.887	ng	98
38) 1-Methylnaphthalene	12.557	142	652629	11.413	ng	100
40) 1,2,4,5-Tetrachloroben...	12.710	216	389280	11.006	ng	99
41) Hexachlorocyclopentadiene	12.693	237	218471	11.616	ng	99
43) 2,4,6-Trichlorophenol	12.951	196	252367	10.967	ng	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010622\
 Data File : BP008705.D
 Acq On : 06 Jan 2022 12:33
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Jan 06 16:18:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 06 16:16:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.022	196	277730	10.931	ng	98
46) 1,1'-Biphenyl	13.346	154	951935	11.464	ng	100
47) 2-Chloronaphthalene	13.387	162	722663	11.105	ng	99
48) 2-Nitroaniline	13.587	65	165192	11.579	ng	96
49) Acenaphthylene	14.234	152	1146097	11.465	ng	100
50) Dimethylphthalate	13.969	163	963790	11.146	ng	99
51) 2,6-Dinitrotoluene	14.081	165	197723	11.144	ng	90
52) Acenaphthene	14.575	154	692172	10.583	ng	100
53) 3-Nitroaniline	14.410	138	202886	10.986	ng	100
54) 2,4-Dinitrophenol	14.622	184	85839	10.862	ng	97
55) Dibenzofuran	14.910	168	1156095	11.026	ng	99
56) 4-Nitrophenol	14.722	139	164958	11.343	ng	93
57) 2,4-Dinitrotoluene	14.869	165	267172	11.285	ng #	89
58) Fluorene	15.557	166	940489	11.511	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.140	232	252783	11.105	ng #	88
60) Diethylphthalate	15.334	149	949386	11.232	ng	98
61) 4-Chlorophenyl-phenyle...	15.557	204	498778	10.958	ng	99
62) 4-Nitroaniline	15.569	138	215206	11.475	ng	95
63) Azobenzene	15.845	77	763020	11.185	ng	94
65) 4,6-Dinitro-2-methylph...	15.640	198	126502	10.074	ng	95
66) n-Nitrosodiphenylamine	15.769	169	838063	11.120	ng	99
67) 4-Bromophenyl-phenylether	16.451	248	313831	11.485	ng	99
68) Hexachlorobenzene	16.569	284	362784	10.816	ng	96
69) Atrazine	16.722	200	326274	11.573	ng	97
70) Pentachlorophenol	16.916	266	202837	11.103	ng	99
71) Phenanthrene	17.304	178	1527247	11.344	ng	98
72) Anthracene	17.398	178	1555482	11.857	ng	99
73) Carbazole	17.663	167	1417085	11.006	ng	99
74) Di-n-butylphthalate	18.222	149	1634633	11.604	ng	99
75) Fluoranthene	19.269	202	1848289	11.935	ng	97
77) Benzidine	19.445	184	1214257	15.554	ng	98
78) Pyrene	19.622	202	1919995	11.358	ng	98
80) Butylbenzylphthalate	20.498	149	768328	11.396	ng	97
81) Benzo(a)anthracene	21.333	228	1896715	11.252	ng	97
82) 3,3'-Dichlorobenzidine	21.263	252	836137	13.490	ng	98
83) Chrysene	21.386	228	1842681	11.690	ng	95
84) Bis(2-ethylhexyl)phtha...	21.263	149	1119915	10.996	ng	98
85) Di-n-octyl phthalate	22.192	149	1968289	12.485	ng	99
87) Indeno(1,2,3-cd)pyrene	26.292	276	2548877	12.875	ng	98
88) Benzo(b)fluoranthene	23.027	252	2138099	11.280	ng	100
89) Benzo(k)fluoranthene	23.080	252	1945139	10.323	ng	98
90) Benzo(a)pyrene	23.657	252	2032951	13.304	ng	99
91) Dibenzo(a,h)anthracene	26.310	278	2174429	13.020	ng	99
92) Benzo(g,h,i)perylene	27.062	276	2142935	13.722	ng	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.881 min Scan# 81
 Delta R.T. 0.000 min
 Lab File: BP008705.D
 Acq: 06 Jan 2022 12:33




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010


Tgt Ion:152 Resp: 335615
 Ion Ratio Lower Upper
 152 100
 150 152.5 127.0 190.6
 115 54.4 43.7 65.5



#2
 1,4-Dioxane
 Concen: 12.241 ng
 RT: 3.364 min Scan# 64
 Delta R.T. 0.000 min
 Lab File: BP008705.D
 Acq: 06 Jan 2022 12:33

Tgt Ion: 88 Resp: 90923
 Ion Ratio Lower Upper
 88 100
 58 63.5 50.6 75.8
 43 25.8 18.2 27.4





#3

Pyridine

Concen: 9.273 ng

RT: 3.770 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33

Instrument :

BNA_P

ClientSampleId :

SSTDICC010





Tgt Ion: 79 Resp: 187897

Ion Ratio Lower Upper

79 100

52 56.6 45.1 67.7

51 34.9 23.1 34.7#



#4

n-Nitrosodimethylamine

Concen: 13.640 ng

RT: 3.670 min Scan# 116

Delta R.T. 0.000 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33




Tgt Ion: 42 Resp: 89244


Ion Ratio Lower Upper

42 100

74 142.6 131.1 196.7

44 11.6 6.2 9.2#









#5
2-Fluorophenol
Concen: 24.984 ng
RT: 5.458 min Scan# 41
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:112 Resp: 459386
Ion Ratio Lower Upper
112 100
64 50.4 39.7 59.5
63 26.9 20.0 30.0



#6
Aniline
Concen: 12.934 ng
RT: 7.205 min Scan# 717
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion: 93 Resp: 384726
Ion Ratio Lower Upper
93 100
66 34.7 27.8 41.6
65 17.6 14.0 21.0









#7
Phenol-d6
Concen: 23.691 ng
RT: 7.046 min Scan# 691
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion: 99 Resp: 606317
Ion Ratio Lower Upper
99 100
42 15.6 10.8 16.2
71 32.0 23.8 35.6



#8
2-Chlorophenol
Concen: 11.821 ng
RT: 7.446 min Scan# 758
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion: 128 Resp: 256024
Ion Ratio Lower Upper
128 100
130 32.5 12.6 52.6
64 38.7 18.6 58.6









#9
Benzaldehyde
Concen: 14.372 ng
RT: 7.016 min Scan# 685
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion: 77 Resp: 209350
Ion Ratio Lower Upper
77 100
105 101.5 84.2 124.2
106 98.6 80.6 120.6



#10
Phenol
Concen: 12.054 ng
RT: 7.075 min Scan# 695
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion: 94 Resp: 311436
Ion Ratio Lower Upper
94 100
65 25.0 4.1 44.1
66 33.0 12.4 52.4









#11
bis(2-Chloroethyl)ether
Concen: 11.698 ng
RT: 7.299 min Scan# 711
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 93 Resp: 243762




Ion	Ratio	Lower	Upper
93	100		
63	67.7	46.0	86.0
95	33.1	12.2	52.2




#12
1,3-Dichlorobenzene
Concen: 11.732 ng
RT: 7.769 min Scan# 813
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion: 146 Resp: 297440

Ion	Ratio	Lower	Upper
146	100		
148	63.8	51.9	77.9
75	26.4	20.5	30.7









#13
1,4-Dichlorobenzene
Concen: 11.908 ng
RT: 7.916 min Scan# 838
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:146 Resp: 307086
Ion Ratio Lower Upper
146 100
148 65.8 52.3 78.5
111 37.8 29.8 44.8



#14
1,2-Dichlorobenzene
Concen: 11.769 ng
RT: 8.234 min Scan# 892
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:146 Resp: 296498
Ion Ratio Lower Upper
146 100
148 66.2 52.2 78.2
111 38.7 31.1 46.7









#15
Benzyl Alcohol
Concen: 12.294 ng
RT: 8.122 min Scan# 81
Delta R.T. 0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion: 79 Resp: 225371
Ion Ratio Lower Upper
79 100
108 82.6 67.8 101.8
77 65.2 50.7 76.1



#16
2,2'-oxybis(1-Chloropropane)
Concen: 12.115 ng
RT: 8.416 min Scan# 923
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion: 45 Resp: 287259
Ion Ratio Lower Upper
45 100
77 17.1 0.0 37.4
79 14.2 0.0 34.1









#17
2-Methylphenol
Concen: 11.833 ng
RT: 8.328 min Scan# 908
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:	107	Resp:	212757
Ion	Ratio	Lower	Upper
107	100		
108	114.5	90.0	135.0
77	45.2	32.7	49.1
79	44.3	32.1	48.1



#18
Hexachloroethane
Concen: 11.968 ng
RT: 8.963 min Scan# 1016
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:	117	Resp:	100620
Ion	Ratio	Lower	Upper
117	100		
119	101.1	76.9	115.3
201	121.4	93.7	140.5









#19
n-Nitroso-di-n-propylamine
Concen: 12.396 ng
RT: 8.687 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 70 Resp: 202590




Ion	Ratio	Lower	Upper
70	100		
42	53.3	38.2	57.2
101	11.0	8.7	13.1
130	23.5	21.0	31.6




#20
3+4-Methylphenols
Concen: 11.884 ng
RT: 8.652 min Scan# 963
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion: 107 Resp: 300188

Ion	Ratio	Lower	Upper
107	100		
108	85.9	69.8	109.8
77	31.3	10.4	50.4
79	25.6	6.4	46.4









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.681 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:136 Resp: 1496984
Ion Ratio Lower Upper
136 100
137 11.0 9.0 13.4
54 6.1 4.9 7.3
68 5.3 4.3 6.5



#22
Acetophenone
Concen: 11.822 ng
RT: 8.699 min Scan# 971
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:105 Resp: 424357
Ion Ratio Lower Upper
105 100
71 4.2 5.3 7.9#
51 23.7 16.3 24.5
120 24.4 19.2 28.8









#23
Nitrobenzene-d5
Concen: 22.088 ng
RT: 9.034 min Scan# 1028
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion: 82 Resp: 555821
Ion Ratio Lower Upper
82 100
128 46.6 39.4 59.2
54 43.6 32.6 48.8



#24
Nitrobenzene
Concen: 11.854 ng
RT: 9.075 min Scan# 1035
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion: 77 Resp: 284088
Ion Ratio Lower Upper
77 100
123 48.2 41.4 62.2
65 12.1 9.8 14.8









#25
Isophorone
Concen: 12.153 ng
RT: 9.610 min Scan# 1126
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion: 82 Resp: 571310
Ion Ratio Lower Upper
82 100
95 7.3 5.4 8.2
138 20.2 16.6 24.8



#26
2-Nitrophenol
Concen: 12.428 ng
RT: 9.793 min Scan# 1157
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion: 139 Resp: 141464
Ion Ratio Lower Upper
139 100
109 29.1 20.7 31.1
65 43.3 31.6 47.4









#27
2,4-Dimethylphenol
Concen: 13.026 ng
RT: 9.857 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:122 Resp: 267564
Ion Ratio Lower Upper
122 100
107 110.8 89.6 134.4
121 57.2 47.8 71.8



#28
bis(2-Chloroethoxy)methane
Concen: 10.883 ng
RT: 10.093 min Scan# 1208
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion: 93 Resp: 354191
Ion Ratio Lower Upper
93 100
95 33.0 25.8 38.8
123 12.1 9.7 14.5









#29
2,4-Dichlorophenol
Concen: 11.335 ng
RT: 10.334 min Scan# 1249
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:162 Resp: 281784
Ion Ratio Lower Upper
162 100
164 64.1 44.9 84.9
98 30.2 11.2 51.2



#30
1,2,4-Trichlorobenzene
Concen: 10.731 ng
RT: 10.546 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:180 Resp: 314568
Ion Ratio Lower Upper
180 100
182 94.1 76.7 115.1
145 28.0 22.4 33.6





#31

Naphthalene

Concen: 11.426 ng

RT: 10.728 min Scan# 1186

Delta R.T. -0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33

Instrument :

BNA_P

ClientSampleId :

SSTDICC010





Tgt Ion:128 Resp: 887786

Ion Ratio Lower Upper

128 100

129 11.2 8.8 13.2

127 13.3 10.4 15.6



#32

Benzoic acid

Concen: 11.898 ng

RT: 9.963 min Scan# 1186

Delta R.T. -0.047 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33




Tgt Ion:122 Resp: 154357


Ion Ratio Lower Upper

122 100

105 115.2 100.1 140.1

77 85.4 65.5 105.5









#33
4-Chloroaniline
Concen: 11.842 ng
RT: 10.840 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:	127	Resp:	394546
Ion	Ratio	Lower	Upper
127	100		
129	32.3	26.0	39.0
65	25.9	20.5	30.7
92	17.1	14.1	21.1



#34
Hexachlorobutadiene
Concen: 10.744 ng
RT: 11.022 min Scan# 1366
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:	225	Resp:	191194
Ion	Ratio	Lower	Upper
225	100		
223	61.9	50.1	75.1
227	62.6	51.5	77.3









#35
Caprolactam
Concen: 12.217 ng
RT: 11.610 min Scan# 1466
Delta R.T. -0.023 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:113 Resp: 99827
Ion Ratio Lower Upper
113 100
55 137.3 107.1 147.1
56 99.2 74.1 114.1



#36
4-Chloro-3-methylphenol
Concen: 11.754 ng
RT: 11.969 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:107 Resp: 292089
Ion Ratio Lower Upper
107 100
144 28.8 23.2 34.8
142 92.9 70.1 105.1









#37
2-Methylnaphthalene
Concen: 11.887 ng
RT: 12.340 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:142 Resp: 695188
Ion Ratio Lower Upper
142 100
141 88.9 70.5 105.7
115 33.1 24.6 36.8



#38
1-Methylnaphthalene
Concen: 11.413 ng
RT: 12.557 min Scan# 1627
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:142 Resp: 652629
Ion Ratio Lower Upper
142 100
141 91.1 73.2 109.8
116 3.7 2.7 4.1









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.510 min Scan# 1959
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:164 Resp: 1122668
Ion Ratio Lower Upper
164 100
162 98.9 79.9 119.9
160 44.3 35.5 53.3



#40
1,2,4,5-Tetrachlorobenzene
Concen: 11.006 ng
RT: 12.710 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:216 Resp: 389280
Ion Ratio Lower Upper
216 100
214 78.0 63.1 94.7
179 19.8 16.0 24.0
108 15.9 11.9 17.9









#41
Hexachlorocyclopentadiene
Concen: 11.616 ng
RT: 12.693 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:	237	Resp:	218471
Ion	Ratio	Lower	Upper
237	100		
235	62.3	42.9	82.9
272	13.1	0.0	33.8



#42
2,4,6-Tribromophenol
Concen: 25.036 ng
RT: 15.998 min Scan# 2212
Delta R.T. -0.012 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:	330	Resp:	380272
Ion	Ratio	Lower	Upper
330	100		
332	96.7	77.0	115.4
141	29.5	21.3	31.9









#43
2,4,6-Trichlorophenol
Concen: 10.967 ng
RT: 12.951 min Scan# 1694
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:196 Resp: 252367
Ion Ratio Lower Upper
196 100
198 98.6 77.8 116.8
200 31.6 24.7 37.1



#44
2,4,5-Trichlorophenol
Concen: 10.931 ng
RT: 13.022 min Scan# 1706
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:196 Resp: 277730
Ion Ratio Lower Upper
196 100
198 96.4 77.7 116.5
97 41.0 29.6 44.4
132 23.4 18.7 28.1









#45
2-Fluorobiphenyl
Concen: 22.732 ng
RT: 13.140 min Scan# 1726
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:172 Resp: 1761431
Ion Ratio Lower Upper
172 100
171 34.1 29.3 43.9
170 22.9 19.5 29.3



#46
1,1'-Biphenyl
Concen: 11.464 ng
RT: 13.346 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:154 Resp: 951935
Ion Ratio Lower Upper
154 100
153 40.1 20.3 60.3
76 10.7 0.0 31.1









#47
2-Chloronaphthalene
Concen: 11.105 ng
RT: 13.387 min Scan# 1768
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:162 Resp: 722663
Ion Ratio Lower Upper
162 100
127 34.6 28.4 42.6
164 32.7 26.8 40.2



#48
2-Nitroaniline
Concen: 11.579 ng
RT: 13.587 min Scan# 1802
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion: 65 Resp: 165192
Ion Ratio Lower Upper
65 100
92 76.0 62.2 93.4
138 126.5 105.8 158.8









#49
Acenaphthylene
Concen: 11.465 ng
RT: 14.234 min Scan# 1912
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:152 Resp: 1146097
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1
153 13.2 10.6 16.0



#50
Dimethylphthalate
Concen: 11.146 ng
RT: 13.969 min Scan# 1867
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:163 Resp: 963790
Ion Ratio Lower Upper
163 100
194 4.4 3.8 5.6
164 10.1 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 11.144 ng
RT: 14.081 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:165 Resp: 197723
Ion Ratio Lower Upper
165 100
63 45.7 31.8 47.8
89 51.3 36.1 54.1



#52
Acenaphthene
Concen: 10.583 ng
RT: 14.575 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:154 Resp: 692172
Ion Ratio Lower Upper
154 100
153 111.8 89.4 134.2
152 53.7 42.6 63.8









#53
3-Nitroaniline
Concen: 10.986 ng
RT: 14.410 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:138 Resp: 202886
Ion Ratio Lower Upper
138 100
108 10.1 8.0 12.0
92 102.4 81.7 122.5



#54
2,4-Dinitrophenol
Concen: 10.862 ng
RT: 14.622 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:184 Resp: 85839
Ion Ratio Lower Upper
184 100
63 48.6 35.6 53.4
154 54.3 43.4 65.0









#55
Dibenzofuran
Concen: 11.026 ng
RT: 14.910 min Scan# 20
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:168 Resp: 1156095
Ion Ratio Lower Upper
168 100
139 34.5 27.4 41.0
169 13.3 10.9 16.3



#56
4-Nitrophenol
Concen: 11.343 ng
RT: 14.722 min Scan# 1995
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:139 Resp: 164958
Ion Ratio Lower Upper
139 100
109 70.3 42.2 82.2
65 80.8 57.3 97.3





#57

2,4-Dinitrotoluene

Concen: 11.285 ng

RT: 14.869 min Scan# 2021

Delta R.T. -0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33

Instrument :

BNA_P

ClientSampleId :

SSTDICC010

Tgt Ion:165 Resp: 267172





Ion Ratio Lower Upper

165 100

63 35.2 23.3 34.9#

89 63.3 44.2 66.4

182 2.9 2.6 4.0



#58

Fluorene

Concen: 11.511 ng

RT: 15.557 min Scan# 2137

Delta R.T. -0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33




Tgt Ion:166 Resp: 940489


Ion Ratio Lower Upper

166 100

165 98.2 77.2 115.8

167 13.7 11.2 16.8









#59
2,3,4,6-Tetrachlorophenol
Concen: 11.105 ng
RT: 15.140 min Scan# 2067
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:232 Resp: 252783
Ion Ratio Lower Upper
232 100
131 38.3 25.2 37.8#
130 2.3 1.4 2.2#
166 28.2 17.8 26.6#



#60
Diethylphthalate
Concen: 11.232 ng
RT: 15.334 min Scan# 2099
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:149 Resp: 949386
Ion Ratio Lower Upper
149 100
177 23.4 19.9 29.9
150 12.6 10.3 15.5









#61
4-Chlorophenyl-phenylether
Concen: 10.958 ng
RT: 15.557 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:204 Resp: 498778
Ion Ratio Lower Upper
204 100
206 33.5 26.3 39.5
141 49.4 39.1 58.7



#62
4-Nitroaniline
Concen: 11.475 ng
RT: 15.569 min Scan# 2139
Delta R.T. -0.012 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:138 Resp: 215206
Ion Ratio Lower Upper
138 100
92 46.3 24.3 64.3
108 70.1 44.8 84.8





#63

Azobenzene

Concen: 11.185 ng

RT: 15.845 min Scan# 2186

Delta R.T. -0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33

Instrument :

BNA_P

ClientSampleId :

SSTDICC010

Tgt Ion: 77 Resp: 763020





Ion Ratio Lower Upper

77 100

182 31.0 15.3 55.3

105 20.9 3.2 43.2

51 26.2 3.4 43.4



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.263 min Scan# 2427

Delta R.T. -0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33




Tgt Ion:188 Resp: 2497260


Ion Ratio Lower Upper

188 100

94 7.5 6.8 10.2

80 7.9 7.2 10.8









#65
4,6-Dinitro-2-methylphenol
Concen: 10.074 ng
RT: 15.640 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:198 Resp: 126502
Ion Ratio Lower Upper
198 100
51 30.5 6.8 46.8
105 36.5 14.4 54.4



#66
n-Nitrosodiphenylamine
Concen: 11.120 ng
RT: 15.769 min Scan# 2173
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:169 Resp: 838063
Ion Ratio Lower Upper
169 100
168 65.9 53.3 79.9
167 35.9 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 11.485 ng
RT: 16.451 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:248 Resp: 313831
Ion Ratio Lower Upper
248 100
250 97.3 77.6 116.4
141 61.8 47.4 71.2



#68
Hexachlorobenzene
Concen: 10.816 ng
RT: 16.569 min Scan# 2309
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:284 Resp: 362784
Ion Ratio Lower Upper
284 100
142 32.2 22.4 33.6
249 30.0 24.4 36.6





#69

Atrazine

Concen: 11.573 ng

RT: 16.722 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33

Instrument :

BNA_P

ClientSampleId :

SSTDICC010





Tgt Ion:200 Resp: 326274

Ion Ratio Lower Upper

200 100

173 24.9 6.5 46.5

215 51.1 29.6 69.6



#70

Pentachlorophenol

Concen: 11.103 ng

RT: 16.916 min Scan# 2368

Delta R.T. -0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33




Tgt Ion:266 Resp: 202837


Ion Ratio Lower Upper

266 100

268 63.9 50.9 76.3

264 64.6 50.3 75.5









#71
Phenanthrene
Concen: 11.344 ng
RT: 17.304 min Scan# 2434
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:178 Resp: 1527247
Ion Ratio Lower Upper
178 100
176 19.3 16.0 24.0
179 15.4 13.0 19.4



#72
Anthracene
Concen: 11.857 ng
RT: 17.398 min Scan# 2450
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:178 Resp: 1555482
Ion Ratio Lower Upper
178 100
176 18.9 15.1 22.7
179 15.4 12.8 19.2









#73
 Carbazole
 Concen: 11.006 ng
 RT: 17.663 min Scan# 2496
 Delta R.T. -0.006 min
 Lab File: BP008705.D
 Acq: 06 Jan 2022 12:33




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010


Tgt Ion:167 Resp: 1417085
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.7 26.5
 139 12.3 9.8 14.6



#74
 Di-n-butylphthalate
 Concen: 11.604 ng
 RT: 18.222 min Scan# 2590
 Delta R.T. -0.006 min
 Lab File: BP008705.D
 Acq: 06 Jan 2022 12:33

Tgt Ion:149 Resp: 1634633
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.9 11.9
 104 4.6 3.7 5.5









#75
Fluoranthene
Concen: 11.935 ng
RT: 19.269 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:202 Resp: 1848289
Ion Ratio Lower Upper
202 100
101 9.4 0.0 30.7
203 17.8 0.0 38.9



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.345 min Scan# 3121
Delta R.T. -0.012 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:240 Resp: 2598397
Ion Ratio Lower Upper
240 100
120 9.5 7.7 11.5
236 25.7 21.2 31.8









#77
Benzidine
Concen: 15.554 ng
RT: 19.445 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:184 Resp: 1214257
Ion Ratio Lower Upper
184 100
185 14.1 12.2 18.2
183 11.6 9.5 14.3



#78
Pyrene
Concen: 11.358 ng
RT: 19.622 min Scan# 2828
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:202 Resp: 1919995
Ion Ratio Lower Upper
202 100
200 21.3 18.0 27.0
203 17.8 15.0 22.6





#79

Terphenyl-d14

Concen: 25.817 ng

RT: 19.822 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33

Instrument :

BNA_P

ClientSampleId :

SSTDICC010





Tgt Ion:244 Resp: 3285504

Ion Ratio Lower Upper

244 100

212 6.7 6.5 9.7

122 8.3 9.2 13.8#



#80

Butylbenzylphthalate

Concen: 11.396 ng

RT: 20.498 min Scan# 2977

Delta R.T. -0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33




Tgt Ion:149 Resp: 768328


Ion Ratio Lower Upper

149 100

91 60.6 50.6 76.0

206 24.0 19.7 29.5









#81
Benzo(a)anthracene
Concen: 11.252 ng
RT: 21.333 min Scan# 3119
Delta R.T. -0.006 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:228 Resp: 1896715
Ion Ratio Lower Upper
228 100
226 27.3 23.3 34.9
229 20.0 17.0 25.6



#82
3,3'-Dichlorobenzidine
Concen: 13.490 ng
RT: 21.263 min Scan# 3107
Delta R.T. -0.012 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:252 Resp: 836137
Ion Ratio Lower Upper
252 100
254 63.7 52.4 78.6
126 8.5 7.3 10.9





#83

Chrysene

Concen: 11.690 ng

RT: 21.386 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33

Instrument :

BNA_P

ClientSampleId :

SSTDICC010





Tgt Ion:228 Resp: 1842681

Ion Ratio Lower Upper

228 100

226 29.8 26.0 39.0

229 19.1 16.9 25.3



#84

Bis(2-ethylhexyl)phthalate

Concen: 10.996 ng

RT: 21.263 min Scan# 3107

Delta R.T. -0.006 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33




Tgt Ion:149 Resp: 1119915


Ion Ratio Lower Upper

149 100

167 30.1 25.0 37.4

279 6.0 4.8 7.2









#85
Di-n-octyl phthalate
Concen: 12.485 ng
RT: 22.192 min Scan# 31
Delta R.T. -0.012 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:149 Resp: 1968289
Ion Ratio Lower Upper
149 100
167 1.8 1.4 2.0
43 6.4 5.4 8.2



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.769 min Scan# 3533
Delta R.T. -0.018 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:264 Resp: 2947329
Ion Ratio Lower Upper
264 100
260 24.3 19.1 28.7
265 22.5 17.1 25.7





#87

Indeno(1,2,3-cd)pyrene

Concen: 12.875 ng

RT: 26.292 min Scan# 3407

Delta R.T. -0.035 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33

Instrument :

BNA_P

ClientSampleId :

SSTDICC010





Tgt Ion:276 Resp: 2548877

Ion Ratio Lower Upper

276 100

138 20.4 17.8 26.6

277 25.5 20.4 30.6



#88

Benzo(b)fluoranthene

Concen: 11.280 ng

RT: 23.027 min Scan# 3407

Delta R.T. -0.018 min

Lab File: BP008705.D

Acq: 06 Jan 2022 12:33




Tgt Ion:252 Resp: 2138099


Ion Ratio Lower Upper

252 100

253 22.4 18.0 27.0

125 8.3 6.9 10.3









#89
Benzo(k)fluoranthene
Concen: 10.323 ng
RT: 23.080 min Scan# 3419
Delta R.T. -0.017 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:252 Resp: 1945139
Ion Ratio Lower Upper
252 100
253 22.0 18.2 27.2
125 8.1 7.3 10.9



#90
Benzo(a)pyrene
Concen: 13.304 ng
RT: 23.657 min Scan# 3514
Delta R.T. -0.023 min
Lab File: BP008705.D
Acq: 06 Jan 2022 12:33

Tgt Ion:252 Resp: 2032951
Ion Ratio Lower Upper
252 100
253 21.8 17.8 26.8
125 9.1 7.8 11.6









#91
 Dibenzo(a,h)anthracene
 Concen: 13.020 ng
 RT: 26.310 min Scan# 3971
 Delta R.T. -0.035 min
 Lab File: BP008705.D
 Acq: 06 Jan 2022 12:33

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Tgt Ion:278 Resp: 2174429
 Ion Ratio Lower Upper
 278 100
 139 13.6 11.5 17.3
 279 24.4 19.1 28.7



#92
 Benzo(g,h,i)perylene
 Concen: 13.722 ng
 RT: 27.062 min Scan# 4093
 Delta R.T. -0.047 min
 Lab File: BP008705.D
 Acq: 06 Jan 2022 12:33

Tgt Ion:276 Resp: 2142935
 Ion Ratio Lower Upper
 276 100
 277 23.9 19.2 28.8
 138 17.3 15.0 22.4

