

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011023\
 Data File : BP013433.D
 Acq On : 10 Jan 2023 18:08
 Operator : CG/JU
 Sample : PB150072BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Jan 10 23:23:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 06 01:03:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	1232160	20.000	ng/ul	0.00
20) Naphthalene-d8	10.728	136	4889051	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.563	164	3297326	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.322	188	7672075	20.000	ng/ul	0.00
79) Chrysene-d12	21.410	240	5442068	20.000	ng/ul	0.00
88) Perylene-d12	23.868	264	4943993	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	157	0.005	ng/uL	-0.09
4) Pyridine-d5	3.746	84	180	0.002	ng/ul	0.00
7) Phenol-d5	7.081	99	17	0.000	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.258	67	191	0.003	ng/ul	0.01
11) 2-Chlorophenol-d4	7.440	132	35	0.000	ng/ul	-0.01
15) 4-Methylphenol-d8	8.634	113	18	0.000	ng/ul	0.00
21) Nitrobenzene-d5	9.087	128	69	0.002	ng/ul	0.00
24) 2-Nitrophenol-d4	9.793	143	55	0.001	ng/ul	-0.02
28) 2,4-Dichlorophenol-d3	10.346	165	17	0.000	ng/ul	0.00
31) 4-Chloroaniline-d4	10.863	131	11	0.000	ng/ul	-0.01
46) Dimethylphthalate-d6	13.981	166	28	0.000	ng/ul	0.00
49) Acenaphthylene-d8	14.251	160	153	0.001	ng/ul	0.00
54) 4-Nitrophenol-d4	14.763	143	14	0.000	ng/ul	0.00
60) Fluorene-d10	15.563	176	2476	0.012	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.681	200	7	0.000	ng/ul	0.00
73) Anthracene-d10	17.322	188	7672075	22.624	ng/ul	-0.10
81) Pyrene-d10	19.657	212	879	0.003	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.868	264	4105731	16.906	ng/ul	0.16

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011023\
Data File : BP013433.D
Acq On : 10 Jan 2023 18:08
Operator : CG/JU
Sample : PB150072BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Jan 10 23:23:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010523.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jan 06 01:03:32 2023
Response via : Initial Calibration

