

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011122\
 Data File : BP008763.D
 Acq On : 11 Jan 2022 12:42
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 12 00:19:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 12 00:18:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	39.555	1.1	116	0.00
3	Pyridine	40.000	39.661	0.8	112	0.00
4	n-Nitrosodimethylamine	40.000	40.808	-2.0	118	0.00
5 S	2-Fluorophenol	80.000	78.060	2.4	113	0.00
6	Aniline	40.000	38.980	2.6	113	0.00
7 S	Phenol-d6	80.000	77.422	3.2	111	0.00
8	2-Chlorophenol	40.000	39.118	2.2	113	0.00
9	Benzaldehyde	40.000	37.680	5.8	109	0.00
10 C	Phenol	40.000	39.173	2.1	113	0.00
11	bis(2-Chloroethyl)ether	40.000	39.465	1.3	115	0.00
12	1,3-Dichlorobenzene	40.000	38.922	2.7	113	0.00
13 C	1,4-Dichlorobenzene	40.000	38.862	2.8	114	0.00
14	1,2-Dichlorobenzene	40.000	38.887	2.8	114	0.00
15	Benzyl Alcohol	40.000	38.431	3.9	110	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.972	2.6	115	0.00
17	2-Methylphenol	40.000	38.527	3.7	110	0.00
18	Hexachloroethane	40.000	39.256	1.9	114	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.537	6.2	109	0.00
20	3+4-Methylphenols	40.000	38.145	4.6	110	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	101	0.00
22	Acetophenone	40.000	37.930	5.2	109	0.00
23 S	Nitrobenzene-d5	80.000	77.940	2.6	110	0.00
24	Nitrobenzene	40.000	39.388	1.5	112	0.00
25	Isophorone	40.000	38.265	4.3	108	0.00
26 C	2-Nitrophenol	40.000	40.069	-0.2	110	0.00
27	2,4-Dimethylphenol	40.000	38.519	3.7	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.835	2.9	111	0.00
29 C	2,4-Dichlorophenol	40.000	38.576	3.6	107	0.00
30	1,2,4-Trichlorobenzene	40.000	38.795	3.0	110	0.00
31	Naphthalene	40.000	38.828	2.9	110	0.00
32	Benzoic acid	40.000	38.682	3.3	106	0.00
33	4-Chloroaniline	40.000	37.973	5.1	107	0.00
34 C	Hexachlorobutadiene	40.000	38.556	3.6	111	0.00
35	Caprolactam	40.000	36.639	8.4	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.693	5.8	107	0.00
37	2-Methylnaphthalene	40.000	37.711	5.7	107	0.00
38	1-Methylnaphthalene	40.000	37.708	5.7	107	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.184	2.0	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.090	2.3	103	0.00
42 S	2,4,6-Tribromophenol	80.000	71.269	10.9	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.982	2.5	104	0.00
44	2,4,5-Trichlorophenol	40.000	38.500	3.8	102	0.00
45 S	2-Fluorobiphenyl	80.000	76.185	4.8	103	0.00
46	1,1'-Biphenyl	40.000	38.823	2.9	105	0.00
47	2-Chloronaphthalene	40.000	39.118	2.2	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011122\
 Data File : BP008763.D
 Acq On : 11 Jan 2022 12:42
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 12 00:19:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 12 00:18:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.962	-2.4	106	0.00
49	Acenaphthylene	40.000	38.872	2.8	104	0.00
50	Dimethylphthalate	40.000	37.263	6.8	102	0.00
51	2,6-Dinitrotoluene	40.000	38.537	3.7	104	0.00
52 C	Acenaphthene	40.000	38.741	3.1	103	0.00
53	3-Nitroaniline	40.000	39.517	1.2	103	0.00
54 P	2,4-Dinitrophenol	40.000	37.093	7.3	95	0.00
55	Dibenzofuran	40.000	38.333	4.2	103	0.00
56 P	4-Nitrophenol	40.000	38.373	4.1	99	0.00
57	2,4-Dinitrotoluene	40.000	39.013	2.5	103	0.00
58	Fluorene	40.000	38.345	4.1	113	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.542	6.1	100	0.00
60	Diethylphthalate	40.000	37.480	6.3	101	0.00
61	4-Chlorophenyl-phenylether	40.000	37.898	5.3	111	0.00
62	4-Nitroaniline	40.000	39.359	1.6	112	0.00
63	Azobenzene	40.000	40.500	-1.3	122	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.007	-0.0	102	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.779	0.6	111	0.00
67	4-Bromophenyl-phenylether	40.000	38.690	3.3	105	0.00
68	Hexachlorobenzene	40.000	37.659	5.9	101	0.00
69	Atrazine	40.000	38.617	3.5	106	0.00
70 C	Pentachlorophenol	40.000	37.946	5.1	98	0.00
71	Phenanthrene	40.000	39.023	2.4	108	0.00
72	Anthracene	40.000	39.242	1.9	108	0.00
73	Carbazole	40.000	38.685	3.3	106	0.00
74	Di-n-butylphthalate	40.000	38.878	2.8	106	0.00
75 C	Fluoranthene	40.000	38.956	2.6	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	91	0.00
77	Benzydine	40.000	39.168	2.1	99	0.00
78	Pyrene	40.000	40.254	-0.6	104	0.00
79 S	Terphenyl-d14	80.000	81.806	-2.3	99	0.00
80	Butylbenzylphthalate	40.000	39.689	0.8	101	0.00
81	Benzo(a)anthracene	40.000	39.372	1.6	100	0.00
82	3,3'-Dichlorobenzidine	40.000	37.557	6.1	95	0.00
83	Chrysene	40.000	39.149	2.1	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.413	1.5	101	0.00
85 c	Di-n-octyl phthalate	40.000	39.304	1.7	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	81	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.120	7.2	83	0.00
88	Benzo(b)fluoranthene	40.000	41.713	-4.3	97	0.00
89	Benzo(k)fluoranthene	40.000	40.516	-1.3	91	0.00
90 C	Benzo(a)pyrene	40.000	40.439	-1.1	92	0.00
91	Dibenzo(a,h)anthracene	40.000	37.095	7.3	83	0.00
92	Benzo(g,h,i)perylene	40.000	35.971	10.1	81	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011122\
Data File : BP008763.D
Acq On : 11 Jan 2022 12:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jan 12 00:19:09 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 12 00:18:56 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0