

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011123\
 Data File : BP013452.D
 Acq On : 11 Jan 2023 19:54
 Operator : CG/JU
 Sample : 01084-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 MH-2

Quant Time: Jan 12 02:20:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 12 02:16:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	566373	20.000	ng	0.00
21) Naphthalene-d8	10.734	136	2427219	20.000	ng	0.00
39) Acenaphthene-d10	14.569	164	1655938	20.000	ng	0.00
64) Phenanthrene-d10	17.328	188	3995070	20.000	ng	0.01
76) Chrysene-d12	21.416	240	3129424	20.000	ng	0.00
86) Perylene-d12	23.886	264	3236509	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	4577915	147.685	ng	0.01
7) Phenol-d6	7.087	99	6107947	150.884	ng	0.01
23) Nitrobenzene-d5	9.093	82	3517787	79.454	ng	0.00
42) 2,4,6-Tribromophenol	16.063	330	3577265	179.173	ng	0.00
45) 2-Fluorobiphenyl	13.187	172	8566241	70.239	ng	0.00
79) Terphenyl-d14	19.875	244	13033813	82.169	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	21.304	149	57529	3.270	ng	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011123\
Data File : BP013452.D
Acq On : 11 Jan 2023 19:54
Operator : CG/JU
Sample : 01084-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
MH-2

Quant Time: Jan 12 02:20:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 12 02:16:32 2023
Response via : Initial Calibration

