



Benzaldehyde

Response Ratio



$R = 2.334e-002 A^2 + 5.557e-001 A + 1.392e-001$
Coef of Det (r^2) = 0.990551 Curve Fit: Quadratic
Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP011221.M
Calibration Table Last Updated: Tue Jan 12 15:19:32 2021


Benzoic acid



$R = 4.229e-002 A^2 + 1.236e-001 A + 4.381e-003$
 Coef of Det (r^2) = 0.995562 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP011221.M
 Calibration Table Last Updated: Tue Jan 12 15:19:32 2021


Hexachlorocyclopentadiene

Response Ratio



$R = 7.229e-002 A^2 + 1.301e-001 A + 1.547e-002$
Coef of Det (r^2) = 0.996580 Curve Fit: Quadratic
Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP011221.M
Calibration Table Last Updated: Tue Jan 12 15:19:32 2021

2,4-Dinitrophenol



$R = 4.762e-002 A^2 + 7.494e-002 A + 1.122e-002$
 Coef of Det (r^2) = 0.996488 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP011221.M
 Calibration Table Last Updated: Tue Jan 12 15:19:32 2021