



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011221\
 Data File : BP004497.D
 Acq On : 12 Jan 2021 10:07
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP011221.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jan 12 15:19:32 2021



AutoFind: Scans 2461, 2462, 2463; Background Corrected with Scan 2454

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.4	34898	PASS
68	69	0.00	2	1.6	656	PASS
69	198	0.00	100	27.6	41221	PASS
70	69	0.00	2	0.5	209	PASS
127	198	10	80	39.8	59416	PASS
197	198	0.00	2	0.3	499	PASS
198	198	100	100	100.0	149138	PASS
199	198	5	9	6.7	9930	PASS
275	198	10	60	29.4	43902	PASS
365	198	1	100	3.9	5762	PASS
441	198	0.01	100	15.7	23351	PASS
442	442	100	100	100.0	151765	PASS
443	442	15	24	19.1	29036	PASS