

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011223\
 Data File : BP013467.D
 Acq On : 12 Jan 2023 20:17
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP011223

Quant Time: Jan 13 03:50:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 13 03:32:57 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	38.956	2.6	100	0.00
3	Pyridine	40.000	40.170	-0.4	100	0.00
4	n-Nitrosodimethylamine	40.000	39.664	0.8	99	0.00
5 S	2-Fluorophenol	80.000	78.645	1.7	99	0.00
6	Aniline	40.000	40.097	-0.2	100	0.00
7 S	Phenol-d6	80.000	80.458	-0.6	100	0.00
8	2-Chlorophenol	40.000	40.028	-0.1	100	0.00
9	Benzaldehyde	40.000	38.937	2.7	106	0.00
10 C	Phenol	40.000	40.376	-0.9	100	0.00
11	bis(2-Chloroethyl)ether	40.000	40.022	-0.1	100	0.00
12	1,3-Dichlorobenzene	40.000	39.318	1.7	99	0.00
13 C	1,4-Dichlorobenzene	40.000	39.380	1.5	99	0.00
14	1,2-Dichlorobenzene	40.000	39.441	1.4	100	0.00
15	Benzyl Alcohol	40.000	40.591	-1.5	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.536	1.2	99	0.00
17	2-Methylphenol	40.000	40.339	-0.8	100	0.00
18	Hexachloroethane	40.000	39.263	1.8	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.065	-2.7	101	0.00
20	3+4-Methylphenols	40.000	40.459	-1.1	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	39.546	1.1	100	0.00
23 S	Nitrobenzene-d5	80.000	80.198	-0.2	100	0.00
24	Nitrobenzene	40.000	39.734	0.7	100	0.00
25	Isophorone	40.000	39.584	1.0	100	0.00
26 C	2-Nitrophenol	40.000	40.597	-1.5	101	0.00
27	2,4-Dimethylphenol	40.000	40.111	-0.3	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.379	1.6	100	0.00
29 C	2,4-Dichlorophenol	40.000	40.213	-0.5	100	0.00
30	1,2,4-Trichlorobenzene	40.000	39.423	1.4	101	0.00
31	Naphthalene	40.000	39.360	1.6	100	0.00
32	Benzoic acid	40.000	41.856	-4.6	100	0.00
33	4-Chloroaniline	40.000	40.085	-0.2	100	0.00
34 C	Hexachlorobutadiene	40.000	39.551	1.1	101	0.00
35	Caprolactam	40.000	39.936	0.2	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.937	0.2	100	0.00
37	2-Methylnaphthalene	40.000	39.897	0.3	101	0.00
38	1-Methylnaphthalene	40.000	39.492	1.3	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.754	0.6	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.452	-3.6	99	0.00
42 S	2,4,6-Tribromophenol	80.000	80.389	-0.5	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.095	-0.2	101	0.00
44	2,4,5-Trichlorophenol	40.000	40.252	-0.6	101	0.00
45 S	2-Fluorobiphenyl	80.000	79.825	0.2	100	0.00
46	1,1'-Biphenyl	40.000	39.545	1.1	100	0.00
47	2-Chloronaphthalene	40.000	39.587	1.0	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011223\
 Data File : BP013467.D
 Acq On : 12 Jan 2023 20:17
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP011223

Quant Time: Jan 13 03:50:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 13 03:32:57 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.963	-2.4	101	0.00
49	Acenaphthylene	40.000	39.733	0.7	101	0.00
50	Dimethylphthalate	40.000	39.651	0.9	101	0.00
51	2,6-Dinitrotoluene	40.000	40.485	-1.2	101	0.00
52 C	Acenaphthene	40.000	39.690	0.8	101	0.00
53	3-Nitroaniline	40.000	41.227	-3.1	102	0.00
54 P	2,4-Dinitrophenol	40.000	41.094	-2.7	102	0.00
55	Dibenzofuran	40.000	39.578	1.1	101	0.00
56 P	4-Nitrophenol	40.000	41.082	-2.7	100	0.00
57	2,4-Dinitrotoluene	40.000	41.077	-2.7	101	0.00
58	Fluorene	40.000	39.958	0.1	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.843	0.4	100	0.00
60	Diethylphthalate	40.000	39.852	0.4	101	0.00
61	4-Chlorophenyl-phenylether	40.000	39.637	0.9	100	0.00
62	4-Nitroaniline	40.000	41.485	-3.7	101	0.00
63	Azobenzene	40.000	39.860	0.4	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.882	-4.7	102	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.533	1.2	101	0.00
67	4-Bromophenyl-phenylether	40.000	39.648	0.9	101	0.00
68	Hexachlorobenzene	40.000	38.956	2.6	100	0.00
69	Atrazine	40.000	39.704	0.7	102	0.00
70 C	Pentachlorophenol	40.000	41.385	-3.5	102	0.00
71	Phenanthrene	40.000	39.587	1.0	101	0.00
72	Anthracene	40.000	39.838	0.4	101	0.00
73	Carbazole	40.000	39.878	0.3	101	0.00
74	Di-n-butylphthalate	40.000	41.213	-3.0	101	0.00
75 C	Fluoranthene	40.000	40.542	-1.4	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzidine	40.000	42.805	-7.0	102	0.00
78	Pyrene	40.000	40.758	-1.9	100	0.00
79 S	Terphenyl-d14	80.000	74.442	6.9	103	0.00
80	Butylbenzylphthalate	40.000	40.074	-0.2	100	0.00
81	Benzo(a)anthracene	40.000	40.420	-1.1	100	0.00
82	3,3'-Dichlorobenzidine	40.000	40.054	-0.1	101	0.00
83	Chrysene	40.000	40.887	-2.2	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.266	-0.7	102	0.00
85 c	Di-n-octyl phthalate	40.000	41.239	-3.1	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.076	2.3	97	0.00
88	Benzo(b)fluoranthene	40.000	39.919	0.2	103	-0.01
89	Benzo(k)fluoranthene	40.000	39.365	1.6	97	0.00
90 C	Benzo(a)pyrene	40.000	39.460	1.3	100	0.00
91	Dibenzo(a,h)anthracene	40.000	39.215	2.0	97	0.00
92	Benzo(g,h,i)perylene	40.000	38.743	3.1	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011223\
Data File : BP013467.D
Acq On : 12 Jan 2023 20:17
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP011223

Quant Time: Jan 13 03:50:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 13 03:32:57 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0