

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011224\
 Data File : BP019352.D
 Acq On : 13 Jan 2024 04:19
 Operator : MA/JU
 Sample : PB158383BL
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK383

Quant Time: Jan 13 04:50:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 09 03:05:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	101693	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.569	136	411779	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.422	164	316116	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.181	188	757436	20.000	ng/u1	0.00
79) Chrysene-d12	21.286	240	785693	20.000	ng/u1	0.00
88) Perylene-d12	23.639	264	816773	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.217	96	17157	6.380	ng/uL	0.00
4) Pyridine-d5	3.634	84	259398	33.356	ng/u1	0.00
7) Phenol-d5	6.964	99	302150	31.087	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.116	67	192965	33.192	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.311	132	215973	32.347	ng/u1	0.00
15) 4-Methylphenol-d8	8.505	113	243456	31.573	ng/u1	-0.01
21) Nitrobenzene-d5	8.946	128	114961	33.185	ng/u1	0.00
24) 2-Nitrophenol-d4	9.663	143	139303	34.608	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.205	165	261547	32.162	ng/u1	0.00
31) 4-Chloroaniline-d4	10.722	131	336755	32.783	ng/u1	0.00
46) Dimethylphthalate-d6	13.840	166	950496	34.034	ng/u1	0.00
49) Acenaphthylene-d8	14.116	160	977779	32.661	ng/u1	0.00
54) 4-Nitrophenol-d4	14.657	143	138787	32.362	ng/u1	-0.01
60) Fluorene-d10	15.416	176	803139	35.125	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	174801	31.259	ng/u1	0.00
73) Anthracene-d10	17.275	188	1299492	34.399	ng/u1	-0.01
81) Pyrene-d10	19.528	212	1697604	33.135	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.486	264	1588566	35.769	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011224\
Data File : BP019352.D
Acq On : 13 Jan 2024 04:19
Operator : MA/JU
Sample : PB158383BL
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK383

Quant Time: Jan 13 04:50:56 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 09 03:05:24 2024
Response via : Initial Calibration

