



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011321\
 Data File : BP004528.D
 Acq On : 13 Jan 2021 09:00
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP011221.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jan 12 15:19:32 2021



AutoFind: Scans 2462, 2463, 2464; Background Corrected with Scan 2454

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.7	32081	PASS
68	69	0.00	2	1.9	696	PASS
69	198	0.00	100	24.3	37541	PASS
70	69	0.00	2	0.6	224	PASS
127	198	10	80	38.1	58972	PASS
197	198	0.00	2	0.3	475	PASS
198	198	100	100	100.0	154649	PASS
199	198	5	9	6.7	10435	PASS
275	198	10	60	31.7	49068	PASS
365	198	1	100	4.4	6853	PASS
441	198	0.01	100	20.4	31561	PASS
442	442	100	100	100.0	203755	PASS
443	442	15	24	19.0	38715	PASS