

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011623\
 Data File : BP013474.D
 Acq On : 16 Jan 2023 10:08
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 17 03:40:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 13 03:32:57 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.446	0.458	-2.7	108	0.00
3	Pyridine	1.392	1.340	3.7	98	0.00
4	n-Nitrosodimethylamine	0.667	0.675	-1.2	104	0.00
5 S	2-Fluorophenol	1.139	1.146	-0.6	104	0.00
6	Aniline	2.151	2.104	2.2	101	0.00
7 S	Phenol-d6	1.609	1.593	1.0	101	0.00
8	2-Chlorophenol	1.318	1.298	1.5	101	0.00
9	Benzaldehyde	0.972	0.921	5.2	106	0.00
10 C	Phenol	1.661	1.651	0.6	101	0.00
11	bis(2-Chloroethyl)ether	1.347	1.316	2.3	100	0.00
12	1,3-Dichlorobenzene	1.388	1.378	0.7	102	0.00
13 C	1,4-Dichlorobenzene	1.415	1.396	1.3	102	0.00
14	1,2-Dichlorobenzene	1.382	1.342	2.9	101	0.00
15	Benzyl Alcohol	1.240	1.205	2.8	98	0.00
16	2,2'-oxybis(1-Chloropropane	2.083	2.033	2.4	101	0.00
17	2-Methylphenol	1.230	1.191	3.2	99	0.00
18	Hexachloroethane	0.501	0.498	0.6	103	0.00
19 P	n-Nitroso-di-n-propylamine	1.149	1.130	1.7	100	0.00
20	3+4-Methylphenols	1.709	1.655	3.2	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
22	Acetophenone	0.512	0.505	1.4	98	0.00
23 S	Nitrobenzene-d5	0.376	0.376	0.0	99	0.00
24	Nitrobenzene	0.389	0.387	0.5	99	0.00
25	Isophorone	0.775	0.765	1.3	99	0.00
26 C	2-Nitrophenol	0.192	0.191	0.5	97	0.00
27	2,4-Dimethylphenol	0.317	0.317	0.0	99	0.00
28	bis(2-Chloroethoxy)methane	0.457	0.454	0.7	100	0.00
29 C	2,4-Dichlorophenol	0.333	0.335	-0.6	100	0.00
30	1,2,4-Trichlorobenzene	0.352	0.348	1.1	100	0.00
31	Naphthalene	1.058	1.039	1.8	99	0.00
32	Benzoic acid	0.264	0.278	-5.3	99	0.00
33	4-Chloroaniline	0.483	0.478	1.0	98	0.00
34 C	Hexachlorobutadiene	0.222	0.222	0.0	101	0.00
35	Caprolactam	0.118	0.118	0.0	100	0.00
36 C	4-Chloro-3-methylphenol	0.362	0.361	0.3	99	0.00
37	2-Methylnaphthalene	0.794	0.786	1.0	99	0.00
38	1-Methylnaphthalene	0.749	0.742	0.9	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.656	0.651	0.8	99	0.00
41 P	Hexachlorocyclopentadiene	0.342	0.354	-3.5	99	0.00
42 S	2,4,6-Tribromophenol	0.312	0.317	-1.6	102	0.00
43 C	2,4,6-Trichlorophenol	0.451	0.449	0.4	100	0.00
44	2,4,5-Trichlorophenol	0.529	0.535	-1.1	101	0.00
45 S	2-Fluorobiphenyl	1.368	1.368	0.0	100	0.00
46	1,1'-Biphenyl	1.512	1.494	1.2	99	0.00
47	2-Chloronaphthalene	1.160	1.138	1.9	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011623\
 Data File : BP013474.D
 Acq On : 16 Jan 2023 10:08
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 17 03:40:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 13 03:32:57 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.382	0.383	-0.3	99	0.00
49	Acenaphthylene	1.912	1.891	1.1	100	0.00
50	Dimethylphthalate	1.586	1.588	-0.1	102	0.00
51	2,6-Dinitrotoluene	0.346	0.350	-1.2	101	0.00
52 C	Acenaphthene	1.132	1.118	1.2	100	0.00
53	3-Nitroaniline	0.370	0.377	-1.9	100	0.00
54 P	2,4-Dinitrophenol	0.235	0.243	-3.4	102	0.00
55	Dibenzofuran	1.877	1.865	0.6	101	0.00
56 P	4-Nitrophenol	0.297	0.310	-4.4	102	0.00
57	2,4-Dinitrotoluene	0.473	0.488	-3.2	101	0.00
58	Fluorene	1.481	1.490	-0.6	101	0.00
59	2,3,4,6-Tetrachlorophenol	0.457	0.460	-0.7	101	0.00
60	Diethylphthalate	1.552	1.566	-0.9	102	0.00
61	4-Chlorophenyl-phenylether	0.787	0.800	-1.7	102	0.00
62	4-Nitroaniline	0.379	0.396	-4.5	101	0.00
63	Azobenzene	1.425	1.447	-1.5	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	0.137	0.142	-3.6	102	0.00
66 c	n-Nitrosodiphenylamine	0.588	0.581	1.2	102	0.00
67	4-Bromophenyl-phenylether	0.230	0.228	0.9	102	0.00
68	Hexachlorobenzene	0.251	0.248	1.2	102	0.00
69	Atrazine	0.215	0.217	-0.9	104	0.00
70 C	Pentachlorophenol	0.184	0.191	-3.8	103	0.00
71	Phenanthrene	1.055	1.047	0.8	102	0.00
72	Anthracene	1.074	1.078	-0.4	103	0.00
73	Carbazole	0.978	0.984	-0.6	103	0.00
74	Di-n-butylphthalate	1.135	1.195	-5.3	104	0.00
75 C	Fluoranthene	1.268	1.307	-3.1	104	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
77	Benzydine	0.594	0.633	-6.6	106	0.00
78	Pyrene	1.315	1.321	-0.5	104	0.00
79 S	Terphenyl-d14	0.978	0.905	7.5	107	0.00
80	Butylbenzylphthalate	0.569	0.560	1.6	103	0.00
81	Benzo(a)anthracene	1.347	1.353	-0.4	104	0.00
82	3,3'-Dichlorobenzidine	0.503	0.499	0.8	105	0.00
83	Chrysene	1.273	1.278	-0.4	104	0.00
84	Bis(2-ethylhexyl)phthalate	0.823	0.815	1.0	104	0.00
85 c	Di-n-octyl phthalate	1.397	1.413	-1.1	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	1.446	1.372	5.1	94	0.01
88	Benzo(b)fluoranthene	1.187	1.174	1.1	102	0.00
89	Benzo(k)fluoranthene	1.153	1.185	-2.8	100	0.00
90 C	Benzo(a)pyrene	1.167	1.156	0.9	100	0.00
91	Dibenzo(a,h)anthracene	1.204	1.144	5.0	94	0.01
92	Benzo(g,h,i)perylene	1.188	1.103	7.2	92	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011623\
Data File : BP013474.D
Acq On : 16 Jan 2023 10:08
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jan 17 03:40:09 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 13 03:32:57 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0