

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011723\
 Data File : BP013489.D
 Acq On : 17 Jan 2023 10:44
 Operator : CG/JU
 Sample : PB150211BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB150211BL

Quant Time: Jan 18 03:10:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 13 03:32:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.904	152	403128	20.000	ng	0.00
21) Naphthalene-d8	10.716	136	1656038	20.000	ng	0.00
39) Acenaphthene-d10	14.557	164	1122407	20.000	ng	0.00
64) Phenanthrene-d10	17.316	188	2675734	20.000	ng	0.00
76) Chrysene-d12	21.404	240	2800932	20.000	ng	0.00
86) Perylene-d12	23.868	264	3127506	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.463	112	3447816	150.238	ng	0.00
7) Phenol-d6	7.069	99	4530593	139.664	ng	0.00
23) Nitrobenzene-d5	9.075	82	2630232	84.381	ng	0.00
42) 2,4,6-Tribromophenol	16.051	330	2337459	133.415	ng	0.00
45) 2-Fluorobiphenyl	13.175	172	6744103	87.836	ng	0.00
79) Terphenyl-d14	19.869	244	11807421	86.184	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011723\
Data File : BP013489.D
Acq On : 17 Jan 2023 10:44
Operator : CG/JU
Sample : PB150211BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB150211BL

Quant Time: Jan 18 03:10:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 13 03:32:57 2023
Response via : Initial Calibration

