

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011822\
 Data File : BP008834.D
 Acq On : 18 Jan 2022 11:23
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 19 03:35:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 14 18:05:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	38.355	4.1	112	0.00
3	Pyridine	40.000	43.293	-8.2	119	0.00
4	n-Nitrosodimethylamine	40.000	43.128	-7.8	119	0.00
5 S	2-Fluorophenol	80.000	81.107	-1.4	114	0.00
6	Aniline	40.000	43.552	-8.9	119	0.00
7 S	Phenol-d6	80.000	86.784	-8.5	118	0.00
8	2-Chlorophenol	40.000	41.909	-4.8	115	0.00
9	Benzaldehyde	40.000	40.913	-2.3	113	0.00
10 C	Phenol	40.000	43.281	-8.2	118	0.00
11	bis(2-Chloroethyl)ether	40.000	42.470	-6.2	117	0.00
12	1,3-Dichlorobenzene	40.000	40.182	-0.5	112	0.00
13 C	1,4-Dichlorobenzene	40.000	40.510	-1.3	113	0.00
14	1,2-Dichlorobenzene	40.000	40.820	-2.1	114	0.00
15	Benzyl Alcohol	40.000	45.981	-15.0	123	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.309	-8.3	118	0.00
17	2-Methylphenol	40.000	44.207	-10.5	120	0.00
18	Hexachloroethane	40.000	41.266	-3.2	113	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	46.847	-17.1	124	0.00
20	3+4-Methylphenols	40.000	45.578	-13.9	122	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	41.424	-3.6	122	0.00
23 S	Nitrobenzene-d5	80.000	83.310	-4.1	122	0.00
24	Nitrobenzene	40.000	41.383	-3.5	122	0.00
25	Isophorone	40.000	44.179	-10.4	128	0.00
26 C	2-Nitrophenol	40.000	42.700	-6.8	123	0.00
27	2,4-Dimethylphenol	40.000	42.279	-5.7	123	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.135	-5.3	124	0.00
29 C	2,4-Dichlorophenol	40.000	41.895	-4.7	122	0.00
30	1,2,4-Trichlorobenzene	40.000	39.323	1.7	118	0.00
31	Naphthalene	40.000	40.153	-0.4	120	0.00
32	Benzoic acid	40.000	48.824	-22.1	133	0.01
33	4-Chloroaniline	40.000	43.244	-8.1	126	0.00
34 C	Hexachlorobutadiene	40.000	38.883	2.8	116	0.00
35	Caprolactam	40.000	50.366	-25.9#	145	0.01
36 C	4-Chloro-3-methylphenol	40.000	45.310	-13.3	131	0.00
37	2-Methylnaphthalene	40.000	42.413	-6.0	125	0.00
38	1-Methylnaphthalene	40.000	42.696	-6.7	125	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	117	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.745	5.6	124	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.211	7.0	113	0.00
42 S	2,4,6-Tribromophenol	80.000	88.269	-10.3	144	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.757	-1.9	131	0.00
44	2,4,5-Trichlorophenol	40.000	40.988	-2.5	132	0.00
45 S	2-Fluorobiphenyl	80.000	76.270	4.7	127	0.00
46	1,1'-Biphenyl	40.000	38.751	3.1	128	0.00
47	2-Chloronaphthalene	40.000	38.805	3.0	128	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011822\
 Data File : BP008834.D
 Acq On : 18 Jan 2022 11:23
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 19 03:35:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 14 18:05:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.834	-12.1	143	0.00
49	Acenaphthylene	40.000	40.313	-0.8	132	0.00
50	Dimethylphthalate	40.000	41.852	-4.6	138	0.00
51	2,6-Dinitrotoluene	40.000	43.472	-8.7	140	0.00
52 C	Acenaphthene	40.000	40.523	-1.3	133	0.00
53	3-Nitroaniline	40.000	45.330	-13.3	146	0.00
54 P	2,4-Dinitrophenol	40.000	50.294	-25.7#	156	0.00
55	Dibenzofuran	40.000	40.745	-1.9	134	0.00
56 P	4-Nitrophenol	40.000	48.310	-20.8	152	0.00
57	2,4-Dinitrotoluene	40.000	47.098	-17.7	149	0.00
58	Fluorene	40.000	42.007	-5.0	137	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.326	-8.3	140	0.00
60	Diethylphthalate	40.000	43.530	-8.8	143	0.00
61	4-Chlorophenyl-phenylether	40.000	41.927	-4.8	136	0.00
62	4-Nitroaniline	40.000	48.053	-20.1	154	0.00
63	Azobenzene	40.000	44.248	-10.6	142	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	132	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.676	-14.2	157	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.533	1.2	144	0.00
67	4-Bromophenyl-phenylether	40.000	39.306	1.7	141	0.00
68	Hexachlorobenzene	40.000	38.655	3.4	141	0.00
69	Atrazine	40.000	42.565	-6.4	155	0.00
70 C	Pentachlorophenol	40.000	42.686	-6.7	153	0.00
71	Phenanthrene	40.000	40.053	-0.1	148	0.00
72	Anthracene	40.000	41.313	-3.3	151	0.00
73	Carbazole	40.000	42.614	-6.5	158	0.00
74	Di-n-butylphthalate	40.000	43.049	-7.6	155	0.00
75 C	Fluoranthene	40.000	43.333	-8.3	164	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	152	0.00
77	Benzydine	40.000	44.864	-12.2	170	0.00
78	Pyrene	40.000	40.590	-1.5	166	0.00
79 S	Terphenyl-d14	80.000	77.008	3.7	158	0.00
80	Butylbenzylphthalate	40.000	42.359	-5.9	174	0.00
81	Benzo(a)anthracene	40.000	40.368	-0.9	172	0.00
82	3,3'-Dichlorobenzidine	40.000	41.641	-4.1	174	0.00
83	Chrysene	40.000	39.995	0.0	170	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.819	-2.0	167	0.00
85 c	Di-n-octyl phthalate	40.000	40.565	-1.4	183	0.00
86 I	Perylene-d12	20.000	20.000	0.0	137	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	34.148	14.6	132	0.00
88	Benzo(b)fluoranthene	40.000	43.596	-9.0	169	0.00
89	Benzo(k)fluoranthene	40.000	42.596	-6.5	170	0.00
90 C	Benzo(a)pyrene	40.000	41.603	-4.0	162	0.00
91	Dibenzo(a,h)anthracene	40.000	34.008	15.0	131	0.00
92	Benzo(g,h,i)perylene	40.000	32.101	19.7	123	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011822\
Data File : BP008834.D
Acq On : 18 Jan 2022 11:23
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jan 19 03:35:06 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 14 18:05:00 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0