



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011822\
 Data File : BP008833.D
 Acq On : 18 Jan 2022 10:15
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 14 18:05:00 2022






AutoFind: Scans 2500, 2501, 2502; Background Corrected with Scan 2488

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.7	125156	PASS
68	69	0.00	2	1.6	2208	PASS
69	198	0.00	100	28.4	138125	PASS
70	69	0.00	2	0.5	681	PASS
127	198	10	80	40.7	198219	PASS
197	198	0.00	2	0.3	1641	PASS
198	198	100	100	100.0	486677	PASS
199	198	5	9	6.7	32749	PASS
275	198	10	60	27.9	135836	PASS
365	198	1	100	3.5	17216	PASS
441	198	0.01	100	10.6	51528	PASS
442	442	100	100	100.0	334052	PASS
443	442	15	24	19.4	64699	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011822\
 Data File : BP008833.D
 Acq On : 18 Jan 2022 10:15
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Jan 19 04:05:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 14 18:05:00 2022
 Response via : Initial Calibration



TIC: BP008833.D\data.ms

(70) Pentachlorophenol (C)

16.904min (-0.006) 281380.81 ng




response 546885

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.60	63.91
264.00	62.90	63.76
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011822\
Data File : BP008833.D
Acq On : 18 Jan 2022 10:15
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Jan 19 04:05:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 14 18:05:00 2022
Response via : Initial Calibration



TIC: BP008833.D\data.ms

(77) Benzidine

19.439min (+ 0.000) 116283.68 ng

response 2986549

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	13.93
183.00	11.90	11.76
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/18/2022	BNA_P	<u>BP008833.D</u>
Compound Name	Response	Retention Time
DDT	1635943	20.645
DDD	45106	20.216
DDE	2899	19.722
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
48005	1683948	2.85

Instrument :
 BNA_P
ClientSampleId :
 DFTPP