

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012025\
 Data File : BP023690.D
 Acq On : 21 Jan 2025 19:29
 Operator : RC/JU
 Sample : Q1108-14
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YE8E3

Quant Time: Jan 21 22:04:44 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP011425.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 20 21:50:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	649538	20.000	ng/ul	0.00
20) Naphthalene-d8	10.581	136	2548431	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.422	164	1506432	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.216	188	2933734	20.000	ng/ul	-0.01
79) Chrysene-d12	21.651	240	2912190	20.000	ng/ul	-0.01
88) Perylene-d12	25.039	264	3066360	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.317	96	42018	2.385	ng/uL	0.00
4) Pyridine-d5	3.723	84	438564	8.930	ng/ul	0.00
7) Phenol-d5	6.963	99	645106	11.536	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	449558	13.723	ng/ul	0.00
11) 2-Chlorophenol-d4	7.340	132	582394	13.188	ng/ul	0.00
15) 4-Methylphenol-d8	8.493	113	324894	7.490	ng/ul	0.00
21) Nitrobenzene-d5	8.946	128	271552	13.961	ng/ul	0.00
24) 2-Nitrophenol-d4	9.663	143	222016	17.389	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.199	165	508988	13.106	ng/ul	0.00
31) 4-Chloroaniline-d4	10.704	131	616976	9.952	ng/ul	0.00
46) Dimethylphthalate-d6	13.822	166	1830722	15.615	ng/ul	0.00
49) Acenaphthylene-d8	14.110	160	1908203	13.682	ng/ul	0.00
54) 4-Nitrophenol-d4	14.604	143	199030	10.213	ng/ul	0.00
60) Fluorene-d10	15.422	176	1371230	13.514	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.534	200	90864	9.519	ng/ul	-0.01
73) Anthracene-d10	17.310	188	2203034	13.911	ng/ul	-0.02
81) Pyrene-d10	19.675	212	2145576	13.299	ng/ul	-0.02
92) Benzo(a)pyrene-d12	24.798	264	1795957	10.777	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012025\
Data File : BP023690.D
Acq On : 21 Jan 2025 19:29
Operator : RC/JU
Sample : Q1108-14
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
YE8E3

Quant Time: Jan 21 22:04:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP011425.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 20 21:50:54 2025
Response via : Initial Calibration

