

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012025\
 Data File : BP023686.D
 Acq On : 21 Jan 2025 16:48
 Operator : RC/JU
 Sample : PB166145BL
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK145

Quant Time: Jan 21 22:36:35 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP011425.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 20 21:50:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	635264	20.000	ng/u1	0.00
20) Naphthalene-d8	10.581	136	2437287	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.422	164	1375367	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.228	188	2717957	20.000	ng/u1	0.00
79) Chrysene-d12	21.651	240	2840008	20.000	ng/u1	-0.01
88) Perylene-d12	25.033	264	3130353	20.000	ng/u1	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.311	96	56076	3.255	ng/uL	0.00
4) Pyridine-d5	3.717	84	779161	16.222	ng/u1	0.00
7) Phenol-d5	6.964	99	867919	15.869	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	504322	15.741	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	700811	16.225	ng/u1	0.00
15) 4-Methylphenol-d8	8.493	113	647749	15.269	ng/u1	0.00
21) Nitrobenzene-d5	8.946	128	305408	16.417	ng/u1	0.00
24) 2-Nitrophenol-d4	9.663	143	246669	20.201	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.199	165	557874	15.020	ng/u1	0.00
31) 4-Chloroaniline-d4	10.705	131	820606	13.840	ng/u1	0.00
46) Dimethylphthalate-d6	13.828	166	1591341	14.867	ng/u1	0.00
49) Acenaphthylene-d8	14.116	160	1947994	15.298	ng/u1	0.00
54) 4-Nitrophenol-d4	14.604	143	217755	12.239	ng/u1	0.00
60) Fluorene-d10	15.434	176	1434932	15.490	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.540	200	86383	9.768	ng/u1	0.00
73) Anthracene-d10	17.322	188	2275955	15.512	ng/u1	0.00
81) Pyrene-d10	19.686	212	2601995	16.538	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.810	264	2653933	15.601	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012025\
Data File : BP023686.D
Acq On : 21 Jan 2025 16:48
Operator : RC/JU
Sample : PB166145BL
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK145

Quant Time: Jan 21 22:36:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP011425.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 20 21:50:54 2025
Response via : Initial Calibration

