

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012025\
 Data File : BP023693.D
 Acq On : 21 Jan 2025 21:31
 Operator : RC/JU
 Sample : Q1108-17
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YE8E6

Quant Time: Jan 21 22:05:28 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP011425.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 20 21:50:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	638110	20.000	ng/u1	0.00
20) Naphthalene-d8	10.581	136	2431877	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.422	164	1332982	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	2458604	20.000	ng/u1	-0.02
79) Chrysene-d12	21.657	240	2646396	20.000	ng/u1	0.00
88) Perylene-d12	25.039	264	2923439	20.000	ng/u1	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.317	96	47967	2.772	ng/uL	0.00
4) Pyridine-d5	3.723	84	137657	2.853	ng/u1	0.00
7) Phenol-d5	6.964	99	852242	15.513	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	524262	16.290	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	683197	15.747	ng/u1	0.00
15) 4-Methylphenol-d8	8.493	113	652399	15.310	ng/u1	0.00
21) Nitrobenzene-d5	8.946	128	311043	16.757	ng/u1	0.00
24) 2-Nitrophenol-d4	9.663	143	230161	18.891	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.199	165	548556	14.802	ng/u1	0.00
31) 4-Chloroaniline-d4	10.704	131	866667	14.649	ng/u1	0.00
46) Dimethylphthalate-d6	13.828	166	1562463	15.061	ng/u1	0.00
49) Acenaphthylene-d8	14.116	160	1996057	16.174	ng/u1	0.00
54) 4-Nitrophenol-d4	14.598	143	166091	9.632	ng/u1	0.00
60) Fluorene-d10	15.428	176	1460245	16.264	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.534	200	42042	5.255	ng/u1	-0.01
73) Anthracene-d10	17.316	188	2139211	16.118	ng/u1	-0.01
81) Pyrene-d10	19.680	212	2436006	16.616	ng/u1	-0.02
92) Benzo(a)pyrene-d12	24.798	264	1839120	11.576	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012025\
Data File : BP023693.D
Acq On : 21 Jan 2025 21:31
Operator : RC/JU
Sample : Q1108-17
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
YE8E6

Quant Time: Jan 21 22:05:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP011425.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 20 21:50:54 2025
Response via : Initial Calibration

