

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012324\
 Data File : BP019428.D
 Acq On : 23 Jan 2024 19:34
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Jan 24 00:06:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 22 11:49:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	161909	20.000	ng/u1	0.00
20) Naphthalene-d8	10.563	136	666412	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.416	164	447599	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.169	188	1021756	20.000	ng/u1	0.00
79) Chrysene-d12	21.274	240	831443	20.000	ng/u1	0.00
88) Perylene-d12	23.621	264	847350	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.216	96	25720	6.007	ng/uL	0.00
4) Pyridine-d5	3.634	84	191159	15.439	ng/u1	0.00
7) Phenol-d5	6.957	99	258194	16.685	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.110	67	159628	17.246	ng/u1	0.00
11) 2-Chlorophenol-d4	7.304	132	192897	18.146	ng/u1	0.00
15) 4-Methylphenol-d8	8.499	113	213582	17.397	ng/u1	0.00
21) Nitrobenzene-d5	8.940	128	100287	17.888	ng/u1	0.00
24) 2-Nitrophenol-d4	9.657	143	120809	18.545	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.198	165	240783	18.295	ng/u1	0.00
31) 4-Chloroaniline-d4	10.716	131	274338	16.502	ng/u1	0.00
46) Dimethylphthalate-d6	13.834	166	738959	18.687	ng/u1	0.00
49) Acenaphthylene-d8	14.104	160	783589	18.485	ng/u1	0.00
54) 4-Nitrophenol-d4	14.657	143	91058	14.995	ng/u1	0.00
60) Fluorene-d10	15.410	176	582525	17.993	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.551	200	130677	17.323	ng/u1	0.00
73) Anthracene-d10	17.269	188	933832	18.325	ng/u1	0.00
81) Pyrene-d10	19.516	212	1052252	19.409	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.468	264	810562	17.592	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.252	88	28182	6.354	ng/uL	95
5) Pyridine	3.652	79	197933	15.737	ng/u1	93
6) Benzaldehyde	6.922	77	150827	22.357	ng/u1	96
8) Phenol	6.987	94	260551	16.713	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.204	93	217306	17.377	ng/u1	98
12) 2-Chlorophenol	7.340	128	197395	17.966	ng/u1	96
13) 2-Methylphenol	8.234	108	202874	17.349	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.299	45	164457	11.052	ng/u1	98
16) Acetophenone	8.604	105	347692	17.780	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.587	70	188449	18.495	ng/u1	98
18) 4-Methylphenol	8.563	108	215222	17.289	ng/u1	97
19) Hexachloroethane	8.834	117	91477	18.433	ng/u1	97
22) Nitrobenzene	8.981	77	280834	17.906	ng/u1	95
23) Isophorone	9.504	82	539617	17.747	ng/u1	99
25) 2-Nitrophenol	9.687	139	123493	18.047	ng/u1	97
26) 2,4-Dimethylphenol	9.757	107	261052	17.694	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.993	93	306353	17.518	ng/u1	99
29) 2,4-Dichlorophenol	10.222	162	232622	18.341	ng/u1	98
30) Naphthalene	10.616	128	669200	17.821	ng/u1	99
32) 4-Chloroaniline	10.740	127	272751	16.770	ng/u1	100
33) Hexachlorobutadiene	10.887	225	209838	19.176	ng/u1	99
34) Caprolactam	11.534	113	66449	16.706	ng/u1	91
35) 4-Chloro-3-methylphenol	11.881	107	251956	17.893	ng/u1	98
36) 2-Methylnaphthalene	12.228	142	493193	18.051	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012324\
 Data File : BP019428.D
 Acq On : 23 Jan 2024 19:34
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Jan 24 00:06:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 22 11:49:04 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.451	142	479831	17.949	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.598	216	376107	19.390	ng/ul	99
40) Hexachlorocyclopentadiene	12.569	237	228629	18.365	ng/ul	98
41) 2,4,6-Trichlorophenol	12.851	196	228706	19.160	ng/ul	99
42) 2,4,5-Trichlorophenol	12.928	196	241257	18.860	ng/ul	100
43) 1,1'-Biphenyl	13.245	154	682075	18.447	ng/ul	100
44) 2-Chloronaphthalene	13.287	162	561833	18.789	ng/ul	98
45) 2-Nitroaniline	13.510	65	159559	18.603	ng/ul	95
47) Dimethylphthalate	13.881	163	729756	18.568	ng/ul	100
48) 2,6-Dinitrotoluene	14.010	165	147890	19.030	ng/ul	97
50) Acenaphthylene	14.134	152	860395	18.531	ng/ul	99
51) 3-Nitroaniline	14.339	138	113964	16.812	ng/ul	100
52) Acenaphthene	14.481	153	542212	17.824	ng/ul	99
53) 2,4-Dinitrophenol	14.557	184	71412	14.014	ng/ul	96
55) 4-Nitrophenol	14.669	109	100750	15.364	ng/ul	97
56) Dibenzofuran	14.816	168	788222	18.098	ng/ul	99
57) 2,4-Dinitrotoluene	14.804	165	205568	18.853	ng/ul	92
58) 2,3,4,6-Tetrachlorophenol	15.045	232	215989	19.020	ng/ul	99
59) Diethylphthalate	15.245	149	673269	18.307	ng/ul	99
61) Fluorene	15.469	166	644051	18.054	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.463	204	394121	18.513	ng/ul	99
63) 4-Nitroaniline	15.510	138	91676	17.445	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.569	198	137423	17.188	ng/ul	95
67) N-Nitrosodiphenylamine	15.681	169	542618	18.015	ng/ul	99
68) 4-Bromophenyl-phenylether	16.357	248	259579	18.599	ng/ul	95
69) Hexachlorobenzene	16.469	284	300326	18.672	ng/ul	97
70) Atrazine	16.645	200	243874	18.574	ng/ul	99
71) Pentachlorophenol	16.828	266	180846	18.103	ng/ul	99
72) Phenanthrene	17.216	178	1021134	17.792	ng/ul	99
74) Anthracene	17.304	178	1052452	18.089	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.210	216	363975	19.239	ng/ul	98
76) Pentachlorobenzene	14.733	250	347986	18.935	ng/ul	99
77) Carbazole	17.586	167	859723	17.368	ng/ul	99
78) Di-n-butylphthalate	18.133	149	1111123	18.757	ng/ul	99
80) Fluoranthene	19.186	202	1244485	19.419	ng/ul	99
82) Pyrene	19.545	202	1241370	19.085	ng/ul	97
83) Butylbenzylphthalate	20.421	149	428867	19.218	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.198	252	329538	17.456	ng/ul	100
85) Benzo(a)anthracene	21.257	228	1134183	17.523	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.180	149	587507	18.341	ng/ul	99
87) Chrysene	21.310	228	1026791	17.355	ng/ul	99
89) Di-n-octyl phthalate	22.086	149	879869	15.213	ng/ul	100
90) Benzo(b)fluoranthene	22.904	252	1007318	16.635	ng/ul	99
91) Benzo(k)fluoranthene	22.951	252	990674	16.864	ng/ul	100
93) Benzo(a)pyrene	23.515	252	939622	17.268	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.056	276	1180536	18.345	ng/ul	98
95) Dibenzo(a,h)anthracene	26.068	278	979646	18.390	ng/ul	99
96) Benzo(g,h,i)perylene	26.803	276	960575	18.587	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012324\
 Data File : BP019428.D
 Acq On : 23 Jan 2024 19:34
 Operator : MA/JU
 Sample : SSTCCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Jan 24 00:06:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 22 11:49:04 2024
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 6.354 ng/uL
 RT: 3.252 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BP019428.D
 Acq: 23 Jan 2024 19:34




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion	88	Ratio	100	Lower	Upper
88	100				
43	27.0		22.0	33.0	
58	63.5		54.6	82.0	



#5
 Pyridine
 Concen: 15.737 ng/uL
 RT: 3.652 min Scan# 96
 Delta R.T. -0.000 min
 Lab File: BP019428.D
 Acq: 23 Jan 2024 19:34

Tgt Ion	79	Ratio	100	Lower	Upper
79	100				
52	71.9		50.7	76.1	
51	34.0		26.8	40.2	









#6
Benzaldehyde
Concen: 22.357 ng/ul
RT: 6.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 150827
Ion Ratio Lower Upper
77 100
105 92.4 75.3 112.9
106 84.4 71.5 107.3



#8
Phenol
Concen: 16.713 ng/ul
RT: 6.987 min Scan# 663
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34


Tgt Ion: 94 Resp: 260551
Ion Ratio Lower Upper
94 100
65 30.0 25.3 37.9
66 45.5 36.8 55.2








#10
 Bis(2-Chloroethyl)ether
 Concen: 17.377 ng/ul
 RT: 7.204 min Scan# 700
 Delta R.T. -0.000 min
 Lab File: BP019428.D
 Acq: 23 Jan 2024 19:34

Instrument :
 BNA_P
 ClientSampleId :







Tgt Ion: 93 Resp: 217306
 Ion Ratio Lower Upper
 93 100
 63 74.3 60.7 91.1
 95 31.7 26.2 39.2



#12
 2-Chlorophenol
 Concen: 17.966 ng/ul
 RT: 7.340 min Scan# 723
 Delta R.T. -0.000 min
 Lab File: BP019428.D
 Acq: 23 Jan 2024 19:34

Tgt Ion:128 Resp: 197395
 Ion Ratio Lower Upper
 128 100
 64 49.0 41.3 61.9
 130 29.5 25.9 38.9









#13
2-Methylphenol
Concen: 17.349 ng/ul
RT: 8.234 min Scan# 875
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 202874
Ion Ratio Lower Upper
108 100
107 88.7 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 11.052 ng/ul
RT: 8.299 min Scan# 886
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion: 45 Resp: 164457
Ion Ratio Lower Upper
45 100
77 18.7 14.2 21.4
79 14.1 10.5 15.7









#16
Acetophenone
Concen: 17.780 ng/ul
RT: 8.604 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 347692
Ion Ratio Lower Upper
105 100
77 81.3 66.1 99.1
51 26.6 20.9 31.3



#17
N-Nitroso-di-n-propylamine
Concen: 18.495 ng/ul
RT: 8.587 min Scan# 935
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion: 70 Resp: 188449
Ion Ratio Lower Upper
70 100
42 49.6 38.5 57.7
101 9.3 7.9 11.9
130 18.1 14.3 21.5









#18
4-Methylphenol
Concen: 17.289 ng/ul
RT: 8.563 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 215222
Ion Ratio Lower Upper
108 100
107 112.9 92.6 139.0



#19
Hexachloroethane
Concen: 18.433 ng/ul
RT: 8.834 min Scan# 977
Delta R.T. -0.006 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion:117 Resp: 91477
Ion Ratio Lower Upper
117 100
201 130.5 101.8 152.6
199 81.6 62.6 93.8









#22
Nitrobenzene
Concen: 17.906 ng/ul
RT: 8.981 min Scan# 1002
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 280834
Ion Ratio Lower Upper
77 100
123 40.8 30.0 45.0
65 14.3 10.5 15.7



#23
Isophorone
Concen: 17.747 ng/ul
RT: 9.504 min Scan# 1091
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion: 82 Resp: 539617
Ion Ratio Lower Upper
82 100
95 8.5 6.6 10.0
138 15.9 13.0 19.4









#25
2-Nitrophenol
Concen: 18.047 ng/ul
RT: 9.687 min Scan# 1122
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 123493
Ion Ratio Lower Upper
139 100
65 59.2 45.6 68.4
109 43.4 32.6 49.0



#26
2,4-Dimethylphenol
Concen: 17.694 ng/ul
RT: 9.757 min Scan# 1134
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion:107 Resp: 261052
Ion Ratio Lower Upper
107 100
121 45.9 35.3 52.9
122 74.8 60.6 91.0





#27

Bis(2-Chloroethoxy)methane

Concen: 17.518 ng/ul

RT: 9.993 min Scan# 1174

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion: 93 Resp: 306353

Ion Ratio Lower Upper

93 100

95 32.1 25.9 38.9

123 10.6 7.9 11.9



#29

2,4-Dichlorophenol

Concen: 18.341 ng/ul

RT: 10.222 min Scan# 1213

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:162 Resp: 232622


Ion Ratio Lower Upper

162 100

164 67.0 52.3 78.5

98 37.1 28.3 42.5





#30

Naphthalene

Concen: 17.821 ng/ul

RT: 10.616 min Scan# 1279

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:128 Resp: 669200

Ion Ratio Lower Upper

128 100

129 11.2 8.6 12.8

127 13.5 10.6 15.8



#32

4-Chloroaniline

Concen: 16.770 ng/ul

RT: 10.740 min Scan# 1301

Delta R.T. -0.000 min

Lab File: BP019428.D




Acq: 23 Jan 2024 19:34


Tgt Ion:127 Resp: 272751

Ion Ratio Lower Upper

127 100

129 32.1 25.8 38.8





#33

Hexachlorobutadiene

Concen: 19.176 ng/ul

RT: 10.887 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:225 Resp: 209838

Ion Ratio Lower Upper

225 100

223 63.3 49.9 74.9

227 64.5 51.0 76.6



#34

Caprolactam

Concen: 16.706 ng/ul

RT: 11.534 min Scan# 1436

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:113 Resp: 66449


Ion Ratio Lower Upper

113 100

55 156.6 133.3 199.9

56 121.7 108.0 162.0









#35
4-Chloro-3-methylphenol
Concen: 17.893 ng/ul
RT: 11.881 min Scan# 1495
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 251956
Ion Ratio Lower Upper
107 100
144 23.2 17.8 26.8
142 72.2 56.3 84.5



#36
2-Methylnaphthalene
Concen: 18.051 ng/ul
RT: 12.228 min Scan# 1554
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34


Tgt Ion:142 Resp: 493193
Ion Ratio Lower Upper
142 100
141 90.5 71.1 106.7








#37
1-Methylnaphthalene
Concen: 17.949 ng/ul
RT: 12.451 min Scan# 1592
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:142 Resp: 479831
Ion Ratio Lower Upper
142 100
141 94.8 74.4 111.6
116 4.8 3.9 5.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.390 ng/ul
RT: 12.598 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion:216 Resp: 376107
Ion Ratio Lower Upper
216 100
214 79.3 63.6 95.4
179 19.5 15.1 22.7
108 12.8 11.3 16.9





#40

Hexachlorocyclopentadiene

Concen: 18.365 ng/ul

RT: 12.569 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:237 Resp: 228629

Ion Ratio Lower Upper

237 100

235 61.6 50.0 75.0

272 12.9 11.4 17.2



#41

2,4,6-Trichlorophenol

Concen: 19.160 ng/ul

RT: 12.851 min Scan# 1660

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:196 Resp: 228706


Ion Ratio Lower Upper

196 100

198 98.5 78.2 117.2

200 32.1 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 18.860 ng/ul

RT: 12.928 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:196 Resp: 241257

Ion Ratio Lower Upper

196 100

198 98.5 78.9 118.3

200 31.6 25.0 37.6



#43

1,1'-Biphenyl

Concen: 18.447 ng/ul

RT: 13.245 min Scan# 1727

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:154 Resp: 682075


Ion Ratio Lower Upper

154 100

153 40.8 32.6 48.8

76 11.8 9.3 13.9









#44
 2-Chloronaphthalene
 Concen: 18.789 ng/ul
 RT: 13.287 min Scan# 1734
 Delta R.T. -0.000 min
 Lab File: BP019428.D
 Acq: 23 Jan 2024 19:34




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:	162	Resp:	561833
Ion Ratio	Lower	Upper	
162	100		
164	32.9	26.6	40.0
127	38.5	29.4	44.2



#45
 2-Nitroaniline
 Concen: 18.603 ng/ul
 RT: 13.510 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: BP019428.D
 Acq: 23 Jan 2024 19:34

Tgt Ion:	65	Resp:	159559
Ion Ratio	Lower	Upper	
65	100		
92	70.2	52.1	78.1
138	96.6	74.8	112.2









#47
Dimethylphthalate
Concen: 18.568 ng/ul
RT: 13.881 min Scan# 1835
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 729756
Ion Ratio Lower Upper
163 100
194 4.5 3.4 5.2
164 10.2 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 19.030 ng/ul
RT: 14.010 min Scan# 1857
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion:165 Resp: 147890
Ion Ratio Lower Upper
165 100
89 59.6 49.9 74.9
121 19.4 16.5 24.7





#50

Acenaphthylene

Concen: 18.531 ng/ul

RT: 14.134 min Scan# 1878

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:152 Resp: 860395

Ion Ratio Lower Upper

152 100

151 19.9 15.8 23.6

153 13.1 10.3 15.5



#51

3-Nitroaniline

Concen: 16.812 ng/ul

RT: 14.339 min Scan# 1913

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:138 Resp: 113964


Ion Ratio Lower Upper

138 100

108 12.5 10.6 15.8

92 132.5 105.8 158.6





#52

Acenaphthene

Concen: 17.824 ng/ul

RT: 14.481 min Scan# 1937

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:153 Resp: 542212

Ion Ratio Lower Upper

153 100

152 47.0 38.2 57.2

154 87.2 68.9 103.3



#53

2,4-Dinitrophenol

Concen: 14.014 ng/ul

RT: 14.557 min Scan# 1950

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:184 Resp: 71412


Ion Ratio Lower Upper

184 100

63 58.1 43.6 65.4

154 68.1 52.1 78.1









#55
4-Nitrophenol
Concen: 15.364 ng/ul
RT: 14.669 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 100750
Ion Ratio Lower Upper
109 100
139 93.8 73.2 109.8
65 110.2 90.9 136.3



#56
Dibenzofuran
Concen: 18.098 ng/ul
RT: 14.816 min Scan# 1994
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion:168 Resp: 788222
Ion Ratio Lower Upper
168 100
139 37.4 30.3 45.5









#57
2,4-Dinitrotoluene
Concen: 18.853 ng/ul
RT: 14.804 min Scan# 1992
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 205568
Ion Ratio Lower Upper
165 100
63 57.2 41.3 61.9
182 2.9 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.020 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. -0.006 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34


Tgt Ion:232 Resp: 215989
Ion Ratio Lower Upper
232 100
131 35.3 27.9 41.9
130 2.2 1.8 2.8
166 27.2 20.9 31.3








#59
Diethylphthalate
Concen: 18.307 ng/ul
RT: 15.245 min Scan# 2067
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:149 Resp: 673269
Ion Ratio Lower Upper
149 100
177 24.1 18.8 28.2
150 12.6 10.2 15.4



#61
Fluorene
Concen: 18.054 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion:166 Resp: 644051
Ion Ratio Lower Upper
166 100
165 99.6 78.6 118.0
167 13.6 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 18.513 ng/ul
RT: 15.463 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 394121
Ion Ratio Lower Upper
204 100
206 32.7 27.0 40.4
141 51.7 41.5 62.3



#63
4-Nitroaniline
Concen: 17.445 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion:138 Resp: 91676
Ion Ratio Lower Upper
138 100
92 58.4 48.9 73.3
108 111.9 90.4 135.6









#66
4,6-Dinitro-2-methylphenol
Concen: 17.188 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 137423
Ion Ratio Lower Upper
198 100
182 3.3 3.2 4.8
77 22.9 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 18.015 ng/ul
RT: 15.681 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34


Tgt Ion:169 Resp: 542618
Ion Ratio Lower Upper
169 100
168 67.5 54.8 82.2
167 36.3 29.3 43.9








#68
4-Bromophenyl-phenylether
Concen: 18.599 ng/ul
RT: 16.357 min Scan# 2256
Delta R.T. -0.006 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:248 Resp: 259579
Ion Ratio Lower Upper
248 100
250 98.6 75.6 113.4
141 65.7 48.7 73.1



#69
Hexachlorobenzene
Concen: 18.672 ng/ul
RT: 16.469 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion:284 Resp: 300326
Ion Ratio Lower Upper
284 100
142 26.5 23.6 35.4
249 31.9 26.1 39.1





#70

Atrazine

Concen: 18.574 ng/ul

RT: 16.645 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 243874

Ion Ratio Lower Upper

200 100

173 23.8 19.4 29.0

215 48.9 39.7 59.5



#71

Pentachlorophenol

Concen: 18.103 ng/ul

RT: 16.828 min Scan# 2336

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:266 Resp: 180846


Ion Ratio Lower Upper

266 100

264 63.5 49.4 74.2

268 64.6 51.6 77.4





#72

Phenanthrene

Concen: 17.792 ng/ul

RT: 17.216 min Scan# 2402

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:178 Resp: 1021134

Ion Ratio Lower Upper

178 100

179 15.2 12.1 18.1

176 18.9 15.4 23.2



#74

Anthracene

Concen: 18.089 ng/ul

RT: 17.304 min Scan# 2417

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:178 Resp: 1052452


Ion Ratio Lower Upper

178 100

179 16.1 12.3 18.5

176 18.7 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.239 ng/uL
RT: 13.210 min Scan# 1721
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 363975
Ion Ratio Lower Upper
216 100
214 78.9 61.8 92.8
218 49.7 39.0 58.6



#76
Pentachlorobenzene
Concen: 18.935 ng/uL
RT: 14.733 min Scan# 1980
Delta R.T. -0.000 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion:250 Resp: 347986
Ion Ratio Lower Upper
250 100
248 62.3 50.8 76.2
252 63.2 50.2 75.4









#77
 Carbazole
 Concen: 17.368 ng/ul
 RT: 17.586 min Scan# 2465
 Delta R.T. -0.000 min
 Lab File: BP019428.D
 Acq: 23 Jan 2024 19:34




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:167 Resp: 859723
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.3 25.9
 139 12.2 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 18.757 ng/ul
 RT: 18.133 min Scan# 2558
 Delta R.T. -0.006 min
 Lab File: BP019428.D
 Acq: 23 Jan 2024 19:34

Tgt Ion:149 Resp: 1111123
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.0 10.6
 104 6.2 4.6 6.8





#80

Fluoranthene

Concen: 19.419 ng/ul

RT: 19.186 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:202 Resp: 1244485

Ion Ratio Lower Upper

202 100

101 7.1 5.9 8.9

100 5.3 4.6 7.0



#82

Pyrene

Concen: 19.085 ng/ul

RT: 19.545 min Scan# 2798

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:202 Resp: 1241370


Ion Ratio Lower Upper

202 100

101 7.8 7.2 10.8

100 6.5 5.8 8.6





#83

Butylbenzylphthalate

Concen: 19.218 ng/ul

RT: 20.421 min Scan# 2947

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 428867

Ion Ratio Lower Upper

149 100

91 76.4 61.7 92.5

206 25.4 18.6 27.8



#84

3,3'-Dichlorobenzidine

Concen: 17.456 ng/ul

RT: 21.198 min Scan# 3079

Delta R.T. -0.000 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:252 Resp: 329538


Ion Ratio Lower Upper

252 100

254 63.4 50.6 76.0

126 6.7 6.0 9.0





#85

Benzo(a)anthracene

Concen: 17.523 ng/ul

RT: 21.257 min Scan# 3089

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 1134183

Ion Ratio Lower Upper

228 100

229 19.9 15.8 23.6

226 27.7 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.341 ng/ul

RT: 21.180 min Scan# 3076

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:149 Resp: 587507


Ion Ratio Lower Upper

149 100

167 28.0 22.9 34.3

279 5.8 4.6 7.0





#87

Chrysene

Concen: 17.355 ng/ul

RT: 21.310 min Scan# 3098

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 1026791

Ion Ratio Lower Upper

228 100

226 30.7 23.9 35.9

229 19.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 15.213 ng/ul




RT: 22.086 min Scan# 3230


Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34


Tgt Ion:149 Resp: 879869








#90
Benzo(b)fluoranthene
Concen: 16.635 ng/ul
RT: 22.904 min Scan# 3369
Delta R.T. -0.006 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:252 Resp: 1007318
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 5.5 4.9 7.3



#91
Benzo(k)fluoranthene
Concen: 16.864 ng/ul
RT: 22.951 min Scan# 3377
Delta R.T. -0.006 min
Lab File: BP019428.D
Acq: 23 Jan 2024 19:34

Tgt Ion:252 Resp: 990674
Ion Ratio Lower Upper
252 100
253 21.6 17.3 25.9
125 5.6 4.8 7.2





#93

Benzo(a)pyrene

Concen: 17.268 ng/ul

RT: 23.515 min Scan# 3474

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 939622

Ion Ratio Lower Upper

252 100

253 22.0 17.3 25.9

125 6.1 5.2 7.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.345 ng/ul

RT: 26.056 min Scan# 3905

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34




Tgt Ion:276 Resp: 1180536


Ion Ratio Lower Upper

276 100

138 13.1 12.2 18.4

277 25.0 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 18.390 ng/ul

RT: 26.068 min Scan# 3907

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 979646

Ion Ratio Lower Upper

278 100

139 7.8 7.4 11.2

279 23.4 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 18.587 ng/ul

RT: 26.803 min Scan# 4032

Delta R.T. -0.006 min

Lab File: BP019428.D

Acq: 23 Jan 2024 19:34

Tgt Ion:276 Resp: 960575

Ion Ratio Lower Upper

276 100

138 11.9 10.7 16.1

277 22.8 19.2 28.8

