

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012225\
 Data File : BP023703.D
 Acq On : 22 Jan 2025 14:50
 Operator : RC/JU
 Sample : SST02075
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020611

Quant Time: Jan 22 23:35:43 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012225.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 22 23:30:27 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	545666	20.000	ng/u1	0.00
20) Naphthalene-d8	10.575	136	2267664	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.422	164	1425282	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	2780710	20.000	ng/u1	0.00
79) Chrysene-d12	21.651	240	3077238	20.000	ng/u1	0.00
88) Perylene-d12	25.033	264	3128986	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	114690	8.466	ng/uL	0.00
4) Pyridine-d5	3.705	84	819260	20.956	ng/u1	0.00
7) Phenol-d5	6.957	99	944134	20.448	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	559186	21.441	ng/u1	0.00
11) 2-Chlorophenol-d4	7.328	132	751518	21.410	ng/u1	0.00
15) 4-Methylphenol-d8	8.487	113	743816	20.487	ng/u1	0.00
21) Nitrobenzene-d5	8.940	128	361890	21.181	ng/u1	0.00
24) 2-Nitrophenol-d4	9.657	143	323474	20.651	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.193	165	701028	21.292	ng/u1	0.00
31) 4-Chloroaniline-d4	10.698	131	1107672	21.037	ng/u1	0.00
46) Dimethylphthalate-d6	13.822	166	2145947	20.936	ng/u1	0.00
49) Acenaphthylene-d8	14.110	160	2542746	21.434	ng/u1	0.00
54) 4-Nitrophenol-d4	14.604	143	414401	20.120	ng/u1	0.00
60) Fluorene-d10	15.422	176	1821259	20.757	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.534	200	236315	17.257	ng/u1	0.00
73) Anthracene-d10	17.310	188	2900789	21.729	ng/u1	0.00
81) Pyrene-d10	19.680	212	3276379	20.924	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.809	264	3377627	21.566	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	121968	8.510	ng/uL	100
5) Pyridine	3.728	79	840625	21.269	ng/u1	100
6) Benzaldehyde	6.940	77	583495	25.092	ng/u1	100
8) Phenol	6.981	94	995548	20.606	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.222	93	794772	21.146	ng/u1	100
12) 2-Chlorophenol	7.363	128	795274	21.372	ng/u1	100
13) 2-Methylphenol	8.222	108	720413	20.548	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.310	45	839900	21.544	ng/u1	100
16) Acetophenone	8.604	105	1238590	21.242	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.593	70	576621	21.375	ng/u1	100
18) 4-Methylphenol	8.546	108	804863	20.647	ng/u1	100
19) Hexachloroethane	8.875	117	323657	21.282	ng/u1	100
22) Nitrobenzene	8.981	77	847017	21.219	ng/u1	100
23) Isophorone	9.504	82	1620591	21.345	ng/u1	100
25) 2-Nitrophenol	9.687	139	388954	21.150	ng/u1	100
26) 2,4-Dimethylphenol	9.746	107	823226	21.599	ng/u1	100
27) Bis(2-Chloroethoxy)met...	9.981	93	1038130	21.101	ng/u1	100
29) 2,4-Dichlorophenol	10.222	162	718201	21.100	ng/u1	100
30) Naphthalene	10.628	128	2572842	21.083	ng/u1	100
32) 4-Chloroaniline	10.722	127	1063782	21.162	ng/u1	100
33) Hexachlorobutadiene	10.922	225	449536	20.981	ng/u1	100
34) Caprolactam	11.481	113	251101	20.383	ng/u1	100
35) 4-Chloro-3-methylphenol	11.845	107	754013	21.384	ng/u1	100
36) 2-Methylnaphthalene	12.234	142	1736741	21.075	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012225\
 Data File : BP023703.D
 Acq On : 22 Jan 2025 14:50
 Operator : RC/JU
 Sample : SSTD02075
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020611

Quant Time: Jan 22 23:35:43 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012225.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 22 23:30:27 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.451	142	1746057	21.196	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.604	216	891148	21.050	ng/ul	100
40) Hexachlorocyclopentadiene	12.592	237	423671	19.641	ng/ul	100
41) 2,4,6-Trichlorophenol	12.840	196	531014	21.334	ng/ul	100
42) 2,4,5-Trichlorophenol	12.910	196	579755	21.088	ng/ul	100
43) 1,1'-Biphenyl	13.245	154	2278860	21.250	ng/ul	100
44) 2-Chloronaphthalene	13.287	162	1772817	21.232	ng/ul	100
45) 2-Nitroaniline	13.481	65	422619	21.313	ng/ul	100
47) Dimethylphthalate	13.869	163	2166942	21.058	ng/ul	100
48) 2,6-Dinitrotoluene	13.987	165	392134	21.365	ng/ul	100
50) Acenaphthylene	14.139	152	2735966	21.271	ng/ul	100
51) 3-Nitroaniline	14.310	138	453926	21.192	ng/ul	100
52) Acenaphthene	14.486	153	1932633	21.251	ng/ul	100
53) 2,4-Dinitrophenol	14.516	184	136962	15.200	ng/ul	100
55) 4-Nitrophenol	14.616	109	315796	20.893	ng/ul	100
56) Dibenzofuran	14.828	168	2651203	21.195	ng/ul	100
57) 2,4-Dinitrotoluene	14.775	165	585045	21.306	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.051	232	471620	20.888	ng/ul	100
59) Diethylphthalate	15.257	149	2129839	21.044	ng/ul	100
61) Fluorene	15.481	166	2241017	21.499	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.481	204	1082908	21.224	ng/ul	100
63) 4-Nitroaniline	15.486	138	503686	23.112	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.545	198	277208	18.000	ng/ul	100
67) N-Nitrosodiphenylamine	15.692	169	1805762	21.621	ng/ul	100
68) 4-Bromophenyl-phenylether	16.386	248	631179	21.201	ng/ul	100
69) Hexachlorobenzene	16.504	284	756117	20.871	ng/ul	100
70) Atrazine	16.663	200	650642	21.121	ng/ul	100
71) Pentachlorophenol	16.851	266	422471	19.305	ng/ul	100
72) Phenanthrene	17.251	178	3376364	21.491	ng/ul	100
74) Anthracene	17.345	178	3453682	22.131	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.210	216	859183	21.297	ng/ul	100
76) Pentachlorobenzene	14.745	250	908188	21.192	ng/ul	100
77) Carbazole	17.622	167	3122419	22.793	ng/ul	100
78) Di-n-butylphthalate	18.222	149	3476053	22.286	ng/ul	100
80) Fluoranthene	19.333	202	3843814	21.589	ng/ul	100
82) Pyrene	19.704	202	4166787	21.548	ng/ul	100
83) Butylbenzylphthalate	20.663	149	1452523	20.828	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.551	252	1079947	19.886	ng/ul	100
85) Benzo(a)anthracene	21.633	228	4218354	21.623	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.592	149	2220122	21.347	ng/ul	100
87) Chrysene	21.698	228	3964596	21.707	ng/ul	100
89) Di-n-octyl phthalate	22.898	149	3207604	20.463	ng/ul	100
90) Benzo(b)fluoranthene	23.974	252	4016496	21.608	ng/ul	100
91) Benzo(k)fluoranthene	24.039	252	4131101	21.559	ng/ul	100
93) Benzo(a)pyrene	24.880	252	3779609	21.605	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	28.886	276	3829449	17.639	ng/ul	100
95) Dibenzo(a,h)anthracene	28.968	278	3809506	21.616	ng/ul	100
96) Benzo(g,h,i)perylene	30.068	276	3579646	21.405	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012225\
 Data File : BP023703.D
 Acq On : 22 Jan 2025 14:50
 Operator : RC/JU
 Sample : SSTD02075
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020611

Quant Time: Jan 22 23:35:43 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012225.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 22 23:30:27 2025
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.510 ng/uL
 RT: 3.334 min Scan# 42
 Delta R.T. 0.000 min
 Lab File: BP023703.D
 Acq: 22 Jan 2025 14:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020611


Tgt Ion: 88 Resp: 121968
 Ion Ratio Lower Upper
 88 100
 43 22.9 18.3 27.5
 58 63.4 50.7 76.1



#5
 Pyridine
 Concen: 21.269 ng/uL
 RT: 3.728 min Scan# 109
 Delta R.T. 0.000 min
 Lab File: BP023703.D
 Acq: 22 Jan 2025 14:50

Tgt Ion: 79 Resp: 840625
 Ion Ratio Lower Upper
 79 100
 52 56.2 45.0 67.4
 51 27.8 22.2 33.4









#6
Benzaldehyde
Concen: 25.092 ng/ul
RT: 6.940 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion: 77 Resp: 583495
Ion Ratio Lower Upper
77 100
105 101.3 81.0 121.6
106 98.3 78.6 118.0



#8
Phenol
Concen: 20.606 ng/ul
RT: 6.981 min Scan# 662
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion: 94 Resp: 995548
Ion Ratio Lower Upper
94 100
65 25.2 20.2 30.2
66 32.5 26.0 39.0









#10
Bis(2-Chloroethyl)ether
Concen: 21.146 ng/ul
RT: 7.222 min Scan# 703
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion: 93 Resp: 794772
Ion Ratio Lower Upper
93 100
63 66.1 52.9 79.3
95 32.6 26.1 39.1



#12
2-Chlorophenol
Concen: 21.372 ng/ul
RT: 7.363 min Scan# 727
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:128 Resp: 795274
Ion Ratio Lower Upper
128 100
64 40.5 32.4 48.6
130 32.8 26.2 39.4









#13
2-Methylphenol
Concen: 20.548 ng/ul
RT: 8.222 min Scan# 873
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:108 Resp: 720413
Ion Ratio Lower Upper
108 100
107 87.0 69.6 104.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.544 ng/ul
RT: 8.310 min Scan# 888
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion: 45 Resp: 839900
Ion Ratio Lower Upper
45 100
77 18.5 14.8 22.2
79 13.3 10.6 16.0









#16
Acetophenone
Concen: 21.242 ng/ul
RT: 8.604 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:105 Resp: 1238590
Ion Ratio Lower Upper
105 100
77 72.6 58.1 87.1
51 20.5 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 21.375 ng/ul
RT: 8.593 min Scan# 936
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion: 70 Resp: 576621
Ion Ratio Lower Upper
70 100
42 47.1 37.7 56.5
101 10.5 8.4 12.6
130 22.9 18.3 27.5









#18
4-Methylphenol
Concen: 20.647 ng/ul
RT: 8.546 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:108 Resp: 804863
Ion Ratio Lower Upper
108 100
107 114.0 91.2 136.8



#19
Hexachloroethane
Concen: 21.282 ng/ul
RT: 8.875 min Scan# 984
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:117 Resp: 323657
Ion Ratio Lower Upper
117 100
201 95.7 76.6 114.8
199 60.9 48.7 73.1









#22
Nitrobenzene
Concen: 21.219 ng/ul
RT: 8.981 min Scan# 1002
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion: 77 Resp: 847017
Ion Ratio Lower Upper
77 100
123 47.5 38.0 57.0
65 12.2 9.8 14.6



#23
Isophorone
Concen: 21.345 ng/ul
RT: 9.504 min Scan# 1091
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion: 82 Resp: 1620591
Ion Ratio Lower Upper
82 100
95 6.5 5.2 7.8
138 18.6 14.9 22.3









#25
2-Nitrophenol
Concen: 21.150 ng/ul
RT: 9.687 min Scan# 1122
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:139 Resp: 388954
Ion Ratio Lower Upper
139 100
65 46.8 37.4 56.2
109 28.5 22.8 34.2



#26
2,4-Dimethylphenol
Concen: 21.599 ng/ul
RT: 9.746 min Scan# 1132
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:107 Resp: 823226
Ion Ratio Lower Upper
107 100
121 52.5 42.0 63.0
122 87.8 70.2 105.4





#27

Bis(2-Chloroethoxy)methane

Concen: 21.101 ng/ul

RT: 9.981 min Scan# 1172

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion: 93 Resp: 1038130

Ion Ratio Lower Upper

93 100

95 32.3 25.8 38.8

123 11.8 9.4 14.2



#29

2,4-Dichlorophenol

Concen: 21.100 ng/ul

RT: 10.222 min Scan# 1213

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion: 162 Resp: 718201


Ion Ratio Lower Upper

162 100

164 65.4 52.3 78.5

98 35.4 28.3 42.5









#30
Naphthalene
Concen: 21.083 ng/ul
RT: 10.628 min Scan# 1282
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:128 Resp: 2572842
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 12.8 10.2 15.4



#32
4-Chloroaniline
Concen: 21.162 ng/ul
RT: 10.722 min Scan# 1298
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:127 Resp: 1063782
Ion Ratio Lower Upper
127 100
129 32.8 26.2 39.4









#33
Hexachlorobutadiene
Concen: 20.981 ng/ul
RT: 10.922 min Scan# 1332
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:225 Resp: 449536
Ion Ratio Lower Upper
225 100
223 63.6 50.9 76.3
227 64.8 51.8 77.8



#34
Caprolactam
Concen: 20.383 ng/ul
RT: 11.481 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:113 Resp: 251101
Ion Ratio Lower Upper
113 100
55 132.2 105.8 158.6
56 99.1 79.3 118.9









#35
4-Chloro-3-methylphenol
Concen: 21.384 ng/ul
RT: 11.845 min Scan# 1489
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:107 Resp: 754013
Ion Ratio Lower Upper
107 100
144 26.6 21.3 31.9
142 81.9 65.5 98.3



#36
2-Methylnaphthalene
Concen: 21.075 ng/ul
RT: 12.234 min Scan# 1555
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:142 Resp: 1736741
Ion Ratio Lower Upper
142 100
141 88.1 70.5 105.7





#37

1-Methylnaphthalene

Concen: 21.196 ng/ul

RT: 12.451 min Scan# 1592

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:142 Resp: 1746057

Ion Ratio Lower Upper

142 100

141 90.1 72.1 108.1

116 3.6 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.050 ng/ul

RT: 12.604 min Scan# 1618

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Tgt Ion:216 Resp: 891148




Ion Ratio Lower Upper


216 100

214 79.0 63.2 94.8

179 21.1 16.9 25.3

108 17.9 14.3 21.5









#40
Hexachlorocyclopentadiene
Concen: 19.641 ng/ul
RT: 12.592 min Scan# 1616
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:237 Resp: 423671
Ion Ratio Lower Upper
237 100
235 63.0 50.4 75.6
272 13.6 10.9 16.3



#41
2,4,6-Trichlorophenol
Concen: 21.334 ng/ul
RT: 12.840 min Scan# 1658
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:196 Resp: 531014
Ion Ratio Lower Upper
196 100
198 97.0 77.6 116.4
200 30.6 24.5 36.7





#42

2,4,5-Trichlorophenol

Concen: 21.088 ng/ul

RT: 12.910 min Scan# 1670

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:196 Resp: 579755

Ion Ratio Lower Upper

196 100

198 96.8 77.4 116.2

200 31.2 25.0 37.4



#43

1,1'-Biphenyl

Concen: 21.250 ng/ul

RT: 13.245 min Scan# 1727

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion:154 Resp: 2278860


Ion Ratio Lower Upper

154 100

153 40.2 32.2 48.2

76 12.7 10.2 15.2









#44
2-Chloronaphthalene
Concen: 21.232 ng/ul
RT: 13.287 min Scan# 1734
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:162 Resp: 1772817
Ion Ratio Lower Upper
162 100
164 32.7 26.2 39.2
127 38.1 30.5 45.7



#45
2-Nitroaniline
Concen: 21.313 ng/ul
RT: 13.481 min Scan# 1767
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion: 65 Resp: 422619
Ion Ratio Lower Upper
65 100
92 76.2 61.0 91.4
138 117.7 94.2 141.2









#47
Dimethylphthalate
Concen: 21.058 ng/ul
RT: 13.869 min Scan# 1833
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:163 Resp: 2166942
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 21.365 ng/ul
RT: 13.987 min Scan# 1853
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:165 Resp: 392134
Ion Ratio Lower Upper
165 100
89 56.0 44.8 67.2
121 21.2 17.0 25.4





#50

Acenaphthylene

Concen: 21.271 ng/ul

RT: 14.139 min Scan# 1879

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:152 Resp: 2735966

Ion Ratio Lower Upper

152 100

151 19.8 15.8 23.8

153 13.2 10.6 15.8



#51

3-Nitroaniline

Concen: 21.192 ng/ul

RT: 14.310 min Scan# 1908

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion:138 Resp: 453926


Ion Ratio Lower Upper

138 100

108 12.1 9.7 14.5

92 112.2 89.8 134.6





#52

Acenaphthene

Concen: 21.251 ng/ul

RT: 14.486 min Scan# 1938

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:153 Resp: 1932633

Ion Ratio Lower Upper

153 100

152 48.9 39.1 58.7

154 89.2 71.4 107.0



#53

2,4-Dinitrophenol

Concen: 15.200 ng/ul

RT: 14.516 min Scan# 1943

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion:184 Resp: 136962


Ion Ratio Lower Upper

184 100

63 59.8 47.8 71.8

154 65.1 52.1 78.1









#55
4-Nitrophenol
Concen: 20.893 ng/ul
RT: 14.616 min Scan# 1960
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:109 Resp: 315796
Ion Ratio Lower Upper
109 100
139 142.9 114.3 171.5
65 128.8 103.0 154.6



#56
Dibenzofuran
Concen: 21.195 ng/ul
RT: 14.828 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:168 Resp: 2651203
Ion Ratio Lower Upper
168 100
139 36.6 29.3 43.9









#57
2,4-Dinitrotoluene
Concen: 21.306 ng/ul
RT: 14.775 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:165 Resp: 585045
Ion Ratio Lower Upper
165 100
63 38.8 31.0 46.6
182 2.8 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.888 ng/ul
RT: 15.051 min Scan# 2034
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:232 Resp: 471620
Ion Ratio Lower Upper
232 100
131 45.1 36.1 54.1
130 2.3 1.8 2.8
166 30.3 24.2 36.4









#59
Diethylphthalate
Concen: 21.044 ng/ul
RT: 15.257 min Scan# 2069
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:149 Resp: 2129839
Ion Ratio Lower Upper
149 100
177 22.1 17.7 26.5
150 12.2 9.8 14.6



#61
Fluorene
Concen: 21.499 ng/ul
RT: 15.481 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:166 Resp: 2241017
Ion Ratio Lower Upper
166 100
165 97.3 77.8 116.8
167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.224 ng/ul
RT: 15.481 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:204 Resp: 1082908
Ion Ratio Lower Upper
204 100
206 32.7 26.2 39.2
141 56.8 45.4 68.2



#63
4-Nitroaniline
Concen: 23.112 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:138 Resp: 503686
Ion Ratio Lower Upper
138 100
92 46.2 37.0 55.4
108 68.0 54.4 81.6









#66
4,6-Dinitro-2-methylphenol
Concen: 18.000 ng/ul
RT: 15.545 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:198 Resp: 277208
Ion Ratio Lower Upper
198 100
182 4.9 3.9 5.9
77 25.4 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 21.621 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:169 Resp: 1805762
Ion Ratio Lower Upper
169 100
168 67.0 53.6 80.4
167 36.5 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 21.201 ng/ul
RT: 16.386 min Scan# 2261
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:248 Resp: 631179
Ion Ratio Lower Upper
248 100
250 97.1 77.7 116.5
141 68.8 55.0 82.6



#69
Hexachlorobenzene
Concen: 20.871 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:284 Resp: 756117
Ion Ratio Lower Upper
284 100
142 35.1 28.1 42.1
249 29.8 23.8 35.8





#70

Atrazine

Concen: 21.121 ng/ul

RT: 16.663 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:200 Resp: 650642

Ion Ratio Lower Upper

200 100

173 27.2 21.8 32.6

215 49.1 39.3 58.9



#71

Pentachlorophenol

Concen: 19.305 ng/ul

RT: 16.851 min Scan# 2340

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion:266 Resp: 422471


Ion Ratio Lower Upper

266 100

264 63.6 50.9 76.3

268 65.3 52.2 78.4





#72

Phenanthrene

Concen: 21.491 ng/ul

RT: 17.251 min Scan# 2408

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:178 Resp: 3376364

Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 19.6 15.7 23.5



#74

Anthracene

Concen: 22.131 ng/ul

RT: 17.345 min Scan# 2424

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion:178 Resp: 3453682


Ion Ratio Lower Upper

178 100

179 15.5 12.4 18.6

176 18.8 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.297 ng/uL
RT: 13.210 min Scan# 1721
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:216 Resp: 859183
Ion Ratio Lower Upper
216 100
214 76.9 61.5 92.3
218 48.0 38.4 57.6



#76
Pentachlorobenzene
Concen: 21.192 ng/uL
RT: 14.745 min Scan# 1982
Delta R.T. 0.000 min
Lab File: BP023703.D
Acq: 22 Jan 2025 14:50

Tgt Ion:250 Resp: 908188
Ion Ratio Lower Upper
250 100
248 62.7 50.2 75.2
252 63.8 51.0 76.6









#77
 Carbazole
 Concen: 22.793 ng/ul
 RT: 17.622 min Scan# 2471
 Delta R.T. 0.000 min
 Lab File: BP023703.D
 Acq: 22 Jan 2025 14:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020611


Tgt Ion:167 Resp: 3122419
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 13.2 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 22.286 ng/ul
 RT: 18.222 min Scan# 2573
 Delta R.T. 0.000 min
 Lab File: BP023703.D
 Acq: 22 Jan 2025 14:50

Tgt Ion:149 Resp: 3476053
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 5.2 4.2 6.2





#80

Fluoranthene

Concen: 21.589 ng/ul

RT: 19.333 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:202 Resp: 3843814

Ion Ratio Lower Upper

202 100

101 11.5 9.2 13.8

100 8.8 7.0 10.6



#82

Pyrene

Concen: 21.548 ng/ul

RT: 19.704 min Scan# 2825

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion:202 Resp: 4166787


Ion Ratio Lower Upper

202 100

101 14.5 11.6 17.4

100 12.0 9.6 14.4





#83

Butylbenzylphthalate

Concen: 20.828 ng/ul

RT: 20.663 min Scan# 2988

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:149 Resp: 1452523

Ion Ratio Lower Upper

149 100

91 70.7 56.6 84.8

206 20.4 16.3 24.5



#84

3,3'-Dichlorobenzidine

Concen: 19.886 ng/ul

RT: 21.551 min Scan# 3139

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion:252 Resp: 1079947


Ion Ratio Lower Upper

252 100

254 65.1 52.1 78.1

126 11.7 9.4 14.0





#85

Benzo(a)anthracene

Concen: 21.623 ng/ul

RT: 21.633 min Scan# 3153

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:228 Resp: 4218354

Ion Ratio Lower Upper

228 100

229 19.8 15.8 23.8

226 27.3 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.347 ng/ul

RT: 21.592 min Scan# 3146

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion:149 Resp: 2220122


Ion Ratio Lower Upper

149 100

167 27.1 21.7 32.5

279 4.3 3.4 5.2





#87

Chrysene

Concen: 21.707 ng/ul

RT: 21.698 min Scan# 3164

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:228 Resp: 3964596

Ion Ratio Lower Upper

228 100

226 30.4 24.3 36.5

229 19.7 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 20.463 ng/ul




RT: 22.898 min Scan# 3368


Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Tgt Ion:149 Resp: 3207604





#90

Benzo(b)fluoranthene

Concen: 21.608 ng/ul

RT: 23.974 min Scan# 3551

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:252 Resp: 4016496

Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 10.3 8.2 12.4



#91

Benzo(k)fluoranthene

Concen: 21.559 ng/ul

RT: 24.039 min Scan# 3562

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion:252 Resp: 4131101


Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 10.8 8.6 13.0





#93

Benzo(a)pyrene

Concen: 21.605 ng/ul

RT: 24.880 min Scan# 3705

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:252 Resp: 3779609

Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 11.4 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.639 ng/ul

RT: 28.886 min Scan# 4386

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50




Tgt Ion:276 Resp: 3829449


Ion Ratio Lower Upper

276 100

138 20.8 16.6 25.0

277 24.0 19.2 28.8





#95

Dibenzo(a,h)anthracene

Concen: 21.616 ng/ul

RT: 28.968 min Scan# 4400

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:278 Resp: 3809506

Ion Ratio Lower Upper

278 100

139 16.8 13.4 20.2

279 24.3 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 21.405 ng/ul

RT: 30.068 min Scan# 4587

Delta R.T. 0.000 min

Lab File: BP023703.D

Acq: 22 Jan 2025 14:50

Tgt Ion:276 Resp: 3579646

Ion Ratio Lower Upper

276 100

138 23.1 18.5 27.7

277 23.8 19.0 28.6

