

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012324\
 Data File : BP019427.D
 Acq On : 23 Jan 2024 19:00
 Operator : MA/JU
 Sample : P1157-13
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AK3

Quant Time: Jan 24 00:15:07 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 22 11:49:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	156222	20.000	ng/u1	0.00
20) Naphthalene-d8	10.563	136	617568	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.416	164	432841	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.168	188	940812	20.000	ng/u1	0.00
79) Chrysene-d12	21.274	240	716361	20.000	ng/u1	0.00
88) Perylene-d12	23.621	264	747430	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/uL	
4) Pyridine-d5	0.000	84	0d	0.000	ng/u1	
7) Phenol-d5	0.000	99	0d	0.000	ng/u1	
9) Bis-(2-Chloroethyl)eth...	0.000	67	0d	0.000	ng/u1	
11) 2-Chlorophenol-d4	0.000	132	0d	0.000	ng/u1	
15) 4-Methylphenol-d8	0.000	113	0d	0.000	ng/u1	
21) Nitrobenzene-d5	0.000	128	0d	0.000	ng/u1	
24) 2-Nitrophenol-d4	0.000	143	0d	0.000	ng/u1	
28) 2,4-Dichlorophenol-d3	0.000	165	0d	0.000	ng/u1	
31) 4-Chloroaniline-d4	0.000	131	0d	0.000	ng/u1	
46) Dimethylphthalate-d6	0.000	166	0d	0.000	ng/u1	
49) Acenaphthylene-d8	0.000	160	0d	0.000	ng/u1	
54) 4-Nitrophenol-d4	0.000	143	0d	0.000	ng/u1	
60) Fluorene-d10	0.000	176	0d	0.000	ng/u1	
65) 4,6-Dinitro-2-methylph...	0.000	200	0d	0.000	ng/u1	
73) Anthracene-d10	0.000	188	0d	0.000	ng/u1	
81) Pyrene-d10	0.000	212	0d	0.000	ng/u1	
92) Benzo(a)pyrene-d12	0.000	264	0d	0.000	ng/u1	
Target Compounds						
					Qvalue	
72) Phenanthrene	17.210	178	213909	4.048	ng/u1	98
80) Fluoranthene	19.192	202	412970	7.479	ng/u1	98
82) Pyrene	19.545	202	193093	3.446	ng/u1	97
85) Benzo(a)anthracene	21.256	228	118384	2.123	ng/u1	96
87) Chrysene	21.309	228	142290	2.791	ng/u1	99
90) Benzo(b)fluoranthene	22.909	252	91296	1.709	ng/u1#	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012324\
Data File : BP019427.D
Acq On : 23 Jan 2024 19:00
Operator : MA/JU
Sample : P1157-13
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AK3

Quant Time: Jan 24 00:15:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 22 11:49:04 2024
Response via : Initial Calibration

