

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012423\
 Data File : BP013575.D
 Acq On : 24 Jan 2023 17:36
 Operator : CG/JU
 Sample : 01190-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PE-1D

Quant Time: Jan 25 03:49:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 25 03:43:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.157	152	387834	20.000	ng	0.00
21) Naphthalene-d8	10.993	136	1550918	20.000	ng	0.00
39) Acenaphthene-d10	14.792	164	908170	20.000	ng	0.00
64) Phenanthrene-d10	17.545	188	1920208	20.000	ng	0.00
76) Chrysene-d12	21.621	240	1438705	20.000	ng	0.00
86) Perylene-d12	24.210	264	1217195	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.687	112	2416150	106.740	ng	0.00
7) Phenol-d6	7.305	99	3095090	101.595	ng	0.00
23) Nitrobenzene-d5	9.334	82	1735223	73.862	ng	0.00
42) 2,4,6-Tribromophenol	16.281	330	1025664	112.595	ng	0.00
45) 2-Fluorobiphenyl	13.422	172	4121180	62.290	ng	0.00
79) Terphenyl-d14	20.080	244	5828866	72.886	ng	0.00
Target Compounds						
						Qvalue
73) Carbazole	17.939	167	246108	2.526	ng	99
74) Di-n-butylphthalate	18.480	149	485833	4.480	ng	99
75) Fluoranthene	19.533	202	505215	4.447	ng	99
78) Pyrene	19.886	202	573105	5.210	ng	98
80) Butylbenzylphthalate	20.745	149	1071858	30.602	ng	98
81) Benzo(a)anthracene	21.604	228	1318399	13.154	ng	100
83) Chrysene	21.663	228	1456066	15.189	ng	99
84) Bis(2-ethylhexyl)phtha...	21.516	149	1235566	22.042	ng	100
85) Di-n-octyl phthalate	22.504	149	391753	4.335	ng	# 100
88) Benzo(b)fluoranthene	23.410	252	217860	2.910	ng	99
89) Benzo(k)fluoranthene	23.462	252	203739	2.668	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012423\
Data File : BP013575.D
Acq On : 24 Jan 2023 17:36
Operator : CG/JU
Sample : 01190-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PE-1D

Quant Time: Jan 25 03:49:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 25 03:43:27 2023
Response via : Initial Calibration

