

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012423\
 Data File : BP013588.D
 Acq On : 25 Jan 2023 01:01
 Operator : CG/JU
 Sample : 01244-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 VNJ-223

Quant Time: Jan 25 03:54:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 25 03:43:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.157	152	398645	20.000	ng	0.00
21) Naphthalene-d8	10.987	136	1569036	20.000	ng	0.00
39) Acenaphthene-d10	14.792	164	852915	20.000	ng	0.00
64) Phenanthrene-d10	17.545	188	1563381	20.000	ng	0.00
76) Chrysene-d12	21.621	240	1242656	20.000	ng	0.00
86) Perylene-d12	24.204	264	1587922	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.687	112	2390606	102.747	ng	0.00
7) Phenol-d6	7.299	99	3086798	98.575	ng	0.00
23) Nitrobenzene-d5	9.328	82	1660050	69.846	ng	0.00
42) 2,4,6-Tribromophenol	16.281	330	919055	107.428	ng	0.00
45) 2-Fluorobiphenyl	13.416	172	3761929	60.543	ng	-0.01
79) Terphenyl-d14	20.074	244	3755199	54.364	ng	0.00
Target Compounds						
						Qvalue
75) Fluoranthene	19.533	202	316143	3.418	ng	98
78) Pyrene	19.886	202	311548	3.279	ng	99
81) Benzo(a)anthracene	21.604	228	224811	2.597	ng	97
83) Chrysene	21.657	228	270258	3.264	ng	98
84) Bis(2-ethylhexyl)phtha...	21.510	149	97932	2.023	ng	99
88) Benzo(b)fluoranthene	23.404	252	235848	2.414	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012423\
Data File : BP013588.D
Acq On : 25 Jan 2023 01:01
Operator : CG/JU
Sample : 01244-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
VNJ-223

Quant Time: Jan 25 03:54:29 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 25 03:43:27 2023
Response via : Initial Calibration

