

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012522\
 Data File : BP008912.D
 Acq On : 25 Jan 2022 17:42
 Operator : CG/JU
 Sample : N1245-06
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SB-3(3-3.5)

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 01/26/2022
 Supervised By :Jagrut Upadhyay 01/26/2022

Quant Time: Jan 26 01:58:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 26 01:52:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	214441	20.000	ng	0.00
21) Naphthalene-d8	10.634	136	787676	20.000	ng	0.00
39) Acenaphthene-d10	14.475	164	494566	20.000	ng	0.00
64) Phenanthrene-d10	17.233	188	1024634	20.000	ng	0.00
76) Chrysene-d12	21.322	240	1113980	20.000	ng	-0.01
86) Perylene-d12	23.739	264	1286957	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	1599109	115.318	ng	0.00
7) Phenol-d6	7.016	99	1935434	115.259	ng	0.00
23) Nitrobenzene-d5	8.993	82	1166438	84.074	ng	0.00
42) 2,4,6-Tribromophenol	15.969	330	868759	120.679	ng	-0.01
45) 2-Fluorobiphenyl	13.098	172	2801537	74.702	ng	0.00
79) Terphenyl-d14	19.792	244	4187729	63.454	ng	-0.01
Target Compounds						Qvalue
75) Fluoranthene	19.245	202	217429	3.389	ng	97
78) Pyrene	19.598	202	203603	2.619	ng	98
81) Benzo(a)anthracene	21.310	228	171458	2.288	ng	99
83) Chrysene	21.357	228	167978m	2.312	ng	
88) Benzo(b)fluoranthene	23.004	252	231153	2.690	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012522\
Data File : BP008912.D
Acq On : 25 Jan 2022 17:42
Operator : CG/JU
Sample : N1245-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SB-3(3-3.5)

Quant Time: Jan 26 01:58:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 26 01:52:51 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 01/26/2022
Supervised By :Jagrut Upadhyay 01/26/2022

