

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012624\
 Data File : BP019451.D
 Acq On : 26 Jan 2024 15:25
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP012624

Quant Time: Jan 26 23:49:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 26 23:43:31 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	38.391	4.0	97	0.00
3	Pyridine	40.000	38.800	3.0	98	0.00
4	n-Nitrosodimethylamine	40.000	41.737	-4.3	102	0.00
5 S	2-Fluorophenol	80.000	79.100	1.1	101	0.00
6	Aniline	40.000	40.159	-0.4	100	0.00
7 S	Phenol-d6	80.000	80.998	-1.2	101	0.00
8	2-Chlorophenol	40.000	40.677	-1.7	103	0.00
9	Benzaldehyde	40.000	44.452	-11.1	102	0.00
10 C	Phenol	40.000	40.411	-1.0	101	0.00
11	bis(2-Chloroethyl)ether	40.000	40.242	-0.6	101	0.00
12	1,3-Dichlorobenzene	40.000	38.979	2.6	100	0.00
13 C	1,4-Dichlorobenzene	40.000	39.596	1.0	101	0.00
14	1,2-Dichlorobenzene	40.000	39.461	1.3	101	0.00
15	Benzyl Alcohol	40.000	41.557	-3.9	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.729	0.7	100	0.00
17	2-Methylphenol	40.000	40.046	-0.1	99	0.00
18	Hexachloroethane	40.000	40.913	-2.3	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.047	-2.6	101	0.00
20	3+4-Methylphenols	40.000	41.029	-2.6	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	39.214	2.0	102	0.00
23 S	Nitrobenzene-d5	80.000	84.228	-5.3	106	0.00
24	Nitrobenzene	40.000	41.325	-3.3	105	0.00
25	Isophorone	40.000	39.108	2.2	101	0.00
26 C	2-Nitrophenol	40.000	43.398	-8.5	116	0.00
27	2,4-Dimethylphenol	40.000	40.029	-0.1	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.037	2.4	101	0.00
29 C	2,4-Dichlorophenol	40.000	41.215	-3.0	104	0.00
30	1,2,4-Trichlorobenzene	40.000	39.365	1.6	102	0.00
31	Naphthalene	40.000	38.848	2.9	101	0.00
32	Benzoic acid	40.000	41.304	-3.3	116	0.00
33	4-Chloroaniline	40.000	38.822	2.9	100	0.00
34 C	Hexachlorobutadiene	40.000	39.679	0.8	103	0.00
35	Caprolactam	40.000	39.913	0.2	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.245	-3.1	105	0.00
37	2-Methylnaphthalene	40.000	39.286	1.8	102	0.00
38	1-Methylnaphthalene	40.000	39.502	1.2	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.235	-0.6	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.298	-0.7	103	0.00
42 S	2,4,6-Tribromophenol	80.000	84.857	-6.1	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.189	-5.5	106	0.00
44	2,4,5-Trichlorophenol	40.000	42.754	-6.9	108	0.00
45 S	2-Fluorobiphenyl	80.000	79.311	0.9	102	0.00
46	1,1'-Biphenyl	40.000	39.698	0.8	103	0.00
47	2-Chloronaphthalene	40.000	39.786	0.5	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012624\
 Data File : BP019451.D
 Acq On : 26 Jan 2024 15:25
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP012624

Quant Time: Jan 26 23:49:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 26 23:43:31 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.412	-1.0	109	0.00
49	Acenaphthylene	40.000	39.398	1.5	102	0.00
50	Dimethylphthalate	40.000	39.652	0.9	103	0.00
51	2,6-Dinitrotoluene	40.000	40.464	-1.2	106	0.00
52 C	Acenaphthene	40.000	39.659	0.9	102	0.00
53	3-Nitroaniline	40.000	43.660	-9.1	107	0.00
54 P	2,4-Dinitrophenol	40.000	44.158	-10.4	119	0.00
55	Dibenzofuran	40.000	39.680	0.8	103	0.00
56 P	4-Nitrophenol	40.000	42.475	-6.2	109	0.00
57	2,4-Dinitrotoluene	40.000	40.714	-1.8	108	0.00
58	Fluorene	40.000	39.511	1.2	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.154	-5.4	103	0.00
60	Diethylphthalate	40.000	39.542	1.1	103	0.00
61	4-Chlorophenyl-phenylether	40.000	39.636	0.9	103	0.00
62	4-Nitroaniline	40.000	43.172	-7.9	106	0.00
63	Azobenzene	40.000	39.287	1.8	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.891	-4.7	118	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.255	-0.6	103	0.00
67	4-Bromophenyl-phenylether	40.000	40.810	-2.0	104	0.00
68	Hexachlorobenzene	40.000	40.235	-0.6	104	0.00
69	Atrazine	40.000	39.346	1.6	102	0.00
70 C	Pentachlorophenol	40.000	44.158	-10.4	108	0.00
71	Phenanthrene	40.000	39.524	1.2	102	0.00
72	Anthracene	40.000	39.620	1.0	101	0.00
73	Carbazole	40.000	39.379	1.6	101	0.00
74	Di-n-butylphthalate	40.000	40.395	-1.0	103	0.00
75 C	Fluoranthene	40.000	39.431	1.4	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	42.783	-7.0	101	0.00
78	Pyrene	40.000	40.468	-1.2	101	0.00
79 S	Terphenyl-d14	80.000	82.224	-2.8	102	0.00
80	Butylbenzylphthalate	40.000	42.586	-6.5	104	0.00
81	Benzo(a)anthracene	40.000	39.243	1.9	100	0.00
82	3,3'-Dichlorobenzidine	40.000	38.815	3.0	99	0.00
83	Chrysene	40.000	39.239	1.9	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.529	-1.3	102	0.00
85 c	Di-n-octyl phthalate	40.000	39.838	0.4	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.336	9.2	97	-0.01
88	Benzo(b)fluoranthene	40.000	40.744	-1.9	99	0.00
89	Benzo(k)fluoranthene	40.000	40.679	-1.7	99	0.00
90 C	Benzo(a)pyrene	40.000	39.950	0.1	98	0.00
91	Dibenzo(a,h)anthracene	40.000	36.357	9.1	98	0.00
92	Benzo(g,h,i)perylene	40.000	35.109	12.2	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012624\
Data File : BP019451.D
Acq On : 26 Jan 2024 15:25
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP012624

Quant Time: Jan 26 23:49:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 26 23:43:31 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0