

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP012624\  
 Data File : BP019445.D  
 Acq On : 26 Jan 2024 12:01  
 Operator : MA/JU  
 Sample : SSTDICC010  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICC010

Manual Integrations  
 APPROVED

Reviewed By :Jagrut Upadhyay 01/29/2024  
 Supervised By :mohammad ahmed 01/30/2024

Quant Time: Jan 26 23:30:32 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP012624.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Jan 26 23:25:28 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| -----                         |        |      |          |        |       |          |
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.928  | 152  | 71063    | 20.000 | ng    | 0.00     |
| 21) Naphthalene-d8            | 10.728 | 136  | 313932   | 20.000 | ng    | 0.00     |
| 39) Acenaphthene-d10          | 14.557 | 164  | 230435   | 20.000 | ng    | 0.00     |
| 64) Phenanthrene-d10          | 17.310 | 188  | 547177   | 20.000 | ng    | 0.00     |
| 76) Chrysene-d12              | 21.410 | 240  | 574311   | 20.000 | ng    | 0.00     |
| 86) Perylene-d12              | 23.839 | 264  | 662136   | 20.000 | ng    | 0.00     |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 5) 2-Fluorophenol             | 5.493  | 112  | 90562    | 20.049 | ng    | 0.00     |
| 7) Phenol-d6                  | 7.087  | 99   | 131705   | 19.740 | ng    | 0.00     |
| 23) Nitrobenzene-d5           | 9.093  | 82   | 127301   | 18.263 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol      | 16.051 | 330  | 54921    | 18.404 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl          | 13.187 | 172  | 345132   | 19.851 | ng    | 0.00     |
| 79) Terphenyl-d14             | 19.880 | 244  | 673585   | 19.229 | ng    | 0.00     |
| Target Compounds              |        |      |          |        |       |          |
|                               |        |      |          |        |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.399  | 88   | 18175    | 10.200 | ng    | 97       |
| 3) Pyridine                   | 3.793  | 79   | 52437    | 9.923  | ng    | 96       |
| 4) n-Nitrosodimethylamine     | 3.722  | 42   | 23264    | 9.501  | ng    | # 99     |
| 6) Aniline                    | 7.252  | 93   | 68275    | 10.068 | ng    | 99       |
| 8) 2-Chlorophenol             | 7.487  | 128  | 46943    | 9.892  | ng    | 95       |
| 9) Benzaldehyde               | 7.069  | 77   | 47939    | 8.222  | ng    | 97       |
| 10) Phenol                    | 7.116  | 94   | 65608    | 9.829  | ng    | 96       |
| 11) bis(2-Chloroethyl)ether   | 7.363  | 93   | 52915    | 10.018 | ng    | 97       |
| 12) 1,3-Dichlorobenzene       | 7.810  | 146  | 57515    | 10.151 | ng    | 94       |
| 13) 1,4-Dichlorobenzene       | 7.963  | 146  | 59017    | 10.314 | ng    | 97       |
| 14) 1,2-Dichlorobenzene       | 8.275  | 146  | 55668    | 10.005 | ng    | 97       |
| 15) Benzyl Alcohol            | 8.169  | 79   | 56000    | 9.829  | ng    | 97       |
| 16) 2,2'-oxybis(1-Chloropr... | 8.463  | 45   | 75391    | 10.252 | ng    | 96       |
| 17) 2-Methylphenol            | 8.369  | 107  | 46167    | 10.085 | ng    | 98       |
| 18) Hexachloroethane          | 8.999  | 117  | 24046    | 10.501 | ng    | 96       |
| 19) n-Nitroso-di-n-propyla... | 8.734  | 70   | 53081    | 10.298 | ng    | 99       |
| 20) 3+4-Methylphenols         | 8.693  | 107  | 62222    | 9.829  | ng    | 97       |
| 22) Acetophenone              | 8.752  | 105  | 92800    | 9.780  | ng    | 98       |
| 24) Nitrobenzene              | 9.134  | 77   | 67653    | 9.271  | ng    | 98       |
| 25) Isophorone                | 9.657  | 82   | 149902   | 10.015 | ng    | 100      |
| 26) 2-Nitrophenol             | 9.840  | 139  | 17414    | 9.928  | ng    | 98       |
| 27) 2,4-Dimethylphenol        | 9.904  | 122  | 36694    | 9.805  | ng    | 94       |
| 28) bis(2-Chloroethoxy)met... | 10.152 | 93   | 80977    | 10.082 | ng    | 98       |
| 29) 2,4-Dichlorophenol        | 10.369 | 162  | 47894    | 9.362  | ng    | 99       |
| 30) 1,2,4-Trichlorobenzene    | 10.587 | 180  | 59213    | 9.746  | ng    | 95       |
| 31) Naphthalene               | 10.775 | 128  | 175165   | 9.942  | ng    | 99       |
| 32) Benzoic acid              | 9.975  | 122  | 23273m   | 12.313 | ng    |          |
| 33) 4-Chloroaniline           | 10.893 | 127  | 69801    | 9.799  | ng    | 99       |
| 34) Hexachlorobutadiene       | 11.051 | 225  | 41596    | 9.759  | ng    | 99       |
| 35) Caprolactam               | 11.651 | 113  | 19874    | 9.844  | ng    | 95       |
| 36) 4-Chloro-3-methylphenol   | 12.010 | 107  | 65380    | 9.545  | ng    | 99       |
| 37) 2-Methylnaphthalene       | 12.381 | 142  | 128591   | 10.110 | ng    | 99       |
| 38) 1-Methylnaphthalene       | 12.604 | 142  | 124679   | 9.885  | ng    | 98       |
| 40) 1,2,4,5-Tetrachloroben... | 12.746 | 216  | 72843    | 9.702  | ng    | 99       |
| 41) Hexachlorocyclopentadiene | 12.722 | 237  | 29966    | 9.382  | ng    | 97       |
| 43) 2,4,6-Trichlorophenol     | 12.987 | 196  | 42516    | 9.112  | ng    | 98       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP012624\  
 Data File : BP019445.D  
 Acq On : 26 Jan 2024 12:01  
 Operator : MA/JU  
 Sample : SSTDICC010  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

## Instrument :

BNA\_P

## ClientSampleId :

SSTDICC010

## Manual Integrations

## APPROVED

Reviewed By :Jagrut Upadhyay 01/29/2024  
 Supervised By :mohammad ahmed 01/30/2024

Quant Time: Jan 26 23:30:32 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP012624.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION


QLast Update : Fri Jan 26 23:25:28 2024

Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol     | 13.057 | 196  | 48326    | 9.250  | ng    | 96       |
| 46) 1,1'-Biphenyl             | 13.398 | 154  | 173283   | 9.834  | ng    | 99       |
| 47) 2-Chloronaphthalene       | 13.434 | 162  | 140693   | 9.889  | ng    | 97       |
| 48) 2-Nitroaniline            | 13.640 | 65   | 34079    | 10.498 | ng    | 97       |
| 49) Acenaphthylene            | 14.281 | 152  | 215158   | 9.968  | ng    | 98       |
| 50) Dimethylphthalate         | 14.028 | 163  | 189329   | 9.939  | ng    | 99       |
| 51) 2,6-Dinitrotoluene        | 14.145 | 165  | 31912    | 10.261 | ng    | 97       |
| 52) Acenaphthene              | 14.622 | 154  | 132560   | 9.912  | ng    | 99       |
| 53) 3-Nitroaniline            | 14.469 | 138  | 31419    | 8.672  | ng #  | 89       |
| 54) 2,4-Dinitrophenol         | 14.669 | 184  | 12033    | 10.625 | ng    | 97       |
| 55) Dibenzofuran              | 14.957 | 168  | 217906   | 9.982  | ng    | 100      |
| 56) 4-Nitrophenol             | 14.763 | 139  | 25708    | 8.159  | ng    | 94       |
| 57) 2,4-Dinitrotoluene        | 14.928 | 165  | 40247    | 10.252 | ng    | 99       |
| 58) Fluorene                  | 15.604 | 166  | 183258   | 10.095 | ng    | 98       |
| 59) 2,3,4,6-Tetrachlorophenol | 15.181 | 232  | 41546    | 9.069  | ng    | 98       |
| 60) Diethylphthalate          | 15.392 | 149  | 204620   | 10.060 | ng    | 99       |
| 61) 4-Chlorophenyl-phenyle... | 15.610 | 204  | 96847    | 9.997  | ng    | 99       |
| 62) 4-Nitroaniline            | 15.628 | 138  | 35407    | 8.763  | ng    | 96       |
| 63) Azobenzene                | 15.904 | 77   | 206122   | 10.057 | ng    | 98       |
| 65) 4,6-Dinitro-2-methylph... | 15.687 | 198  | 16208    | 12.861 | ng    | 94       |
| 66) n-Nitrosodiphenylamine    | 15.822 | 169  | 157796   | 9.903  | ng    | 97       |
| 67) 4-Bromophenyl-phenylether | 16.510 | 248  | 60666    | 9.862  | ng    | 99       |
| 68) Hexachlorobenzene         | 16.604 | 284  | 69072    | 9.862  | ng    | 96       |
| 69) Atrazine                  | 16.786 | 200  | 63058    | 10.544 | ng    | 97       |
| 70) Pentachlorophenol         | 16.951 | 266  | 36620    | 8.550  | ng    | 98       |
| 71) Phenanthrene              | 17.357 | 178  | 294567   | 10.113 | ng    | 100      |
| 72) Anthracene                | 17.445 | 178  | 292526   | 10.022 | ng    | 99       |
| 73) Carbazole                 | 17.722 | 167  | 288446   | 10.213 | ng    | 99       |
| 74) Di-n-butylphthalate       | 18.286 | 149  | 357417   | 9.928  | ng    | 99       |
| 75) Fluoranthene              | 19.322 | 202  | 385101   | 10.171 | ng    | 99       |
| 77) Benzidine                 | 19.510 | 184  | 77049    | 8.076  | ng    | 99       |
| 78) Pyrene                    | 19.675 | 202  | 396242   | 9.544  | ng    | 100      |
| 80) Butylbenzylphthalate      | 20.575 | 149  | 154023   | 9.127  | ng    | 99       |
| 81) Benzo(a)anthracene        | 21.392 | 228  | 405800   | 9.851  | ng    | 99       |
| 82) 3,3'-Dichlorobenzidine    | 21.327 | 252  | 141735   | 10.165 | ng    | 98       |
| 83) Chrysene                  | 21.445 | 228  | 388576   | 10.048 | ng    | 99       |
| 84) Bis(2-ethylhexyl)phtha... | 21.345 | 149  | 251020   | 9.614  | ng    | 99       |
| 85) Di-n-octyl phthalate      | 22.298 | 149  | 422207   | 9.621  | ng    | 99       |
| 87) Indeno(1,2,3-cd)pyrene    | 26.374 | 276  | 463448   | 10.681 | ng    | 99       |
| 88) Benzo(b)fluoranthene      | 23.092 | 252  | 425427   | 9.600  | ng    | 99       |
| 89) Benzo(k)fluoranthene      | 23.145 | 252  | 403773   | 9.581  | ng    | 99       |
| 90) Benzo(a)pyrene            | 23.727 | 252  | 369097   | 9.742  | ng    | 98       |
| 91) Dibenzo(a,h)anthracene    | 26.409 | 278  | 388397   | 10.748 | ng    | 99       |
| 92) Benzo(g,h,i)perylene      | 27.151 | 276  | 384374   | 11.062 | ng    | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 7.928 min Scan# 81  
Delta R.T. 0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :





SSTDICC010

Tgt Ion:152 Resp: 7106  
Ion Ratio Lower Upper  
152 100  
150 151.0 123.5 185.3  
115 66.3 55.1 82.7

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#2  
1,4-Dioxane  
Concen: 10.200 ng  
RT: 3.399 min Scan# 53  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01


Tgt Ion: 88 Resp: 18175  
Ion Ratio Lower Upper  
88 100  
58 79.5 61.8 92.6  
43 33.4 25.6 38.4





#3  
Pyridine  
Concen: 9.923 ng  
RT: 3.793 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010




Tgt Ion: 79 Resp: 5243  
Ion Ratio Lower Upper  
79 100  
52 78.5 59.2 88.8  
51 40.1 31.0 46.4

Manual Integrations  
APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024




#4  
n-Nitrosodimethylamine  
Concen: 9.501 ng  
RT: 3.722 min Scan# 108  
Delta R.T. 0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01



Tgt Ion: 42 Resp: 23264  
Ion Ratio Lower Upper  
42 100  
74 117.8 94.1 141.1  
44 13.5 5.4 8.0#









#5  
2-Fluorophenol  
Concen: 20.049 ng  
RT: 5.493 min Scan# 409  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion:112 Resp: 90562  
Ion Ratio Lower Upper  
112 100  
64 64.3 53.1 79.7  
63 34.3 29.4 44.0




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#6  
Aniline  
Concen: 10.068 ng  
RT: 7.252 min Scan# 708  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion: 93 Resp: 68275  
Ion Ratio Lower Upper  
93 100  
66 45.0 35.8 53.8  
65 22.9 17.7 26.5









#7  
Phenol-d6  
Concen: 19.740 ng  
RT: 7.087 min Scan# 681  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion: 99 Resp: 13170  
Ion Ratio Lower Upper  
99 100  
42 20.9 16.4 24.6  
71 43.7 34.6 52.0




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#8  
2-Chlorophenol  
Concen: 9.892 ng  
RT: 7.487 min Scan# 748  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:128 Resp: 46943  
Ion Ratio Lower Upper  
128 100  
130 30.0 12.2 52.2  
64 53.6 37.4 77.4









#9  
Benzaldehyde  
Concen: 8.222 ng  
RT: 7.069 min Scan# 61  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion: 77 Resp: 47939  
Ion Ratio Lower Upper  
77 100  
105 86.3 69.0 109.0  
106 81.9 65.1 105.1




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#10  
Phenol  
Concen: 9.829 ng  
RT: 7.116 min Scan# 685  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01


Tgt Ion: 94 Resp: 65608  
Ion Ratio Lower Upper  
94 100  
65 30.8 12.0 52.0  
66 42.8 26.3 66.3





#11  
bis(2-Chloroethyl)ether  
Concen: 10.018 ng  
RT: 7.363 min Scan# 727  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010



Tgt Ion: 93 Resp: 52915  
Ion Ratio Lower Upper  
93 100  
63 76.5 59.4 99.4  
95 34.0 12.6 52.6




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#12  
1,3-Dichlorobenzene  
Concen: 10.151 ng  
RT: 7.810 min Scan# 803  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:146 Resp: 57515  
Ion Ratio Lower Upper  
146 100  
148 60.2 52.5 78.7  
75 39.2 33.2 49.8









#13  
1,4-Dichlorobenzene  
Concen: 10.314 ng  
RT: 7.963 min Scan# 81  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion:146 Resp: 5901  
Ion Ratio Lower Upper  
146 100  
148 61.7 51.8 77.6  
111 46.0 37.6 56.4




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#14  
1,2-Dichlorobenzene  
Concen: 10.005 ng  
RT: 8.275 min Scan# 882  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:146 Resp: 55668  
Ion Ratio Lower Upper  
146 100  
148 62.0 52.6 78.8  
111 49.1 39.1 58.7









#15  
Benzyl Alcohol  
Concen: 9.829 ng  
RT: 8.169 min Scan# 80  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion: 79 Resp: 56000  
Ion Ratio Lower Upper  
79 100  
108 66.0 51.5 77.3  
77 60.9 50.6 76.0




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 10.252 ng  
RT: 8.463 min Scan# 914  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion: 45 Resp: 75391  
Ion Ratio Lower Upper  
45 100  
77 14.2 0.0 36.2  
79 11.5 0.0 32.5





#17  
2-Methylphenol  
Concen: 10.085 ng  
RT: 8.369 min Scan# 898  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :





SSTDICC010

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 107   | Resp: | 4616  |
| Ion Ratio | Lower | Upper |       |
| 107       | 100   |       |       |
| 108       | 111.6 | 88.8  | 133.2 |
| 77        | 56.8  | 44.6  | 67.0  |
| 79        | 51.3  | 44.6  | 66.8  |

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#18  
Hexachloroethane  
Concen: 10.501 ng  
RT: 8.999 min Scan# 1005  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 117   | Resp: | 24046 |
| Ion Ratio | Lower | Upper |       |
| 117       | 100   |       |       |
| 119       | 92.4  | 78.3  | 117.5 |
| 201       | 108.1 | 89.0  | 133.4 |






#19  
n-Nitroso-di-n-propylamine  
Concen: 10.298 ng  
RT: 8.734 min Scan# 910  
Delta R.T. -0.012 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :

SSTDICC010






Tgt Ion: 70 Resp: 5308  
Ion Ratio Lower Upper  
70 100  
42 56.5 45.4 68.0  
101 9.0 7.4 11.2  
130 16.7 14.3 21.5


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024




#20  
3+4-Methylphenols  
Concen: 9.829 ng  
RT: 8.693 min Scan# 953  
Delta R.T. -0.012 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01



Tgt Ion:107 Resp: 62222  
Ion Ratio Lower Upper  
107 100  
108 83.3 67.6 107.6  
77 40.5 19.7 59.7  
79 35.9 15.6 55.6






#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 10.728 min Scan# 1299  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :

SSTDICC010






Tgt Ion:136 Resp: 31393  
Ion Ratio Lower Upper  
136 100  
137 10.6 8.8 13.2  
54 8.8 7.8 11.6  
68 5.9 5.2 7.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#22  
Acetophenone  
Concen: 9.780 ng  
RT: 8.752 min Scan# 963  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:105 Resp: 92800  
Ion Ratio Lower Upper  
105 100  
71 4.1 4.1 6.1  
51 30.3 23.8 35.6  
120 20.5 15.7 23.5





#23

Nitrobenzene-d5

Concen: 18.263 ng

RT: 9.093 min Scan# 1021

Delta R.T. -0.006 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :

SSTDICC010

Tgt Ion: 82 Resp: 12730

Ion Ratio Lower Upper

82 100





128 32.2 28.2 42.2

54 51.7 41.1 61.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#24

Nitrobenzene

Concen: 9.271 ng

RT: 9.134 min Scan# 1028

Delta R.T. -0.006 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01




Tgt Ion: 77 Resp: 67653


Ion Ratio Lower Upper

77 100

123 39.0 29.7 44.5

65 14.3 11.4 17.2









#25  
Isophorone  
Concen: 10.015 ng  
RT: 9.657 min Scan# 1117  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion: 82 Resp: 149902  
Ion Ratio Lower Upper  
82 100  
95 8.3 6.6 9.8  
138 14.6 11.8 17.8




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#26  
2-Nitrophenol  
Concen: 9.928 ng  
RT: 9.840 min Scan# 1148  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:139 Resp: 17414  
Ion Ratio Lower Upper  
139 100  
109 39.1 33.5 50.3  
65 62.7 49.7 74.5





#27

2,4-Dimethylphenol

Concen: 9.805 ng

RT: 9.904 min Scan# 1159

Delta R.T. -0.000 min

Lab File: BP019445.D


Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :

SSTDICC010



Tgt Ion:122 Resp: 36694

Ion Ratio Lower Upper

122 100




107 125.5 107.4 161.2

121 60.6 46.6 70.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#28

bis(2-Chloroethoxy)methane

Concen: 10.082 ng

RT: 10.152 min Scan# 1201

Delta R.T. -0.006 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01




Tgt Ion: 93 Resp: 80977


Ion Ratio Lower Upper

93 100

95 33.9 25.8 38.8


123 10.4 7.9 11.9





#29  
2,4-Dichlorophenol  
Concen: 9.362 ng  
RT: 10.369 min Scan# 1238  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010




Tgt Ion:162 Resp: 47894  
Ion Ratio Lower Upper  
162 100  
164 64.8 45.0 85.0  
98 39.3 21.0 61.0

Manual Integrations  
APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024




#30  
1,2,4-Trichlorobenzene  
Concen: 9.746 ng  
RT: 10.587 min Scan# 1275  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01



Tgt Ion:180 Resp: 59213  
Ion Ratio Lower Upper  
180 100  
182 99.9 75.1 112.7  
145 31.1 24.1 36.1





#31  
Naphthalene  
Concen: 9.942 ng  
RT: 10.775 min Scan# 1171  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNP\_P

ClientSampleId :





SSTDIC010

Tgt Ion:128 Resp: 17516  
Ion Ratio Lower Upper  
128 100  
129 10.7 8.4 12.6  
127 13.1 10.6 16.0

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#32  
Benzoic acid  
Concen: 12.313 ng m  
RT: 9.975 min Scan# 1171  
Delta R.T. -0.077 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:122 Resp: 23273  
Ion Ratio Lower Upper  
122 100  
105 131.7 115.0 155.0  
77 98.4 91.6 131.6





#33  
4-Chloroaniline  
Concen: 9.799 ng  
RT: 10.893 min Scan# 1327  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :





SSTDICC010

Tgt Ion:127 Resp: 6980  
Ion Ratio Lower Upper  
127 100  
129 31.8 25.9 38.9  
65 38.4 30.9 46.3  
92 24.6 20.4 30.6

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#34  
Hexachlorobutadiene  
Concen: 9.759 ng  
RT: 11.051 min Scan# 1354  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01


Tgt Ion:225 Resp: 41596  
Ion Ratio Lower Upper  
225 100  
223 63.9 51.7 77.5  
227 64.3 50.7 76.1





#35  
Caprolactam  
Concen: 9.844 ng  
RT: 11.651 min Scan# 1461  
Delta R.T. -0.030 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010



Tgt Ion:113 Resp: 19874  
Ion Ratio Lower Upper  
113 100  
55 186.9 156.9 196.9  
56 148.8 126.3 166.3




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#36  
4-Chloro-3-methylphenol  
Concen: 9.545 ng  
RT: 12.010 min Scan# 1517  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:107 Resp: 65380  
Ion Ratio Lower Upper  
107 100  
144 22.2 17.8 26.6  
142 69.4 55.0 82.4









#37  
2-Methylnaphthalene  
Concen: 10.110 ng  
RT: 12.381 min Scan# 1580  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion:142 Resp: 12859  
Ion Ratio Lower Upper  
142 100  
141 91.1 72.5 108.7  
115 41.6 34.4 51.6




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#38  
1-Methylnaphthalene  
Concen: 9.885 ng  
RT: 12.604 min Scan# 1618  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:142 Resp: 124679  
Ion Ratio Lower Upper  
142 100  
141 94.9 74.4 111.6  
116 4.6 3.8 5.6






#39  
Acenaphthene-d10  
Concen: 20.000 ng  
RT: 14.557 min Scan# 1950  
Delta R.T. 0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNP\_P

ClientSampleId :

SSTDICC010






Tgt Ion:164 Resp: 230435  
Ion Ratio Lower Upper  
164 100  
162 101.6 80.8 121.2  
160 44.9 35.0 52.6


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024




#40  
1,2,4,5-Tetrachlorobenzene  
Concen: 9.702 ng  
RT: 12.746 min Scan# 1642  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Tgt Ion:216 Resp: 72843  
Ion Ratio Lower Upper  
216 100  
214 78.7 62.6 93.8  
179 21.3 17.1 25.7  
108 19.3 15.0 22.4





#41  
Hexachlorocyclopentadiene  
Concen: 9.382 ng  
RT: 12.722 min Scan# 1638  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010



Tgt Ion: 237 Resp: 2996  
Ion Ratio Lower Upper  
237 100  
235 60.1 42.8 82.8  
272 11.5 0.0 32.2




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#42  
2,4,6-Tribromophenol  
Concen: 18.404 ng  
RT: 16.051 min Scan# 2204  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01


Tgt Ion: 330 Resp: 54921  
Ion Ratio Lower Upper  
330 100  
332 96.9 77.1 115.7  
141 38.9 30.9 46.3





#43  
2,4,6-Trichlorophenol  
Concen: 9.112 ng  
RT: 12.987 min Scan# 1683  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010




Tgt Ion:196 Resp: 42510  
Ion Ratio Lower Upper  
196 100  
198 96.8 77.4 116.0  
200 34.2 24.4 36.6

Manual Integrations  
APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024




#44  
2,4,5-Trichlorophenol  
Concen: 9.250 ng  
RT: 13.057 min Scan# 1695  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01



Tgt Ion:196 Resp: 48326  
Ion Ratio Lower Upper  
196 100  
198 95.3 77.7 116.5  
97 55.5 48.0 72.0  
132 26.0 22.7 34.1





#45  
2-Fluorobiphenyl  
Concen: 19.851 ng  
RT: 13.187 min Scan# 1718  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :





SSTDICC010

Tgt Ion:172 Resp: 34513  
Ion Ratio Lower Upper  
172 100  
171 35.0 27.9 41.9  
170 23.5 18.7 28.1

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#46  
1,1'-Biphenyl  
Concen: 9.834 ng  
RT: 13.398 min Scan# 1753  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:154 Resp: 173283  
Ion Ratio Lower Upper  
154 100  
153 40.9 20.3 60.3  
76 13.0 0.0 33.3





#47  
2-Chloronaphthalene  
Concen: 9.889 ng  
RT: 13.434 min Scan# 1759  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :





SSTDICC010

Tgt Ion:162 Resp: 14069  
Ion Ratio Lower Upper  
162 100  
127 40.8 34.1 51.1  
164 34.0 26.2 39.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#48  
2-Nitroaniline  
Concen: 10.498 ng  
RT: 13.640 min Scan# 1794  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion: 65 Resp: 34079  
Ion Ratio Lower Upper  
65 100  
92 68.1 51.9 77.9  
138 85.0 67.0 100.6





#49

Acenaphthylene

Concen: 9.968 ng

RT: 14.281 min Scan# 1903

Delta R.T. -0.000 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :





SSTDICC010

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 152   | Resp: | 215158 |
| Ion Ratio | Lower | Upper |        |
| 152       | 100   |       |        |
| 151       | 19.0  | 15.8  | 23.6   |
| 153       | 12.3  | 10.5  | 15.7   |

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#50

Dimethylphthalate

Concen: 9.939 ng




RT: 14.028 min Scan# 1860


Delta R.T. -0.006 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 163   | Resp: | 189329 |
| Ion Ratio | Lower | Upper |        |
| 163       | 100   |       |        |
| 194       | 4.3   | 3.1   | 4.7    |
| 164       | 10.1  | 7.8   | 11.8   |





#51  
2,6-Dinitrotoluene  
Concen: 10.261 ng  
RT: 14.145 min Scan# 1881  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :





SSTDICC010

Tgt Ion:165 Resp: 31912  
Ion Ratio Lower Upper  
165 100  
63 68.7 55.6 83.4  
89 70.4 53.0 79.6

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#52  
Acenaphthene  
Concen: 9.912 ng  
RT: 14.622 min Scan# 1961  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01


Tgt Ion:154 Resp: 132560  
Ion Ratio Lower Upper  
154 100  
153 113.5 91.0 136.6  
152 51.9 42.6 63.8





#53  
3-Nitroaniline  
Concen: 8.672 ng  
RT: 14.469 min Scan# 1936  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010



Tgt Ion:138 Resp: 31419  
Ion Ratio Lower Upper  
138 100  
108 25.5 12.0 18.0  
92 148.4 109.6 164.4




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#54  
2,4-Dinitrophenol  
Concen: 10.625 ng  
RT: 14.669 min Scan# 1969  
Delta R.T. -0.012 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01


Tgt Ion:184 Resp: 12033  
Ion Ratio Lower Upper  
184 100  
63 65.5 54.3 81.5  
154 62.3 51.1 76.7





#55  
Dibenzenofuran  
Concen: 9.982 ng  
RT: 14.957 min Scan# 2019  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010




Tgt Ion:168 Resp: 217900  
Ion Ratio Lower Upper  
168 100  
139 40.2 32.0 48.0  
169 13.2 10.2 15.4

Manual Integrations  
APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024




#56  
4-Nitrophenol  
Concen: 8.159 ng  
RT: 14.763 min Scan# 1985  
Delta R.T. -0.012 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Tgt Ion:139 Resp: 25708  
Ion Ratio Lower Upper  
139 100  
109 92.8 80.4 120.4  
65 120.7 105.5 145.5





#57  
2,4-Dinitrotoluene  
Concen: 10.252 ng  
RT: 14.928 min Scan# 2013  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010




Tgt Ion:165 Resp: 4024  
Ion Ratio Lower Upper  
165 100  
63 52.1 41.4 62.0  
89 82.0 64.2 96.4  
182 2.9 2.2 3.4

Manual Integrations  
APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024




#58  
Fluorene  
Concen: 10.095 ng  
RT: 15.604 min Scan# 2128  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Tgt Ion:166 Resp: 183258  
Ion Ratio Lower Upper  
166 100  
165 95.5 78.1 117.1  
167 13.9 10.7 16.1





#59  
2,3,4,6-Tetrachlorophenol  
Concen: 9.069 ng  
RT: 15.181 min Scan# 2056  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
Client Sample Id :  
SSTDICC010




Tgt Ion:232 Resp: 41540  
Ion Ratio Lower Upper  
232 100  
131 46.2 35.7 53.5  
130 2.9 2.0 3.0  
166 31.8 24.9 37.3

Manual Integrations  
APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024




#60  
Diethylphthalate  
Concen: 10.060 ng  
RT: 15.392 min Scan# 2092  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Tgt Ion:149 Resp: 204620  
Ion Ratio Lower Upper  
149 100  
177 22.8 17.8 26.8  
150 12.8 10.2 15.2





#61  
4-Chlorophenyl-phenylether  
Concen: 9.997 ng  
RT: 15.610 min Scan# 2129  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010




Tgt Ion:204 Resp: 9684  
Ion Ratio Lower Upper  
204 100  
206 33.8 26.4 39.6  
141 62.6 50.2 75.4

Manual Integrations  
APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024




#62  
4-Nitroaniline  
Concen: 8.763 ng  
RT: 15.628 min Scan# 2132  
Delta R.T. -0.012 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01



Tgt Ion:138 Resp: 35407  
Ion Ratio Lower Upper  
138 100  
92 71.4 45.3 85.3  
108 108.2 90.2 130.2





#63  
Azobenzene  
Concen: 10.057 ng  
RT: 15.904 min Scan# 2179  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :





SSTDICC010

Tgt Ion: 77 Resp: 20612  
Ion Ratio Lower Upper  
77 100  
182 23.4 2.4 42.4  
105 20.6 0.0 39.6  
51 31.3 12.1 52.1

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#64  
Phenanthrene-d10  
Concen: 20.000 ng  
RT: 17.310 min Scan# 2418  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01


Tgt Ion:188 Resp: 547177  
Ion Ratio Lower Upper  
188 100  
94 8.7 6.2 9.2  
80 9.7 8.1 12.1





#65  
4,6-Dinitro-2-methylphenol  
Concen: 12.861 ng  
RT: 15.687 min Scan# 2142  
Delta R.T. -0.005 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010



Tgt Ion:198 Resp: 16208  
Ion Ratio Lower Upper  
198 100  
51 44.8 32.3 72.3  
105 43.4 24.1 64.1




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#66  
n-Nitrosodiphenylamine  
Concen: 9.903 ng  
RT: 15.822 min Scan# 2165  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:169 Resp: 157796  
Ion Ratio Lower Upper  
169 100  
168 68.6 52.5 78.7  
167 36.9 28.6 43.0









#67  
4-Bromophenyl-phenylether  
Concen: 9.862 ng  
RT: 16.510 min Scan# 2282  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion:248 Resp: 60660  
Ion Ratio Lower Upper  
248 100  
250 97.5 77.0 115.4  
141 78.8 63.8 95.6




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#68  
Hexachlorobenzene  
Concen: 9.862 ng  
RT: 16.604 min Scan# 2298  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:284 Resp: 69072  
Ion Ratio Lower Upper  
284 100  
142 31.0 28.2 42.4  
249 33.8 27.0 40.4









#69  
Atrazine  
Concen: 10.544 ng  
RT: 16.786 min Scan# 21  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion:200 Resp: 63058  
Ion Ratio Lower Upper  
200 100  
173 25.9 4.2 44.2  
215 47.0 28.5 68.5




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#70  
Pentachlorophenol  
Concen: 8.550 ng  
RT: 16.951 min Scan# 2357  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:266 Resp: 36620  
Ion Ratio Lower Upper  
266 100  
268 63.9 50.9 76.3  
264 61.2 51.0 76.4









#71  
Phenanthrene  
Concen: 10.113 ng  
RT: 17.357 min Scan# 2426  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument : BNA\_P  
ClientSampleId : SSTDICC010

|             |                |
|-------------|----------------|
| Tgt Ion:178 | Resp: 29456    |
| Ion Ratio   | Lower Upper    |
| 178         | 100            |
| 176         | 18.5 15.0 22.4 |
| 179         | 15.3 12.2 18.4 |




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#72  
Anthracene  
Concen: 10.022 ng  
RT: 17.445 min Scan# 2441  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01


|             |                |
|-------------|----------------|
| Tgt Ion:178 | Resp: 292526   |
| Ion Ratio   | Lower Upper    |
| 178         | 100            |
| 176         | 18.3 14.7 22.1 |
| 179         | 15.8 12.2 18.2 |





#73  
 Carbazole  
 Concen: 10.213 ng  
 RT: 17.722 min Scan# 2488  
 Delta R.T. -0.000 min  
 Lab File: BP019445.D  
 Acq: 26 Jan 2024 12:01




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICC010




Tgt Ion:167 Resp: 288440  
 Ion Ratio Lower Upper  
 167 100  
 166 21.3 17.1 25.7  
 139 13.0 10.8 16.2

Manual Integrations  
 APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
 Supervised By :mohammad ahmed 01/30/2024




#74  
 Di-n-butylphthalate  
 Concen: 9.928 ng  
 RT: 18.286 min Scan# 2584  
 Delta R.T. -0.006 min  
 Lab File: BP019445.D  
 Acq: 26 Jan 2024 12:01




Tgt Ion:149 Resp: 357417  
 Ion Ratio Lower Upper  
 149 100  
 150 9.7 7.4 11.0  
 104 5.9 4.9 7.3





#75  
Fluoranthene  
Concen: 10.171 ng  
RT: 19.322 min Scan# 21  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010



Tgt Ion:202 Resp: 38510  
Ion Ratio Lower Upper  
202 100  
101 9.2 0.0 29.1  
203 17.8 0.0 37.0




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#76  
Chrysene-d12  
Concen: 20.000 ng  
RT: 21.410 min Scan# 3115  
Delta R.T. -0.000 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01


Tgt Ion:240 Resp: 574311  
Ion Ratio Lower Upper  
240 100  
120 7.9 6.9 10.3  
236 25.9 19.8 29.8





#77  
Benzidine  
Concen: 8.076 ng  
RT: 19.510 min Scan# 21  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010




Tgt Ion:184 Resp: 77049  
Ion Ratio Lower Upper  
184 100  
185 14.2 11.0 16.4  
183 12.2 9.4 14.2

Manual Integrations  
APPROVED



Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024




#78  
Pyrene  
Concen: 9.544 ng  
RT: 19.675 min Scan# 2820  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01



Tgt Ion:202 Resp: 396242  
Ion Ratio Lower Upper  
202 100  
200 21.0 16.5 24.7  
203 17.5 13.9 20.9





#79

Terphenyl-d14

Concen: 19.229 ng

RT: 19.880 min Scan# 2856

Delta R.T. -0.006 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :

SSTDICC010





Tgt Ion:244 Resp: 67358

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 244 | 100   |       |       |
| 212 | 7.4   | 5.8   | 8.8   |
| 122 | 8.9   | 6.6   | 10.0  |

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#80

Butylbenzylphthalate

Concen: 9.127 ng

RT: 20.575 min Scan# 2973




Delta R.T. -0.000 min


Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

Tgt Ion:149 Resp: 154023

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 149 | 100   |       |       |
| 91  | 76.9  | 62.5  | 93.7  |
| 206 | 22.5  | 17.7  | 26.5  |









#81  
Benzo(a)anthracene  
Concen: 9.851 ng  
RT: 21.392 min Scan# 3113  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion:228 Resp: 405800  
Ion Ratio Lower Upper  
228 100  
226 27.4 21.4 32.2  
229 19.8 15.7 23.5




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#82  
3,3'-Dichlorobenzidine  
Concen: 10.165 ng  
RT: 21.327 min Scan# 3101  
Delta R.T. -0.006 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:252 Resp: 141735  
Ion Ratio Lower Upper  
252 100  
254 64.4 50.4 75.6  
126 10.9 7.5 11.3





#83

Chrysene

Concen: 10.048 ng

RT: 21.445 min Scan# 3122

Delta R.T. -0.006 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :





SSTDICC010

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 228   | Resp: | 388570 |
| Ion Ratio | Lower | Upper |        |
| 228       | 100   |       |        |
| 226       | 29.6  | 23.9  | 35.9   |
| 229       | 19.8  | 15.4  | 23.2   |

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#84

Bis(2-ethylhexyl)phthalate

Concen: 9.614 ng




RT: 21.345 min Scan# 3104


Delta R.T. -0.000 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 149   | Resp: | 251020 |
| Ion Ratio | Lower | Upper |        |
| 149       | 100   |       |        |
| 167       | 27.9  | 22.0  | 33.0   |
| 279       | 4.5   | 3.7   | 5.5    |





#85

Di-n-octyl phthalate

Concen: 9.621 ng

RT: 22.298 min Scan# 31

Delta R.T. -0.000 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :

SSTDICC010





Tgt Ion:149 Resp: 42220

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 149 | 100   |       |       |
| 167 | 1.6   | 1.3   | 1.9   |
| 43  | 10.8  | 8.2   | 12.2  |

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#86

Perylene-d12

Concen: 20.000 ng

RT: 23.839 min Scan# 3528




Delta R.T. -0.006 min


Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

Tgt Ion:264 Resp: 662136

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 264 | 100   |       |       |
| 260 | 23.6  | 18.6  | 28.0  |
| 265 | 21.7  | 16.4  | 24.6  |





#87  
 Indeno(1,2,3-cd)pyrene  
 Concen: 10.681 ng  
 RT: 26.374 min Scan# 3959  
 Delta R.T. -0.018 min  
 Lab File: BP019445.D  
 Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P





ClientSampleId :

SSTDICC010

Tgt Ion: 276 Resp: 463448  
 Ion Ratio Lower Upper  
 276 100  
 138 20.4 15.5 23.3  
 277 25.8 20.6 30.8




Manual Integrations  
 APPROVED


Reviewed By : Jagrut Upadhyay 01/29/2024  
 Supervised By : mohammad ahmed 01/30/2024



#88  
 Benzo(b)fluoranthene  
 Concen: 9.600 ng  
 RT: 23.092 min Scan# 3401  
 Delta R.T. -0.012 min  
 Lab File: BP019445.D  
 Acq: 26 Jan 2024 12:01

Tgt Ion: 252 Resp: 425427  
 Ion Ratio Lower Upper  
 252 100  
 253 21.4 17.6 26.4  
 125 8.1 6.5 9.7









#89  
Benzo(k)fluoranthene  
Concen: 9.581 ng  
RT: 23.145 min Scan# 3412  
Delta R.T. -0.012 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICC010

Tgt Ion:252 Resp: 40377  
Ion Ratio Lower Upper  
252 100  
253 21.6 17.6 26.4  
125 8.4 5.9 8.9




Manual Integrations  
APPROVED


Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#90  
Benzo(a)pyrene  
Concen: 9.742 ng  
RT: 23.727 min Scan# 3509  
Delta R.T. -0.012 min  
Lab File: BP019445.D  
Acq: 26 Jan 2024 12:01

Tgt Ion:252 Resp: 369097  
Ion Ratio Lower Upper  
252 100  
253 22.4 17.0 25.4  
125 8.3 6.6 10.0





#91

Dibenzo(a,h)anthracene

Concen: 10.748 ng

RT: 26.409 min Scan# 3966

Delta R.T. -0.006 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

Instrument :

BNA\_P

ClientSampleId :





SSTDICC010

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 278   | Resp: | 38839 |
| Ion Ratio | Lower | Upper |       |
| 278       | 100   |       |       |
| 139       | 12.4  | 10.2  | 15.4  |
| 279       | 23.7  | 19.1  | 28.7  |

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 01/29/2024  
Supervised By :mohammad ahmed 01/30/2024



#92

Benzo(g,h,i)perylene

Concen: 11.062 ng

RT: 27.151 min Scan# 4091

Delta R.T. -0.012 min

Lab File: BP019445.D

Acq: 26 Jan 2024 12:01

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 276   | Resp: | 384374 |
| Ion Ratio | Lower | Upper |        |
| 276       | 100   |       |        |
| 277       | 22.8  | 18.9  | 28.3   |
| 138       | 17.2  | 13.9  | 20.9   |

