

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012723\
 Data File : BP013638.D
 Acq On : 28 Jan 2023 02:48
 Operator : CG/JU
 Sample : 01322-01
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EO-01-1-26-2023

Quant Time: Jan 29 23:57:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 27 03:27:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.158	152	187419	20.000	ng	0.00
21) Naphthalene-d8	10.987	136	671277	20.000	ng	0.00
39) Acenaphthene-d10	14.793	164	416862	20.000	ng	0.00
64) Phenanthrene-d10	17.551	188	922706	20.000	ng	0.00
76) Chrysene-d12	21.639	240	925484	20.000	ng	0.00
86) Perylene-d12	24.239	264	1124217	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.687	112	970392	92.782	ng	0.00
7) Phenol-d6	7.299	99	1264407	89.492	ng	0.00
23) Nitrobenzene-d5	9.328	82	712919	58.356	ng	0.00
42) 2,4,6-Tribromophenol	16.287	330	586503	92.411	ng	0.00
45) 2-Fluorobiphenyl	13.416	172	1760940	60.332	ng	0.00
79) Terphenyl-d14	20.086	244	2679179	51.084	ng	0.00
Target Compounds						
						Qvalue
31) Naphthalene	11.034	128	109098	3.198	ng	98
37) 2-Methylnaphthalene	12.634	142	239239	9.177	ng	99
38) 1-Methylnaphthalene	12.852	142	246857	10.117	ng	95
48) 2-Nitroaniline	13.875	65	315	3.970	ng	# 51
49) Acenaphthylene	14.522	152	732108	18.723	ng	98
52) Acenaphthene	14.857	154	89069	3.559	ng	98
55) Dibenzofuran	15.187	168	132800	3.388	ng	# 69
56) 4-Nitrophenol	14.998	139	12460	2.361	ng	# 46
58) Fluorene	15.840	166	508009	16.949	ng	99
71) Phenanthrene	17.598	178	5412687	114.377	ng	99
72) Anthracene	17.687	178	1653784	34.585	ng	100
73) Carbazole	17.945	167	254119	5.964	ng	98
74) Di-n-butylphthalate	18.457	149	6787	3.729	ng	# 1
75) Fluoranthene	19.545	202	3723328	63.081	ng	98
78) Pyrene	19.904	202	6284680	87.743	ng	99
80) Butylbenzylphthalate	20.739	149	2035	5.544	ng	# 12
81) Benzo(a)anthracene	21.622	228	2918877	44.788	ng	96
82) 3,3'-Dichlorobenzidine	21.516	252	4002	6.565	ng	# 74
83) Chrysene	21.680	228	2693422	43.985	ng	99
84) Bis(2-ethylhexyl)phtha...	21.522	149	5982	3.752	ng	# 30
87) Indeno(1,2,3-cd)pyrene	26.980	276	1045315	13.226	ng	96
88) Benzo(b)fluoranthene	23.445	252	2707135	37.507	ng	99
89) Benzo(k)fluoranthene	23.445	252	2707135	37.782	ng	98
90) Benzo(a)pyrene	24.121	252	2079048	30.604	ng	98
91) Dibenzo(a,h)anthracene	26.986	278	373874	5.685	ng	# 91
92) Benzo(g,h,i)perylene	27.839	276	1384079	21.374	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012723\
Data File : BP013638.D
Acq On : 28 Jan 2023 02:48
Operator : CG/JU
Sample : 01322-01
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EO-01-1-26-2023

Quant Time: Jan 29 23:57:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 27 03:27:54 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.158 min Scan# 845
Delta R.T. -0.000 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48




Instrument :
BNA_P
ClientSampleId :
EO-01-1-26-2023


Tgt Ion:152 Resp: 187419
Ion Ratio Lower Upper
152 100
150 150.7 126.2 189.2
115 55.2 45.1 67.7



#5
2-Fluorophenol
Concen: 92.782 ng
RT: 5.687 min Scan# 425
Delta R.T. 0.006 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48

Tgt Ion:112 Resp: 970392
Ion Ratio Lower Upper
112 100
64 56.7 45.5 68.3
63 30.9 25.1 37.7









#7
Phenol-d6
Concen: 89.492 ng
RT: 7.299 min Scan# 699
Delta R.T. 0.000 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48

Instrument :
BNA_P
ClientSampleId :
EO-01-1-26-2023

Tgt Ion: 99 Resp: 1264407




Ion	Ratio	Lower	Upper
99	100		
42	17.2	13.4	20.2
71	37.8	29.9	44.9




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.987 min Scan# 1326
Delta R.T. 0.000 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48

Tgt Ion: 136 Resp: 671277

Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.6	13.0
54	7.3	6.1	9.1
68	6.0	5.0	7.4









#23
Nitrobenzene-d5
Concen: 58.356 ng
RT: 9.328 min Scan# 1044
Delta R.T. -0.006 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48




Instrument :
BNA_P
ClientSampleId :
EO-01-1-26-2023


Tgt Ion: 82 Resp: 712919
Ion Ratio Lower Upper
82 100
128 40.4 31.8 47.8
54 45.0 36.2 54.2



#31
Naphthalene
Concen: 3.198 ng
RT: 11.034 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48

Tgt Ion: 128 Resp: 109098
Ion Ratio Lower Upper
128 100
129 13.0 8.9 13.3
127 13.0 10.4 15.6









#37
2-Methylnaphthalene
Concen: 9.177 ng
RT: 12.634 min Scan# 1606
Delta R.T. 0.000 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48




Instrument :
BNA_P
ClientSampleId :
EO-01-1-26-2023


Tgt Ion:142 Resp: 239239
Ion Ratio Lower Upper
142 100
141 89.3 70.8 106.2
115 36.6 30.6 46.0



#38
1-Methylnaphthalene
Concen: 10.117 ng
RT: 12.852 min Scan# 1643
Delta R.T. 0.000 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48

Tgt Ion:142 Resp: 246857
Ion Ratio Lower Upper
142 100
141 89.7 75.5 113.3
116 4.0 3.3 4.9





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.793 min Scan# 1973

Delta R.T. 0.000 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:164 Resp: 416862

Ion Ratio Lower Upper

164 100

162 103.0 82.3 123.5

160 45.5 36.6 54.8



#42

2,4,6-Tribromophenol

Concen: 92.411 ng

RT: 16.287 min Scan# 2227

Delta R.T. 0.006 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48




Tgt Ion:330 Resp: 586503


Ion Ratio Lower Upper

330 100

332 96.6 77.0 115.4

141 28.8 23.3 34.9









#45
2-Fluorobiphenyl
Concen: 60.332 ng
RT: 13.416 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48




Instrument :
BNA_P
ClientSampleId :
EO-01-1-26-2023


Tgt Ion:172 Resp: 1760940
Ion Ratio Lower Upper
172 100
171 34.7 28.2 42.4
170 23.3 19.1 28.7



#48
2-Nitroaniline
Concen: 3.970 ng
RT: 13.875 min Scan# 1817
Delta R.T. 0.006 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48

Tgt Ion: 65 Resp: 315
Ion Ratio Lower Upper
65 100
92 64.9 59.5 89.3
138 30.1 87.1 130.7#





#49

Acenaphthylene

Concen: 18.723 ng

RT: 14.522 min Scan# 1926

Delta R.T. 0.006 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:152 Resp: 732108

Ion Ratio Lower Upper

152 100

151 20.4 16.0 24.0

153 14.3 10.6 16.0



#52

Acenaphthene

Concen: 3.559 ng

RT: 14.857 min Scan# 1984

Delta R.T. 0.000 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48




Tgt Ion:154 Resp: 89069


Ion Ratio Lower Upper

154 100

153 113.3 91.0 136.6

152 59.1 44.0 66.0









#55
Dibenzofuran
Concen: 3.388 ng
RT: 15.187 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48




Instrument :
BNA_P
ClientSampleId :
EO-01-1-26-2023


Tgt Ion:168 Resp: 132800
Ion Ratio Lower Upper
168 100
139 43.1 29.0 43.4
169 46.1 10.6 15.8#



#56
4-Nitrophenol
Concen: 2.361 ng
RT: 14.998 min Scan# 2008
Delta R.T. 0.017 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48

Tgt Ion:139 Resp: 12460
Ion Ratio Lower Upper
139 100
109 60.2 69.5 109.5#
65 27.6 83.1 123.1#





#58

Fluorene

Concen: 16.949 ng

RT: 15.840 min Scan# 2151

Delta R.T. 0.000 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:166 Resp: 508009

Ion Ratio Lower Upper

166 100

165 99.5 79.4 119.0

167 14.9 10.9 16.3



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.551 min Scan# 2442

Delta R.T. 0.000 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48




Tgt Ion:188 Resp: 922706


Ion Ratio Lower Upper

188 100

94 7.7 6.1 9.1

80 8.3 6.6 10.0





#71

Phenanthrene

Concen: 114.377 ng

RT: 17.598 min Scan# 2449

Delta R.T. 0.006 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:178 Resp: 5412687

Ion Ratio Lower Upper

178 100

176 19.8 15.3 22.9

179 15.7 12.2 18.4



#72

Anthracene

Concen: 34.585 ng

RT: 17.687 min Scan# 2465

Delta R.T. 0.006 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48




Tgt Ion:178 Resp: 1653784


Ion Ratio Lower Upper

178 100

176 18.5 15.0 22.4

179 15.6 12.3 18.5









#73
 Carbazole
 Concen: 5.964 ng
 RT: 17.945 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BP013638.D
 Acq: 28 Jan 2023 02:48




Instrument :
 BNA_P
 ClientSampleId :
 EO-01-1-26-2023


Tgt Ion:167 Resp: 254119
 Ion Ratio Lower Upper
 167 100
 166 22.9 17.4 26.2
 139 13.1 9.8 14.6



#74
 Di-n-butylphthalate
 Concen: 3.729 ng
 RT: 18.457 min Scan# 2596
 Delta R.T. -0.024 min
 Lab File: BP013638.D
 Acq: 28 Jan 2023 02:48

Tgt Ion:149 Resp: 6787
 Ion Ratio Lower Upper
 149 100
 150 464.9 7.6 11.4#
 104 18.5 4.4 6.6#





#75

Fluoranthene

Concen: 63.081 ng

RT: 19.545 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:202 Resp: 3723328

Ion Ratio Lower Upper

202 100

101 7.9 0.0 28.9

203 18.0 0.0 37.6



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.639 min Scan# 3137

Delta R.T. 0.006 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48




Tgt Ion:240 Resp: 925484


Ion Ratio Lower Upper

240 100

120 7.2 6.7 10.1

236 26.5 20.7 31.1





#78

Pyrene

Concen: 87.743 ng

RT: 19.904 min Scan# 2842

Delta R.T. 0.012 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:202 Resp: 6284680

Ion Ratio Lower Upper

202 100

200 21.5 17.0 25.6

203 18.6 14.2 21.2



#79

Terphenyl-d14

Concen: 51.084 ng

RT: 20.086 min Scan# 2873

Delta R.T. 0.006 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48




Tgt Ion:244 Resp: 2679179


Ion Ratio Lower Upper

244 100

212 7.3 5.9 8.9

122 6.6 5.3 7.9





#80

Butylbenzylphthalate

Concen: 5.544 ng

RT: 20.739 min Scan# 2984

Delta R.T. -0.012 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:149 Resp: 2035

Ion Ratio Lower Upper

149 100

91 130.0 57.0 85.6#

206 92.9 19.0 28.6#



#81

Benzo(a)anthracene

Concen: 44.788 ng

RT: 21.622 min Scan# 3134

Delta R.T. 0.006 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48




Tgt Ion:228 Resp: 2918877


Ion Ratio Lower Upper

228 100

226 29.5 22.0 33.0

229 21.4 15.9 23.9





#82

3,3'-Dichlorobenzidine

Concen: 6.565 ng

RT: 21.516 min Scan# 3116

Delta R.T. -0.023 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:252 Resp: 4002

Ion Ratio Lower Upper

252 100

254 54.8 52.3 78.5

126 53.1 6.2 9.4#



#83

Chrysene

Concen: 43.985 ng

RT: 21.680 min Scan# 3144

Delta R.T. 0.011 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48




Tgt Ion:228 Resp: 2693422


Ion Ratio Lower Upper

228 100

226 31.1 24.8 37.2

229 20.8 15.8 23.8





#84

Bis(2-ethylhexyl)phthalate

Concen: 3.752 ng

RT: 21.522 min Scan# 3116

Delta R.T. 0.006 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:149 Resp: 5982

Ion Ratio Lower Upper

149 100

167 63.0 22.2 33.4#

279 32.7 3.8 5.6#



#86

Perylene-d12

Concen: 20.000 ng

RT: 24.239 min Scan# 3579

Delta R.T. 0.018 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48




Tgt Ion:264 Resp: 1124217


Ion Ratio Lower Upper

264 100

260 24.1 18.8 28.2

265 25.0 17.4 26.0





#87

Indeno(1,2,3-cd)pyrene

Concen: 13.226 ng

RT: 26.980 min Scan# 4045

Delta R.T. 0.018 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:276 Resp: 1045315

Ion Ratio Lower Upper

276 100

138 16.6 16.4 24.6

277 25.4 20.2 30.2



#88

Benzo(b)fluoranthene

Concen: 37.507 ng

RT: 23.445 min Scan# 3444

Delta R.T. 0.018 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48




Tgt Ion:252 Resp: 2707135


Ion Ratio Lower Upper

252 100

253 23.0 17.8 26.8

125 8.1 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 37.782 ng
RT: 23.445 min Scan# 34
Delta R.T. -0.035 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48




Instrument :
BNA_P
ClientSampleId :
EO-01-1-26-2023


Tgt Ion:252 Resp: 2707135
Ion Ratio Lower Upper
252 100
253 23.0 17.6 26.4
125 8.1 5.9 8.9



#90
Benzo(a)pyrene
Concen: 30.604 ng
RT: 24.121 min Scan# 3559
Delta R.T. 0.012 min
Lab File: BP013638.D
Acq: 28 Jan 2023 02:48

Tgt Ion:252 Resp: 2079048
Ion Ratio Lower Upper
252 100
253 22.5 17.2 25.8
125 7.7 6.8 10.2





#91

Dibenzo(a,h)anthracene

Concen: 5.685 ng

RT: 26.986 min Scan# 4045

Delta R.T. 0.006 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Instrument :

BNA_P

ClientSampleId :

EO-01-1-26-2023





Tgt Ion:278 Resp: 373874

Ion Ratio Lower Upper

278 100

139 15.9 10.6 15.8#

279 28.8 18.7 28.1#



#92

Benzo(g,h,i)perylene

Concen: 21.374 ng

RT: 27.839 min Scan# 4191

Delta R.T. 0.029 min

Lab File: BP013638.D

Acq: 28 Jan 2023 02:48

Tgt Ion:276 Resp: 1384079

Ion Ratio Lower Upper

276 100

277 24.7 18.9 28.3

138 16.5 14.3 21.5

